

Brunswick County Public Utilities - NC

PO Box 249
Bolivia, NC 28422-0249

Leland, NC

Client Project# Northeast Water Plant
Samples Received: 4/5/2024

Analytical Report 0424-750

PFAS by Isotope Dilution (non-potable water)

Custom List

Report Issue Date: 5/2/2024

I certify that to the best of my knowledge all analytical data presented in this report have been checked for completeness, accuracy, errors and legibility in addition to having been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s). This analytical report was prepared in Portable Document Format (.PDF) and contains 24 pages. This report shall not be reproduced except in full without approval of the laboratory. This will provide assurance that parts of the report are not taken out of context.

Amendment(s):

Signature:



Amanda Valois, QA Associate



Enthalpy Analytical, LLC – Ultratrace
Christina Kurnath, Project Manager
chkurnath@montrose-env.com / www.enthalpy.com
O: 910-876-6895
2714 Exchange Drive, Wilmington, NC 28405

Summary of Results



Enthalpy Analytical

Job No.: 0424-750-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC Northeast Water Plant Leland, NC

Summary

	Compound	CAS	040524-S01 ng/L	040524-E01 ng/L	
Acids	PFPtA	422-64-0	5.18	4.83	
	PFBA	375-22-4	ND U	ND U	
	PFPeA	2706-90-3	2.94	3.00	
	PFHxA	307-24-4	3.55	3.35	
	PFHpA	375-85-9	1.71	1.57	
	PFOA	335-67-1	4.23	4.08	
	PFNA	375-95-1	0.427 J	0.427 J	
	PFDA	335-76-2	0.188 J	0.168 J	
	PFUnDA	2058-94-8	ND U	ND U	
	PFDoDA	307-55-1	ND U	ND U	
	PFTtDA	72629-94-8	ND U	ND U	
	PFTeDA	376-06-7	ND U	ND U	
	PFHxDA	67905-19-5	ND U	ND U	
	Sulfonates	PFBS	375-73-5	2.68	2.38
		PFPeS	2706-91-4	0.321 J	0.368 J
PFHxS		355-46-4	1.67	1.87	
PFHpS		375-92-8	0.158 L	0.0917 L	
PFOS		1763-23-1	7.82	7.75	
PFNS		68259-12-1	ND U	ND U	
PFDS		335-77-3	ND U	ND U	
4:2 FTS		757124-72-4	ND U	ND U	
6:2 FTS		27619-97-2	0.0973 L	0.0666 L	
8:2 FTS		39108-34-4	ND U	ND U	
10:2 FTS		120226-60-0	ND U	ND U	
Sulfonamidos		FBSA	30334-69-1	0.311 J	0.229 L
	N-EtFOSA	4151-50-2	ND U	ND U	
	N-EtFOSAA	2991-50-6	ND U	ND U	
	N-EtFOSE	1691-99-2	ND U	ND U	
	N-MeFOSA	31506-32-8	ND U	ND U	
	N-MeFOSAA	2355-31-9	ND U	ND U	
	N-MeFOSE	24448-09-7	ND U	ND U	
	PFOSA	754-91-6	ND U	ND U	
PFECAs	ADONA	919005-14-4	ND U	ND U	
	EVE Acid	69087-46-3	ND U	ND U	
	HFPO-DA	13252-13-6	1.33	1.22	
	Hydro-EVE Acid	773804-62-9	0.0124 L	ND U	
	NFDHA	151772-58-6	ND U	ND U	
	PEPA	267239-61-2	0.678	0.421 J	
	PFECA-G	801212-59-9	ND U	ND U	
	PFMOAA	674-13-5	7.70	7.94	
	PFMOBA	863090-89-5	ND U	ND U	
	PFMOPrA	377-73-1	ND U	ND U	
	PFO2HxA	39492-88-1	ND U	ND U	
	PFO3OA	39492-89-2	ND U	ND U	
	PFO4DA	39492-90-5	ND U	ND U	
	PFO5DA	39492-91-6	ND U	ND U	
	PMPA	13140-29-9	2.52	2.40	
	R-EVE	2416366-22-6	2.81	2.06	
	PFESAs	11Cl-PF3OUds	763051-92-9	ND U	ND U
		9Cl-PF3ONS	756426-58-1	ND U	ND U
Hydrolyzed PSDA		2416366-19-1	0.794	0.588	
Nafion Byproduct 1 (PS Acid)		29311-67-9	ND U	ND U	
Nafion Byproduct 2 (Hydro-PS Acid)		749836-20-2	0.0697 L	0.0654 L	
NVHOS		1132933-86-8	ND U	ND U	
PFEESA		113507-82-7	ND U	ND U	
R-PSDA		2416366-18-0	1.63 L	1.16 L	
R-PSDCA		241636-21-5	ND U	ND U	

Detailed Results

Enthalpy Analytical

Job No.: 0424-750-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC Northeast Water Plant Leland, NC

Details

Sample Name 040524-S01
 Sampling Site
 Enthalpy ID 0424-750-001-1 Prep Batch EU17218
 Matrix aqueous Analyst rappelle
 Sampling Date 2024-04-05 09:50 Instrument Sauron
 Received Date 2024-04-05 Sample Vol mL 289.17
 Prep Date 2024-04-10 09:00 Extract Vol mL 0.4
 AnalysisDate 2024-04-11 00:38 Split Factor N/A
 SampleType Sample Method Code EU-047-NPW
 Bottle ID A

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	F110424005	5.18	0.484	1.06			
	PFBA	375-22-4	S100424037	ND	0.220	0.553			U
	PFPeA	2706-90-3	S100424037	2.94	0.158	0.553			
	PFHxA	307-24-4	S100424037	3.55	0.185	0.553			
	PFHpA	375-85-9	S100424037	1.71	0.194	0.553			
	PFOA	335-67-1	S100424037	4.23	0.127	0.553			
	PFNA	375-95-1	S100424037	0.427	0.125	0.553			J
	PFDA	335-76-2	S100424037	0.188	0.158	0.553			J
	PFUnDA	2058-94-8	S100424037	ND	0.125	0.553			U
	PFDoDA	307-55-1	S100424037	ND	0.225	0.553			U
	PFTrDA	72629-94-8	S100424037	ND	0.183	0.553			U
	PFTeDA	376-06-7	S100424037	ND	0.211	0.553			U
	PFHxDA	67905-19-5	S100424037	ND	0.294	0.553			U
Sulfonates	PFBS	375-73-5	S100424037	2.68	0.294	0.553			
	PFPeS	2706-91-4	S100424037	0.321	0.114	0.521			J
	PFHxS	355-46-4	S100424037	1.67	0.427	0.507			
	PFHpS	375-92-8	S100424037	0.158	0.268	0.527			L
	PFOS	1763-23-1	S100424037	7.82	0.292	0.513			
	PFNS	68259-12-1	S100424037	ND	0.172	0.533			U
	PFDS	335-77-3	S100424037	ND	0.290	0.533			U
	4:2 FTS	757124-72-4	S100424037	ND	0.0718	0.518			U
	6:2 FTS	27619-97-2	S100424037	0.0973	0.261	0.527			L
	8:2 FTS	39108-34-4	S100424037	ND	0.124	0.530			U
10:2 FTS	120226-60-0	S100424037	ND	0.424	0.553			U	
Sulfonamidos	FBSA	30334-69-1	S100424037	0.311	0.263	0.553			J
	N-EtFOSA	4151-50-2	S100424037	ND	0.342	0.553			U
	N-EtFOSAA	2991-50-6	S100424037	ND	0.225	0.553			U
	N-EtFOSE	1691-99-2	S100424037	ND	0.847	2.49			U
	N-MeFOSA	31506-32-8	S100424037	ND	0.228	0.553			U
	N-MeFOSAA	2355-31-9	S100424037	ND	0.155	0.553			U
	N-MeFOSE	24448-09-7	S100424037	ND	0.526	2.49			U
	PFOSA	754-91-6	S100424037	ND	0.0776	0.553			U
PFECAs	ADONA	919005-14-4	S100424037	ND	0.150	0.524			U
	EVE Acid	69087-46-3	S100424037	ND	0.176	1.24			U
	HFPO-DA	13252-13-6	S100424037	1.33	0.0586	0.553			
	Hydro-EVE Acid	773804-62-9	S100424037	0.0124	0.182	0.553			L
	NFDHA	151772-58-6	S100424037	ND	0.116	0.553			U
	PEPA	267239-61-2	S100424037	0.678	0.104	0.553			
	PFECA-G	801212-59-9	S100424037	ND	0.0738	0.553			U
	PFMOAA	674-13-5	S100424037	7.70	0.280	0.553			
	PFMOBA	863090-89-5	S100424037	ND	0.929	1.24			U
	PFMOPrA	377-73-1	S100424037	ND	0.197	0.553			U
	PFO2HxA	39492-88-1	S100424037	ND	0.178	0.553			U
	PFO3OA	39492-89-2	S100424037	ND	0.254	0.553			U
	PFO4DA	39492-90-5	S100424037	ND	0.437	2.77			U
	PFO5DA	39492-91-6	S100424037	ND	0.443	2.77			U
	PMPA	13140-29-9	S100424037	2.52	0.130	0.553			
R-EVE	2416366-22-6	S100424037	2.81	0.918	1.24				
PFESAs	11Cl-PF3OUdS	763051-92-9	S100424037	ND	0.261	0.521			U
	9Cl-PF3ONS	756426-58-1	S100424037	ND	0.354	0.515			U
	Hydrolyzed PSDA	2416366-19-1	S100424037	0.794	0.368	0.553			
	Nafion Byproduct 1 (PS Acid)	29311-67-9	S100424037	ND	0.296	0.553			U
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	S100424037	0.0697	0.458	0.553			L
	NVHOS	1132933-86-8	S100424037	ND	0.0852	0.553			U
	PFEESA	113507-82-7	S100424037	ND	0.166	0.553			U
R-PSDA	2416366-18-0	S100424037	1.63	2.44	2.44			L	
R-PSDCA	241636-21-5	S100424037	ND	0.233	0.553			U	
ES	MPFBA		S100424037				20-150%	116.3%	
	M5PFPeA		S100424037				20-150%	169.9%	Q
	M3PFBS		S100424037				20-150%	184.5%	Q

Enthalpy Analytical

Job No.: 0424-750-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC Northeast Water Plant Leland, NC

Details

Sample Name 040524-S01
 Sampling Site
 Enthalpy ID 0424-750-001-1 Prep Batch EU17218
 Matrix aqueous Analyst rappelle
 Sampling Date 2024-04-05 09:50 Instrument Sauron
 Received Date 2024-04-05 Sample Vol mL 289.17
 Prep Date 2024-04-10 09:00 Extract Vol mL 0.4
 AnalysisDate 2024-04-11 00:38 Split Factor N/A
 SampleType Sample Method Code EU-047-NPW
 Bottle ID A

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M2-4:2 FTS		S100424037				20-150%	117.7%	
M5PFHxA		S100424037				20-150%	95.0%	
M3HFPO-DA		S100424037				20-150%	94.2%	
M4PFHpA		S100424037				20-150%	96.7%	
M3PFHxS		S100424037				20-150%	114.1%	
M2-6:2 FTS		S100424037				20-150%	127.5%	
M8PFOA		S100424037				20-150%	112.1%	
M9PFNA		S100424037				20-150%	114.1%	
M8PFOS		S100424037				20-150%	117.3%	
M2-8:2 FTS		S100424037				20-150%	137.2%	
M8FOSA-I		S100424037				20-150%	109.2%	
M6PFDA		S100424037				20-150%	112.3%	
d3-N-MeFOSAA		S100424037				20-150%	112.7%	
d5-N-EtFOSAA		S100424037				20-150%	104.6%	
M7PFUdA		S100424037				20-150%	107.9%	
MPFDoA		S100424037				20-150%	99.8%	
M2PFTeDA		S100424037				20-150%	65.0%	
d3-N-MeFOSA		S100424037				10-200%	28.1%	
d5-N-EtFOSA		S100424037				10-200%	25.1%	
d7-N-MeFOSE		S100424037				10-200%	67.2%	
d9-N-EtFOSE		S100424037				10-200%	60.7%	
13C3-PFPtA		F110424005				20-150%	14.5%	Q

Enthalpy Analytical

Job No.: 0424-750-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC Northeast Water Plant Leland, NC

Details

Sample Name 040524-E01
 Sampling Site
 Enthalpy ID 0424-750-002-1 Prep Batch EU17218
 Matrix aqueous Analyst rappelle
 Sampling Date 2024-04-05 09:50 Instrument Sauron
 Received Date 2024-04-05 Sample Vol mL 281.83
 Prep Date 2024-04-10 09:00 Extract Vol mL 0.4
 AnalysisDate 2024-04-11 01:01 Split Factor N/A
 SampleType Sample Method Code EU-047-NPW
 Bottle ID A

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPtA	422-64-0	F110424006	4.83	0.497	1.08			
	PFBA	375-22-4	S100424038	ND	0.225	0.568			U
	PFPeA	2706-90-3	S100424038	3.00	0.162	0.568			
	PFFhxA	307-24-4	S100424038	3.35	0.190	0.568			
	PFFHpA	375-85-9	S100424038	1.57	0.199	0.568			
	PFOA	335-67-1	S100424038	4.08	0.130	0.568			
	PFNA	375-95-1	S100424038	0.427	0.128	0.568			J
	PFDA	335-76-2	S100424038	0.168	0.162	0.568			J
	PFUnDA	2058-94-8	S100424038	ND	0.128	0.568			U
	PFDODA	307-55-1	S100424038	ND	0.231	0.568			U
	PFTTrDA	72629-94-8	S100424038	ND	0.188	0.568			U
	PFTeDA	376-06-7	S100424038	ND	0.216	0.568			U
	PFFhxDA	67905-19-5	S100424038	ND	0.302	0.568			U
Sulfonates	PFBS	375-73-5	S100424038	2.38	0.302	0.568			
	PFPeS	2706-91-4	S100424038	0.368	0.117	0.535			J
	PFFhXS	355-46-4	S100424038	1.87	0.438	0.520			
	PFFHpS	375-92-8	S100424038	0.0917	0.275	0.541			L
	PFOS	1763-23-1	S100424038	7.75	0.300	0.526			
	PFNS	68259-12-1	S100424038	ND	0.176	0.547			U
	PFDS	335-77-3	S100424038	ND	0.298	0.547			U
	4:2 FTS	757124-72-4	S100424038	ND	0.0736	0.532			U
	6:2 FTS	27619-97-2	S100424038	0.0666	0.268	0.541			L
	8:2 FTS	39108-34-4	S100424038	ND	0.127	0.544			U
10:2 FTS	120226-60-0	S100424038	ND	0.435	0.568			U	
Sulfonamidos	FBSA	30334-69-1	S100424038	0.229	0.270	0.568			L
	N-EtFOSA	4151-50-2	S100424038	ND	0.351	0.568			U
	N-EtFOSAA	2991-50-6	S100424038	ND	0.231	0.568			U
	N-EtFOSE	1691-99-2	S100424038	ND	0.869	2.55			U
	N-MeFOSA	31506-32-8	S100424038	ND	0.234	0.568			U
	N-MeFOSAA	2355-31-9	S100424038	ND	0.159	0.568			U
	N-MeFOSE	24448-09-7	S100424038	ND	0.539	2.55			U
	PFOSA	754-91-6	S100424038	ND	0.0797	0.568			U
PFECAs	ADONA	919005-14-4	S100424038	ND	0.154	0.538			U
	EVE Acid	69087-46-3	S100424038	ND	0.181	1.28			U
	HFPO-DA	13252-13-6	S100424038	1.22	0.0601	0.568			
	Hydro-EVE Acid	773804-62-9	S100424038	ND	0.186	0.568			U
	NFDHA	151772-58-6	S100424038	ND	0.119	0.568			U
	PEPA	267239-61-2	S100424038	0.421	0.106	0.568			J
	PFECA-G	801212-59-9	S100424038	ND	0.0758	0.568			U
	PFMOAA	674-13-5	S100424038	7.94	0.287	0.568			
	PFMOBA	863090-89-5	S100424038	ND	0.953	1.28			U
	PFMOPrA	377-73-1	S100424038	ND	0.202	0.568			U
	PFO2HxA	39492-88-1	S100424038	ND	0.183	0.568			U
	PFO3OA	39492-89-2	S100424038	ND	0.261	0.568			U
	PFO4DA	39492-90-5	S100424038	ND	0.449	2.84			U
	PFO5DA	39492-91-6	S100424038	ND	0.454	2.84			U
	PMPA	13140-29-9	S100424038	2.40	0.134	0.568			
	R-EVE	2416366-22-6	S100424038	2.06	0.942	1.28			
	PFESAs	11Cl-PF3OUdS	763051-92-9	S100424038	ND	0.268	0.535		
9Cl-PF3ONS		756426-58-1	S100424038	ND	0.364	0.529			U
Hydrolyzed PSDA		2416366-19-1	S100424038	0.588	0.378	0.568			
Nafion Byproduct 1 (PS Acid)		29311-67-9	S100424038	ND	0.303	0.568			U
Nafion Byproduct 2 (Hydro-PS Acid)		749836-20-2	S100424038	0.0654	0.470	0.568			L
NVHOS		1132933-86-8	S100424038	ND	0.0875	0.568			U
PFEESA		113507-82-7	S100424038	ND	0.171	0.568			U
R-PSDA	2416366-18-0	S100424038	1.16	2.50	2.50			L	
R-PSDCA	241636-21-5	S100424038	ND	0.240	0.568			U	
ES	MPFBA		S100424038				20-150%	111.4%	
	M5PFPeA		S100424038				20-150%	147.5%	
	M3PFBS		S100424038				20-150%	179.4%	Q

Enthalpy Analytical

Job No.: 0424-750-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC Northeast Water Plant Leland, NC

Details

Sample Name 040524-E01
 Sampling Site
 Enthalpy ID 0424-750-002-1 Prep Batch EU17218
 Matrix aqueous Analyst rappelle
 Sampling Date 2024-04-05 09:50 Instrument Sauron
 Received Date 2024-04-05 Sample Vol mL 281.83
 Prep Date 2024-04-10 09:00 Extract Vol mL 0.4
 AnalysisDate 2024-04-11 01:01 Split Factor N/A
 SampleType Sample Method Code EU-047-NPW
 Bottle ID A

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M2-4:2 FTS		S100424038				20-150%	156.5%	Q
M5PFHxA		S100424038				20-150%	103.6%	
M3HFPO-DA		S100424038				20-150%	99.1%	
M4PFHpA		S100424038				20-150%	103.0%	
M3PFHxS		S100424038				20-150%	100.7%	
M2-6:2 FTS		S100424038				20-150%	139.7%	
M8PFOA		S100424038				20-150%	118.9%	
M9PFNA		S100424038				20-150%	123.7%	
M8PFOS		S100424038				20-150%	115.6%	
M2-8:2 FTS		S100424038				20-150%	156.6%	Q
M8FOSA-I		S100424038				20-150%	113.9%	
M6PFDA		S100424038				20-150%	117.2%	
d3-N-MeFOSAA		S100424038				20-150%	117.0%	
d5-N-EtFOSAA		S100424038				20-150%	113.0%	
M7PFUdA		S100424038				20-150%	114.8%	
MPFDoA		S100424038				20-150%	107.1%	
M2PFTeDA		S100424038				20-150%	82.3%	
d3-N-MeFOSA		S100424038				10-200%	20.5%	
d5-N-EtFOSA		S100424038				10-200%	21.2%	
d7-N-MeFOSE		S100424038				10-200%	72.4%	
d9-N-EtFOSE		S100424038				10-200%	70.3%	
13C3-PFPtA		F110424006				20-150%	13.6%	Q

QC Data



Enthalpy Analytical

Job No.: 0424-750-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC Northeast Water Plant Leland, NC

Details

Sample Name MB_17218_PFAS
 Sampling Site
 Enthalpy ID MB_17218_PFAS Prep Batch EU17218
 Matrix aqueous Analyst rappelle
 Sampling Date Instrument Sauron
 Received Date Sample Vol mL 250
 Prep Date 2024-04-10 09:00 Extract Vol mL 0.4
 AnalysisDate 2024-04-10 21:59 Split Factor N/A
 SampleType Blank Method Code EU-047-NPW
 Bottle ID -

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	F110424003	ND	0.560	1.22			U
	PFBA	375-22-4	S100424030	ND	0.254	0.640			U
	PFPeA	2706-90-3	S100424030	ND	0.183	0.640			U
	PFHxA	307-24-4	S100424030	ND	0.214	0.640			U
	PFHpA	375-85-9	S100424030	ND	0.224	0.640			U
	PFOA	335-67-1	S100424030	ND	0.146	0.640			U
	PFNA	375-95-1	S100424030	ND	0.145	0.640			U
	PFDA	335-76-2	S100424030	ND	0.183	0.640			U
	PFUnDA	2058-94-8	S100424030	ND	0.145	0.640			U
	PFDODA	307-55-1	S100424030	ND	0.260	0.640			U
	PFTrDA	72629-94-8	S100424030	ND	0.212	0.640			U
	PFTeDA	376-06-7	S100424030	ND	0.244	0.640			U
	PFHxDA	67905-19-5	S100424030	ND	0.340	0.640			U
Sulfonates	PFBS	375-73-5	S100424030	ND	0.340	0.640			U
	PFPeS	2706-91-4	S100424030	ND	0.131	0.603			U
	PFHxS	355-46-4	S100424030	ND	0.494	0.586			U
	PFHpS	375-92-8	S100424030	ND	0.310	0.610			U
	PFOS	1763-23-1	S100424030	ND	0.338	0.593			U
	PFNS	68259-12-1	S100424030	ND	0.199	0.616			U
	PFDS	335-77-3	S100424030	ND	0.336	0.616			U
	4:2 FTS	757124-72-4	S100424030	ND	0.0830	0.600			U
	6:2 FTS	27619-97-2	S100424030	ND	0.302	0.610			U
	8:2 FTS	39108-34-4	S100424030	ND	0.143	0.613			U
10:2 FTS	120226-60-0	S100424030	ND	0.490	0.640			U	
Sulfonamidos	FBSA	30334-69-1	S100424030	ND	0.304	0.640			U
	N-EtFOSA	4151-50-2	S100424030	ND	0.396	0.640			U
	N-EtFOSAA	2991-50-6	S100424030	ND	0.260	0.640			U
	N-EtFOSE	1691-99-2	S100424030	ND	0.980	2.88			U
	N-MeFOSA	31506-32-8	S100424030	ND	0.264	0.640			U
	N-MeFOSAA	2355-31-9	S100424030	ND	0.180	0.640			U
	N-MeFOSE	24448-09-7	S100424030	ND	0.608	2.88			U
	PFOSA	754-91-6	S100424030	ND	0.0898	0.640			U
PFECAs	ADONA	919005-14-4	S100424030	ND	0.173	0.606			U
	EVE Acid	69087-46-3	S100424030	ND	0.204	1.44			U
	HFPO-DA	13252-13-6	S100424030	ND	0.0678	0.640			U
	Hydro-EVE Acid	773804-62-9	S100424030	ND	0.210	0.640			U
	NFDHA	151772-58-6	S100424030	ND	0.135	0.640			U
	PEPA	267239-61-2	S100424030	ND	0.120	0.640			U
	PFECA-G	801212-59-9	S100424030	ND	0.0854	0.640			U
	PFMOAA	674-13-5	S100424030	ND	0.324	0.640			U
	PFMOBA	863090-89-5	S100424030	ND	1.07	1.44			U
	PFMOPrA	377-73-1	S100424030	ND	0.228	0.640			U
	PFO2HxA	39492-88-1	S100424030	ND	0.206	0.640			U
	PFO3OA	39492-89-2	S100424030	ND	0.294	0.640			U
	PFO4DA	39492-90-5	S100424030	ND	0.506	3.20			U
	PFO5DA	39492-91-6	S100424030	ND	0.512	3.20			U
	PMPA	13140-29-9	S100424030	ND	0.151	0.640			U
	R-EVE	2416366-22-6	S100424030	ND	1.06	1.44			U
	PFESAs	11Cl-PF3OUds	763051-92-9	S100424030	ND	0.302	0.603		
9Cl-PF3ONS		756426-58-1	S100424030	ND	0.410	0.596			U
Hydrolyzed PSDA		2416366-19-1	S100424030	ND	0.426	0.640			U
Nafion Byproduct 1 (PS Acid)		29311-67-9	S100424030	ND	0.342	0.640			U
Nafion Byproduct 2 (Hydro-PS Acid)		749836-20-2	S100424030	ND	0.530	0.640			U
NVHOS		1132933-86-8	S100424030	ND	0.0986	0.640			U
PFEEASA		113507-82-7	S100424030	ND	0.192	0.640			U
R-PSDA	2416366-18-0	S100424030	ND	2.82	2.82			U	
R-PSDCA	241636-21-5	S100424030	ND	0.270	0.640			U	
ES	MPFBA		S100424030				20-150%	111.6%	
	M5PFPeA		S100424030				20-150%	108.9%	
	M3PFBS		S100424030				20-150%	95.1%	

Enthalpy Analytical

Job No.: 0424-750-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC Northeast Water Plant Leland, NC

Details

Sample Name MB_17218_PFAS
 Sampling Site
 Enthalpy ID MB_17218_PFAS Prep Batch EU17218
 Matrix aqueous Analyst rappelle
 Sampling Date Instrument Sauron
 Received Date Sample Vol mL 250
 Prep Date 2024-04-10 09:00 Extract Vol mL 0.4
 AnalysisDate 2024-04-10 21:59 Split Factor N/A
 SampleType Blank Method Code EU-047-NPW
 Bottle ID -

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M2-4:2 FTS		S100424030				20-150%	107.0%	
M5PFHxA		S100424030				20-150%	97.6%	
M3HFPO-DA		S100424030				20-150%	88.6%	
M4PFHpA		S100424030				20-150%	96.4%	
M3PFHxS		S100424030				20-150%	105.1%	
M2-6:2 FTS		S100424030				20-150%	133.4%	
M8PFOA		S100424030				20-150%	113.1%	
M9PFNA		S100424030				20-150%	112.2%	
M8PFOS		S100424030				20-150%	109.3%	
M2-8:2 FTS		S100424030				20-150%	154.1%	Q
M8FOSA-I		S100424030				20-150%	105.7%	
M6PFDA		S100424030				20-150%	114.2%	
d3-N-MeFOSAA		S100424030				20-150%	112.9%	
d5-N-EiFOSAA		S100424030				20-150%	107.4%	
M7PFUdA		S100424030				20-150%	112.4%	
MPFDoA		S100424030				20-150%	111.4%	
M2PFTeDA		S100424030				20-150%	73.4%	
d3-N-MeFOSA		S100424030				10-200%	19.1%	
d5-N-EiFOSA		S100424030				10-200%	19.8%	
d7-N-MeFOSE		S100424030				10-200%	71.8%	
d9-N-EiFOSE		S100424030				10-200%	67.0%	
13C3-PFPtA		F110424003				20-150%	17.7%	Q

Enthalpy Analytical

Job No.: 0424-750-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC Northeast Water Plant Leland, NC

Enthalpy ID	OPR_17218_PFAS	Prep Batch	EU17218	Sample Vol (mL)	250
Sample Name	OPR_17218_PFAS	Prep Date	2024-04-10 09:00	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	2024-04-10 22:22	Split Factor	N/A
Sampling Date		Analyst	rappelle	Method Code	EU-047-NPW
Received Date		Instrument	Sauron	Sample Type	Control
		Bottle ID	-		

	Compound	CAS	InjFileName	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	S100424031	16.5	0.254	0.640	69.1-122%	82.3%	
	PFPeA	2706-90-3	S100424031	15.9	0.183	0.640	68.5-121%	79.7%	
	PFHxA	307-24-4	S100424031	16.5	0.214	0.640	68.3-121%	82.6%	
	PFHpA	375-85-9	S100424031	15.9	0.224	0.640	62.4-128%	79.5%	
	PFOA	335-67-1	S100424031	16.0	0.146	0.640	66.3-124%	80.0%	
	PFNA	375-95-1	S100424031	16.0	0.145	0.640	70.5-120%	79.9%	
	PFDA	335-76-2	S100424031	15.5	0.183	0.640	68.9-117%	77.6%	
	PFUnDA	2058-94-8	S100424031	16.0	0.145	0.640	58.1-132%	80.1%	
	PFDoDA	307-55-1	S100424031	16.5	0.260	0.640	52.1-140%	82.4%	
	PFTeDA	72629-94-8	S100424031	23.5	0.212	0.640	65-144%	117.3%	
	PFTeDA	376-06-7	S100424031	15.0	0.244	0.640	36.1-161%	75.2%	
Sulfonates	PFBS	375-73-5	S100424031	15.2	0.340	0.640	67.5-111.6%	85.8%	
	PFPeS	2706-91-4	S100424031	17.0	0.131	0.603	51.8-142%	90.3%	
	PFHxS	355-46-4	S100424031	13.9	0.494	0.586	59.6-128%	75.8%	
	PFHpS	375-92-8	S100424031	15.7	0.310	0.610	46.9-157%	82.3%	
	PFOS	1763-23-1	S100424031	14.7	0.338	0.593	59.2-132%	79.0%	
	PFNS	68259-12-1	S100424031	15.2	0.199	0.616	53.9-133%	78.8%	
	PFDS	335-77-3	S100424031	14.8	0.336	0.616	38.1-142%	76.9%	
	4:2 FTS	757124-72-4	S100424031	14.1	0.0830	0.600	61.9-131%	75.2%	
	6:2 FTS	27619-97-2	S100424031	15.6	0.302	0.610	62.3-129%	82.1%	
	8:2 FTS	39108-34-4	S100424031	15.3	0.143	0.613	37.5-159%	79.5%	
Sulfonamidos	N-EtFOSAA	2991-50-6	S100424031	15.9	0.260	0.640	61.5-133%	79.7%	
	N-MeFOSAA	2355-31-9	S100424031	16.1	0.180	0.640	57.3-138%	80.5%	
	PFOSA	754-91-6	S100424031	13.4	0.0898	0.640	49.1-143%	67.0%	
PFECAs	HFPO-DA	13252-13-6	S100424031	18.4	0.0678	0.640	57.2-130%	92.1%	
ES	MPFBA		S100424031				20-150%	113.3%	
	M5PFPeA		S100424031				20-150%	110.7%	
	M3PFBS		S100424031				20-150%	91.5%	
	M2-4:2 FTS		S100424031				20-150%	111.1%	
	M5PFHxA		S100424031				20-150%	93.6%	
	M3HFPO-DA		S100424031				20-150%	87.7%	
	M4PFHpA		S100424031				20-150%	92.7%	
	M3PFHxS		S100424031				20-150%	104.0%	
	M2-6:2 FTS		S100424031				20-150%	129.6%	
	M8PFOA		S100424031				20-150%	104.1%	
	M9PFNA		S100424031				20-150%	96.2%	
	M8PFOS		S100424031				20-150%	99.9%	
	M2-8:2 FTS		S100424031				20-150%	136.8%	
	M8FOSA-I		S100424031				20-150%	86.1%	
	M6PFDA		S100424031				20-150%	100.1%	
	d3-N-MeFOSAA		S100424031				20-150%	98.7%	
	d5-N-EtFOSAA		S100424031				20-150%	94.8%	
	M7PFUdA		S100424031				20-150%	99.3%	
	MPFDaA		S100424031				20-150%	96.3%	
	M2PFTeDA		S100424031				20-150%	53.3%	

Narrative Summary



Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0424-750-1 PFAS by Isotope Dilution (non-potable water)
Client ID.	Northeast Water Plant Site: Leland, NC

1. Custody

Jack Sullivan received the samples at 8 °C after being relinquished by Brunswick County Public Utilities - NC.

The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC.

Samples received above 6 °C, client notified in order acknowledgment documentation.

Table 1 - Sample Inventory

EU Lab Sample ID	Client Sample ID	Matrix	Received
0424-750-001-1	040524-S01	aqueous	2024-04-05
0424-750-002-1	040524-E01	aqueous	2024-04-05

2. Methods and Analytes

A list of analytes of interest and corresponding methods of analysis is shown in Table 3. Abbreviations are defined in the listed Appendices.

Table 3 - Methods and Analytes

EU Method	Analytes	Cleanup Method
EU-047	Custom List	ENVI-Carb

3. Analysis

The samples were analyzed using Waters Acquity UPLC equipped with Xevo TQ MS (LC/MS/MS "Sauron").

4. Calibration

In the initial calibration, the analytes exhibited R² of ≥ 0.99 . The reported analytes in the calibration standards, continuing calibration (concal) and Initial Calibration Verification (ICV) met the accuracy criterion for native analytes.

5. QC Notes

Except where noted below, the QC sample analyses passed all method criteria.

Select extraction standards (ES) fell outside method recovery criteria in the QC samples. Target analytes are quantified based on their ratio to their labeled standard analogs. When detected at a signal-to-noise above 10:1 the ES peak area is used to quantify its respective target analyte using accepted isotope dilution principles. The data is reported without adverse impact.

The samples were extracted within the 28-day from collection holding time and analyzed within the 28-day from extraction to analysis holding time required by the method.

Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0424-750-1 PFAS by Isotope Dilution (non-potable water)
Client ID.	Northeast Water Plant Site: Leland, NC

6. Reporting Notes

The results presented in this report are representative of the samples as provided to the laboratory.

This report provides all results including detections below LOD following client instruction.

Some labeled extraction standards in the sample analyses fell outside the control limits for ES recovery, as denoted by the "Q" qualifier. The target analytes are quantified based on their ratio to their labeled standard analogs. As a result, low or high labeled standard recovery do not cause any change to ratios or contribute any additional error in the measurement of the target analytes. The data have been accepted and reported with no further actions.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, LLC in Wilmington NC is accredited by the Louisiana Department of Environmental Quality to the 2009 TNI Standard under certificate number 05075.



General Reporting Notes – Data Qualifiers

The following are general reporting notes that are applicable to all Enthalpy Analytical, LLC - Wilmington, NC data reports, unless specifically noted otherwise.

General Data Qualifiers

- Ac - Alternate calculation flag indicates the es recovery was calculated using the opening concal when either of the following situations is encountered in the data processing software: the ES recovery is over 400% or the JS is not detected.
- B – The analyte was found in the method blank, at a concentration that was at least 10% of the amount in the sample.
- Cxx – Two or more congeners co-elute. In EDDs, C denotes the lowest IUPAC congener in a co-elution group and additional co-eluters for the group ('xx') are shown with the number of the lowest IUPAC co-eluter.
- E – The reported concentration exceeds the calibration range (upper point of the calibration curve). For HRMS data, this condition does not imply additional measurement uncertainty. For LC-MS/MS data, these values should be considered as having measurement uncertainty higher than values within the calibration range.
- EDL – Estimated Detection Level: The EDL is unique to isotope dilution methods and reflects the conditions of analysis at the time of analysis, including the equipment used. Where the MDL is a static value, the EDL is a dynamic value.
- EMPC – Estimated Maximum Possible Concentration: EMPC is specific to Dioxin/Furan tests to indicate the determined ion-abundance ratio was outside the allowed theoretical range (usually due to being near the detection limit, although it can very rarely be caused by a co-eluting interference). The EMPC concentration is adjusted to reflect the value at the theoretical ion-abundance ratio.
- I/IR – The ion ratio between the primary and secondary ions was observed to be outside the method criteria. The analyte concentration may be inaccurate due to interference.
- J – The analyte has a concentration below the minimum calibration level (LOQ value) but greater than the LOD. These values should be considered as having measurement uncertainty higher than values within the calibration range
- L - For reports containing PFAS analytes only, this flag indicates that an analyte has a concentration below the Minimum Detection Limit (MDL) . The reported concentration is not recommended for regulatory use as the analyte signal may have a signal-to-noise ratio less than the criteria deemed necessary to be considered a detected analyte.
- LOD – Limit of Detection: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOD. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the minimum detection limit (MDL). The LOD is adjusted for sample weight or volume.

General Reporting Notes – Data Qualifiers

- LOQ – Limit of Quantitation: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOQ. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the reporting limit (RL). The LOQ is adjusted for sample weight or volume.
- <LOD() – Analyte was not found at a concentration high enough to be reported as detected. It is reported as less than the LOD, and the LOD is given in the parentheses.
- <LOQ() – Analyte was not found at a concentration high enough to be reported as above the QSM-defined LOQ or TNI defined Reporting Limit. It is reported as less than the LOQ, and the LOQ is given in the parentheses.
- ND – Indicates a non-detect.
- NR – Indicates a value that is not reportable due to issues observed in sample preparation or analysis.
- PR – The associated congener(s) is(are) poorly resolved.
- QI – Indicates the presence of a quantitative interference.
- RL – Reporting Limit. Lowest reportable value. The level is higher than the MDL.
- SI – Denotes “Single Ion Mode” and is utilized for PCBs where the secondary ion trace has a significantly elevated noise level due to background PFK. Responses for such peaks are calculated using an EMPC approach based solely on the primary ion area(s) and may be considered estimates.
- U – The analyte was not detected.
- V / Q – The labeled standard recovery is not within method control limits.
- X – Indicates the result is from re-injection/repeat/second-column analysis.

Lab Identifiers/ Data Attributes

- AR – Indicates use of the archived portion of the sample extract.
- CU – Indicates a sample that required additional clean-up prior to HRMS injection/processing.
- D – Dilution Data. Result was obtained from the analysis of a dilution. The number that follows the “D” indicates the dilution factor.
- DE – Indicates a dilution performed with the addition of ES (Extraction Standard) solution.
- DUP – Designation for a duplicate sample.
- MS – Designation for a matrix spike.
- MSD – Designation for a matrix spike duplicate.



General Reporting Notes – Data Qualifiers

- R – Indicates a re-extraction of the sample.
- RJ – Indicates a reinjection of the sample extract.
- S – Indicates a sample split. The number that follows the “S” indicates the split factor.
- SAT – Indicates an analyte saturated the detector.

PFAS Compound Acronym List		
Acronym	CAS #	Compound Name
Target Analytes		
* Analyte is not accredited for SOP EU047 # Method 537.1 Accredited ^ Method 533 Accredited		
^ PFBA	375-22-4	Perfluorobutanoic Acid
^ PFPeA	2706-90-3	Perfluoropentanoic Acid
#, ^ PFHxA	307-24-4	Perfluorohexanoic Acid
#, ^ PFHpA	375-85-9	Perfluoroheptanoic Acid
#, ^ PFOA	335-67-1	Perfluorooctanoic Acid
#, ^ PFNA	375-95-1	Perfluorononanoic Acid
#, ^ PFDA	335-76-2	Perfluorodecanoic acid
#, ^ PFUnA (PFUnDA)	2058-94-8	Perfluoroundecanoic acid
#, ^ PFDoA (PFDoDA)	307-55-1	Perfluorododecanoic acid
# PFTrDA (PFTriA)	72629-94-8	Perfluorotridecanoic acid
# PFTeDA (PFTA)	376-06-7	Perfluorotetradecanoic acid
#, ^ PFBS	375-73-5	Perfluorobutane sulfonic acid
^ PFPeS	2706-91-4	Perfluoropentane sulfonic acid
#, ^ PFHxS	355-46-4	Perfluorohexane sulfonic acid
^ PFHpS	375-92-8	Perfluoroheptane sulfonic acid
#, ^ PFOS	1763-23-1	Perfluorooctane sulfonic acid
PFNS	68259-12-1	Perfluorononane sulfonic acid
PFDS	335-77-3	Perfluorodecane sulfonic acid
^ 4:2 FTS	757124-72-4	4:2 fluorotelomer sulfonic acid
^ 6:2 FTS	27619-97-2	6:2 fluorotelomer sulfonic acid
^ 8:2 FTS	39108-34-4	8:2 fluorotelomer sulfonic acid
PFOSA (FOSA)	754-91-6	Perfluorooctane sulfonamide
# N-MeFOSAA	2355-31-9	N-methyl perfluorooctane sulfonamido acetic acid
# N-EtFOSAA	2991-50-6	N-ethyl perfluorooctane sulfonamido acetic acid
#, ^ HFPO-DA	13252-13-6	2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy)-propanoic acid (Gen-X)
*, #, ^ 11Cl-PF3OUdS	763051-92-9	11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid
*, #, ^ 9Cl-PF3ONS	756426-58-1	9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid
*, #, ^ ADONA	919005-14-4	4,8-dioxa-3H-perfluorononanoic acid
*, ^ PFEESA	113507-82-7	Perfluoro(2-ethoxyethane)sulphonic acid
*, ^ PFMOBA (PFMBA)	863090-89-5	Perfluoro-4-methoxybutanic acid
*, ^ NFDHA	151772-58-6	Nonafluoro-3,6-dioxaheptanoic acid
*, ^ PFMOPrA (PFMPA)	377-73-1	Perfluoro-3-methoxypropanoic acid
* PFMOAA	674-13-5	Perfluoro-2-methoxyacetic acid
* PFO2HxA	39492-88-1	Perfluoro (3,5-dioxaheptanoic) acid
* PFO3OA	39492-89-2	Perfluoro (3,5,7-trioxaoctanoic) acid
* PFO4DA	39492-90-5	Perfluoro (3,5,7,9-tetraoxadecanoic) acid
* PFO5DA	39492-91-6	Perfluoro(3,5,7,9,11-pentaoxadodecanoic) acid
* Nafion Byproduct 1	29311-67-9	Nafion Byproduct 1
* Nafion Byproduct 2	749836-20-2	Nafion Byproduct 2
* PEPA	267239-61-2	Perfluoro-2-ethoxypropanoic acid
* PMPA	13140-29-9	Perfluoro-2-methoxypropanoic acid
* 10:2 FTS	120226-60-0	Fluorotelomer sulfonate 10:2
* N-EtFOSA	4151-50-2	N-ethylperfluoro-1-octanesulfonamide
* N-EtFOSE	1691-99-2	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol
* N-MeFOSA	31506-32-8	N-methylperfluoro-1-octanesulfonamide
* N-MeFOSE	24448-09-7	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol
* PFECA-G	801212-59-9	4-(Heptafluoroisopropoxy)hexafluorobutanoic acid
* PFHxDA	67905-19-5	Perfluorohexadecanoic acid
* R-PSDA (Nafion Byproduct 4)	2416366-18-0	Perfluoro-4-(2-sulfoethoxy)pentanoic acid



PFAS Compound Acronym List		
Acronym	CAS #	Compound Name
Target Analytes		
* Analyte is not accredited for SOP EU047 # Method 537.1 Accredited ^ Method 533 Accredited		
* Hydrolyzed PSDA (Nafion Byproduct 5)	2416366-19-1	2-fluoro-2-[1,1,2,3,3,3-hexafluoro-2-(1,1,2,2-tetrafluoro-2-sulfoethoxy)propoxy]-acetic acid
* R-PSDCA (Nafion Byproduct 6)	2416366-21-5	1,1,2,2-tetrafluoro-2-[1,2,2,3,3-pentafluoro-1-(trifluoromethyl)propoxy] ethanesulfonic acid
* EVE Acid	69087-46-3	2,2,3,3-tetrafluoro-3-({1,1,1,2,3,3-hexafluoro-3-[(1,2,2-trifluoroethenyl)oxy]propan-2-yl}oxy)propionic acid
* FBSA	30334-69-1	Perfluorobutylsulfonamide
* Hydro-EVE Acid	773804-62-9	2,2,3,3-Tetrafluoro-3-({1,1,1,2,3,3-hexafluoro-3-(1,2,2,2-tetrafluoroethoxy)propan-2-yl}oxy)propanoic acid
* R-EVE Acid	2416366-22-6	4-(2-carboxy-1,1,2,2-tetrafluoroethoxy)-2,2,3,3,4,5,5,5-octafluoro-pentanoic acid
* NVHOS	1132933-86-8	Perfluoroethoxysulfonic acid
* PFDoS	79780-39-5	Perfluorododecane sulfonic acid
* PFOA	16517-11-6	Perfluorooctadecanoic acid
* 3:3 FTCA	356-02-5	2H,2H,3H,3H-Perfluorohexanoic acid
* 5:3 FTCA	914637-49-3	2H,2H,3H,3H-Perfluorooctanoic acid
* 7:3 FTCA	812-70-4	2H,2H,3H,3H-Perfluorodecanoic acid
* N-AP-FHxSA	50598-28-2	N-(3-(Dimethylamino)propyl)tridecafluoro-1-hexanesulfonamide
* N-CMAmP-6:2 FOSA	34455-29-3	N-(Carboxymethyl)-N,N-dimethyl-3-(((3,3,4,4,5,5,6,6,7,7,8,8,8-tridecafluorooctyl)sulfonyl)amino)1-propanaminium

Sample Custody



0924 - 750



Chain of Custody Record

Enthalpy Ultratrace Job#: _____ COC Page 1 of 1

Special Handling:

- Standard Turn Around Time
- Rush Turn Around Time -- Date Needed _____
- All Fast TATs Subject to Approval by Enthalpy Analytical, Inc.
- All Samples Disposed of After 6 months Unless Otherwise Instructed.

Enthalpy Analytical-Wilmington, NC has added enhancements to standard methods to improve accuracy, precision and permit an assessment of laboratory performance in the context of your specific data needs. For more information email Cindy.James@enthalpy.com.

Client Name: BRUNSWICK COUNTY UTILITIES
 Project Manager: GLENN WALKER
 Report To: SAME

Project Number: _____
 Site Name: NORTHWEST WATER PLANT
 Location: LELAND N.C.

PO#: _____
 Telephone#: _____
 Email: _____

This Chain of Custody is applicable to Non-Air samples. Standard TAT differ per analysis and are provided by request.

Client Special Instructions: _____
 Matrix: GW-Groundwater, WW-Wastewater, NW-Non-Potable Water, DW-Drinking Water, S-Soil, SL-Sludge, BT-Biological Tissue, O-Other
 Type: G=Grab C=Composite Q=Quality Control

Sample ID	Date	Time	Sample Volume	Type	Matrix	Sample Containers				Analyses							Notes:		
						# of Bottles	# of Jars	# of Bags	# Other	Method 1613	Method 8290	Method 1668A/B/C PCE	PFAS by LC/MS/MS	PAHs by HRGC/HRMS	Sample on Hold	Method 23		ALL PFAS	
040524-SO1	4/5/2024	0950AM	250 ml	G	NW	2												X	Please Add PFPrA and
040524-EO1	4/5/2024	0950AM	250 ml	G	DW	2												X	PFHpA To The Testing.
																			Mark Hager Knows About
																			This If you Have Questions

Relinquished By:	Date: <u>4/5/2024</u>	Received By:	Date: <u>4-5-24</u>	Time: <u>11:03</u>	Sample Temperature Upon Receipt: <input checked="" type="checkbox"/> Iced <input type="checkbox"/> Ambient °C <u>8.0</u>
					<input type="checkbox"/> Iced <input type="checkbox"/> Ambient °C _____
					<input type="checkbox"/> Iced <input type="checkbox"/> Ambient °C _____

2714 Exchange Drive • Wilmington, NC 28405 • (910) 212-5858 • www.enthalpy.com

JOB ID: 0424-750

Date / Time: 04/05/24 11:03

Initials: J.T.S

OR

Client: Brunswick County Utilities

Temp °C: 8.0

Thermometer ID: T12

Cooler 1 of 1

Received via

- FedEx
- UPS
- DHL
- USPS
- Courier
- Other

Check one

On ice:

Melted ice:

Ambient:

Check one

in a Box:

in a Cooler:

Cooler in Box:

	Yes	No
Cooler seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Sample seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Good condition:	<input checked="" type="checkbox"/>	<input type="checkbox"/>

Comment:

Temp °C: []

Thermometer ID: []

Cooler [] of []

Received via

- FedEx
- UPS
- DHL
- USPS
- Courier
- Other

Check one

On ice:

Melted ice:

Ambient:

Check one

in a Box:

in a Cooler:

Cooler in Box:

	Yes	No
Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>
Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>
Good condition:	<input type="checkbox"/>	<input type="checkbox"/>

Comment:

Temp °C: []

Thermometer ID: []

Cooler [] of []

Received via

- FedEx
- UPS
- DHL
- USPS
- Courier
- Other

Check one

On ice:

Melted ice:

Ambient:

Check one

in a Box:

in a Cooler:

Cooler in Box:

	Yes	No
Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>
Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>
Good condition:	<input type="checkbox"/>	<input type="checkbox"/>

Comment:

**This Is The Last Page
Of This Report.**

