

# Brunswick County Public Utilities - NC

PO Box 249  
Bolivia, NC 28422-0249

## LELAND, N.C.

Client Project# NORTHWEST WATER PLANT  
Samples Received: 8/9/2024

### Analytical Report 0824-772

#### PFAS by Isotope Dilution (non-potable water)

Report Issue Date: 9/20/2024

I certify that to the best of my knowledge all analytical data presented in this report have been checked for completeness, accuracy, errors and legibility in addition to having been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s). This analytical report was prepared in Portable Document Format (.PDF) and contains 24 pages. This report shall not be reproduced except in full without approval of the laboratory. This will provide assurance that parts of the report are not taken out of context.

Amendment(s):

Signature:



Laura Boivin, QA Associate II



Enthalpy Analytical, LLC – Wilmington  
Brittany Neff, Project Manager  
brneff@montrose-env.com / www.enthalpy.com  
O: (910)876-6894  
2714 Exchange Drive, Wilmington, NC 28405

# Summary of Results

## Enthalpy Analytical

Job No.: 0824-772-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND, N.C.

### Summary

	Compound	CAS	080924-S01 ng/L	080924-E01 ng/L	
Acids	PFBA	375-22-4	3.69	3.27	
	PFPeA	2706-90-3	5.25	6.44	
	PFHxA	307-24-4	3.98	5.83	
	PFHpA	375-85-9	1.69	2.41	
	PFOA	335-67-1	3.21	4.13	
	PFNA	375-95-1	0.433 J	0.606	
	PFDA	335-76-2	0.407 J	0.361 J	
	PFUnDA	2058-94-8	ND U	0.00133 L	
	PFDODA	307-55-1	ND U	ND U	
	PFTrDA	72629-94-8	ND U	ND U	
	PFTeDA	376-06-7	ND U	ND U	
	PFHxDA	67905-19-5	ND U	0.0313 L	
	Sulfonates	PFBS	375-73-5	2.45	3.62
		PFPeS	2706-91-4	0.300 J	0.545
PFHxS		355-46-4	1.51	2.56	
PFHpS		375-92-8	0.0543 L	0.126 L	
PFOS		1763-23-1	8.55	9.69	
PFNS		68259-12-1	ND U	ND U	
PFDS		335-77-3	ND U	ND U	
4:2 FTS		757124-72-4	ND U	ND U	
6:2 FTS		27619-97-2	ND U	0.214 L	
8:2 FTS		39108-34-4	ND U	ND U	
10:2 FTS		120226-60-0	ND U	ND U	
Sulfonamidos		FBSA	30334-69-1	0.247 L	0.442 J
		N-EtFOSA	4151-50-2	NR	ND U
	N-EtFOSAA	2991-50-6	ND U	0.0190 L	
	N-EtFOSE	1691-99-2	ND U	ND U	
	N-MeFOSA	31506-32-8	NR	ND U	
	N-MeFOSAA	2355-31-9	ND U	ND U	
	N-MeFOSE	24448-09-7	ND U	ND U	
	PFOSA	754-91-6	0.162 J	ND U	
	PFECAs	ADONA	919005-14-4	ND U	ND U
EVE Acid		69087-46-3	ND U	ND U	
HFPO-DA		13252-13-6	1.48	1.70	
Hydro-EVE Acid		773804-62-9	0.00584 L	0.0498 L	
NFDHA		151772-58-6	ND U	ND U	
PEPA		267239-61-2	1.14	1.25	
PFECA-G		801212-59-9	ND U	ND U	
PFMOAA		674-13-5	7.48	10.1	
PFMOBA		863090-89-5	ND U	ND U	
PFMOPrA		377-73-1	ND U	ND U	
PFO2HxA		39492-88-1	ND U	3.14	
PFO3OA		39492-89-2	ND U	ND U	
PFO4DA		39492-90-5	ND U	ND U	
PFO5DA		39492-91-6	ND U	ND U	
PMPA		13140-29-9	3.04	4.57	
R-EVE		2416366-22-6	5.22	5.78	
PFESAs		11Cl-PF3OUds	763051-92-9	ND U	ND U
		9Cl-PF3ONS	756426-58-1	ND U	ND U
		Hydrolyzed PSDA	2416366-19-1	0.564	0.915
	Nafion Byproduct 1 (PS Acid)	29311-67-9	ND U	ND U	
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	0.284 L	0.278 L	
	NVHOS	1132933-86-8	ND U	ND U	
	PFEESA	113507-82-7	ND U	ND U	
	R-PSDA	2416366-18-0	2.68	2.64	
R-PSDCA	2416366-21-5	ND U	ND U		

# Detailed Results

# Enthalpy Analytical

Job No.: 0824-772-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND, N.C.

Enthalpy ID	0824-772-001-2	Prep Batch	eu17987	Sample Vol (mL)	290.35
Sample Name	080924-S01	Prep Date	2024-08-19 09:00	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	2024-08-21 21:31	Split Factor	N/A
Sampling Date	2024-08-09 13:00	Analyst	bmay	Method Code	EU-U4 / - NPW
Received Date	2024-08-09	Instrument	Sauron	Sample Type	Sample
		Bottle ID	B		

	Compound	CAS	InjFileName	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	S210824015	3.69	0.219	0.551				
	PFPeA	2706-90-3	S210824015	5.25	0.158	0.551				
	PFHxA	307-24-4	S210824015	3.98	0.184	0.551				
	PFHpA	375-85-9	S210824015	1.69	0.193	0.551				
	PFOA	335-67-1	S210824015	3.21	0.126	0.551				
	PFNA	375-95-1	S210824015	0.433	0.125	0.551			J	
	PFDA	335-76-2	S210824015	0.407	0.158	0.551			J	
	PFUnDA	2058-94-8	S210824015	ND	0.125	0.551			U	
	PFDoDA	307-55-1	S210824015	ND	0.224	0.551			U	
	PFTTrDA	72629-94-8	S210824015	ND	0.183	0.551			U	
	PFTeDA	376-06-7	S210824015	ND	0.210	0.551			U	
	PFHxDA	67905-19-5	S210824015	ND	0.293	0.551			U	
	Sulfonates	PFBS	375-73-5	S210824015	2.45	0.293	0.551			
		PFPeS	2706-91-4	S210824015	0.300	0.113	0.519			J
PFHxS		355-46-4	S210824015	1.51	0.425	0.505				
PFHpS		375-92-8	S210824015	0.0543	0.267	0.525			L	
PFOS		1763-23-1	S210824015	8.55	0.291	0.510				
PFNS		68259-12-1	S210824015	ND	0.171	0.531			U	
PFDS		335-77-3	S210824015	ND	0.289	0.531			U	
4:2 FTS		757124-72-4	S210824015	ND	0.0715	0.516			U	
6:2 FTS		27619-97-2	S210824015	ND	0.260	0.525			U	
8:2 FTS		39108-34-4	S210824015	ND	0.123	0.528			U	
10:2 FTS	120226-60-0	S210824015	ND	0.422	0.551			U		
Sulfonamidos	FBSA	30334-69-1	S210824015	0.247	0.262	0.551			L	
	N-EtFOSA	4151-50-2	S210824015	NR	0.341	0.551				
	N-EtFOSAA	2991-50-6	S210824015	ND	0.224	0.551			U	
	N-EtFOSE	1691-99-2	S210824015	ND	0.844	2.48			U	
	N-MeFOSA	31506-32-8	S210824015	NR	0.227	0.551				
	N-MeFOSAA	2355-31-9	S210824015	ND	0.155	0.551			U	
	N-MeFOSE	24448-09-7	S210824015	ND	0.524	2.48			U	
	PFOSA	754-91-6	P260824021	0.162	0.0773	0.551			J	
PFECAs	ADONA	919005-14-4	S210824015	ND	0.149	0.522			U	
	EVE Acid	69087-46-3	S210824015	ND	0.176	1.24			U	
	HFPO-DA	13252-13-6	S210824015	1.48	0.0584	0.551				
	Hydro-EVE Acid	773804-62-9	S210824015	0.00584	0.181	0.551			L	
	NFDHA	151772-58-6	S210824015	ND	0.116	0.551			U	
	PEPA	267239-61-2	S210824015	1.14	0.103	0.551				
	PFECA-G	801212-59-9	S210824015	ND	0.0735	0.551			U	
	PFMOAA	674-13-5	P260824021	7.48	0.279	0.551				
	PFMOBA	863090-89-5	S210824015	ND	0.925	1.24			U	
	PFMOPrA	377-73-1	S210824015	ND	0.196	0.551			U	
	PFO2HxA	39492-88-1	S210824015	ND	0.177	0.551			U	
	PFO3OA	39492-89-2	S210824015	ND	0.253	0.551			U	
	PFO4DA	39492-90-5	S210824015	ND	0.436	2.76			U	
	PFO5DA	39492-91-6	S210824015	ND	0.441	2.76			U	
	PMPA	13140-29-9	S210824015	3.04	0.130	0.551				
	R-EVE	2416366-22-6	S210824015	5.22	0.914	1.24				
PFESAs	11Cl-PF3OUtS	763051-92-9	S210824015	ND	0.260	0.519			U	
	9Cl-PF3ONS	756426-58-1	S210824015	ND	0.353	0.513			U	
	Hydrolyzed PSDA	2416366-19-1	S210824015	0.564	0.367	0.551				
	Nafion Byproduct 1 (PS Acid)	29311-67-9	S210824015	ND	0.294	0.551			U	
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	S210824015	0.284	0.456	0.551			L	
	NVHOS	1132933-86-8	S210824015	ND	0.0849	0.551			U	
	PFEESA	113507-82-7	S210824015	ND	0.166	0.551			U	
	R-PSDA	2416366-18-0	S210824015	2.68	2.43	2.43				
	R-PSDCA	2416366-21-5	S210824015	ND	0.232	0.551			U	
ES	MPFBA		S210824015				20-150%	88.3%		
	M5PFPeA		S210824015				20-150%	218%	Q	
	M3PFBS		S210824015				20-150%	233%	Q	
	M2-4:2 FTS		S210824015				20-150%	155%	Q	
	M5PFHxA		S210824015				20-150%	118%		
	M3HFPO-DA		S210824015				20-150%	120%		
	M4PFHpA		S210824015				20-150%	114%		

**Enthalpy Analytical**

Job No.: 0824-772-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND, N.C.

ES	M3PFHxS		S210824015			20-150%	119%
	M2-6:2 FTS		S210824015			20-150%	86.7%
	M8PFOA		S210824015			20-150%	82.4%
	M9PFNA		S210824015			20-150%	59.0%
	M8PFOS		S210824015			20-150%	81.3%
	M2-8:2 FTS		S210824015			20-150%	62.5%
	M8FOSA-I		P260824021			20-150%	47.6%
	M6PFDA		S210824015			20-150%	80.0%
	d3-N-MeFOSAA		S210824015			20-150%	55.2%
	d5-N-EtFOSAA		S210824015			20-150%	48.8%
	M7PFUDa		S210824015			20-150%	52.7%
	MPFDoA		S210824015			20-150%	26.4%
	M2PFTeDA		S210824015			20-150%	5.02% Q
	d3-N-MeFOSA		S210824015			10-200%	NR UQ
	d5-N-EtFOSA		S210824015			10-200%	NR UQ
	d7-N-MeFOSE		S210824015			10-200%	8.88% Q
	d9-N-EtFOSE		S210824015			10-200%	5.81% Q

# Enthalpy Analytical

Job No.: 0824-772-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND, N.C.

Enthalpy ID	0824-772-002-2	Prep Batch	eu17987	Sample Vol (mL)	286.08
Sample Name	080924-E01	Prep Date	2024-08-19 09:00	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	2024-08-21 21:54	Split Factor	N/A
Sampling Date	2024-08-09 13:00	Analyst	bmay	Method Code	EU-U4 / - NPW
Received Date	2024-08-09	Instrument	Sauron	Sample Type	Sample
		Bottle ID	B		

	Compound	CAS	InjFileName	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	S210824016	3.27	0.222	0.559				
	PFPeA	2706-90-3	S210824016	6.44	0.160	0.559				
	PFHxA	307-24-4	S210824016	5.83	0.187	0.559				
	PFHpA	375-85-9	S210824016	2.41	0.196	0.559				
	PFOA	335-67-1	S210824016	4.13	0.128	0.559				
	PFNA	375-95-1	S210824016	0.606	0.126	0.559				
	PFDA	335-76-2	S210824016	0.361	0.160	0.559			J	
	PFUnDA	2058-94-8	S210824016	0.00133	0.126	0.559			L	
	PFDoDA	307-55-1	S210824016	ND	0.227	0.559			U	
	PFTTrDA	72629-94-8	S210824016	ND	0.185	0.559			U	
	PFTeDA	376-06-7	S210824016	ND	0.213	0.559			U	
	PFHxDA	67905-19-5	S210824016	0.0313	0.297	0.559			L	
	Sulfonates	PFBS	375-73-5	S210824016	3.62	0.297	0.559			
		PFPeS	2706-91-4	S210824016	0.545	0.115	0.527			
PFHxS		355-46-4	S210824016	2.56	0.432	0.512				
PFHpS		375-92-8	S210824016	0.126	0.271	0.533			L	
PFOS		1763-23-1	S210824016	9.69	0.295	0.518				
PFNS		68259-12-1	S210824016	ND	0.174	0.539			U	
PFDS		335-77-3	S210824016	ND	0.294	0.539			U	
4:2 FTS		757124-72-4	S210824016	ND	0.0725	0.524			U	
6:2 FTS		27619-97-2	S210824016	0.214	0.264	0.533			L	
8:2 FTS		39108-34-4	S210824016	ND	0.125	0.536			U	
10:2 FTS	120226-60-0	S210824016	ND	0.428	0.559			U		
Sulfonamidos	FBSA	30334-69-1	S210824016	0.442	0.266	0.559			J	
	N-EtFOSA	4151-50-2	S210824016	ND	0.346	0.559			U	
	N-EtFOSAA	2991-50-6	S210824016	0.0190	0.227	0.559			L	
	N-EtFOSE	1691-99-2	S210824016	ND	0.856	2.52			U	
	N-MeFOSA	31506-32-8	S210824016	ND	0.231	0.559			U	
	N-MeFOSAA	2355-31-9	S210824016	ND	0.157	0.559			U	
	N-MeFOSE	24448-09-7	S210824016	ND	0.531	2.52			U	
	PFOSA	754-91-6	P260824022	ND	0.0785	0.559			U	
PFECAs	ADONA	919005-14-4	S210824016	ND	0.152	0.530			U	
	EVE Acid	69087-46-3	S210824016	ND	0.178	1.26			U	
	HFPO-DA	13252-13-6	S210824016	1.70	0.0592	0.559				
	Hydro-EVE Acid	773804-62-9	S210824016	0.0498	0.184	0.559			L	
	NFDHA	151772-58-6	S210824016	ND	0.118	0.559			U	
	PEPA	267239-61-2	S210824016	1.25	0.105	0.559				
	PFECA-G	801212-59-9	S210824016	ND	0.0746	0.559			U	
	PFMOAA	674-13-5	P260824022	10.1	0.283	0.559				
	PFMOBA	863090-89-5	S210824016	ND	0.939	1.26			U	
	PFMOPrA	377-73-1	S210824016	ND	0.199	0.559			U	
	PFO2HxA	39492-88-1	S210824016	3.14	0.180	0.559				
	PFO3OA	39492-89-2	S210824016	ND	0.257	0.559			U	
	PFO4DA	39492-90-5	S210824016	ND	0.442	2.80			U	
	PFO5DA	39492-91-6	S210824016	ND	0.447	2.80			U	
	PMPA	13140-29-9	S210824016	4.57	0.132	0.559				
	R-EVE	2416366-22-6	S210824016	5.78	0.928	1.26				
PFESAs	11Cl-PF3OUtS	763051-92-9	S210824016	ND	0.264	0.527			U	
	9Cl-PF3ONS	756426-58-1	S210824016	ND	0.358	0.521			U	
	Hydrolyzed PSDA	2416366-19-1	S210824016	0.915	0.372	0.559				
	Nafion Byproduct 1 (PS Acid)	29311-67-9	S210824016	ND	0.299	0.559			U	
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	S210824016	0.278	0.463	0.559			L	
	NVHOS	1132933-86-8	S210824016	ND	0.0862	0.559			U	
	PFEESA	113507-82-7	S210824016	ND	0.168	0.559			U	
	R-PSDA	2416366-18-0	S210824016	2.64	2.46	2.46				
	R-PSDCA	2416366-21-5	S210824016	ND	0.236	0.559			U	
ES	MPFBA		S210824016				20-150%	89.8%		
	M5PFPeA		S210824016				20-150%	199%	Q	
	M3PFBS		S210824016				20-150%	193%	Q	
	M2-4:2 FTS		S210824016				20-150%	79.9%		
	M5PFHxA		S210824016				20-150%	80.9%		
	M3HFPO-DA		S210824016				20-150%	81.2%		
	M4PFHpA		S210824016				20-150%	81.2%		

## Enthalpy Analytical

Job No.: 0824-772-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND, N.C.

ES	M3PFHxS		S210824016			20-150%	70.4%
	M2-6:2 FTS		S210824016			20-150%	54.6%
	M8PFOA		S210824016			20-150%	77.4%
	M9PFNA		S210824016			20-150%	64.3%
	M8PFOS		S210824016			20-150%	72.4%
	M2-8:2 FTS		S210824016			20-150%	59.6%
	M8FOSA-I		P260824022			20-150%	41.3%
	M6PFDA		S210824016			20-150%	70.2%
	d3-N-MeFOSAA		S210824016			20-150%	56.3%
	d5-N-EtFOSAA		S210824016			20-150%	56.9%
	M7PFUdA		S210824016			20-150%	58.6%
	MPFDoA		S210824016			20-150%	49.5%
	M2PFTeDA		S210824016			20-150%	41.3%
	d3-N-MeFOSA		S210824016			10-200%	1.41% Q
	d5-N-EtFOSA		S210824016			10-200%	1.32% Q
	d7-N-MeFOSE		S210824016			10-200%	20.5%
	d9-N-EtFOSE		S210824016			10-200%	17.1%

# QC Data

# Enthalpy Analytical

Job No.: 0824-772-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND, N.C.

Enthalpy ID	MB_17987_PFAS	Prep Batch	eu17987	Sample Vol (mL)	250
Sample Name	MB_17987_PFAS	Prep Date	2024-08-19 09:00	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	2024-08-21 20:46	Split Factor	N/A
Sampling Date		Analyst	bmay	Method Code	EU-U4/- NPW
Received Date		Instrument	Sauron	Sample Type	Blank
		Bottle ID	-		

	Compound	CAS	InjFileName	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	S210824013	ND	0.254	0.640			U	
	PFPeA	2706-90-3	S210824013	ND	0.183	0.640			U	
	PFHxA	307-24-4	S210824013	ND	0.214	0.640			U	
	PFHpA	375-85-9	S210824013	ND	0.224	0.640			U	
	PFOA	335-67-1	S210824013	ND	0.146	0.640			U	
	PFNA	375-95-1	S210824013	ND	0.145	0.640			U	
	PFDA	335-76-2	S210824013	ND	0.183	0.640			U	
	PFUnDA	2058-94-8	S210824013	ND	0.145	0.640			U	
	PFDoDA	307-55-1	S210824013	ND	0.260	0.640			U	
	PFTriDA	72629-94-8	S210824013	ND	0.212	0.640			U	
	PFTeDA	376-06-7	S210824013	ND	0.244	0.640			U	
	PFHxDA	67905-19-5	S210824013	ND	0.340	0.640			U	
	Sulfonates	PFBS	375-73-5	S210824013	ND	0.340	0.640			U
		PFPeS	2706-91-4	S210824013	ND	0.131	0.603			U
		PFHxS	355-46-4	S210824013	ND	0.494	0.586			U
PFHpS		375-92-8	S210824013	ND	0.310	0.610			U	
PFOS		1763-23-1	S210824013	ND	0.338	0.593			U	
PFNS		68259-12-1	S210824013	ND	0.199	0.616			U	
PFDS		335-77-3	S210824013	ND	0.336	0.616			U	
4:2 FTS		757124-72-4	S210824013	ND	0.0830	0.600			U	
6:2 FTS		27619-97-2	S210824013	ND	0.302	0.610			U	
8:2 FTS		39108-34-4	S210824013	ND	0.143	0.613			U	
10:2 FTS	120226-60-0	S210824013	ND	0.490	0.640			U		
Sulfonamidos	FBSA	30334-69-1	S210824013	ND	0.304	0.640			U	
	N-EtFOSA	4151-50-2	S210824013	ND	0.396	0.640			U	
	N-EtFOSAA	2991-50-6	S210824013	ND	0.260	0.640			U	
	N-EtFOSE	1691-99-2	S210824013	ND	0.980	2.88			U	
	N-MeFOSA	31506-32-8	S210824013	ND	0.264	0.640			U	
	N-MeFOSAA	2355-31-9	S210824013	ND	0.180	0.640			U	
	N-MeFOSE	24448-09-7	S210824013	ND	0.608	2.88			U	
	PFOSA	754-91-6	P260824019	ND	0.0898	0.640			U	
PFECAs	ADONA	919005-14-4	S210824013	ND	0.173	0.606			U	
	EVE Acid	69087-46-3	S210824013	ND	0.204	1.44			U	
	HFPO-DA	13252-13-6	S210824013	ND	0.0678	0.640			U	
	Hydro-EVE Acid	773804-62-9	S210824013	ND	0.210	0.640			U	
	NFDHA	151772-58-6	S210824013	ND	0.135	0.640			U	
	PEPA	267239-61-2	S210824013	ND	0.120	0.640			U	
	PFECA-G	801212-59-9	S210824013	ND	0.0854	0.640			U	
	PFMOAA	674-13-5	P260824019	ND	0.324	0.640			U	
	PFMOBA	863090-89-5	S210824013	ND	1.07	1.44			U	
	PFMOPrA	377-73-1	S210824013	ND	0.228	0.640			U	
	PFO2HxA	39492-88-1	S210824013	ND	0.206	0.640			U	
	PFO3OA	39492-89-2	S210824013	ND	0.294	0.640			U	
	PFO4DA	39492-90-5	S210824013	ND	0.506	3.20			U	
	PFOSDA	39492-91-6	S210824013	ND	0.512	3.20			U	
	PMPA	13140-29-9	S210824013	ND	0.151	0.640			U	
	R-EVE	2416366-22-6	S210824013	ND	1.06	1.44			U	
	PFESAs	11Cl-PF3OUtS	763051-92-9	S210824013	ND	0.302	0.603			U
		9Cl-PF3ONS	756426-58-1	S210824013	ND	0.410	0.596			U
Hydrolyzed PSDA		2416366-19-1	S210824013	ND	0.426	0.640			U	
Nafion Byproduct 1 (PS Acid)		29311-67-9	S210824013	ND	0.342	0.640			U	
Nafion Byproduct 2 (Hydro-PS Acid)		749836-20-2	S210824013	ND	0.530	0.640			U	
NVHOS		1132933-86-8	S210824013	ND	0.0986	0.640			U	
PFEESA		113507-82-7	S210824013	ND	0.192	0.640			U	
R-PSDA		2416366-18-0	S210824013	ND	2.82	2.82			U	
R-PSDCA		2416366-21-5	S210824013	ND	0.270	0.640			U	
ES	MPFBA		S210824013				20-150%	87.5%		
	M5PFPeA		S210824013				20-150%	106%		
	M3PFBS		S210824013				20-150%	75.6%		
	M2-4:2 FTS		S210824013				20-150%	74.1%		
	M5PFHxA		S210824013				20-150%	85.1%		
	M3HFPO-DA		S210824013				20-150%	86.1%		
	M4PFHpA		S210824013				20-150%	86.8%		

### Enthalpy Analytical

Job No.: 0824-772-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND, N.C.

ES	M3PFHxS		S210824013			20-150%	77.0%
	M2-6:2 FTS		S210824013			20-150%	66.2%
	M8PFOA		S210824013			20-150%	79.5%
	M9PFNA		S210824013			20-150%	61.1%
	M8PFOS		S210824013			20-150%	69.5%
	M2-8:2 FTS		S210824013			20-150%	54.5%
	M8FOSA-I		P260824019			20-150%	44.4%
	M6PFDA		S210824013			20-150%	66.5%
	d3-N-MeFOSAA		S210824013			20-150%	52.3%
	d5-N-EtFOSAA		S210824013			20-150%	50.8%
	M7PFUdA		S210824013			20-150%	55.5%
	MPFDoA		S210824013			20-150%	46.8%
	M2PFTeDA		S210824013			20-150%	21.3%
	d3-N-MeFOSA		S210824013			10-200%	1.42% Q
	d5-N-EtFOSA		S210824013			10-200%	1.50% Q
	d7-N-MeFOSE		S210824013			10-200%	35.4%
	d9-N-EtFOSE		S210824013			10-200%	29.3%

# Enthalpy Analytical

Job No.: 0824-772-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND, N.C.

Enthalpy ID	OPR_17987_PFAS	Prep Batch	eu17987	Sample Vol (mL)	250
Sample Name	OPR_17987_PFAS	Prep Date	2024-08-19 09:00	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	2024-08-21 21:08	Split Factor	N/A
Sampling Date		Analyst	bmay	Method Code	EU-047-NPW
Received Date		Instrument	Sauron	Sample Type	Control
		Bottle ID	-		

	Compound	CAS	InjFileName	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	S210824014	18.9	0.254	0.640	69.1-122%	94.4%	
	PFPeA	2706-90-3	S210824014	18.9	0.183	0.640	68.5-121%	94.3%	
	PFHxA	307-24-4	S210824014	20.9	0.214	0.640	68.3-121%	104%	
	PFHpA	375-85-9	S210824014	18.5	0.224	0.640	62.4-128%	92.4%	
	PFOA	335-67-1	S210824014	19.3	0.146	0.640	66.3-124%	96.7%	
	PFNA	375-95-1	S210824014	20.2	0.145	0.640	70.5-120%	101%	
	PFDA	335-76-2	S210824014	17.3	0.183	0.640	68.9-117%	86.4%	
	PFUnDA	2058-94-8	S210824014	19.6	0.145	0.640	58.1-132%	98.0%	
	PFDoDA	307-55-1	S210824014	18.3	0.260	0.640	52.1-140%	91.7%	
	PFTrDA	72629-94-8	S210824014	32.5	0.212	0.640	65-144%	162%	Q
	PFTeDA	376-06-7	S210824014	16.1	0.244	0.640	36.1-161%	80.7%	
Sulfonates	PFBS	375-73-5	S210824014	17.3	0.340	0.640	67.5-111.6%	97.8%	
	PFPeS	2706-91-4	S210824014	20.8	0.131	0.603	51.8-142%	111%	
	PFHxS	355-46-4	S210824014	16.2	0.494	0.586	59.6-128%	88.8%	
	PFHpS	375-92-8	S210824014	20.2	0.310	0.610	46.9-157%	106%	
	PFOS	1763-23-1	S210824014	16.5	0.338	0.593	59.2-132%	89.1%	
	PFNS	68259-12-1	S210824014	16.0	0.199	0.616	53.9-133%	83.4%	
	PFDS	335-77-3	S210824014	14.1	0.336	0.616	38.1-142%	73.0%	
	4:2 FTS	757124-72-4	S210824014	17.3	0.0830	0.600	61.9-131%	92.1%	
	6:2 FTS	27619-97-2	S210824014	18.9	0.302	0.610	62.3-129%	99.2%	
8:2 FTS	39108-34-4	S210824014	18.5	0.143	0.613	37.5-159%	96.6%		
Sulfonamidos	N-EtFOSAA	2991-50-6	S210824014	19.6	0.260	0.640	61.5-133%	98.0%	
	N-MeFOSAA	2355-31-9	S210824014	20.3	0.180	0.640	57.3-138%	101%	
	PFOSA	754-91-6	P260824020	19.2	0.0898	0.640	49.1-143%	95.9%	
PFECAs	HFPO-DA	13252-13-6	S210824014	16.8	0.0678	0.640	57.2-130%	83.8%	
ES	MPFBA		S210824014				20-150%	87.2%	
	M5PFPeA		S210824014				20-150%	114%	
	M3PFBS		S210824014				20-150%	78.5%	
	M2-4:2 FTS		S210824014				20-150%	84.6%	
	M5PFHxA		S210824014				20-150%	91.8%	
	M3HFPO-DA		S210824014				20-150%	91.1%	
	M4PFHpA		S210824014				20-150%	88.5%	
	M3PFHxS		S210824014				20-150%	89.6%	
	M2-6:2 FTS		S210824014				20-150%	70.4%	
	M8PFOA		S210824014				20-150%	75.8%	
	M9PFNA		S210824014				20-150%	60.3%	
	M8PFOS		S210824014				20-150%	72.6%	
	M2-8:2 FTS		S210824014				20-150%	63.2%	
	M8FOSA-I		P260824020				20-150%	50.4%	
	M6PFDA		S210824014				20-150%	73.1%	
	d3-N-MeFOSAA		S210824014				20-150%	58.3%	
	d5-N-EtFOSAA		S210824014				20-150%	55.1%	
	M7PFUDa		S210824014				20-150%	62.2%	
	MPFDa		S210824014				20-150%	53.5%	
	M2PFTeDA		S210824014				20-150%	24.4%	

# Narrative Summary

# Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0824-772
Client ID.	NORTHWEST WATER PLANT Site: LELAND, N.C.

## 1. Custody

Jayson-Shane Santos received the samples at 9.7 °C after being relinquished by Brunswick County Public Utilities - NC.

The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC.

**Table 1 - Sample Inventory**

EU Lab Sample ID	Client Sample ID	Matrix	Received
0824-772-001-2	080924-S01	aqueous	2024-08-09
0824-772-002-2	080924-E01	aqueous	2024-08-09

## 2. Methods and Analytes

A list of analytes of interest and corresponding methods of analysis is shown in Table 3. Abbreviations are defined in the listed Appendices.

**Table 3 - Methods and Analytes**

EU Method	Analytes	Cleanup Method
EU-047	Brunswick List	ENVI-Carb

## 3. Analysis

The samples were analyzed using Waters Acquity UPLC equipped with Xevo TQ MS (LC/MS/MS "Sauron").

The samples were analyzed using more than one extraction batch and analytical sequence to meet method acceptance criteria.

PFPrA did not meet method criteria in the initial batch method blank and ongoing precision recovery (OPR) QC samples. The samples were re-extracted with results that did not meet method criteria in the method blank. Client approved reporting data without this analyte of interest.

## 4. Calibration

In the initial calibration, the reported analytes exhibited  $R^2$  of  $\geq 0.99$ . The reported analytes in the calibration standards, Initial Calibration Verification (ICV) and continuing calibration (concal) met the accuracy criterion for native analytes.

# Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0824-772
Client ID.	NORTHWEST WATER PLANT Site: LELAND, N.C.

## 5. QC Notes

Ongoing Precision Recovery (OPR) control limits have not been established for some analytes of interest.

Except where noted below, the QC sample analyses passed all method criteria.

QC samples that did not meet method acceptance criteria were:

- OPR\_17987\_PFAS (PFTTrDA) exceeds method recovery limits but are not detected >LOQ in the samples: therefore, the data is reportable without adverse impact.

Select surrogates (ES) deviated from method recovery criteria in the method blank (MB) and/or OPR. Target analytes are quantified based on their ratio to their labeled standard analogs. When detected at a signal-to-noise above 10:1 the ES peak area is used to quantify its respective target analyte using accepted isotope dilution principles. The data is reported without adverse impact.

PFAS by Isotope Dilution (non-potable water) samples were extracted within 28 days, and extracts analyzed within 28 days.

## 6. Reporting Notes

The results presented in this report are representative of the samples as provided to the laboratory.

This report provides all results including detections below LOD following client instruction.

Surrogates (ES) d3-N-MeFOSA and d5-N-EtFOSA were not detected (ND) in the initial analysis of sample 080924-S01 extract. Re-extracted results confirmed the ND. Client is aware that these ES and their related analytes N-MeFOSA and N-EtFOSA are not reportable (NR) in sample 080924-S01.

Some labeled extraction standards (ES) in the sample analyses recovered outside the control limits for ES recovery, as denoted by the "Q" qualifier. The target analytes are quantified based on their ratio to their labeled standard analogs. As a result, low or high labeled standard recovery do not cause any change to ratios or contribute any additional error in the measurement of the target analytes. When detected at a signal-to-noise above 10:1 the ES peak area is used to quantify its respective target analyte using accepted isotope dilution principles. The data is reported without adverse impact.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, LLC in Wilmington NC is accredited by the Louisiana Department of Environmental Quality to the 2009 TNI Standard under certificate number 05075.

## General Reporting Notes – Data Qualifiers

The following are general reporting notes that are applicable to all Enthalpy Analytical, LLC - Wilmington, NC data reports, unless specifically noted otherwise.

### General Data Qualifiers

- Ac - Alternate calculation flag indicates the es recovery was calculated using the opening concal when either of the following situations is encountered in the data processing software: the ES recovery is over 400% or the JS is not detected.
- B – The analyte was found in the method blank, at a concentration that was at least 10% of the amount in the sample.
- Cxx – Two or more congeners co-elute. In EDDs, C denotes the lowest IUPAC congener in a co-elution group and additional co-eluters for the group ('xx') are shown with the number of the lowest IUPAC co-eluter.
- E – The reported concentration exceeds the calibration range (upper point of the calibration curve). For HRMS data, this condition does not imply additional measurement uncertainty. For LC-MS/MS data, these values should be considered as having measurement uncertainty higher than values within the calibration range.
- EDL – Estimated Detection Level: The EDL is unique to isotope dilution methods and reflects the conditions of analysis at the time of analysis, including the equipment used. Where the MDL is a static value, the EDL is a dynamic value.
- EMPC – Estimated Maximum Possible Concentration: EMPC is specific to Dioxin/Furan tests to indicate the determined ion-abundance ratio was outside the allowed theoretical range (usually due to being near the detection limit, although it can very rarely be caused by a co-eluting interference). The EMPC concentration is adjusted to reflect the value at the theoretical ion-abundance ratio.
- I/IR – The ion ratio between the primary and secondary ions was observed to be outside the method criteria. The analyte concentration may be inaccurate due to interference.
- J – The analyte has a concentration below the minimum calibration level (LOQ value) but greater than the LOD. These values should be considered as having measurement uncertainty higher than values within the calibration range
- L - For reports containing PFAS analytes only, this flag indicates that an analyte has a concentration below the Minimum Detection Limit (MDL) . The reported concentration is not recommended for regulatory use as the analyte signal may have a signal-to-noise ratio less than the criteria deemed necessary to be considered a detected analyte.
- LOD – Limit of Detection: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOD. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the minimum detection limit (MDL). The LOD is adjusted for sample weight or volume.

## General Reporting Notes – Data Qualifiers

- LOQ – Limit of Quantitation: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOQ. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the reporting limit (RL). The LOQ is adjusted for sample weight or volume.
- <LOD() – Analyte was not found at a concentration high enough to be reported as detected. It is reported as less than the LOD, and the LOD is given in the parentheses.
- <LOQ() – Analyte was not found at a concentration high enough to be reported as above the QSM-defined LOQ or TNI defined Reporting Limit. It is reported as less than the LOQ, and the LOQ is given in the parentheses.
- ND – Indicates a non-detect.
- NR – Indicates a value that is not reportable due to issues observed in sample preparation or analysis.
- PR – The associated congener(s) is(are) poorly resolved.
- QI – Indicates the presence of a quantitative interference.
- RL – Reporting Limit. Lowest reportable value. The level is higher than the MDL.
- SI – Denotes “Single Ion Mode” and is utilized for PCBs where the secondary ion trace has a significantly elevated noise level due to background PFK. Responses for such peaks are calculated using an EMPC approach based solely on the primary ion area(s) and may be considered estimates.
- U – The analyte was not detected.
- V / Q – The labeled standard recovery is not within method control limits.
- X – Indicates the result is from re-injection/repeat/second-column analysis.

### **Lab Identifiers/ Data Attributes**

- AR – Indicates use of the archived portion of the sample extract.
- CU – Indicates a sample that required additional clean-up prior to HRMS injection/processing.
- D – Dilution Data. Result was obtained from the analysis of a dilution. The number that follows the “D” indicates the dilution factor.
- DE – Indicates a dilution performed with the addition of ES (Extraction Standard) solution.
- DUP – Designation for a duplicate sample.
- MS – Designation for a matrix spike.
- MSD – Designation for a matrix spike duplicate.



## General Reporting Notes – Data Qualifiers

- R – Indicates a re-extraction of the sample.
- RJ – Indicates a reinjection of the sample extract.
- S – Indicates a sample split. The number that follows the “S” indicates the split factor.
- SAT – Indicates an analyte saturated the detector.

PFAS Compound Acronym List			
Acronym	CAS #	Compound Name	
* accredited for SOP EU047 / EPA method 1633 # Method 537.1 Accredited ^ Method 533 Accredited ~EPA 1633 extended list			
<b>Target Analytes</b>			
* , ^	PFBA	375-22-4	Perfluorobutanoic Acid
* , # , ^	PFPeA	2706-90-3	Perfluoropentanoic Acid
* , # , ^	PFHxA	307-24-4	Perfluorohexanoic Acid
* , # , ^	PFHpA	375-85-9	Perfluoroheptanoic Acid
* , # , ^	PFOA	335-67-1	Perfluorooctanoic Acid
* , # , ^	PFNA	375-95-1	Perfluorononanoic Acid
* , # , ^	PFDA	335-76-2	Perfluorodecanoic acid
* , # , ^	PFUnA (PFUnDA)	2058-94-8	Perfluoroundecanoic acid
* , #	PFDoA (PFDoDA)	307-55-1	Perfluorododecanoic acid
* , #	PFTriA (PFTriA)	72629-94-8	Perfluorotridecanoic acid
* , # , ^	PFTeDA (PFTA)	376-06-7	Perfluorotetradecanoic acid
* , ^	PFBS	375-73-5	Perfluorobutane sulfonic acid
* , # , ^	PFPeS	2706-91-4	Perfluoropentane sulfonic acid
* , ^	PFHxS	355-46-4	Perfluorohexane sulfonic acid
* , # , ^	PFHpS	375-92-8	Perfluoroheptane sulfonic acid
* , # , ^	PFOS	1763-23-1	Perfluorooctane sulfonic acid
* , ^	PFNS	68259-12-1	Perfluorononane sulfonic acid
* , ^	PFDS	335-77-3	Perfluorodecane sulfonic acid
* , ^	4:2 FTS	757124-72-4	4:2 fluorotelomer sulfonic acid
* , ^	6:2 FTS	27619-97-2	6:2 fluorotelomer sulfonic acid
* , ^	8:2 FTS	39108-34-4	8:2 fluorotelomer sulfonic acid
~	10:2 FTS	120226-60-0	Fluorotelomer sulfonate 10:2
~	FHxSA	41997-13-1	Perfluorohexanesulfonamide
* , #	PFOSA (FOSA)	754-91-6	Perfluorooctane sulfonamide
* , #	N-MeFOSAA	2355-31-9	N-methyl perfluorooctane sulfonamido acetic acid
* , #	N-MeFOSA	31506-32-8	N-methylperfluoro-1-octanesulfonamide
* , #	N-MeFOSE	24448-09-7	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol
* , #	N-EtFOSAA	2991-50-6	N-ethyl perfluorooctane sulfonamido acetic acid
* , #	N-EtFOSA	4151-50-2	N-ethylperfluoro-1-octanesulfonamide
* , #	N-EtFOSE	1691-99-2	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol
* , # , ^	HFPO-DA	13252-13-6	2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy)-propanoic acid (Gen-X)
* , # , ^	11Cl-PF3OUdS	763051-92-9	11-chloroheptafluoro-3-oxadecane-1-sulfonic acid
* , # , ^	9Cl-PF3ONS	756426-58-1	9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid
* , # , ^	ADONA	919005-14-4	4,8-dioxa-3H-perfluorononanoic acid
* , ^	PFEESA	113507-82-7	Perfluoro(2-ethoxyethane)sulphonic acid
* , ^	PFMOBA (PFMBA)	863090-89-5	Perfluoro-4-methoxybutanoic acid
* , ^	NFDHA	151772-58-6	Nonafluoro-3,6-dioxahexanoic acid
* , ^	PFMOPrA (PFMPA)	377-73-1	Perfluoro-3-methoxypropanoic acid
~	PFPrA	422-64-0	2,2,3,3,3-Pentafluoropropionic acid
~	PFPrS (PFPS)	423-41-6	Perfluoropropanesulfonic acid
~	PFMOAA	674-13-5	Perfluoro-2-methoxyacetic acid
~	PFO2HxA	39492-88-1	Perfluoro (3,5-dioxahexanoic) acid
~	PFO3OA	39492-89-2	Perfluoro (3,5,7-trioxaoctanoic) acid
~	PFO4DA	39492-90-5	Perfluoro (3,5,7,9-tetraoxadecanoic) acid
~	PFO5DA	39492-91-6	Perfluoro(3,5,7,9,11-pentaoxadodecanoic) acid
~	Nafion Byproduct 1 (PS Acid)	29311-67-9	Nafion Byproduct 1
~	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	Nafion Byproduct 2
~	PEPA	267239-61-2	Perfluoro-2-ethoxypropanoic acid
~	PMPA	13140-29-9	Perfluoro-2-methoxypropanoic acid

PFAS Compound Acronym List		
Acronym	CAS #	Compound Name
<b>* accredited for SOP EU047 / EPA method 1633 # Method 537.1 Accredited ^ Method 533 Accredited ~EPA 1633 extended list</b>		
~ PFECA-G	801212-59-9	4-(Heptafluoroisopropoxy)hexafluorobutanoic acid
~ PFHxDA	67905-19-5	Perfluorohexadecanoic acid
~ R-PSDA (Nafion Byproduct 4)	2416366-18-0	Perfluoro-4-(2-sulfoethoxy)pentanoic acid
Hydrolyzed PSDA (Nafion Byproduct 5)	2416366-19-1	2-fluoro-2-[1,1,2,3,3,3-hexafluoro-2-(1,1,2-tetrafluoro-2-sulfoethoxy)propoxy]-acetic acid
~ R-PSDCA (Nafion Byproduct 6)	2416366-21-5	1,1,2,2-tetrafluoro-2-[1,2,2,3,3-pentafluoro-1-(trifluoromethyl)propoxy] ethanesulfonic acid
~ EVE Acid	69087-46-3	2,2,3,3-tetrafluoro-3-({1,1,1,2,3,3-hexafluoro-3-[(1,2,2-trifluoroethenyl)oxy]propan-2-yl)oxy}propionic acid
~ FBSA	30334-69-1	Perfluorobutylsulfonamide
~ MeFBSA	68298-12-4	1-Butanesulfonamide; (N-(Methyl)nonafluorobutanesulfonamide)
~ Hydro-EVE Acid	773804-62-9	2,2,3,3-Tetrafluoro-3-[[1,1,1,2,3,3-hexafluoro-3-(1,2,2,2-tetrafluoroethoxy)propan-2-yl]oxy}propanoic acid
~ R-EVE Acid	2416366-22-6	4-(2-carboxy-1,1,2,2-tetrafluoroethoxy)-2,2,3,3,4,5,5,5-octafluoro-pentanoic acid
~ NVHOS	1132933-86-8	Perfluoroethoxysulfonic acid
*~ PFDoS	79780-39-5	Perfluorododecane sulfonic acid
~ PFODA	16517-11-6	Perfluorooctadecanoic acid
* 3:3 FTCA	356-02-5	2H,2H,3H,3H-Perfluorohexanoic acid
* 5:3 FTCA	914637-49-3	2H,2H,3H,3H-Perfluorooctanoic acid
* 7:3 FTCA	812-70-4	2H,2H,3H,3H-Perfluorodecanoic acid
~ N-AP-FHxSA	50598-28-2	N-(3-(Dimethylamino)propyl)tridecafluoro-1-hexanesulfonamide
~ N-CMAmP-6:2 FOSA	34455-29-3	N-(Carboxymethyl)-N,N-dimethyl-3-(((3,3,4,4,5,5,6,6,7,7,8,8,8-tridecafluorooctyl)sulfonyl)amino)1-propanaminium
~ BPAF	1478-61-1	Bisphenol AF
~ HQ-115	90076-65-6	Bis(trifluoromethane)sulfonimide lithium salt

# Sample Custody

0824-772



# Chain of Custody Record

Enthalpy Ultratrace Job#: \_\_\_\_\_ COC Page 1 of 1

**Special Handling:**

- Standard Turn Around Time
- Rush Turn Around Time -- Date Needed \_\_\_\_\_
- All Fast TATs Subject to Approval by Enthalpy Analytical, Inc.
- All Samples Disposed of After 6 months Unless Otherwise Instructed.

Enthalpy Analytical-Wilmington, NC has added enhancements to standard methods to improve accuracy, precision and permit an assessment of laboratory performance in the context of your specific data needs. For more information email CindyJames@enthalpy.com.

Client Name: BRUNSWICK COUNTY UTILITIES  
 Project Manager: GLENN WALKER  
 Report To: SAME

Project Number: \_\_\_\_\_  
 Site Name: NORTHWEST WATER PLANT  
 Location: LELAND N.C.

PO#: \_\_\_\_\_  
 Telephone#: \_\_\_\_\_  
 Email: \_\_\_\_\_

This Chain of Custody is applicable to Non-Air samples. Standard TAT differ per analysis and are provided by request.

**Client Special Instructions:**

Matrix: GW-Groundwater, WW-Wastewater, NW-Non-Potable Water, DW-Drinking Water, S-Soil, SL-Sludge, BT-Biological Tissue, O-Other

Type: G=Grab C=Composite Q=Quality Control

Sample ID	Date	Time	Sample Volume	Type	Matrix	Sample Containers				Analyses:							Notes:		
						# of Bottles	# of Jars	# of Bags	# Other	Method 1613	Method 8290	Method 1668A/B/C PCB	PFAS by LC/MS/MS	PAHs by HRGC/HRMS	Sample on Hold	Method 23		ALL PFAS	
080924-S01	8/9/2024	1:00PM	250 ml	G	NW	2												X	Please Add PFPrA and
080924-E01	8/9/2024	1:00PM	250 ml	G	DW	2												X	PFHpA To The Testing.
																			Mark Hager Knows About
																			This If you Have Questions

Relinquished By:	Date: <u>8/9/2024</u>	Received By:	Date: <u>8/9/24</u>	Time: <u>14:45</u>	Sample Temperature Upon Receipt: <u>9.7°</u>
					<input checked="" type="checkbox"/> Iced <input type="checkbox"/> Ambient °C
					<input type="checkbox"/> Iced <input type="checkbox"/> Ambient °C
					<input type="checkbox"/> Iced <input type="checkbox"/> Ambient °C

JOB ID: 0824-772 Date / Time: 8/1/24 14:45 Initials: S.S.  
 OR  
 Client: Brunswick County Utilities

Cooler 1 of 1

Temp °C: 9.7 Thermometer ID: T15

Received via FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input checked="" type="checkbox"/> Other <input type="checkbox"/>	<i>Check one</i>		<i>Check one</i>			
	On ice:	<input checked="" type="checkbox"/>	in a Box:	<input type="checkbox"/>	Cooler seals:	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No
	Melted ice:	<input type="checkbox"/>	in a Cooler:	<input checked="" type="checkbox"/>	Sample seals:	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No
	Ambient:	<input type="checkbox"/>	Cooler in Box:	<input type="checkbox"/>	Good condition:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
	Comment:					

Cooler  of

Temp °C:  Thermometer ID:

Received via FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input type="checkbox"/> Other <input type="checkbox"/>	<i>Check one</i>		<i>Check one</i>			
	On ice:	<input type="checkbox"/>	in a Box:	<input type="checkbox"/>	Cooler seals:	<input type="checkbox"/> Yes <input type="checkbox"/> No
	Melted ice:	<input type="checkbox"/>	in a Cooler:	<input type="checkbox"/>	Sample seals:	<input type="checkbox"/> Yes <input type="checkbox"/> No
	Ambient:	<input type="checkbox"/>	Cooler in Box:	<input type="checkbox"/>	Good condition:	<input type="checkbox"/> Yes <input type="checkbox"/> No
	Comment:					

Cooler  of

Temp °C:  Thermometer ID:

Received via FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input type="checkbox"/> Other <input type="checkbox"/>	<i>Check one</i>		<i>Check one</i>			
	On ice:	<input type="checkbox"/>	in a Box:	<input type="checkbox"/>	Cooler seals:	<input type="checkbox"/> Yes <input type="checkbox"/> No
	Melted ice:	<input type="checkbox"/>	in a Cooler:	<input type="checkbox"/>	Sample seals:	<input type="checkbox"/> Yes <input type="checkbox"/> No
	Ambient:	<input type="checkbox"/>	Cooler in Box:	<input type="checkbox"/>	Good condition:	<input type="checkbox"/> Yes <input type="checkbox"/> No
	Comment:					

**This Is The Last Page  
Of This Report.**