

Brunswick County Public Utilities - NC

3954 Clearwell Dr NE
Leland, NC 28451

Northwest Water Plant

Leland, NC

Samples Received: 10/27/22

Analytical Report 1022-820

Isotope Dilution Method PFAS



Enthalpy Analytical, LLC – Ultratrace

Mark Hager

O: 910-876-6894/ F: 910-212-6886

mark.hager@enthalpy.com / www.enthalpy.com

2714 Exchange Drive, Wilmington, NC 28405

I certify that to the best of my knowledge all analytical data presented in this report:

- Have been checked for completeness
- Are accurate, error-free, and legible
- Have been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s)

This analytical report was prepared in Portable Document Format (.PDF) and contains _____ pages.

Report Issued Date: _____



Summary of Results

Enthalpy Analytical

Job No.: 1022-820-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC Client Project: N/A Site: Northwest Water Plant

Summary

	Compound	CAS	102622S01 ng/L	102622E01 ng/L
Acids	PFBA	375-22-4	4.79	4.96
	PFPeA	2706-90-3	8.73	8.76
	PFHxA	307-24-4	7.19	7.31
	PFHpA	375-85-9	2.90	2.62
	PFOA	335-67-1	4.47	4.57
	PFNA	375-95-1	0.595 J	0.534 J
	PFDA	335-76-2	0.213 J	0.411 J
	PFUnDA	2058-94-8	ND U	ND U
	PFDoDA	307-55-1	0.154 L	ND U
	PFTTrDA	72629-94-8	ND U	ND U
	PFTeDA	376-06-7	ND U	ND U
Sulfonates	PFBS	375-73-5	4.44	4.46
	PFPeS	2706-91-4	0.625	0.704
	PFHxS	355-46-4	7.07	4.06
	PFHpS	375-92-8	0.172 J	ND U
	PFOS	1763-23-1	7.87	7.84
	PFNS	68259-12-1	ND U	ND U
	PFDS	335-77-3	ND U	ND U
	4:2 FTS	757124-72-4	ND U	ND U
Other	6:2 FTS	27619-97-2	0.408 J	0.423 J
	8:2 FTS	39108-34-4	ND U	ND U
	PFOSA	754-91-6	ND U	0.251 J
	N-MeFOSAA	2355-31-9	ND U	0.0189 L
	N-EtFOSAA	2991-50-6	0.0223 L	ND U
	HFPO-DA	13252-13-6	9.33	9.99
	PFMOAA	674-13-5	41.3	47.5
	PFMOPrA	377-73-1	ND U	ND U
	PFO2HxA	39492-88-1	14.3	13.2
	PFO3OA	39492-89-2	3.43	4.12
	PFO4DA	39492-90-5	0.389 L	0.418 L
	Nafion Byproduct 1	29311-67-9	ND U	ND U
	ADONA	919005-14-4	ND U	ND U
	9Cl-PF3ONS	756426-58-1	ND U	ND U
	11Cl-PF3OUdS	763051-92-9	ND U	ND U
	10:2 FTS	120226-60-0	ND U	ND U
	EVE Acid	69087-46-3	0.00895 L	0.00835 L
	FBSA	30334-69-1	0.778	0.589 J
	Hydro-EVE Acid	773804-62-9	0.294 L	0.299 L
	Hydrolyzed PSDA	2416366-19-1	5.53	10.8
	Nafion Byproduct 2	749836-20-2	0.724	0.612 J
	N-EtFOSA	4151-50-2	ND U	ND U
	N-EtFOSE	1691-99-2	ND U	ND U
	NFDHA	151772-58-6	ND U	ND U
	N-MeFOSA	31506-32-8	ND U	ND U
	N-MeFOSE	24448-09-7	ND U	ND U
	NVHOS	1132933-86-8	7.44	6.75
	PEPA	267239-61-2	2.99	3.08
	PFECA-G	801212-59-9	ND U	ND U
	PFEESA	113507-82-7	ND U	ND U
	PFHxDA	67905-19-5	ND U	ND U
	PFMOBA	863090-89-5	ND U	ND U
	PFO5DA	39492-91-6	ND U	ND U
	PMPA	13140-29-9	11.4	12.7
	R-EVE	2416366-22-6	3.30	3.87
	R-PSDA	2416366-18-0	4.71	7.34
	R-PSDCA	241636-21-5	ND U	ND U

Detailed Results

Enthalpy Analytical

Job No.: 1022-820-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC Client Project: N/A Site: Northwest Water Plant

Enthalpy ID	1022-820-001-1	Prep Batch	EU14227	Sample Vol (mL)	252.75
Sample Name	102622S01	Prep Date	2022-10-28 09:05	Extract Vol (mL)	0.4
Matrix	AQ	Analysis Date	11/4/2022 6:08:10 AM	Split Factor	N/A
Sampling Date	20221026 13:00	Analyst	rappelle	Method Code	WM-026
Received Date	2022-10-27 13:25	Instrument	Aragorn	Sample Type	Sample

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	3026.31	4.79	4.79	0.151	0.633			
	PFFPeA	2706-90-3	5518.08	8.73	8.73	0.168	0.633			
	PFHxA	307-24-4	4542.65	7.19	7.19	0.191	0.633			
	PFHpA	375-85-9	1835.53	2.90	2.90	0.121	0.633			
	PFOA	335-67-1	2824.09	4.47	4.47	0.175	0.633			
	PFNA	375-95-1	376.14	0.595	0.595	0.0753	0.633			J
	PFDA	335-76-2	134.55	0.213	0.213	0.0836	0.633			J
	PFUnDA	2058-94-8	ND	ND	ND	0.183	0.633			U
	PFDoDA	307-55-1	97.06	0.154	0.154	0.200	0.633			L
	PFTriDA	72629-94-8	ND	ND	ND	0.149	0.633			U
Sulfonates	PFTeDA	376-06-7	ND	ND	ND	0.216	0.633			U
	PFBS	375-73-5	2807.75	4.44	4.44	0.351	0.739			
	PFFPeS	2706-91-4	395.14	0.625	0.625	0.204	0.596			
	PFHxS	355-46-4	4465.18	7.07	7.07	0.189	0.580			
	PFHpS	375-92-8	108.50	0.172	0.172	0.134	0.603			J
	PFOS	1763-23-1	4970.09	7.87	7.87	0.158	0.586			
	PFNS	68259-12-1	ND	ND	ND	0.0855	0.610			U
	PFDS	335-77-3	ND	ND	ND	0.190	0.610			U
	4:2 FTS	757124-72-4	ND	ND	ND	0.117	0.593			U
	6:2 FTS	27619-97-2	257.66	0.408	0.408	0.115	0.603			J
Other	8:2 FTS	39108-34-4	ND	ND	ND	0.169	0.606			U
	PFOSA	754-91-6	ND	ND	ND	0.129	0.633			U
	N-MeFOSAA	2355-31-9	ND	ND	ND	0.142	0.633			U
	N-EiFOSAA	2991-50-6	14.11	0.0223	0.0223	0.108	0.633			L
	HFPO-DA	13252-13-6	5896.80	9.33	9.33	0.226	0.633			
	PFMOAA	674-13-5	26123.88	41.3	41.3	1.42	1.42			
	PFMOPrA	377-73-1	ND	ND	ND	0.237	0.633			U
	PFO2HxA	39492-88-1	9011.58	14.3	14.3	1.42	1.42			
	PFO3OA	39492-89-2	2168.69	3.43	3.43	1.42	1.42			
	PFO4DA	39492-90-5	245.76	0.389	0.389	1.50	1.50			L
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.301	0.633			U
	ADONA	919005-14-4	ND	ND	ND	0.119	0.600			U
	9CI-PF3ONS	756426-58-1	ND	ND	ND	0.119	0.590			U
	11CI-PF3OUdS	763051-92-9	ND	ND	ND	0.119	0.596			U
	10:2 FTS	120226-60-0	ND	ND	ND	0.237	0.633			U
	EVE Acid	69087-46-3	5.66	0.00895	0.00895	1.42	1.42			L
	FBSA	30334-69-1	491.38	0.778	0.778	0.237	0.633			
	Hydro-EVE Acid	773804-62-9	185.65	0.294	0.294	1.42	1.42			L
	Hydrolyzed PSDA	2416366-19-1	3496.94	5.53	5.53	1.42	1.42			
	Nafion Byproduct 2	749836-20-2	457.66	0.724	0.724	0.301	0.633			
	N-EiFOSA	4151-50-2	ND	ND	ND	0.237	0.633			U
	N-EiFOSE	1691-99-2	ND	ND	ND	7.12	7.12			U
	NFDHA	151772-58-6	ND	ND	ND	0.237	0.633			U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.237	0.633			U
	N-MeFOSE	24448-09-7	ND	ND	ND	7.12	7.12			U
	NVHOS	1132933-86-8	4703.40	7.44	7.44	1.42	1.42			
	PEPA	267239-61-2	1890.47	2.99	2.99	1.42	1.42			
	PFECA-G	801212-59-9	ND	ND	ND	0.301	1.42			U
	PFEESA	113507-82-7	ND	ND	ND	0.237	0.633			U
	PFHxDA	67905-19-5	ND	ND	ND	1.42	1.42			U
	PFMOBA	863090-89-5	ND	ND	ND	1.42	1.42			U
	PFO5DA	39492-91-6	ND	ND	ND	1.50	1.50			U
	PMPA	13140-29-9	7202.38	11.4	11.4	1.42	1.42			
	R-EVE	2416366-22-6	2083.44	3.30	3.30	1.42	1.42			
	R-PSDA	2416366-18-0	2973.19	4.71	4.71	1.42	1.42			
	R-PSDCA	241636-21-5	ND	ND	ND	1.42	1.42			U
ES	MPFBA		3221.42	5.10				20-150%	64.4%	
	M5PFFPeA		3489.75	5.52				20-150%	69.8%	
	M3PFBS		3193.00	5.05				20-150%	63.9%	
	M2-4:2 FTS		3406.78	5.39				20-150%	68.1%	
	M5PFHxA		2848.23	4.51				20-150%	57.0%	
	M3HFPO-DA		2696.99	4.27				20-150%	53.9%	
	M4PFHpA		2947.19	4.66				20-150%	58.9%	
	M3PFHxS		2401.41	3.80				20-150%	48.0%	
	M2-6:2 FTS		3302.53	5.23				20-150%	66.1%	
	M8PFOA		3036.21	4.81				20-150%	60.7%	
	M9PFNA		2749.50	4.35				20-150%	55.0%	
	M8PFOS		2881.34	4.56				20-150%	57.6%	
	M2-8:2 FTS		3186.34	5.04				20-150%	63.7%	
	M8FOSA-I		3079.90	4.87				20-150%	61.6%	
	M6PFDA		3225.98	5.11				20-150%	64.5%	
	d3-N-MeFOSAA		2255.45	3.57				20-150%	45.1%	
	d5-N-EiFOSAA		1836.94	2.91				20-150%	36.7%	
	M7PFUDa		2949.13	4.67				20-150%	59.0%	
	MPFDaA		2571.26	4.07				20-150%	51.4%	
	M2PFTeDA		1093.93	1.73				20-150%	21.9%	
	d3-N-MeFOSA		3080.15	4.87				10-200%	30.8%	
	d5-N-EiFOSA		2313.84	3.66				10-200%	23.1%	
	d7-N-MeFOSE		2128.13	3.37				10-200%	21.3%	
	d9-N-EiFOSE		1151.78	1.82				10-200%	11.5%	

Enthalpy Analytical

Job No.: 1022-820-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC Client Project: N/A Site: Northwest Water Plant

Enthalpy ID	1022-820-002-1	Prep Batch	EU14227	Sample Vol (mL)	257.56
Sample Name	102622E01	Prep Date	2022-10-28 09:05	Extract Vol (mL)	0.4
Matrix	AQ	Analysis Date	11/4/2022 6:30:50 AM	Split Factor	N/A
Sampling Date	20221026 13:00	Analyst	rappelle	Method Code	WM-026
Received Date	2022-10-27 13:25	Instrument	Aragorn	Sample Type	Sample

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	3192.71	4.96	4.96	0.149	0.621			
	PFFPeA	2706-90-3	5637.52	8.76	8.76	0.165	0.621			
	PFHxA	307-24-4	4705.95	7.31	7.31	0.187	0.621			
	PFHpA	375-85-9	1688.78	2.62	2.62	0.118	0.621			
	PFOA	335-67-1	2942.01	4.57	4.57	0.172	0.621			
	PFNA	375-95-1	343.93	0.534	0.534	0.0739	0.621			J
	PFDA	335-76-2	264.45	0.411	0.411	0.0820	0.621			J
	PFUnDA	2058-94-8	ND	ND	ND	0.180	0.621			U
	PFDoDA	307-55-1	ND	ND	ND	0.196	0.621			U
	PFTriDA	72629-94-8	ND	ND	ND	0.147	0.621			U
Sulfonates	PFTeDA	376-06-7	ND	ND	ND	0.212	0.621			U
	PFBS	375-73-5	2874.97	4.46	4.46	0.345	0.725			
	PFFPeS	2706-91-4	453.39	0.704	0.704	0.200	0.585			
	PFHxS	355-46-4	2612.21	4.06	4.06	0.185	0.569			
	PFHpS	375-92-8	ND	ND	ND	0.131	0.592			U
	PFOS	1763-23-1	5051.09	7.84	7.84	0.155	0.575			
	PFNS	68259-12-1	ND	ND	ND	0.0839	0.598			U
	PFDS	335-77-3	ND	ND	ND	0.186	0.598			U
	4:2 FTS	757124-72-4	ND	ND	ND	0.115	0.582			U
	6:2 FTS	27619-97-2	272.52	0.423	0.423	0.113	0.592			J
Other	8:2 FTS	39108-34-4	ND	ND	ND	0.166	0.595			U
	PFOSA	754-91-6	161.92	0.251	0.251	0.126	0.621			J
	N-MeFOSAA	2355-31-9	12.17	0.0189	0.0189	0.140	0.621			L
	N-EiFOSAA	2991-50-6	ND	ND	ND	0.106	0.621			U
	HFPO-DA	13252-13-6	6434.68	9.99	9.99	0.221	0.621			
	PFMOA	674-13-5	30598.55	47.5	47.5	1.40	1.40			
	PFMOPrA	377-73-1	ND	ND	ND	0.233	0.621			U
	PFO2HxA	39492-88-1	8522.84	13.2	13.2	1.40	1.40			
	PFO3OA	39492-89-2	2653.15	4.12	4.12	1.40	1.40			
	PFO4DA	39492-90-5	268.95	0.418	0.418	1.48	1.48			L
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.295	0.621			U
	ADONA	919005-14-4	ND	ND	ND	0.116	0.589			U
	9Cl-PF3ONS	756426-58-1	ND	ND	ND	0.116	0.579			U
	11Cl-PF3OUdS	763051-92-9	ND	ND	ND	0.116	0.585			U
	10:2 FTS	120226-60-0	ND	ND	ND	0.233	0.621			U
	EVE Acid	69087-46-3	5.38	0.00835	0.00835	1.40	1.40			L
	FBSA	30334-69-1	378.95	0.589	0.589	0.233	0.621			J
	Hydro-EVE Acid	773804-62-9	192.32	0.299	0.299	1.40	1.40			L
	Hydrolyzed PSDA	2416366-19-1	6966.19	10.8	10.8	1.40	1.40			
	Nafion Byproduct 2	749836-20-2	394.16	0.612	0.612	0.295	0.621			J
	N-EiFOSA	4151-50-2	ND	ND	ND	0.233	0.621			U
	N-EiFOSE	1691-99-2	ND	ND	ND	6.99	6.99			U
	NFDHA	151772-58-6	ND	ND	ND	0.233	0.621			U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.233	0.621			U
	N-MeFOSE	24448-09-7	ND	ND	ND	6.99	6.99			U
	NVHOS	1132933-86-8	4348.66	6.75	6.75	1.40	1.40			
	PEPA	267239-61-2	1981.57	3.08	3.08	1.40	1.40			
	PFECA-G	801212-59-9	ND	ND	ND	0.295	1.40			U
	PFEESA	113507-82-7	ND	ND	ND	0.233	0.621			U
	PFHxDA	67905-19-5	ND	ND	ND	1.40	1.40			U
	PFMOBA	863090-89-5	ND	ND	ND	1.40	1.40			U
	PFO5DA	39492-91-6	ND	ND	ND	1.48	1.48			U
	PMPA	13140-29-9	8185.90	12.7	12.7	1.40	1.40			
	R-EVE	2416366-22-6	2490.63	3.87	3.87	1.40	1.40			
	R-PSDA	2416366-18-0	4727.63	7.34	7.34	1.40	1.40			
	R-PSDCA	241636-21-5	ND	ND	ND	1.40	1.40			U
ES	MPFBA		4902.25	7.61				20-150%	98.0%	
	M5PFFPeA		5921.99	9.20				20-150%	118.4%	
	M3PFBS		6086.59	9.45				20-150%	121.7%	
	M2-4:2 FTS		5494.80	8.53				20-150%	109.9%	
	M5PFHxA		4543.28	7.06				20-150%	90.9%	
	M3HFPO-DA		3984.36	6.19				20-150%	79.7%	
	M4PFHpA		4472.86	6.95				20-150%	89.5%	
	M3PFHxS		4616.11	7.17				20-150%	92.3%	
	M2-6:2 FTS		4903.57	7.62				20-150%	98.1%	
	M8PFOA		4206.35	6.53				20-150%	84.1%	
	M9PFNA		4481.33	6.96				20-150%	89.6%	
	M8PFOS		4284.51	6.65				20-150%	85.7%	
	M2-8:2 FTS		4641.57	7.21				20-150%	92.8%	
	M8FOSA-I		4099.48	6.37				20-150%	82.0%	
	M6PFDA		4524.74	7.03				20-150%	90.5%	
	d3-N-MeFOSAA		3887.23	6.04				20-150%	77.7%	
	d5-N-EiFOSAA		4096.07	6.36				20-150%	81.9%	
	M7PFUdA		4464.28	6.93				20-150%	89.3%	
	MPPFDa		4528.44	7.03				20-150%	90.6%	
	M2PFTeDA		3472.66	5.39				20-150%	69.5%	
	d3-N-MeFOSA		3511.78	5.45				10-200%	35.1%	
	d5-N-EiFOSA		3052.25	4.74				10-200%	30.5%	
	d7-N-MeFOSE		7353.96	11.4				10-200%	73.5%	
	d9-N-EiFOSE		6809.13	10.6				10-200%	68.1%	

QC Data

Enthalpy Analytical

Job No.: 1022-820-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC Client Project: N/A Site: Northwest Water Plant

Enthalpy ID	MB-14227-PFAS	Prep Batch	EU14227	Sample Vol (mL)	250
Sample Name	MB-14227-PFAS	Prep Date	2022-10-28 09:05	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	11/4/2022 5:22:50 AM	Split Factor	N/A
Sampling Date		Analyst	rappelle	Method Code	WM-026
Received Date		Instrument	Aragorn	Sample Type	Blank

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	ND	ND	ND	0.153	0.640			U
	PFFPeA	2706-90-3	ND	ND	ND	0.170	0.640			U
	PFHxA	307-24-4	ND	ND	ND	0.193	0.640			U
	PFHpA	375-85-9	ND	ND	ND	0.122	0.640			U
	PFOA	335-67-1	ND	ND	ND	0.177	0.640			U
	PFNA	375-95-1	ND	ND	ND	0.0761	0.640			U
	PFDA	335-76-2	ND	ND	ND	0.0845	0.640			U
	PFUnDA	2058-94-8	ND	ND	ND	0.185	0.640			U
	PFDoDA	307-55-1	ND	ND	ND	0.202	0.640			U
	PFTriDA	72629-94-8	ND	ND	ND	0.151	0.640			U
Sulfonates	PFTeDA	376-06-7	ND	ND	ND	0.218	0.640			U
	PFBS	375-73-5	ND	ND	ND	0.355	0.747			U
	PFFPeS	2706-91-4	ND	ND	ND	0.206	0.603			U
	PFHxS	355-46-4	ND	ND	ND	0.191	0.586			U
	PFHpS	375-92-8	ND	ND	ND	0.135	0.610			U
	PFOS	1763-23-1	ND	ND	ND	0.160	0.593			U
	PFNS	68259-12-1	ND	ND	ND	0.0864	0.616			U
	PFDS	335-77-3	ND	ND	ND	0.192	0.616			U
	4:2 FTS	757124-72-4	ND	ND	ND	0.118	0.600			U
	6:2 FTS	27619-97-2	ND	ND	ND	0.116	0.610			U
Other	8:2 FTS	39108-34-4	ND	ND	ND	0.171	0.613			U
	PFOSA	754-91-6	ND	ND	ND	0.130	0.640			U
	N-MeFOSAA	2355-31-9	ND	ND	ND	0.144	0.640			U
	N-EtFOSAA	2991-50-6	ND	ND	ND	0.109	0.640			U
	HFPO-DA	13252-13-6	ND	ND	ND	0.228	0.640			U
	PFMOA	674-13-5	ND	ND	ND	1.44	1.44			U
	PFMOPrA	377-73-1	ND	ND	ND	0.240	0.640			U
	PFO2HxA	39492-88-1	ND	ND	ND	1.44	1.44			U
	PFO3OA	39492-89-2	ND	ND	ND	1.44	1.44			U
	PFO4DA	39492-90-5	ND	ND	ND	1.52	1.52			U
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.304	0.640			U
	ADONA	919005-14-4	ND	ND	ND	0.120	0.606			U
	9Cl-PF3ONS	756426-58-1	ND	ND	ND	0.120	0.596			U
	11Cl-PF3OUdS	763051-92-9	ND	ND	ND	0.120	0.603			U
	10:2 FTS	120226-60-0	ND	ND	ND	0.240	0.640			U
	EVE Acid	69087-46-3	ND	ND	ND	1.44	1.44			U
	FBSA	30334-69-1	ND	ND	ND	0.240	0.640			U
	Hydro-EVE Acid	773804-62-9	ND	ND	ND	1.44	1.44			U
	Hydrolyzed PSDA	2416366-19-1	ND	ND	ND	1.44	1.44			U
	Nafion Byproduct 2	749836-20-2	ND	ND	ND	0.304	0.640			U
	N-EtFOSA	4151-50-2	ND	ND	ND	0.240	0.640			U
	N-EtFOSE	1691-99-2	ND	ND	ND	7.20	7.20			U
	NFDHA	151772-58-6	ND	ND	ND	0.240	0.640			U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.240	0.640			U
	N-MeFOSE	24448-09-7	ND	ND	ND	7.20	7.20			U
	NVHOS	1132933-86-8	ND	ND	ND	1.44	1.44			U
	PEPA	267239-61-2	ND	ND	ND	1.44	1.44			U
	PFECA-G	801212-59-9	ND	ND	ND	0.304	1.44			U
	PFEESA	113507-82-7	ND	ND	ND	0.240	0.640			U
	PFHxDA	67905-19-5	ND	ND	ND	1.44	1.44			U
	PFMOBA	863090-89-5	ND	ND	ND	1.44	1.44			U
	PFO5DA	39492-91-6	ND	ND	ND	1.52	1.52			U
	PMPA	13140-29-9	ND	ND	ND	1.44	1.44			U
	R-EVE	2416366-22-6	ND	ND	ND	1.44	1.44			U
	R-PSDA	2416366-18-0	ND	ND	ND	1.44	1.44			U
	R-PSDCA	241636-21-5	ND	ND	ND	1.44	1.44			U
ES	MPFBA		2933.87	4.69				20-150%	58.7%	
	M5PFFPeA		2718.14	4.35				20-150%	54.4%	
	M3PFBS		2498.20	4.00				20-150%	50.0%	
	M2-4:2 FTS		3333.54	5.33				20-150%	66.7%	
	M5PFHxA		2910.16	4.66				20-150%	58.2%	
	M3HFPO-DA		2543.94	4.07				20-150%	50.9%	
	M4PFHpA		2714.91	4.34				20-150%	54.3%	
	M3PFHxS		2258.01	3.61				20-150%	45.2%	
	M2-6:2 FTS		3118.98	4.99				20-150%	62.4%	
	M8PFOA		3086.69	4.94				20-150%	61.7%	
	M9PFNA		2837.90	4.54				20-150%	56.8%	
	M8PFOS		3436.56	5.50				20-150%	68.7%	
	M2-8:2 FTS		3320.65	5.31				20-150%	66.4%	
	M8FOSA-I		3349.63	5.36				20-150%	67.0%	
	M6PFDA		3218.71	5.15				20-150%	64.4%	
	d3-N-MeFOSAA		2337.83	3.74				20-150%	46.8%	
	d5-N-EtFOSAA		1766.73	2.83				20-150%	35.3%	
	M7PFUDa		2690.00	4.30				20-150%	53.8%	
	MPPFDaA		2251.50	3.60				20-150%	45.0%	
	M2PFTeDA		886.26	1.42				20-150%	17.7%	Q
	d3-N-MeFOSA		2918.29	4.67				10-200%	29.2%	
	d5-N-EtFOSA		2325.74	3.72				10-200%	23.3%	
	d7-N-MeFOSE		2137.78	3.42				10-200%	21.4%	
	d9-N-EtFOSE		1120.00	1.79				10-200%	11.2%	

Enthalpy Analytical

Job No.: 1022-820-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC Client Project: N/A Site: Northwest Water Plant

Enthalpy ID	OPR-14227-PFAS	Prep Batch	EU14227	Sample Vol (mL)	250
Sample Name	OPR-14227-PFAS	Prep Date	2022-10-28 09:05	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	11/4/2022 5:45:30 AM	Split Factor	N/A
Sampling Date		Analyst	rappelle	Method Code	WM-026
Received Date		Instrument	Aragorn	Sample Type	Control

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	13054.69	20.9	20.9	0.153	0.640	73-129%	104.4%	
	PFPeA	2706-90-3	12563.94	20.1	20.1	0.170	0.640	72-129%	100.5%	
	PFHxA	307-24-4	13464.68	21.5	21.5	0.193	0.640	72-129%	107.7%	
	PFHpA	375-85-9	13842.27	22.1	22.1	0.122	0.640	72-130%	110.7%	
	PFOA	335-67-1	12982.90	20.8	20.8	0.177	0.640	71-133%	103.9%	
	PFNA	375-95-1	14406.37	23.1	23.1	0.0761	0.640	69-130%	115.3%	
	PFDA	335-76-2	13679.48	21.9	21.9	0.0845	0.640	71-129%	109.4%	
	PFUnDA	2058-94-8	13146.26	21.0	21.0	0.185	0.640	69-133%	105.2%	
	PFDoDA	307-55-1	12424.48	19.9	19.9	0.202	0.640	72-134%	99.4%	
	PFTTrDA	72629-94-8	26775.37	42.8	42.8	0.151	0.640	65-144%	214.2%	Q
Sulfonates	PFTeDA	376-06-7	11423.60	18.3	18.3	0.218	0.640	71-132%	91.4%	
	PFBS	375-73-5	10504.17	16.8	16.8	0.355	0.747	72-134%	94.7%	
	PFPeS	2706-91-4	12817.31	20.5	20.5	0.206	0.603	71-127%	109.0%	
	PFHxS	355-46-4	14250.98	22.8	22.8	0.191	0.586	68-131%	124.7%	
	PFHpS	375-92-8	13347.98	21.4	21.4	0.135	0.610	69-134%	112.1%	
	PFOS	1763-23-1	12467.27	19.9	19.9	0.160	0.593	65-140%	107.5%	
	PFNS	68259-12-1	13290.30	21.3	21.3	0.0864	0.616	69-127%	110.5%	
	PFDS	335-77-3	12097.70	19.4	19.4	0.192	0.616	53-142%	100.3%	
	4:2 FTS	757124-72-4	11345.79	18.2	18.2	0.118	0.600	63-143%	96.9%	
	6:2 FTS	27619-97-2	13364.06	21.4	21.4	0.116	0.610	64-140%	112.4%	
Other	8:2 FTS	39108-34-4	12938.68	20.7	20.7	0.171	0.613	67-138%	107.8%	
	PFOSA	754-91-6	10824.21	17.3	17.3	0.130	0.640	67-137%	86.6%	
	N-MeFOSAA	2355-31-9	14101.36	22.6	22.6	0.144	0.640	65-136%	112.8%	
	N-EtFOSAA	2991-50-6	12873.22	20.6	20.6	0.109	0.640	61-135%	103.0%	
ES	HFPO-DA	13252-13-6	14460.10	23.1	23.1	0.228	0.640	70-130%	115.7%	
	MPFBA		3535.42	5.66				20-150%	70.7%	
	M5PFPeA		3602.03	5.76				20-150%	72.0%	
	M3PFBS		3243.86	5.19				20-150%	64.9%	
	M2-4:2 FTS		3836.79	6.14				20-150%	76.7%	
	M5PFHxA		3119.78	4.99				20-150%	62.4%	
	M3HFPO-DA		2771.15	4.43				20-150%	55.4%	
	M4PFHpA		3265.06	5.22				20-150%	65.3%	
	M3PFHxS		3108.87	4.97				20-150%	62.2%	
	M2-6:2 FTS		3648.87	5.84				20-150%	73.0%	
	M8PFOA		3333.46	5.33				20-150%	66.7%	
	M9PFNA		2888.52	4.62				20-150%	57.8%	
	M8PFOS		3474.59	5.56				20-150%	69.5%	
	M2-8:2 FTS		3411.13	5.46				20-150%	68.2%	
	M8FOSA-I		3511.97	5.62				20-150%	70.2%	
	M6PFDA		3231.15	5.17				20-150%	64.6%	
	d3-N-MeFOSAA		2394.03	3.83				20-150%	47.9%	
	d5-N-EtFOSAA		2313.43	3.70				20-150%	46.3%	
	M7PFUdA		3196.50	5.11				20-150%	63.9%	
	MPFDoA		2941.13	4.71				20-150%	58.8%	
	M2PFTeDA		1096.78	1.75				20-150%	21.9%	

Narrative Summary

Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	1022-820-1 PFAS by Isotope Dilution (non-potable water)
Client ID.	N/A Site: Northwest Water Plant

1. Custody

Megan Holden received the samples on October 27, 2022 at 2.5 °C after being relinquished by Brunswick County Public Utilities - NC. The samples were received in good condition.

Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC.

Table 1 - Sample Inventory

EU Lab Sample ID	Client Sample ID	Matrix
1022-820-001-1	102622S01	AQ
1022-820-002-1	102622E01	AQ

2. Methods and Analytes

A list of analytes of interest and corresponding methods of analysis is shown in Table 3. Abbreviations are defined in the listed Appendices.

Table 3 - Methods and Analytes

EU Method	Analytes	Cleanup Method
EU-047	Brunswick PFAS List	ENVI-Carb

3. Analysis

The samples were analyzed using Waters Acquity UPLC equipped with Xevo TQ MS (LC/MS/MS "Aragorn").

For aqueous samples, the sample volume was measured gravimetrically by the laboratory, and spiked with Extraction Standard (ES). The sample was then mixed well and centrifuged, if needed. The samples were then extracted via SPE, and the extracts were cleaned up using ENVI-Carb.

Each final sample extract was transferred to an autosampler vial, spiked with Injection Standard (IS), and brought to a final volume of 400µL prior to analysis.

4. Calibration

In the initial calibration, the reported analytes exhibited R^2 of ≥ 0.99 . The reported analytes in the calibration standards, continuing calibration (concal) and Initial Calibration Verification (ICV) met the 30% accuracy criterion for native analytes.

The Technical Director extended the method criteria for certain non-legacy analytes that do not have their own internal standard and exhibit observed variability during calibration.

Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	1022-820-1 PFAS by Isotope Dilution (non-potable water)
Client ID.	N/A Site: Northwest Water Plant

5. QC Notes

Except where noted below, the QC sample analyses passed all method criteria.

QC samples that did not meet method acceptance criteria were:

OPR-14227-PFAS PFTrDA fell above method recovery criteria. This analyte was not detected in the samples; therefore, the data is reportable without adverse impact.

MB-14227-PFAS M2PFTeDA: Labeled standards in the method blank (MB) that fell above the control limits for ES recovery are denoted by the "Q" qualifier. The target analytes are quantified based on their ratio to their labeled standard analogs. As a result, low or high labeled standard recovery does not cause any change to ratios or contribute any additional error in the measurement of the target analyte(s). The data is reported without adverse impact.

The samples were extracted within the 28-day from collection holding time and analyzed within the 28-day from extraction to analysis holding time required by the method.

6. Reporting Notes

This report provides all results including detections below LOD following client instruction.

The results presented in this report are representative of the samples as provided to the laboratory.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, LLC in Wilmington NC is accredited by the Louisiana Department of Environmental Quality to the 2009 TNI Standard under certificate number 05075.

General Reporting Notes – Data Qualifiers

The following are general reporting notes that are applicable to all Enthalpy Analytical, LLC - Wilmington, NC data reports, unless specifically noted otherwise.

General Data Qualifiers

- B – The analyte was found in the method blank, at a concentration that was at least 10% of the amount in the sample.
- Cxx – Two or more congeners co-elute. In EDDs, C denotes the lowest IUPAC congener in a co-elution group and additional co-eluters for the group ('xx') are shown with the number of the lowest IUPAC co-eluter.
- E – The reported concentration exceeds the calibration range (upper point of the calibration curve). For HRMS data, this condition does not imply additional measurement uncertainty. For LC-MS/MS data, these values should be considered as having measurement uncertainty higher than values within the calibration range.
- EDL – Estimated Detection Level: The EDL is unique to isotope dilution methods and reflects the conditions of analysis at the time of analysis, including the equipment used. Where the MDL is a static value, the EDL is a dynamic value.
- EMPC – Estimated Maximum Possible Concentration: EMPC is specific to Dioxin/Furan tests to indicate the determined ion-abundance ratio was outside the allowed theoretical range (usually due to being near the detection limit, although it can very rarely be caused by a co-eluting interference). The EMPC concentration is adjusted to reflect the value at the theoretical ion-abundance ratio.
- IR – The ion ratio between the primary and secondary ions was observed to be outside the method criteria. The analyte concentration may be inaccurate due to interference.
- J – The analyte has a concentration below the minimum calibration level (LOQ value) but greater than the LOD. These values should be considered as having measurement uncertainty higher than values within the calibration range
- L - Indicates that an analyte has a concentration below the Minimum Detection Limit (MDL). The reported concentration is not recommended for regulatory use as the analyte signal may have a signal-to-noise ratio less than the criteria deemed necessary to be considered a detected analyte.
- LOD – Limit of Detection: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOD. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the minimum detection limit (MDL). The LOD is adjusted for sample weight or volume.
- LOQ – Limit of Quantitation: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOQ. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the reporting limit (RL). The LOQ is adjusted for sample weight or volume.

General Reporting Notes – Data Qualifiers

- <LOD() – Analyte was not found at a concentration high enough to be reported as detected. It is reported as less than the LOD, and the LOD is given in the parentheses.
- <LOQ() – Analyte was not found at a concentration high enough to be reported as above the QSM-defined LOQ or TNI defined Reporting Limit. It is reported as less than the LOQ, and the LOQ is given in the parentheses.
- ND – Indicates a non-detect.
- NR – Indicates a value that is not reportable due to issues observed in sample preparation or analysis.
- PR – The associated congener(s) is(are) poorly resolved.
- QI – Indicates the presence of a quantitative interference.
- RL – Reporting Limit. Lowest reportable value. The level is higher than the MDL.
- SI – Denotes “Single Ion Mode” and is utilized for PCBs where the secondary ion trace has a significantly elevated noise level due to background PFK. Responses for such peaks are calculated using an EMPC approach based solely on the primary ion area(s) and may be considered estimates.
- U – The analyte was not detected.
- V / Q – The labeled standard recovery is not within method control limits.
- X – Indicates the result is from re-injection/repeat/second-column analysis.

Lab Identifiers/ Data Attributes

- AR – Indicates use of the archived portion of the sample extract.
- CU – Indicates a sample that required additional clean-up prior to HRMS injection/processing.
- D – Dilution Data. Result was obtained from the analysis of a dilution. The number that follows the “D” indicates the dilution factor.
- DE – Indicates a dilution performed with the addition of ES (Extraction Standard) solution.
- DUP – Designation for a duplicate sample.
- MS – Designation for a matrix spike.
- MSD – Designation for a matrix spike duplicate.
- R – Indicates a re-extraction of the sample.
- RJ – Indicates a reinjection of the sample extract.



General Reporting Notes – Data Qualifiers

- S – Indicates a sample split. The number that follows the “S” indicates the split factor.
- SAT – Indicates an analyte saturated the detector.

PFAS Compound Acronym List		
Acronym	CAS #	Compound Name
Target Analytes		
* Analyte is not accredited		
PFBA	375-22-4	Perfluorobutanoic Acid
PFPeA	2706-90-3	Perfluoropentanoic Acid
PFHxA	307-24-4	Perfluorohexanoic Acid
PFHpA	375-85-9	Perfluoroheptanoic Acid
PFOA	335-67-1	Perfluorooctanoic Acid
PFNA	375-95-1	Perfluorononanoic Acid
PFDA	335-76-2	Perfluorodecanoic acid
PFUnA (PFUnDA)	2058-94-8	Perfluoroundecanoic acid
PFDoA (PFDoDA)	307-55-1	Perfluorododecanoic acid
PFTriA (PFTriA)	72629-94-8	Perfluorotridecanoic acid
PFTeDA (PFTA)	376-06-7	Perfluorotetradecanoic acid
PFBS	375-73-5	Perfluorobutane sulfonic acid
PFPeS	2706-91-4	Perfluoropentane sulfonic acid
PFHxS	355-46-4	Perfluorohexane sulfonic acid
PFHpS	375-92-8	Perfluoroheptane sulfonic acid
PFOS	1763-23-1	Perfluorooctane sulfonic acid
PFNS	68259-12-1	Perfluorononane sulfonic acid
PFDS	335-77-3	Perfluorodecane sulfonic acid
4:2 FTS	757124-72-4	4:2 fluorotelomer sulfonic acid
6:2 FTS	27619-97-2	6:2 fluorotelomer sulfonic acid
8:2 FTS	39108-34-4	8:2 fluorotelomer sulfonic acid
PFOSA (FOSA)	754-91-6	Perfluorooctane sulfonamide
N-MeFOSAA	2355-31-9	N-methyl perfluorooctane sulfonamido acetic acid
N-EtFOSAA	2991-50-6	N-ethyl perfluorooctane sulfonamido acetic acid
HFPO-DA	13252-13-6	2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy)-propanoic acid (Gen-X)
11CI-PF3OUDS	763051-92-9	11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid
9CI-PF3ONS	756426-58-1	9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid
ADONA	919005-14-4	4,8-dioxa-3H-perfluorononanoic acid
PFEESA	113507-82-7	Perfluoro(2-ethoxyethane)sulphonic acid
PFMOBA (PFMBA)	863090-89-5	Perfluoro-4-methoxybutanic acid
NFDHA	151772-58-6	Nonafluoro-3,6-dioxaheptanoic acid
PFMOPra (PFMPA)	377-73-1	Perfluoro-3-methoxypropanoic acid
* PFMOAA	674-13-5	Perfluoro-2-methoxyacetic acid
* PFO2HxA	39492-88-1	Perfluoro (3,5-dioxaheptanoic) acid
* PFO3OA	39492-89-2	Perfluoro (3,5,7-trioxaoctanoic) acid
* PFO4DA	39492-90-5	Perfluoro (3,5,7,9-tetraoxadecanoic) acid
* PFO5DA	39492-91-6	Perfluoro(3,5,7,9,11-pentaoxadodecanoic) acid
* Nafion Byproduct 1	29311-67-9	Nafion Byproduct 1
* Nafion Byproduct 2	749836-20-2	Nafion Byproduct 2
* PEPA	267239-61-2	Perfluoro-2-ethoxypropanoic acid
* PMPA	13140-29-9	Perfluoro-2-methoxypropanoic acid
* 10:2 FTS	120226-60-0	Fluorotelomer sulfonate 10:2
* N-EtFOSA	4151-50-2	N-ethylperfluoro-1-octanesulfonamide
* N-EtFOSE	1691-99-2	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol
* N-MeFOSA	31506-32-8	N-methylperfluoro-1-octanesulfonamide
* N-MeFOSE	24448-09-7	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol
* PFECA-G	801212-59-9	4-(Heptafluoroisopropoxy)hexafluorobutanoic acid
* PFHxDA	67905-19-5	Perfluorohexadecanoic acid
R-PSDA (Nafion Byproduct 4)	2416366-18-0	Perfluoro-4-(2-sulfoethoxy)pentanoic acid

PFAS Compound Acronym List		
Acronym	CAS #	Compound Name
Target Analytes		
* Analyte is not accredited		
* Hydrolyzed PSDA (Nafion Byproduct 5)	2416366-19-1	2-fluoro-2-[1,1,2,3,3,3-hexafluoro-2-(1,1,2,2-tetrafluoro-2-sulfoethoxy)propoxy]-acetic acid
* R-PSDCA (Nafion Byproduct 6)	2416366-21-5	1,1,2,2-tetrafluoro-2-[1,2,2,3,3-pentafluoro-1-(trifluoromethyl)propoxy] ethanesulfonic acid
* EVE Acid	69087-46-3	2,2,3,3-tetrafluoro-3-({1,1,1,2,3,3-hexafluoro-3-[(1,2,2-trifluoroethenyl)oxy]propan-2-yl}oxy)propionic acid
* FBSA	30334-69-1	Perfluorobutylsulfonamide
* Hydro-EVE Acid	773804-62-9	2,2,3,3-Tetrafluoro-3-({1,1,1,2,3,3-hexafluoro-3-(1,2,2,2-tetrafluoroethoxy)propan-2-yl}oxy)propanoic acid
* R-EVE Acid	2416366-22-6	4-(2-carboxy-1,1,2,2-tetrafluoroethoxy)-2,2,3,3,4,5,5,5-octafluoro-pentanoic acid
* NVHOS	1132933-86-8	Perfluoroethoxysulfonic acid
* PFDoS	79780-39-5	Perfluorododecane sulfonic acid
* PFODA	16517-11-6	Perfluorooctadecanoic acid
* 3:3 FTCA	356-02-5	2H,2H,3H,3H-Perfluorohexanoic acid
* 5:3 FTCA	914637-49-3	2H,2H,3H,3H-Perfluorooctanoic acid
* 7:3 FTCA	812-70-4	2H,2H,3H,3H-Perfluorodecanoic acid
* N-AP-FHxSA	50598-28-2	N-(3-(Dimethylamino)propyl)tridecafluoro-1-hexanesulfonamide
* N-CMAmP-6:2 FOSA	34455-29-3	N-(Carboxymethyl)-N,N-dimethyl-3-(((3,3,4,4,5,5,6,6,7,7,8,8,8-tridecafluorooctyl)sulfonyl)amino)1-propanaminium

Sample Custody



Enthalpy Ultratrace Job#: _____ COC Page 1 of 1

Special Handling:

X Standard Turn Around Time

☐ Rush Turn Around Time -- Date Needed _____

• All Fast TATs Subject to Approval by Enthalpy Analytical, Inc.

- All Samples Disposed of After 6 months Unless Otherwise Instructed.

Enthalpy Analytical, Wilmington, NC has added enhancements to standard methods to improve accuracy, precision and permit an assessment of laboratory performance in the context of your specific data needs. For more information email Cindy.James@enthalpy.com.

Client Name: **BRUNSWICK COUNTY UTILITIES**

Project Manager: **GLENN WALKER**

Report To: **SAME**

Project Number:

Site Name: **NORTHWEST WATER PLANT**

Location: **LELAND N.C.**

PO#:

Telephone#:

Email:

This Chain of Custody is applicable to Non-Air samples. Standard TAT differ per analysis and are provided by request.

Client Special Instructions:

Matrix: GW-Groundwater, WW-Wastewater, NW-Non-Potable Water, DW-Drinking Water, S-Soil, SL-Sludge, BT-Biological Tissue, O-Other

Type: G=Grab C=Composite Q=Quality Control

Courier, cooler, on ice, no seals,
good condition MYL 10.27.22

Relinquished By:

Date:

Phil McCulloch

10/26/2022

Received By:

Date:

Time:

Sample Temperature Upon Receipt:

☐ Iced ☐ Ambient °C 2.5 T10☐ Iced ☐ Ambient °C☐ Iced ☐ Ambient °C

2714 Exchange Drive • Wilmington, NC 28405 • (910) 212-5858 • www.enthalpy.com

**This Is The Last Page
Of This Report.**