

Brunswick County Public Utilities - NC

PO Box 249
Bolivia, NC 28422-0249

LELAND N.C.

Client Project# NORTHWEST WATER PLANT
Samples Received: 12/20/2024

Analytical Report 1224-824

PFAS by Isotope Dilution (non-potable water)

Report Issue Date: 1/13/2025

I certify that to the best of my knowledge all analytical data presented in this report have been checked for completeness, accuracy, errors and legibility in addition to having been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s). This analytical report was prepared in Portable Document Format (.PDF) and contains 27 pages. This report shall not be reproduced except in full without approval of the laboratory. This will provide assurance that parts of the report are not taken out of context.

Amendment(s):

Signature:



Alexandra Mejia, Quality Assurance Associate I



Enthalpy Analytical, LLC – Wilmington
Christina Kurnath, Project Manager
chkurnath@montrose-env.com / www.enthalpy.com
O: 910-876-6895
2714 Exchange Drive, Wilmington, NC 28405

Table of Contents

Case Narrative	
General Reporting Notes	
PFAS Acronym List.....	
Results	
Summary of Results	
Detailed Results	
QC Data	
Method Blank.....	
LCS.....	
Sample Custody.....	
Chain of Custody	
Shipping Documents	

Narrative Summary



Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	1224-824-1
Client ID.	NORTHWEST WATER PLANT Site: LELAND N.C.

1. Custody

Meredith Curtis received the samples at 5.4 °C after being relinquished by Brunswick County Public Utilities - NC.

The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC.

Table 1 - Sample Inventory

EU Lab Sample ID	Client Sample ID	Matrix	Received
1224-824-001-1	122024-S01	aqueous	2024-12-20
1224-824-001-2	122024-S01	aqueous	2024-12-20
1224-824-002-1	122024-E01	aqueous	2024-12-20
1224-824-002-2	122024-E01	aqueous	2024-12-20

2. Methods and Analytes

A list of analytes of interest and corresponding methods of analysis is shown in Table 3. Abbreviations are defined in the listed Appendices.

Table 3 - Methods and Analytes

EU Method	Analytes	Cleanup Method
EU-047	Brunswick List + PFPrA	ENVI-Carb

3. Analysis

The samples were analyzed using Sciex Triple Quad 7500 (LC/MS/MS "Bumblebee").

The samples were analyzed using Waters Acquity UPLC equipped with Xevo TQ MS (LC/MS/MS "Sauron").

4. Calibration

In the initial calibration, the reported analytes exhibited R^2 of ≥ 0.99 . The reported analytes in the calibration standards, Initial Calibration Verification (ICV) and continuing calibration (concal) met the accuracy criterion for native analytes.

Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	1224-824-1
Client ID.	NORTHWEST WATER PLANT Site: LELAND N.C.

The Standards that did not were:

- SID BH57 (PFMOBA)
- SID BH58 (PFMOBA)
- SID BH62 (N-MeFOSA, PFMOAA, PFMOPrA) The alternate supplier of the unlabeled standard solution used in the ICV does not contain select analytes of interest.

Analyte(s) that exceeded method control limits in the concals were not detected >LOQ in the samples. The data is reported without adverse impact.

5. QC Notes

Except where noted below, the QC sample analyses passed all method criteria.

QC samples that did not meet method acceptance criteria were:

- OPR_18746_PFAS (M2PFTeDA)

Surrogate (ES) M2PFTeDA deviated from method recovery criteria in the ongoing precision recovery (OPR) QC sample. Target analytes are quantified based on their ratio to their labeled standard analogs. When detected at a signal-to-noise above 10:1 the ES peak area is used to quantify its respective target analyte using accepted isotope dilution principles. The data is reported without adverse impact.

PFAS by Isotope Dilution (non-potable water) samples were extracted within 28 days, and extracts analyzed within 28 days.

6. Reporting Notes

The results presented in this report are representative of the samples as provided to the laboratory.

Some labeled extraction standards (ES) in the sample analyses fell outside the control limits for ES recovery, as denoted by the "Q" qualifier. The target analytes are quantified based on their ratio to their labeled standard analogs. As a result, low or high labeled standard recovery do not cause any change to ratios or contribute any additional error in the measurement of the target analytes. The data have been accepted and reported with no further actions.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, LLC in Wilmington NC is accredited by the Louisiana Department of Environmental Quality to the 2009 TNI Standard under certificate number 05075.



General Reporting Notes – Data Qualifiers

The following are general reporting notes that are applicable to all Enthalpy Analytical, LLC - Wilmington, NC data reports, unless specifically noted otherwise.

General Data Qualifiers

- Ac - Alternate calculation flag indicates the es recovery was calculated using the opening concal when either of the following situations is encountered in the data processing software: the ES recovery is over 400% or the JS is not detected.
- B – The analyte was found in the method blank, at a concentration that was at least 10% of the amount in the sample.
- Cxx – Two or more congeners co-elute. In EDDs, C denotes the lowest IUPAC congener in a co-elution group and additional co-eluters for the group ('xx') are shown with the number of the lowest IUPAC co-eluter.
- E – The reported concentration exceeds the calibration range (upper point of the calibration curve). For HRMS data, this condition does not imply additional measurement uncertainty. For LC-MS/MS data, these values should be considered as having measurement uncertainty higher than values within the calibration range.
- EDL – Estimated Detection Level: The EDL is unique to isotope dilution methods and reflects the conditions of analysis at the time of analysis, including the equipment used. Where the MDL is a static value, the EDL is a dynamic value.
- EMPC – Estimated Maximum Possible Concentration: EMPC is specific to Dioxin/Furan tests to indicate the determined ion-abundance ratio was outside the allowed theoretical range (usually due to being near the detection limit, although it can very rarely be caused by a co-eluting interference). The EMPC concentration is adjusted to reflect the value at the theoretical ion-abundance ratio.
- I/IR – The ion ratio between the primary and secondary ions was observed to be outside the method criteria. The analyte concentration may be inaccurate due to interference.
- J – The analyte has a concentration below the minimum calibration level (LOQ value) but greater than the LOD. These values should be considered as having measurement uncertainty higher than values within the calibration range
- L - For reports containing PFAS analytes only, this flag indicates that an analyte has a concentration below the Minimum Detection Limit (MDL) . The reported concentration is not recommended for regulatory use as the analyte signal may have a signal-to-noise ratio less than the criteria deemed necessary to be considered a detected analyte.
- LOD – Limit of Detection: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOD. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the minimum detection limit (MDL). The LOD is adjusted for sample weight or volume.

General Reporting Notes – Data Qualifiers

- LOQ – Limit of Quantitation: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOQ. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the reporting limit (RL). The LOQ is adjusted for sample weight or volume.
- <LOD() – Analyte was not found at a concentration high enough to be reported as detected. It is reported as less than the LOD, and the LOD is given in the parentheses.
- <LOQ() – Analyte was not found at a concentration high enough to be reported as above the QSM-defined LOQ or TNI defined Reporting Limit. It is reported as less than the LOQ, and the LOQ is given in the parentheses.
- ND – Indicates a non-detect.
- NR – Indicates a value that is not reportable due to issues observed in sample preparation or analysis.
- PR – The associated congener(s) is(are) poorly resolved.
- QI – Indicates the presence of a quantitative interference.
- RL – Reporting Limit. Lowest reportable value. The level is higher than the MDL.
- SI – Denotes “Single Ion Mode” and is utilized for PCBs where the secondary ion trace has a significantly elevated noise level due to background PFK. Responses for such peaks are calculated using an EMPC approach based solely on the primary ion area(s) and may be considered estimates.
- U – The analyte was not detected.
- V / Q – The labeled standard recovery is not within method control limits.
- X – Indicates the result is from re-injection/repeat/second-column analysis.

Lab Identifiers/ Data Attributes

- AR – Indicates use of the archived portion of the sample extract.
- CU – Indicates a sample that required additional clean-up prior to HRMS injection/processing.
- D – Dilution Data. Result was obtained from the analysis of a dilution. The number that follows the “D” indicates the dilution factor.
- DE – Indicates a dilution performed with the addition of ES (Extraction Standard) solution.
- DUP – Designation for a duplicate sample.
- MS – Designation for a matrix spike.
- MSD – Designation for a matrix spike duplicate.



General Reporting Notes – Data Qualifiers

- R – Indicates a re-extraction of the sample.
- RJ – Indicates a reinjection of the sample extract.
- S – Indicates a sample split. The number that follows the “S” indicates the split factor.
- SAT – Indicates an analyte saturated the detector.

PFAS Compound Acronym List			
Acronym	CAS #	Compound Name	
* accredited for SOP EU047 / EPA method 1633 # Method 537.1 Accredited ^ Method 533 Accredited ~EPA 1633 extended list			
Target Analytes			
* , ^	PFBA	375-22-4	Perfluorobutanoic Acid
* , # , ^	PFPeA	2706-90-3	Perfluoropentanoic Acid
* , # , ^	PFHxA	307-24-4	Perfluorohexanoic Acid
* , # , ^	PFHpA	375-85-9	Perfluoroheptanoic Acid
* , # , ^	PFOA	335-67-1	Perfluorooctanoic Acid
* , # , ^	PFNA	375-95-1	Perfluorononanoic Acid
* , # , ^	PFDA	335-76-2	Perfluorodecanoic acid
* , # , ^	PFUnA (PFUnDA)	2058-94-8	Perfluoroundecanoic acid
* , #	PFDoA (PFDoDA)	307-55-1	Perfluorododecanoic acid
* , #	PFTriA (PFTriA)	72629-94-8	Perfluorotridecanoic acid
* , # , ^	PFTeDA (PFTA)	376-06-7	Perfluorotetradecanoic acid
* , ^	PFBS	375-73-5	Perfluorobutane sulfonic acid
* , # , ^	PFPeS	2706-91-4	Perfluoropentane sulfonic acid
* , ^	PFHxS	355-46-4	Perfluorohexane sulfonic acid
* , # , ^	PFHpS	375-92-8	Perfluoroheptane sulfonic acid
* , # , ^	PFOS	1763-23-1	Perfluorooctane sulfonic acid
*	PFNS	68259-12-1	Perfluorononane sulfonic acid
*	PFDS	335-77-3	Perfluorodecane sulfonic acid
* , ^	4:2 FTS	757124-72-4	4:2 fluorotelomer sulfonic acid
* , ^	6:2 FTS	27619-97-2	6:2 fluorotelomer sulfonic acid
* , ^	8:2 FTS	39108-34-4	8:2 fluorotelomer sulfonic acid
~	10:2 FTS	120226-60-0	Fluorotelomer sulfonate 10:2
~	FHxSA	41997-13-1	Perfluorohexanesulfonamide
*	PFOSA (FOSA)	754-91-6	Perfluorooctane sulfonamide
* , #	N-MeFOSAA	2355-31-9	N-methyl perfluorooctane sulfonamido acetic acid
*	N-MeFOSA	31506-32-8	N-methylperfluoro-1-octanesulfonamide
*	N-MeFOSE	24448-09-7	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol
* , #	N-EtFOSAA	2991-50-6	N-ethyl perfluorooctane sulfonamido acetic acid
*	N-EtFOSA	4151-50-2	N-ethylperfluoro-1-octanesulfonamide
*	N-EtFOSE	1691-99-2	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol
* , # , ^	HFPO-DA	13252-13-6	2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy)-propanoic acid (Gen-X)
* , # , ^	11Cl-PF3OUdS	763051-92-9	11-chloroheptafluoro-3-oxadecane-1-sulfonic acid
* , # , ^	9Cl-PF3ONS	756426-58-1	9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid
* , # , ^	ADONA	919005-14-4	4,8-dioxa-3H-perfluorononanoic acid
* , ^	PFEESA	113507-82-7	Perfluoro(2-ethoxyethane)sulphonic acid
* , ^	PFMOBA (PFMBA)	863090-89-5	Perfluoro-4-methoxybutanoic acid
* , ^	NFDHA	151772-58-6	Nonafluoro-3,6-dioxahexanoic acid
* , ^	PFMOPrA (PFMPA)	377-73-1	Perfluoro-3-methoxypropanoic acid
~	PFPrA	422-64-0	2,2,3,3,3-Pentafluoropropionic acid
~	PFPrS (PFPS)	423-41-6	Perfluoropropanesulfonic acid
~	PFMOAA	674-13-5	Perfluoro-2-methoxyacetic acid
~	PFO2HxA	39492-88-1	Perfluoro (3,5-dioxahexanoic) acid
~	PFO3OA	39492-89-2	Perfluoro (3,5,7-trioxaoctanoic) acid
~	PFO4DA	39492-90-5	Perfluoro (3,5,7,9-tetraoxadecanoic) acid
~	PFO5DA	39492-91-6	Perfluoro(3,5,7,9,11-pentaoxadodecanoic) acid
~	Nafion Byproduct 1 (PS Acid)	29311-67-9	Nafion Byproduct 1
~	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	Nafion Byproduct 2
~	PEPA	267239-61-2	Perfluoro-2-ethoxypropanoic acid
~	PMPA	13140-29-9	Perfluoro-2-methoxypropanoic acid

PFAS Compound Acronym List		
Acronym	CAS #	Compound Name
* accredited for SOP EU047 / EPA method 1633 # Method 537.1 Accredited ^ Method 533 Accredited ~EPA 1633 extended list		
~ PFECA-G	801212-59-9	4-(Heptafluoroisopropoxy)hexafluorobutanoic acid
~ PFHxDA	67905-19-5	Perfluorohexadecanoic acid
~ R-PSDA (Nafion Byproduct 4)	2416366-18-0	Perfluoro-4-(2-sulfoethoxy)pentanoic acid
Hydrolyzed PSDA (Nafion Byproduct 5)	2416366-19-1	2-fluoro-2-[1,1,2,3,3,3-hexafluoro-2-(1,1,2-tetrafluoro-2-sulfoethoxy)propoxy]-acetic acid
~ R-PSDCA (Nafion Byproduct 6)	2416366-21-5	1,1,2,2-tetrafluoro-2-[1,2,2,3,3-pentafluoro-1-(trifluoromethyl)propoxy] ethanesulfonic acid
~ EVE Acid	69087-46-3	2,2,3,3-tetrafluoro-3-({1,1,1,2,3,3-hexafluoro-3-[(1,2,2-trifluoroethenyl)oxy]propan-2-yl)oxy}propionic acid
~ FBSA	30334-69-1	Perfluorobutylsulfonamide
~ MeFBSA	68298-12-4	1-Butanesulfonamide; (N-(Methyl)nonafluorobutanesulfonamide)
~ Hydro-EVE Acid	773804-62-9	2,2,3,3-Tetrafluoro-3-[[1,1,1,2,3,3-hexafluoro-3-(1,2,2,2-tetrafluoroethoxy)propan-2-yl]oxy}propanoic acid
~ R-EVE Acid	2416366-22-6	4-(2-carboxy-1,1,2,2-tetrafluoroethoxy)-2,2,3,3,4,5,5,5-octafluoro-pentanoic acid
~ NVHOS	1132933-86-8	Perfluoroethoxysulfonic acid
*~ PFDoS	79780-39-5	Perfluorododecane sulfonic acid
~ PFODA	16517-11-6	Perfluorooctadecanoic acid
* 3:3 FTCA	356-02-5	2H,2H,3H,3H-Perfluorohexanoic acid
* 5:3 FTCA	914637-49-3	2H,2H,3H,3H-Perfluorooctanoic acid
* 7:3 FTCA	812-70-4	2H,2H,3H,3H-Perfluorodecanoic acid
~ N-AP-FHxSA	50598-28-2	N-(3-(Dimethylamino)propyl)tridecafluoro-1-hexanesulfonamide
~ N-CMAmP-6:2 FOSA	34455-29-3	N-(Carboxymethyl)-N,N-dimethyl-3-(((3,3,4,4,5,5,6,6,7,7,8,8,8-tridecafluorooctyl)sulfonyl)amino)1-propanaminium
~ BPAF	1478-61-1	Bisphenol AF
~ HQ-115	90076-65-6	Bis(trifluoromethane)sulfonimide lithium salt

Results

Enthalpy Analytical

Job No.: 1224-824-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Summary

	Compound	CAS	122024-S01 ng/L	122024-E01 ng/L	
Acids	PFPrA	422-64-0	537 L	460 L	
	PFBA	375-22-4	ND U	ND U	
	PFPeA	2706-90-3	3.10	3.51	
	PFHxA	307-24-4	3.61	3.74	
	PFHpA	375-85-9	1.57	1.59	
	PFOA	335-67-1	2.78	3.04	
	PFNA	375-95-1	0.510 J	0.552 J	
	PFDA	335-76-2	0.214 J	0.163 J	
	PFUnDA	2058-94-8	ND U	ND U	
	PFDoDA	307-55-1	ND U	ND U	
	PFTrDA	72629-94-8	ND U	ND U	
	PFTeDA	376-06-7	ND U	ND U	
	PFHxDA	67905-19-5	ND U	ND U	
	Sulfonates	PFBS	375-73-5	1.60	1.62
PFPeS		2706-91-4	0.0789 L	0.220 J	
PFHxS		355-46-4	1.85	1.62	
PFHpS		375-92-8	ND U	ND U	
PFOS		1763-23-1	5.52	5.33	
PFNS		68259-12-1	ND U	ND U	
PFDS		335-77-3	ND U	ND U	
4:2 FTS		757124-72-4	ND U	ND U	
6:2 FTS		27619-97-2	ND U	0.0140 L	
8:2 FTS		39108-34-4	ND U	ND U	
10:2 FTS		120226-60-0	ND U	ND U	
Sulfonamidos	FBSA	30334-69-1	ND U	ND U	
	N-EtFOSA	4151-50-2	ND U	ND U	
	N-EtFOSAA	2991-50-6	ND U	ND U	
	N-EtFOSE	1691-99-2	ND U	ND U	
	N-MeFOSA	31506-32-8	ND U	ND U	
	N-MeFOSAA	2355-31-9	ND U	ND U	
	N-MeFOSE	24448-09-7	ND U	ND U	
	PFOSA	754-91-6	0.0196 L	ND U	
	PFECAs	ADONA	919005-14-4	ND U	ND U
		EVE Acid	69087-46-3	ND U	ND U
HFPO-DA		13252-13-6	2.30	2.08	
Hydro-EVE Acid		773804-62-9	ND U	ND U	
NFDHA		151772-58-6	ND U	ND U	
PEPA		267239-61-2	2.29	1.26	
PFECA-G		801212-59-9	ND U	ND U	
PFMOAA		674-13-5	27.0	21.9	
PFMOBA		863090-89-5	ND U	ND U	
PFMOPrA		377-73-1	ND U	ND U	
PFO2HxA		39492-88-1	ND U	ND U	
PFO3OA		39492-89-2	ND U	ND U	
PFO4DA		39492-90-5	ND U	ND U	
PFO5DA		39492-91-6	ND U	ND U	
PMPA		13140-29-9	6.51	5.97	
R-EVE		2416366-22-6	4.41	4.55	
PFESAs	11Cl-PF3OUds	763051-92-9	ND U	ND U	
	9Cl-PF3ONS	756426-58-1	ND U	ND U	
	Hydrolyzed PSDA	2416366-19-1	1.89	1.73	
	Nafion Byproduct 1 (PS Acid)	29311-67-9	ND U	ND U	
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	ND U	ND U	
	NVHOS	1132933-86-8	ND U	ND U	
	PFEESA	113507-82-7	ND U	ND U	
	R-PSDA	2416366-18-0	2.59 L	2.15 L	
	R-PSDCA	2416366-21-5	ND U	ND U	

Enthalpy Analytical

Job No.: 1224-824-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	122024-S01	Prep Batch	EU18746
Sampling Site		Analyst	ext-magennaef
Enthalpy ID	1224-824-001-1	Instrument	Sauron
Matrix	aqueous	Sample Vol mL	271.84
Sampling Date	2024-11-27 11:30	Extract Vol mL	0.4
Received Date	2024-12-20	Split Factor	N/A
Prep Date	2024-12-30 09:54	Method Code	EU-047-NPW
AnalysisDate	2024-12-30 23:01		
SampleType	Sample		
Bottle ID	A		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	S301224028	ND	0.234	0.589			U	
	PFPeA	2706-90-3	S301224028	3.10	0.168	0.589				
	PFHxA	307-24-4	S301224028	3.61	0.197	0.589				
	PFFHpA	375-85-9	S301224028	1.57	0.206	0.589				
	PFOA	335-67-1	S301224028	2.78	0.135	0.589				
	PFNA	375-95-1	S301224028	0.510	0.133	0.589			J	
	PFDA	335-76-2	S301224028	0.214	0.168	0.589			J	
	PFUnDA	2058-94-8	S301224028	ND	0.133	0.589			U	
	PFDoDA	307-55-1	S301224028	ND	0.239	0.589			U	
	PFTrDA	72629-94-8	S301224028	ND	0.195	0.589			U	
	PFTeDA	376-06-7	S301224028	ND	0.224	0.589			U	
	PFHxDA	67905-19-5	S301224028	ND	0.313	0.589			U	
	Sulfonates	PFBS	375-73-5	S301224028	1.60	0.313	0.589			
		PFPeS	2706-91-4	S301224028	0.0789	0.121	0.555			L
PFFHxS		355-46-4	S301224028	1.85	0.454	0.539				
PFFHpS		375-92-8	S301224028	ND	0.285	0.561			U	
PFOS		1763-23-1	S301224028	5.52	0.311	0.545				
PFNS		68259-12-1	S301224028	ND	0.183	0.567			U	
PFDS		335-77-3	S301224028	ND	0.309	0.567			U	
4:2 FTS		757124-72-4	S301224028	ND	0.0763	0.551			U	
6:2 FTS		27619-97-2	S301224028	ND	0.278	0.561			U	
8:2 FTS		39108-34-4	S301224028	ND	0.132	0.564			U	
10:2 FTS	120226-60-0	S301224028	ND	0.451	0.589			U		
Sulfonamidos	FBSA	30334-69-1	S301224028	ND	0.280	0.589			U	
	N-EiFOSA	4151-50-2	S301224028	ND	0.364	0.589			U	
	N-EiFOSAA	2991-50-6	S301224028	ND	0.239	0.589			U	
	N-EiFOSE	1691-99-2	S301224028	ND	0.901	2.65			U	
	N-MeFOSA	31506-32-8	S301224028	ND	0.243	0.589			U	
	N-MeFOSAA	2355-31-9	S301224028	ND	0.165	0.589			U	
	N-MeFOSE	24448-09-7	S301224028	ND	0.559	2.65			U	
	PFOSA	754-91-6	S301224028	0.0196	0.0826	0.589			L	
PFECAs	ADONA	919005-14-4	S301224028	ND	0.159	0.558			U	
	EVE Acid	69087-46-3	S301224028	ND	0.188	1.32			U	
	HFPO-DA	13252-13-6	S301224028	2.30	0.0624	0.589				
	Hydro-EVE Acid	773804-62-9	S301224028	ND	0.193	0.589			U	
	NFDHA	151772-58-6	S301224028	ND	0.124	0.589			U	
	PEPA	267239-61-2	S301224028	2.29	0.110	0.589				
	PFECA-G	801212-59-9	S301224028	ND	0.0785	0.589			U	
	PFMOAA	674-13-5	S301224028	27.0	0.298	0.589				
	PFMOBA	863090-89-5	S301224028	ND	0.988	1.32			U	
	PFMOPrA	377-73-1	S301224028	ND	0.210	0.589			U	
	PFO2HxA	39492-88-1	S301224028	ND	0.189	0.589			U	
	PFO3OA	39492-89-2	S301224028	ND	0.270	0.589			U	
	PFO4DA	39492-90-5	S301224028	ND	0.465	2.94			U	
	PFO5DA	39492-91-6	S301224028	ND	0.471	2.94			U	
	PMPA	13140-29-9	S301224028	6.51	0.139	0.589				
	R-EVE	2416366-22-6	S301224028	4.41	0.977	1.32				
	PFESAs	11Cl-PF3OUdS	763051-92-9	S301224028	ND	0.278	0.555			U
9Cl-PF3ONS		756426-58-1	S301224028	ND	0.377	0.548			U	
Hydrolyzed PSDA		2416366-19-1	S301224028	1.89	0.392	0.589				
Nafion Byproduct 1 (PS Acid)		29311-67-9	S301224028	ND	0.315	0.589			U	
Nafion Byproduct 2 (Hydro-PS Acid)		749836-20-2	S301224028	ND	0.487	0.589			U	
NVHOS		1132933-86-8	S301224028	ND	0.0907	0.589			U	
PFEESA		113507-82-7	S301224028	ND	0.177	0.589			U	
R-PSDA		2416366-18-0	S301224028	2.59	2.59	2.59			L	
R-PSDCA	2416366-21-5	S301224028	ND	0.248	0.589			U		
ES	MPFBA		S301224028				20-150%	93.9%		
	M5PFPeA		S301224028				20-150%	204%	Q	
	M3PFBS		S301224028				20-150%	334%	Q	
	M2-4:2 FTS		S301224028				20-150%	158%	Q	
	M5PFFHxA		S301224028				20-150%	96.0%		
	M3HFPO-DA		S301224028				20-150%	83.0%		
	M4PFFHpA		S301224028				20-150%	95.8%		

Enthalpy Analytical

Job No.: 1224-824-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	122024-S01		
Sampling Site			
Enthalpy ID	1224-824-001-1	Prep Batch	EU18746
Matrix	aqueous	Analyst	ext-magennaef
Sampling Date	2024-11-27 11:30	Instrument	Sauron
Received Date	2024-12-20	Sample Vol mL	271.84
Prep Date	2024-12-30 09:54	Extract Vol mL	0.4
AnalysisDate	2024-12-30 23:01	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	A		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M3PFHxS		S301224028				20-150%	113%	
M2-6:2 FTS		S301224028				20-150%	105%	
M8PFOA		S301224028				20-150%	90.0%	
M9PFNA		S301224028				20-150%	90.1%	
M8PFOS		S301224028				20-150%	89.6%	
M2-8:2 FTS		S301224028				20-150%	90.2%	
M8FOSA-I		S301224028				20-150%	68.1%	
M6PFDA		S301224028				20-150%	89.5%	
d3-N-MeFOSAA		S301224028				20-150%	86.0%	
d5-N-EtFOSAA		S301224028				20-150%	73.0%	
M7PFUdA		S301224028				20-150%	66.5%	
MPFDoA		S301224028				20-150%	45.4%	
M2PFTeDA		S301224028				20-150%	11.9%	Q
d3-N-MeFOSA		S301224028				10-200%	2.11%	Q
d5-N-EtFOSA		S301224028				10-200%	1.89%	Q
d7-N-MeFOSE		S301224028				10-200%	30.8%	
d9-N-EtFOSE		S301224028				10-200%	24.1%	

Enthalpy Analytical

Job No.: 1224-824-1 PFAS by Isotope Dilution (non-potable water)
Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name 122024-S01
Sampling Site
Enthalpy ID 1224-824-001-2 Prep Batch eu18749
Matrix aqueous Analyst ext-magennaef
Sampling Date 2024-11-27 11:30 Instrument Bumblebee
Received Date 2024-12-20 Sample Vol mL 0.1
Prep Date 2024-12-30 07:00 Extract Vol mL 0.2
AnalysisDate 2024-12-31 18:24 Split Factor N/A
SampleType Sample Method Code EU-047-NPW
Bottle ID -

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPfA	422-64-0	B311224-12311824	537	700	1530			L
ES	13C3-PFPfA		B311224-12311824				20-150%	110%	

Enthalpy Analytical

Job No.: 1224-824-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	122024-E01	Prep Batch	EU18746
Sampling Site		Analyst	ext-magennaef
Enthalpy ID	1224-824-002-1	Instrument	Sauron
Matrix	aqueous	Sample Vol mL	283.41
Sampling Date	2024-11-27 11:30	Extract Vol mL	0.4
Received Date	2024-12-20	Split Factor	N/A
Prep Date	2024-12-30 09:54	Method Code	EU-047-NPW
AnalysisDate	2024-12-30 23:46		
SampleType	Sample		
Bottle ID	A		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	S301224030	ND	0.224	0.565			U	
	PFPeA	2706-90-3	S301224030	3.51	0.161	0.565				
	PFFhxA	307-24-4	S301224030	3.74	0.189	0.565				
	PFFHpA	375-85-9	S301224030	1.59	0.198	0.565				
	PFOA	335-67-1	S301224030	3.04	0.129	0.565				
	PFNA	375-95-1	S301224030	0.552	0.128	0.565			J	
	PFDA	335-76-2	S301224030	0.163	0.161	0.565			J	
	PFUnDA	2058-94-8	S301224030	ND	0.128	0.565			U	
	PFFDoDA	307-55-1	S301224030	ND	0.229	0.565			U	
	PFTTrDA	72629-94-8	S301224030	ND	0.187	0.565			U	
	PFTeDA	376-06-7	S301224030	ND	0.215	0.565			U	
	PFFhxDA	67905-19-5	S301224030	ND	0.300	0.565			U	
	Sulfonates	PFBS	375-73-5	S301224030	1.62	0.300	0.565			
		PFPeS	2706-91-4	S301224030	0.220	0.116	0.532			J
		PFFhXS	355-46-4	S301224030	1.62	0.436	0.517			
PFFHpS		375-92-8	S301224030	ND	0.273	0.538			U	
PFOs		1763-23-1	S301224030	5.33	0.298	0.523				
PFNS		68259-12-1	S301224030	ND	0.175	0.544			U	
PFDs		335-77-3	S301224030	ND	0.296	0.544			U	
4:2 FTS		757124-72-4	S301224030	ND	0.0732	0.529			U	
6:2 FTS		27619-97-2	S301224030	0.0140	0.266	0.538			L	
8:2 FTS		39108-34-4	S301224030	ND	0.126	0.541			U	
10:2 FTS	120226-60-0	S301224030	ND	0.432	0.565			U		
Sulfonamidos	FBSA	30334-69-1	S301224030	ND	0.268	0.565			U	
	N-EiFOSA	4151-50-2	S301224030	ND	0.349	0.565			U	
	N-EiFOSAA	2991-50-6	S301224030	ND	0.229	0.565			U	
	N-EiFOSE	1691-99-2	S301224030	ND	0.864	2.54			U	
	N-MeFOSA	31506-32-8	S301224030	ND	0.233	0.565			U	
	N-MeFOSAA	2355-31-9	S301224030	ND	0.159	0.565			U	
	N-MeFOSE	24448-09-7	S301224030	ND	0.536	2.54			U	
	PFOsA	754-91-6	S301224030	ND	0.0792	0.565			U	
	PFECAs	ADONA	919005-14-4	S301224030	ND	0.153	0.535			U
		EVE Acid	69087-46-3	S301224030	ND	0.180	1.27			U
HFPO-DA		13252-13-6	S301224030	2.08	0.0598	0.565				
Hydro-EVE Acid		773804-62-9	S301224030	ND	0.185	0.565			U	
NFDHA		151772-58-6	S301224030	ND	0.119	0.565			U	
PEPA		267239-61-2	S301224030	1.26	0.106	0.565				
PFECA-G		801212-59-9	S301224030	ND	0.0753	0.565			U	
PfMOAA		674-13-5	S301224030	21.9	0.286	0.565				
PfMOBA		863090-89-5	S301224030	ND	0.947	1.27			U	
PfMOPrA		377-73-1	S301224030	ND	0.201	0.565			U	
PFO2HxA		39492-88-1	S301224030	ND	0.182	0.565			U	
PFO3OA		39492-89-2	S301224030	ND	0.259	0.565			U	
PFO4DA		39492-90-5	S301224030	ND	0.446	2.82			U	
PFO5DA		39492-91-6	S301224030	ND	0.452	2.82			U	
PMPA		13140-29-9	S301224030	5.97	0.133	0.565				
R-EVE		2416366-22-6	S301224030	4.55	0.937	1.27				
PFESAs		11CI-PF3OUdS	763051-92-9	S301224030	ND	0.266	0.532			U
		9CI-PF3ONS	756426-58-1	S301224030	ND	0.362	0.526			U
		Hydrolyzed PSDA	2416366-19-1	S301224030	1.73	0.376	0.565			
		Nafion Byproduct 1 (PS Acid)	29311-67-9	S301224030	ND	0.302	0.565			U
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	S301224030	ND	0.468	0.565			U	
	NVHOS	1132933-86-8	S301224030	ND	0.0870	0.565			U	
	PFEESA	113507-82-7	S301224030	ND	0.170	0.565			U	
	R-PSDA	2416366-18-0	S301224030	2.15	2.49	2.49			L	
	R-PSDCA	2416366-21-5	S301224030	ND	0.238	0.565			U	
	ES	MPFBA		S301224030				20-150%	75.0%	
M5PFPeA			S301224030				20-150%	136%		
M3PFBS			S301224030				20-150%	238%	Q	
M2-4:2 FTS			S301224030				20-150%	197%	Q	
M5PFFhxA			S301224030				20-150%	86.5%		
M3HFPO-DA			S301224030				20-150%	80.7%		
M4PFFHpA			S301224030				20-150%	96.5%		

Enthalpy Analytical

Job No.: 1224-824-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	122024-E01		
Sampling Site			
Enthalpy ID	1224-824-002-1	Prep Batch	EU18746
Matrix	aqueous	Analyst	ext-magennaef
Sampling Date	2024-11-27 11:30	Instrument	Sauron
Received Date	2024-12-20	Sample Vol mL	283.41
Prep Date	2024-12-30 09:54	Extract Vol mL	0.4
AnalysisDate	2024-12-30 23:46	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	A		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M3PFHxS		S301224030				20-150%	110%	
M2-6:2 FTS		S301224030				20-150%	121%	
M8PFOA		S301224030				20-150%	87.4%	
M9PFNA		S301224030				20-150%	90.3%	
M8PFOS		S301224030				20-150%	89.5%	
M2-8:2 FTS		S301224030				20-150%	100%	
M8FOSA-I		S301224030				20-150%	93.1%	
M6PFDA		S301224030				20-150%	90.6%	
d3-N-MeFOSAA		S301224030				20-150%	91.7%	
d5-N-EtFOSAA		S301224030				20-150%	91.6%	
M7PFUdA		S301224030				20-150%	78.0%	
MPFDoA		S301224030				20-150%	63.5%	
M2PFTeDA		S301224030				20-150%	15.7%	Q
d3-N-MeFOSA		S301224030				10-200%	25.9%	
d5-N-EtFOSA		S301224030				10-200%	18.6%	
d7-N-MeFOSE		S301224030				10-200%	53.5%	
d9-N-EtFOSE		S301224030				10-200%	40.1%	

Enthalpy Analytical

Job No.: 1224-824-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	122024-E01		
Sampling Site			
Enthalpy ID	1224-824-002-2	Prep Batch	eu18749
Matrix	aqueous	Analyst	ext-magennaef
Sampling Date	2024-11-27 11:30	Instrument	Bumblebee
Received Date	2024-12-20	Sample Vol mL	0.1
Prep Date	2024-12-30 07:00	Extract Vol mL	0.2
AnalysisDate	2024-12-31 18:35	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPfA	422-64-0	B311224-12311835	460	700	1530			L
ES	13C3-PFPfA		B311224-12311835				20-150%	110%	

QC Data

Enthalpy Analytical

Job No.: 1224-824-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	MB_18746_PFAS	Prep Batch	EU18746
Sampling Site		Analyst	ext-magennaef
Enthalpy ID	MB_18746_PFAS	Instrument	Sauron
Matrix	aqueous	Sample Vol mL	250
Sampling Date		Extract Vol mL	0.4
Received Date		Split Factor	N/A
Prep Date	2024-12-30 09:54	Method Code	EU-047-NPW
AnalysisDate	2024-12-30 18:52		
SampleType	Blank		
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	S301224017	ND	0.254	0.640			U	
	PFPeA	2706-90-3	S301224017	ND	0.183	0.640			U	
	PFFhxA	307-24-4	S301224017	0.0153	0.214	0.640			L	
	PFFHpA	375-85-9	S301224017	ND	0.224	0.640			U	
	PFOA	335-67-1	S301224017	ND	0.146	0.640			U	
	PFNA	375-95-1	S301224017	ND	0.145	0.640			U	
	PFDA	335-76-2	S301224017	ND	0.183	0.640			U	
	PFUnDA	2058-94-8	S301224017	ND	0.145	0.640			U	
	PFDODA	307-55-1	S301224017	ND	0.260	0.640			U	
	PFTTrDA	72629-94-8	S301224017	ND	0.212	0.640			U	
	PFTeDA	376-06-7	S301224017	ND	0.244	0.640			U	
	PFFhxDA	67905-19-5	S301224017	ND	0.340	0.640			U	
	Sulfonates	PFBS	375-73-5	S301224017	ND	0.340	0.640			U
		PFPeS	2706-91-4	S301224017	ND	0.131	0.603			U
		PFFhXS	355-46-4	S301224017	ND	0.494	0.586			U
		PFFHpS	375-92-8	S301224017	ND	0.310	0.610			U
PFOS		1763-23-1	S301224017	0.129	0.338	0.593			L	
PFNS		68259-12-1	S301224017	ND	0.199	0.616			U	
PFDS		335-77-3	S301224017	ND	0.336	0.616			U	
4:2 FTS		757124-72-4	S301224017	ND	0.0830	0.600			U	
6:2 FTS		27619-97-2	S301224017	ND	0.302	0.610			U	
8:2 FTS		39108-34-4	S301224017	ND	0.143	0.613			U	
10:2 FTS	120226-60-0	S301224017	ND	0.490	0.640			U		
Sulfonamidos	FBSA	30334-69-1	S301224017	ND	0.304	0.640			U	
	N-EiFOSA	4151-50-2	S301224017	ND	0.396	0.640			U	
	N-EiFOSAA	2991-50-6	S301224017	ND	0.260	0.640			U	
	N-EiFOSE	1691-99-2	S301224017	ND	0.980	2.88			U	
	N-MeFOSA	31506-32-8	S301224017	ND	0.264	0.640			U	
	N-MeFOSAA	2355-31-9	S301224017	ND	0.180	0.640			U	
	N-MeFOSE	24448-09-7	S301224017	ND	0.608	2.88			U	
	PFOSA	754-91-6	S301224017	ND	0.0898	0.640			U	
	PFECAs	ADONA	919005-14-4	S301224017	ND	0.173	0.606			U
		EVE Acid	69087-46-3	S301224017	ND	0.204	1.44			U
HFPO-DA		13252-13-6	S301224017	ND	0.0678	0.640			U	
Hydro-EVE Acid		773804-62-9	S301224017	ND	0.210	0.640			U	
NFDHA		151772-58-6	S301224017	ND	0.135	0.640			U	
PEPA		267239-61-2	S301224017	ND	0.120	0.640			U	
PFECA-G		801212-59-9	S301224017	ND	0.0854	0.640			U	
PFMOAA		674-13-5	S301224017	ND	0.324	0.640			U	
PFMOBA		863090-89-5	S301224017	ND	1.07	1.44			U	
PFMOPrA		377-73-1	S301224017	ND	0.228	0.640			U	
PFO2HxA		39492-88-1	S301224017	ND	0.206	0.640			U	
PFO3OA		39492-89-2	S301224017	ND	0.294	0.640			U	
PFO4DA		39492-90-5	S301224017	ND	0.506	3.20			U	
PFO5DA		39492-91-6	S301224017	ND	0.512	3.20			U	
PMPA		13140-29-9	S301224017	ND	0.151	0.640			U	
R-EVE		2416366-22-6	S301224017	ND	1.06	1.44			U	
PFESAs	11Cl-PF3OUdS	763051-92-9	S301224017	ND	0.302	0.603			U	
	9Cl-PF3ONS	756426-58-1	S301224017	ND	0.410	0.596			U	
	Hydrolyzed PSDA	2416366-19-1	S301224017	ND	0.426	0.640			U	
	Nafion Byproduct 1 (PS Acid)	29311-67-9	S301224017	ND	0.342	0.640			U	
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	S301224017	ND	0.530	0.640			U	
	NVHOS	1132933-86-8	S301224017	ND	0.0986	0.640			U	
	PFEESA	113507-82-7	S301224017	ND	0.192	0.640			U	
	R-PSDA	2416366-18-0	S301224017	ND	2.82	2.82			U	
	R-PSDCA	2416366-21-5	S301224017	ND	0.270	0.640			U	
ES	MPFBA		S301224017				20-150%	102%		
	M5PFPeA		S301224017				20-150%	103%		
	M3PFBS		S301224017				20-150%	98.5%		
	M2-4:2 FTS		S301224017				20-150%	133%		
	M5PFFhxA		S301224017				20-150%	112%		
	M3HFPO-DA		S301224017				20-150%	112%		
	M4PFFHpA		S301224017				20-150%	95.0%		

Enthalpy Analytical

Job No.: 1224-824-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	MB_18746_PFAS		
Sampling Site			
Enthalpy ID	MB_18746_PFAS	Prep Batch	EU18746
Matrix	aqueous	Analyst	ext-magennaef
Sampling Date		Instrument	Sauron
Received Date		Sample Vol mL	250
Prep Date	2024-12-30 09:54	Extract Vol mL	0.4
AnalysisDate	2024-12-30 18:52	Split Factor	N/A
SampleType	Blank	Method Code	EU-047-NPW
Bottle ID	-		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M3PFHxS		S301224017				20-150%	107%	
M2-6:2 FTS		S301224017				20-150%	99.3%	
M8PFOA		S301224017				20-150%	100%	
M9PFNA		S301224017				20-150%	101%	
M8PFOS		S301224017				20-150%	97.8%	
M2-8:2 FTS		S301224017				20-150%	92.1%	
M8FOSA-I		S301224017				20-150%	79.7%	
M6PFDA		S301224017				20-150%	103%	
d3-N-MeFOSAA		S301224017				20-150%	91.7%	
d5-N-EtFOSAA		S301224017				20-150%	85.4%	
M7PFUdA		S301224017				20-150%	90.7%	
MPFDoA		S301224017				20-150%	62.0%	
M2PFTeDA		S301224017				20-150%	28.8%	
d3-N-MeFOSA		S301224017				10-200%	12.7%	
d5-N-EtFOSA		S301224017				10-200%	12.1%	
d7-N-MeFOSE		S301224017				10-200%	61.8%	
d9-N-EtFOSE		S301224017				10-200%	50.4%	

Enthalpy Analytical

Job No.: 1224-824-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Details

Sample Name	MB_18749_PFAS		
Sampling Site			
Enthalpy ID	MB_18749_PFAS	Prep Batch	eu18749
Matrix	aqueous	Analyst	ext-magennaef
Sampling Date		Instrument	Bumblebee
Received Date		Sample Vol mL	0.1
Prep Date	2024-12-30 07:00	Extract Vol mL	0.2
AnalysisDate	2024-12-31 16:28	Split Factor	N/A
SampleType	Blank	Method Code	EU-047-NPW
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	B311224-12311628	465	700	1530			L
ES	13C3-PFPrA		B311224-12311628				20-150%	110%	

Enthalpy Analytical

Job No.: 1224-824-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Enthalpy ID	OPR_18746_PFAS	Prep Batch	EU18746	Sample Vol (mL)	250
Sample Name	OPR_18746_PFAS	Prep Date	2024-12-30 09:54	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	2024-12-30 19:14	Split Factor	N/A
Sampling Date		Analyst	ext-magennaef	Method Code	EU-047-NPW
Received Date		Instrument	Sauron	Sample Type	Control
		Bottle ID	-		

	Compound	CAS	InjFileName	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	S301224018	17.3	0.254	0.640	47.9-144%	86.4%	
	PFPeA	2706-90-3	S301224018	18.0	0.183	0.640	41.7-159%	89.8%	
	PFHxA	307-24-4	S301224018	18.9	0.214	0.640	43.2-154%	94.5%	
	PFHpA	375-85-9	S301224018	18.6	0.224	0.640	42.1-155%	93.2%	
	PFOA	335-67-1	S301224018	18.2	0.146	0.640	51.1-148%	90.8%	
	PFNA	375-95-1	S301224018	19.2	0.145	0.640	51.6-153%	96.2%	
	PFDA	335-76-2	S301224018	18.1	0.183	0.640	44.5-156%	90.7%	
	PFUnDA	2058-94-8	S301224018	16.3	0.145	0.640	40.3-156%	81.6%	
	PFDoDA	307-55-1	S301224018	18.3	0.260	0.640	40.4-158%	91.3%	
	PFTrDA	72629-94-8	S301224018	34.6	0.212	0.640	42.2-201%	173%	
	PFTeDA	376-06-7	S301224018	23.1	0.244	0.640	43-162%	116%	
	Sulfonates	PFBS	375-73-5	S301224018	16.2	0.340	0.640	42.7-155%	91.3%
PFPeS		2706-91-4	S301224018	14.8	0.131	0.603	40.3-152%	78.8%	
PFHxS		355-46-4	S301224018	15.2	0.494	0.586	45-148%	83.0%	
PFHpS		375-92-8	S301224018	20.7	0.310	0.610	39.8-166%	109%	
PFOS		1763-23-1	S301224018	16.8	0.338	0.593	59.2-132%	90.4%	
PFNS		68259-12-1	S301224018	14.5	0.199	0.616	38.1-153%	75.4%	
PFDS		335-77-3	S301224018	12.5	0.336	0.616	28.6-148%	64.9%	
4:2 FTS		757124-72-4	S301224018	17.3	0.0830	0.600	41.5-157%	92.3%	
6:2 FTS		27619-97-2	S301224018	17.1	0.302	0.610	44.5-160%	89.9%	
8:2 FTS		39108-34-4	S301224018	16.5	0.143	0.613	39.4-166%	86.2%	
Sulfonamidos	N-EtFOSAA	2991-50-6	S301224018	17.5	0.260	0.640	42.8-156%	87.5%	
	N-MeFOSAA	2355-31-9	S301224018	19.6	0.180	0.640	42-155%	97.8%	
	PFOSA	754-91-6	S301224018	18.9	0.0898	0.640	39.1-158%	94.3%	
PFECAs	HFPO-DA	13252-13-6	S301224018	21.8	0.0678	0.640	61.8-131%	109%	
ES	MPFBA		S301224018				20-150%	96.6%	
	M5PFPeA		S301224018				20-150%	105%	
	M3PFBS		S301224018				20-150%	103%	
	M2-4:2 FTS		S301224018				20-150%	129%	
	M5PFHxA		S301224018				20-150%	109%	
	M3HFPO-DA		S301224018				20-150%	111%	
	M4PFHpA		S301224018				20-150%	90.7%	
	M3PFHxS		S301224018				20-150%	109%	
	M2-6:2 FTS		S301224018				20-150%	108%	
	M8PFOA		S301224018				20-150%	97.5%	
	M9PFNA		S301224018				20-150%	87.7%	
	M8PFOS		S301224018				20-150%	86.6%	
	M2-8:2 FTS		S301224018				20-150%	79.2%	
	M8FOSA-I		S301224018				20-150%	68.2%	
	M6PFDA		S301224018				20-150%	89.6%	
	d3-N-MeFOSAA		S301224018				20-150%	78.6%	
	d5-N-EtFOSAA		S301224018				20-150%	75.8%	
	M7PFUDa		S301224018				20-150%	73.8%	
	MPFDa		S301224018				20-150%	50.7%	
	M2PFTeDA		S301224018				20-150%	17.6%	Q

Enthalpy Analytical

Job No.: 1224-824-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT LELAND N.C.

Enthalpy ID	OPR_18749_PFAS	Prep Batch	eu18749	Sample Vol (mL)	0.08
Sample Name	OPR_18749_PFAS	Prep Date	2024-12-30 07:00	Extract Vol (mL)	0.2
Matrix	aqueous	Analysis Date	2024-12-31 16:39	Split Factor	N/A
Sampling Date		Analyst	ext-magennaef	Method Code	EU-047-NPW
Received Date		Instrument	Bumblebee	Sample Type	Control
		Bottle ID	-		

	Compound	CAS	InjFileName	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	B311224-12311639	20500	875	1910	40-150%	82.1%	
ES	¹³ C3-PFPrA		B311224-12311639				20-150%	101%	

Sample Custody



1224-824

Chain of Custody Record

Enthalpy Ultratrace Job#: _____ COC Page 1 of 1

Special Handling:

- Standard Turn Around Time
- Rush Turn Around Time -- Date Needed _____

• All Fast TATs Subject to Approval by Enthalpy Analytical, Inc.
 • All Samples Disposed of After 6 months Unless Otherwise Instructed.
 Enthalpy Analytical-Wilmington, NC has added enhancements to standard methods to improve accuracy, precision and permit an assessment of laboratory performance in the context of your specific data needs. For more information email Cindy.James@enthalpy.com.

Client Name: <u>BRUNSWICK COUNTY UTILITIES</u>	Project Number: _____	PO#: _____
Project Manager: <u>GLENN WALKER</u>	Site Name: <u>NORTHWEST WATER PLANT</u>	Telephone#: _____
Report To: <u>SAME</u>	Location: <u>LELAND N.C.</u>	Email: _____

This Chain of Custody is applicable to Non-Air samples. Standard TAT differ per analysis and are provided by request.

Client Special Instructions:						Sample Containers				Analyses:						Notes:	
						# of Bottles	# of Jars	# of Bags	# Other	Method 1613	Method 8290	Method 1668A/BIC PCB	PFAS by LC/MS/MS	PAHs by HRGC/HRMS	Sample on Hold		Method 23
Sample ID	Date	Time	Sample Volume	Type	Matrix												
122024-S01	11/27/2024	11:30 AM	250 ml	G	NW	2										X	Please Add PFPa and
122024-E01	11/27/2024	11:30 AM	250 ml	G	DW	2											PFHpA To The Testing.
																	Mark Hager Knows About
																	This If you Have Questions

Relinquished By:	Date:	Received By:	Date:	Time:	Sample Temperature Upon Receipt:
PHIL MCCULLOCH	12/20/2024	<i>Meredith Curtis</i>	12/20/24	14:05	<input checked="" type="checkbox"/> Iced <input type="checkbox"/> Ambient °C <u>5.4</u>
					<input type="checkbox"/> Iced <input type="checkbox"/> Ambient °C _____
					<input type="checkbox"/> Iced <input type="checkbox"/> Ambient °C _____

JOB ID: 1224-824 Date / Time: 12/20/24 14:05 Initials: M.A.C.
 OR
 Client: Brunswick County Public Utilities

Cooler 1 of 1

Temp °C: 5.4 Thermometer ID: T15

Received via FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input checked="" type="checkbox"/> Other <input type="checkbox"/>	<i>Check one</i>		<i>Check one</i>			
	On ice:	<input checked="" type="checkbox"/>	in a Box:	<input type="checkbox"/>	Cooler seals:	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No
	Melted ice:	<input type="checkbox"/>	in a Cooler:	<input checked="" type="checkbox"/>	Sample seals:	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No
	Ambient:	<input type="checkbox"/>	Cooler in Box:	<input type="checkbox"/>	Good condition:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
	Comment:					

Cooler of

Temp °C: Thermometer ID:

Received via FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input type="checkbox"/> Other <input type="checkbox"/>	<i>Check one</i>		<i>Check one</i>			
	On ice:	<input type="checkbox"/>	in a Box:	<input type="checkbox"/>	Cooler seals:	<input type="checkbox"/> Yes <input type="checkbox"/> No
	Melted ice:	<input type="checkbox"/>	in a Cooler:	<input type="checkbox"/>	Sample seals:	<input type="checkbox"/> Yes <input type="checkbox"/> No
	Ambient:	<input type="checkbox"/>	Cooler in Box:	<input type="checkbox"/>	Good condition:	<input type="checkbox"/> Yes <input type="checkbox"/> No
	Comment:					

Cooler of

Temp °C: Thermometer ID:

Received via FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input type="checkbox"/> Other <input type="checkbox"/>	<i>Check one</i>		<i>Check one</i>			
	On ice:	<input type="checkbox"/>	in a Box:	<input type="checkbox"/>	Cooler seals:	<input type="checkbox"/> Yes <input type="checkbox"/> No
	Melted ice:	<input type="checkbox"/>	in a Cooler:	<input type="checkbox"/>	Sample seals:	<input type="checkbox"/> Yes <input type="checkbox"/> No
	Ambient:	<input type="checkbox"/>	Cooler in Box:	<input type="checkbox"/>	Good condition:	<input type="checkbox"/> Yes <input type="checkbox"/> No
	Comment:					