

County of Brunswick


3954 Clearwell Dr NE
Leland, NC 28451

Northwest Water Plant

Leland, NC
Samples Received: 12/10/2021

Analytical Report 1221-748

Isotope Dilution Method PFAS



Enthalpy Analytical, LLC – Ultratrace

Mark Hager

M: 910-520-2526 / F: 910-212-5866

mark.hager@enthalpy.com / www.enthalpy.com

2714 Exchange Drive, Wilmington, NC 28405

I certify that to the best of my knowledge all analytical data presented in this report:

- Have been checked for completeness
- Are accurate, error-free, and legible
- Have been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s)

This analytical report was prepared in Portable Document Format (.PDF) and contains _____ pages.

....."Report Issued Date: _____



Summary of Results



Enthalpy Analytical

Job No.: 1221-748-1 PFAS by Isotope Dilution (non-potable water)

County of Brunswick Client Proj.: N/A Site: Northwest Water Plant, Leland, NC

Summary

	Compound	CAS	121021-SO1 ng/L	121021-EO1 ng/L
Acids	PFBA	375-22-4	7.71	ND U
	PFPeA	2706-90-3	19.0	20.1
	PFHxA	307-24-4	13.6	14.2
	PFHpA	375-85-9	5.11	5.62
	PFOA	335-67-1	9.39	10.4
	PFNA	375-95-1	1.27	1.26
	PFDA	335-76-2	0.781	0.692
	PFUnDA	2058-94-8	0.262 J	0.223 J
	PFDoDA	307-55-1	0.132 LB	ND U
	PFTTrDA	72629-94-8	0.118 L	0.0823 L
	PFTeDA	376-06-7	0.0897 L	ND U
Sulfonates	PFBS	375-73-5	9.49	11.6
	PFPeS	2706-91-4	1.42	1.36
	PFHxS	355-46-4	7.60	7.29
	PFHpS	375-92-8	0.508	ND U
	PFOS	1763-23-1	11.8	14.2
	PFNS	68259-12-1	ND U	ND U
	PFDS	335-77-3	ND U	ND U
	4:2 FTS	757124-72-4	0.259	0.0890 L
	6:2 FTS	27619-97-2	0.391	0.386
	8:2 FTS	39108-34-4	0.0862 L	ND U
other	PFOSA	754-91-6	ND U	0.0554 L
	N-MeFOSAA	2355-31-9	0.310	0.174 J
	N-EiFOSAA	2991-50-6	0.182 J	0.131 J
	HFPO-DA	13252-13-6	15.9	10.5
	PFMOAA	674-13-5	83.8	76.2
	PFMOPrA	377-73-1	0.356	0.277
	PFO2HxA	39492-88-1	13.4	12.9
	PFO3OA	39492-89-2	3.87	2.82
	PFO4DA	39492-90-5	ND U	ND U
	Nafion Byproduct 1	29311-67-9	ND U	ND U
	ADONA	919005-14-4	0.0868 L	0.0238 L
	9CI-PF3ONS	756426-58-1	0.0631 L	0.0699 L
	11CI-PF3OUdS	763051-92-9	0.0912 L	0.105 L
	10:2 FTS	120226-60-0	ND U	ND U
	EVE Acid	69087-46-3	ND U	ND U
	FBSA	30334-69-1	1.39	1.42
	Hydro-EVE Acid	773804-62-9	0.652 LB	0.661 LB
	Hydrolyzed PSDA	2416366-19-1	20.6	20.7
	Nafion Byproduct 2	749836-20-2	ND U	ND U
	N-EiFOSA	4151-50-2	ND U	ND U
	N-EiFOSE	1691-99-2	ND U	ND U
	NFDHA	151772-58-6	ND U	ND U
	N-MeFOSA	31506-32-8	ND U	ND U
	N-MeFOSE	24448-09-7	ND U	ND U
	NVHOS	1132933-86-8	11.7	9.49
	PEPA	267239-61-2	ND U	ND U
	PFECA-G	801212-59-9	0.120 L	0.0297 L
	PFEESA	113507-82-7	0.174 L	0.0678 L
	PFHxDA	67905-19-5	ND U	ND U
	PFMOBA	863090-89-5	ND U	ND U
	PFO5DA	39492-91-6	ND U	ND U
PMPA	13140-29-9	7.24	5.67	
R-EVE	2416366-22-6	21.9	19.3	
R-PSDA	2416366-18-0	42.1	ND U	
R-PSDCA	2416366-21-5	0.0826 LB	0.0964 LB	

Detailed Results

Enthalpy Analytical

Job No.: 1221-748-1 PFAS by Isotope Dilution (non-potable water)
 County of Brunswick Client Proj.: N/A Site: Northwest Water Plant, Leland, NC

Enthalpy ID	1221-748-001-1	Prep Batch	EU12711	Sample Vol (mL)	288.18
Sample Name	121021-SO1	Prep Date	2021-12-13 15:33	Extract Vol (mL)	0.4
Matrix	Aqueous	Analysis Date	2021-12-14 06:41	Dilution Factor	1
Sampling Date	20211210 00:00	Analyst	brneff/tbrooker	Method Code	WM-026
Received Date	2021-12-10 14:37	Instrument	Kill	Sample Type	Sample

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	5553.40	7.71	7.71	0.133	0.264			
	PFPeA	2706-90-3	13701.78	19.0	19.0	0.147	0.264			
	PFHxA	307-24-4	9804.72	13.6	13.6	0.167	0.264			
	PFHpA	375-85-9	3684.99	5.11	5.11	0.106	0.264			
	PFOA	335-67-1	6761.81	9.39	9.39	0.154	0.264			
	PFNA	375-95-1	911.43	1.27	1.27	0.0660	0.264			
	PFDA	335-76-2	562.69	0.781	0.781	0.0733	0.264			
	PFUnDA	2058-94-8	189.07	0.262	0.262	0.160	0.264			J
	PFDoDA	307-55-1	95.25	0.132	0.132	0.175	0.264			LB
	PFTrDA	72629-94-8	85.32	0.118	0.118	0.131	0.264			L
PFTeDA	376-06-7	64.65	0.0897	0.0897	0.189	0.264			L	
Sulfonates	PFBS	375-73-5	6833.85	9.49	9.49	0.308	0.308			
	PFPeS	2706-91-4	1025.57	1.42	1.42	0.179	0.248			
	PFHxS	355-46-4	5475.51	7.60	7.60	0.166	0.242			
	PFHpS	375-92-8	365.68	0.508	0.508	0.117	0.251			
	PFOS	1763-23-1	8479.12	11.8	11.8	0.139	0.244			
	PFNS	68259-12-1	ND	ND	ND	0.0750	0.254			U
	PFDS	335-77-3	ND	ND	ND	0.167	0.254			U
	4:2 FTS	757124-72-4	186.53	0.259	0.259	0.102	0.247			
	6:2 FTS	27619-97-2	281.35	0.391	0.391	0.101	0.251			
	8:2 FTS	39108-34-4	62.10	0.0862	0.0862	0.148	0.253			L
other	PFOSA	754-91-6	ND	ND	ND	0.113	0.264			U
	N-MeFOSAA	2355-31-9	223.60	0.310	0.310	0.125	0.264			
	N-EiFOSAA	2991-50-6	131.40	0.182	0.182	0.0946	0.264			J
	HFPO-DA	13252-13-6	11464.01	15.9	15.9	0.198	0.264			
	PFMOAA	674-13-5	60361.09	83.8	83.8	1.25	1.25			
	PFMOPrA	377-73-1	256.72	0.356	0.356	0.208	0.264			
	PFO2HxA	39492-88-1	9633.77	13.4	13.4	1.25	1.25			
	PFO3OA	39492-89-2	2789.15	3.87	3.87	1.25	1.25			
	PFO4DA	39492-90-5	ND	ND	ND	1.32	1.32			U
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.264	0.264			U
	ADONA	919005-14-4	62.55	0.0868	0.0868	0.104	0.250			L
	9Cl-PF3ONS	756426-58-1	45.47	0.0631	0.0631	0.104	0.246			L
	11Cl-PF3OUds	763051-92-9	65.69	0.0912	0.0912	0.104	0.248			L
	10:2 FTS	120226-60-0	ND	ND	ND	0.208	0.264			U
	EVE Acid	69087-46-3	ND	ND	ND	1.25	1.25			U
	FBSA	30334-69-1	1003.08	1.39	1.39	0.208	0.264			
	Hydro-EVE Acid	773804-62-9	470.09	0.652	0.652	1.25	1.25			LB
	Hydrolyzed PSDA	2416366-19-1	14827.07	20.6	20.6	1.25	1.25			
	Nafion Byproduct 2	749836-20-2	ND	ND	ND	0.264	0.264			U
	N-EiFOSA	4151-50-2	ND	ND	ND	0.208	0.264			U
	N-EiFOSE	1691-99-2	ND	ND	ND	6.25	6.25			U
	NFDHA	151772-58-6	ND	ND	ND	0.208	0.264			U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.208	0.264			U
	N-MeFOSE	24448-09-7	ND	ND	ND	6.25	6.25			U
	NVHOS	1132933-86-8	8430.16	11.7	11.7	1.25	1.25			
	PEPA	267239-61-2	ND	ND	ND	1.25	1.25			U
	PFECA-G	801212-59-9	86.16	0.120	0.120	0.264	1.25			L
	PFEESA	113507-82-7	125.57	0.174	0.174	0.208	0.264			L
	PFHxDA	67905-19-5	ND	ND	ND	1.25	1.25			U
	PFMOBA	863090-89-5	ND	ND	ND	1.25	1.25			U
PFOSDA	39492-91-6	ND	ND	ND	1.32	1.32			U	
PMPA	13140-29-9	5212.99	7.24	7.24	1.25	1.25				
R-EVE	2416366-22-6	15746.98	21.9	21.9	1.25	1.25				
R-PSDA	2416366-18-0	30365.75	42.1	42.1	1.25	1.25				
R-PSDCA	2416366-21-5	59.49	0.0826	0.0826	1.25	1.25			LB	
ES	MPFBA		4161.59	5.78				20-150%	83.2%	
	M5PFPeA		7240.80	10.1				20-150%	144.8%	
	M3PFBS		9500.41	13.2				20-150%	190.0%	Q
	M2-4:2 FTS		5674.52	7.88				20-150%	113.5%	
	M5PFHxA		4977.33	6.91				20-150%	99.5%	
	M3HFPO-DA		3490.69	4.85				20-150%	69.8%	
	M4PFHpA		3879.43	5.38				20-150%	77.6%	
	M3PFHxS		3226.22	4.48				20-150%	64.5%	
	M2-6:2 FTS		2582.70	3.58				20-150%	51.7%	
	M8PFOA		4002.50	5.56				20-150%	80.0%	
	M9PFNA		3822.83	5.31				20-150%	76.5%	
	M8PFOS		3786.08	5.26				20-150%	75.7%	
	M2-8:2 FTS		4094.05	5.68				20-150%	81.9%	
	M8FOSA-I		3367.98	4.67				20-150%	67.4%	
	M6PFDA		3858.67	5.36				20-150%	77.2%	
	d3-N-MeFOSAA		3296.97	4.58				20-150%	65.9%	
	d5-N-EiFOSAA		3701.57	5.14				20-150%	74.0%	
	M7PFUDa		3624.52	5.03				20-150%	72.5%	
	MPFDaA		2868.09	3.98				20-150%	57.4%	
	M2PFTeDA		1785.91	2.48				20-150%	35.7%	


Enthalpy Analytical

Job No.: 1221-748-1 PFAS by Isotope Dilution (non-potable water)
 County of Brunswick Client Proj.: N/A Site: Northwest Water Plant, Leland, NC

Enthalpy ID	1221-748-002-1	Prep Batch	EU12711	Sample Vol (mL)	283.38
Sample Name	121021-EO1	Prep Date	2021-12-13 15:33	Extract Vol (mL)	0.4
Matrix	Aqueous	Analysis Date	2021-12-14 07:04	Dilution Factor	1
Sampling Date	20211210 00:00	Analyst	brneff/tbrooker	Method Code	WM-026
Received Date	2021-12-10 14:37	Instrument	Kill	Sample Type	Sample

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	ND	ND	ND	0.135	0.268			U
	PFPeA	2706-90-3	14274.82	20.1	20.1	0.150	0.268			
	PFHxA	307-24-4	10072.26	14.2	14.2	0.170	0.268			
	PFHpA	375-85-9	3981.13	5.62	5.62	0.108	0.268			
	PFOA	335-67-1	7393.82	10.4	10.4	0.156	0.268			
	PFNA	375-95-1	891.95	1.26	1.26	0.0671	0.268			
	PFDA	335-76-2	490.27	0.692	0.692	0.0745	0.268			
	PFUnDA	2058-94-8	157.89	0.223	0.223	0.163	0.268			J
	PFDoDA	307-55-1	ND	ND	ND	0.178	0.268			U
	PFTrDA	72629-94-8	58.31	0.0823	0.0823	0.133	0.268			L
PFTeDA	376-06-7	ND	ND	ND	0.192	0.268			U	
Sulfonates	PFBS	375-73-5	8252.18	11.6	11.6	0.313	0.313			
	PFPeS	2706-91-4	962.60	1.36	1.36	0.182	0.253			
	PFHxS	355-46-4	5167.43	7.29	7.29	0.169	0.246			
	PFHpS	375-92-8	ND	ND	ND	0.119	0.255			U
	PFOS	1763-23-1	10090.05	14.2	14.2	0.141	0.248			
	PFNS	68259-12-1	ND	ND	ND	0.0762	0.258			U
	PFDS	335-77-3	ND	ND	ND	0.169	0.258			U
	4:2 FTS	757124-72-4	63.09	0.0890	0.0890	0.104	0.251			L
	6:2 FTS	27619-97-2	273.36	0.386	0.386	0.102	0.255			
	8:2 FTS	39108-34-4	ND	ND	ND	0.151	0.257			U
other	PFOSA	754-91-6	39.25	0.0554	0.0554	0.115	0.268			L
	N-MeFOSAA	2355-31-9	123.19	0.174	0.174	0.127	0.268			J
	N-EiFOSAA	2991-50-6	92.65	0.131	0.131	0.0962	0.268			J
	HFPO-DA	13252-13-6	7449.03	10.5	10.5	0.201	0.268			
	PFMOAA	674-13-5	53972.79	76.2	76.2	1.27	1.27			
	PFMOPrA	377-73-1	196.42	0.277	0.277	0.212	0.268			
	PFO2HxA	39492-88-1	9161.35	12.9	12.9	1.27	1.27			
	PFO3OA	39492-89-2	2001.22	2.82	2.82	1.27	1.27			
	PFO4DA	39492-90-5	ND	ND	ND	1.34	1.34			U
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.268	0.268			U
	ADONA	919005-14-4	16.84	0.0238	0.0238	0.106	0.254			L
	9Cl-PF3ONS	756426-58-1	49.52	0.0699	0.0699	0.106	0.250			L
	11Cl-PF3OUdS	763051-92-9	74.35	0.105	0.105	0.106	0.253			L
	10:2 FTS	120226-60-0	ND	ND	ND	0.212	0.268			U
	EVE Acid	69087-46-3	ND	ND	ND	1.27	1.27			U
	FBSA	30334-69-1	1006.66	1.42	1.42	0.212	0.268			
	Hydro-EVE Acid	773804-62-9	468.22	0.661	0.661	1.27	1.27			LB
	Hydrolyzed PSDA	2416366-19-1	14696.75	20.7	20.7	1.27	1.27			
	Nafion Byproduct 2	749836-20-2	ND	ND	ND	0.268	0.268			U
	N-EiFOSA	4151-50-2	ND	ND	ND	0.212	0.268			U
	N-EiFOSE	1691-99-2	ND	ND	ND	6.35	6.35			U
	NFDHA	151772-58-6	ND	ND	ND	0.212	0.268			U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.212	0.268			U
	N-MeFOSE	24448-09-7	ND	ND	ND	6.35	6.35			U
	NVHOS	1132933-86-8	6725.70	9.49	9.49	1.27	1.27			
	PEPA	267239-61-2	ND	ND	ND	1.27	1.27			U
	PFCEA-G	801212-59-9	21.04	0.0297	0.0297	0.268	1.27			L
	PFEEESA	113507-82-7	48.00	0.0678	0.0678	0.212	0.268			L
	PFHXDA	67905-19-5	ND	ND	ND	1.27	1.27			U
	PFMOBA	863090-89-5	ND	ND	ND	1.27	1.27			U
PFOSDA	39492-91-6	ND	ND	ND	1.34	1.34			U	
PMPA	13140-29-9	4014.50	5.67	5.67	1.27	1.27				
R-EVE	2416366-22-6	13641.67	19.3	19.3	1.27	1.27				
R-PSDA	2416366-18-0	ND	ND	ND	1.27	1.27			U	
R-PSDCA	2416366-21-5	68.30	0.0964	0.0964	1.27	1.27			LB	
ES	MPFBA		3829.39	5.41				20-150%	76.6%	
	M5PFPeA		6728.96	9.50				20-150%	134.6%	
	M3PFBS		8098.67	11.4				20-150%	162.0%	Q
	M2-4:2 FTS		6645.10	9.38				20-150%	132.9%	
	M5PFHxA		3873.83	5.47				20-150%	77.5%	
	M3HFPO-DA		3984.67	5.62				20-150%	79.7%	
	M4PFHpA		3814.35	5.38				20-150%	76.3%	
	M3PFHxS		3591.79	5.07				20-150%	71.8%	
	M2-6:2 FTS		4606.33	6.50				20-150%	92.1%	
	M8PFOA		4115.81	5.81				20-150%	82.3%	
	M9PFNA		3623.62	5.11				20-150%	72.5%	
	M8PFOS		3106.31	4.38				20-150%	62.1%	
	M2-8:2 FTS		3090.17	4.36				20-150%	61.8%	
	M8FOSA-I		2944.09	4.16				20-150%	58.9%	
	M6PFDA		3792.47	5.35				20-150%	75.8%	
	d3-N-MeFOSAA		3395.27	4.79				20-150%	67.9%	
	d5-N-EiFOSAA		3554.86	5.02				20-150%	71.1%	
	M7PFUDa		3348.40	4.73				20-150%	67.0%	
	MPFDcA		2819.47	3.98				20-150%	56.4%	
	M2PFTeDA		1752.03	2.47				20-150%	35.0%	

QC Data



Enthalpy Analytical

Job No.: 1221-748-1 PFAS by Isotope Dilution (non-potable water)
 County of Brunswick Client Proj.: N/A Site: Northwest Water Plant, Leland, NC

Enthalpy ID	MB-12711-PFAS	Prep Batch	EU12711	Sample Vol (mL)	250
Sample Name	MB-12711-PFAS	Prep Date	2021-12-13 15:33	Extract Vol (mL)	0.4
Matrix	Aqueous	Analysis Date	2021-12-14 05:55	Dilution Factor	1
Sampling Date		Analyst	brneff/tbrooker	Method Code	WM-026
Received Date		Instrument	Kili	Sample Type	Blank

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	ND	ND	ND	0.153	0.304			U	
	PFPeA	2706-90-3	ND	ND	ND	0.170	0.304			U	
	PFHxA	307-24-4	17.56	0.0281	0.0281	0.193	0.304			L	
	PFHpA	375-85-9	ND	ND	ND	0.122	0.304			U	
	PFOA	335-67-1	11.33	0.0181	0.0181	0.177	0.304			L	
	PFNA	375-95-1	ND	ND	ND	0.0761	0.304			U	
	PFDA	335-76-2	ND	ND	ND	0.0845	0.304			U	
	PFUnDA	2058-94-8	ND	ND	ND	0.185	0.304			U	
	PFDoDA	307-55-1	13.23	0.0212	0.0212	0.202	0.304			L	
	PFTrDA	72629-94-8	ND	ND	ND	0.151	0.304			U	
	PFTeDA	376-06-7	ND	ND	ND	0.218	0.304			U	
	PFBS	375-73-5	ND	ND	ND	0.355	0.355			U	
	PFPeS	2706-91-4	ND	ND	ND	0.206	0.286			U	
	PFHxS	355-46-4	ND	ND	ND	0.191	0.278			U	
PFHpS	375-92-8	ND	ND	ND	0.135	0.290			U		
Sulfonates	PFOS	1763-23-1	ND	ND	ND	0.160	0.282			U	
	PFNS	68259-12-1	ND	ND	ND	0.0864	0.293			U	
	PFDS	335-77-3	ND	ND	ND	0.192	0.293			U	
	4:2 FTS	757124-72-4	ND	ND	ND	0.118	0.285			U	
	6:2 FTS	27619-97-2	ND	ND	ND	0.116	0.290			U	
	8:2 FTS	39108-34-4	ND	ND	ND	0.171	0.291			U	
	PFOSA	754-91-6	ND	ND	ND	0.130	0.304			U	
	N-MeFOSAA	2355-31-9	ND	ND	ND	0.144	0.304			U	
	N-EiFOSAA	2991-50-6	ND	ND	ND	0.109	0.304			U	
	HFPO-DA	13252-13-6	ND	ND	ND	0.228	0.304			U	
other	PFMOAA	674-13-5	ND	ND	ND	1.44	1.44			U	
	PFMOPrA	377-73-1	ND	ND	ND	0.240	0.304			U	
	PFO2HxA	39492-88-1	ND	ND	ND	1.44	1.44			U	
	PFO3OA	39492-89-2	ND	ND	ND	1.44	1.44			U	
	PFO4DA	39492-90-5	ND	ND	ND	1.52	1.52			U	
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.304	0.304			U	
	ADONA	919005-14-4	ND	ND	ND	0.120	0.288			U	
	9Cl-PF3ONS	756426-58-1	ND	ND	ND	0.120	0.283			U	
	11Cl-PF3OUdS	763051-92-9	ND	ND	ND	0.120	0.286			U	
	10:2 FTS	120226-60-0	ND	ND	ND	0.240	0.304			U	
	EVE Acid	69087-46-3	50.66	0.0811	0.0811	1.44	1.44			L	
	FBSA	30334-69-1	ND	ND	ND	0.240	0.304			U	
	Hydro-EVE Acid	773804-62-9	55.73	0.0892	0.0892	1.44	1.44			L	
	Hydrolyzed PSDA	2416366-19-1	ND	ND	ND	1.44	1.44			U	
	Nafion Byproduct 2	749836-20-2	ND	ND	ND	0.304	0.304			U	
	N-EiFOSA	4151-50-2	ND	ND	ND	0.240	0.304			U	
	N-EiFOSE	1691-99-2	ND	ND	ND	7.20	7.20			U	
	NFDHA	151772-58-6	ND	ND	ND	0.240	0.304			U	
	N-MeFOSA	31506-32-8	ND	ND	ND	0.240	0.304			U	
	N-MeFOSE	24448-09-7	ND	ND	ND	7.20	7.20			U	
	NVHOS	1132933-86-8	ND	ND	ND	1.44	1.44			U	
	PEPA	267239-61-2	ND	ND	ND	1.44	1.44			U	
	PFECA-G	801212-59-9	ND	ND	ND	0.304	1.44			U	
	PFEESA	113507-82-7	ND	ND	ND	0.240	0.304			U	
	PFHxD	67905-19-5	ND	ND	ND	1.44	1.44			U	
	PFMOBA	863090-89-5	ND	ND	ND	1.44	1.44			U	
	PFO5DA	39492-91-6	ND	ND	ND	1.52	1.52			U	
	PMPA	13140-29-9	ND	ND	ND	1.44	1.44			U	
	R-EVE	2416366-22-6	ND	ND	ND	1.44	1.44			U	
	R-PSDA	2416366-18-0	ND	ND	ND	1.44	1.44			U	
	R-PSDCA	2416366-21-5	55.27	0.0884	0.0884	1.44	1.44			L	
	ES	MPFBA		3529.93	5.65				20-150%	70.6%	
		M5PFPeA		4007.43	6.41				20-150%	80.1%	
		M3PFBS		2994.14	4.79				20-150%	59.9%	
M2-4:2 FTS			3416.93	5.47				20-150%	68.3%		
M5PFHxA			5186.62	8.30				20-150%	103.7%		
M3HFPO-DA			3373.84	5.40				20-150%	67.5%		
M4PFHpA			3636.19	5.82				20-150%	72.7%		
M3PFHxS			3998.83	6.40				20-150%	80.0%		
M2-6:2 FTS			2551.51	4.08				20-150%	51.0%		
M8PFOA			3991.76	6.39				20-150%	79.8%		
M9PFNA			3452.46	5.52				20-150%	69.0%		
M8PFOS			3425.99	5.48				20-150%	68.5%		
M2-8:2 FTS			4516.96	7.23				20-150%	90.3%		
M8FOSA-I			3642.22	5.83				20-150%	72.8%		
M6PFDA			3974.96	6.36				20-150%	79.5%		
d3-N-MeFOSAA			3153.88	5.05				20-150%	63.1%		
d5-N-EiFOSAA			3155.49	5.05				20-150%	63.1%		
M7PFUdA			3956.64	6.33				20-150%	79.1%		
MPFDoA			3432.34	5.49				20-150%	68.6%		
M2PFTeDA			2302.38	3.68				20-150%	46.0%		

Enthalpy Analytical

Job No.: 1221-748-1 PFAS by Isotope Dilution (non-potable water)
 County of Brunswick Client Proj.: N/A Site: Northwest Water Plant, Leland, NC

Enthalpy ID	MB-12711-PFAS	Prep Batch	EU12711	Sample Vol (mL)	250
Sample Name	MB-12711-PFAS	Prep Date	2021-12-13 15:33	Extract Vol (mL)	0.4
Matrix	Aqueous	Analysis Date	2021-12-14 05:55	Dilution Factor	1
Sampling Date		Analyst	brneff/tbrooker	Method Code	WM-026
Received Date		Instrument	Kili	Sample Type	Blank

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	ND	ND	ND	0.153	0.304			U
	PFPeA	2706-90-3	ND	ND	ND	0.170	0.304			U
	PFHxA	307-24-4	17.56	0.0281	0.0281	0.193	0.304			L
	PFHpA	375-85-9	ND	ND	ND	0.122	0.304			U
	PFOA	335-67-1	11.33	0.0181	0.0181	0.177	0.304			L
	PFNA	375-95-1	ND	ND	ND	0.0761	0.304			U
	PFDA	335-76-2	ND	ND	ND	0.0845	0.304			U
	PFUnDA	2058-94-8	ND	ND	ND	0.185	0.304			U
	PFDoDA	307-55-1	13.23	0.0212	0.0212	0.202	0.304			L
	PFTrDA	72629-94-8	ND	ND	ND	0.151	0.304			U
	PFTeDA	376-06-7	ND	ND	ND	0.218	0.304			U
	PFBS	375-73-5	ND	ND	ND	0.355	0.355			U
	PFPeS	2706-91-4	ND	ND	ND	0.206	0.286			U
	PFHxS	355-46-4	ND	ND	ND	0.191	0.278			U
PFHpS	375-92-8	ND	ND	ND	0.135	0.290			U	
Sulfonates	PFOS	1763-23-1	ND	ND	ND	0.160	0.282			U
	PFNS	68259-12-1	ND	ND	ND	0.0864	0.293			U
	PFDS	335-77-3	ND	ND	ND	0.192	0.293			U
	4:2 FTS	757124-72-4	ND	ND	ND	0.118	0.285			U
	6:2 FTS	27619-97-2	ND	ND	ND	0.116	0.290			U
	8:2 FTS	39108-34-4	ND	ND	ND	0.171	0.291			U
	PFOSA	754-91-6	ND	ND	ND	0.130	0.304			U
	N-MeFOSAA	2355-31-9	ND	ND	ND	0.144	0.304			U
	N-EiFOSAA	2991-50-6	ND	ND	ND	0.109	0.304			U
	HFPO-DA	13252-13-6	ND	ND	ND	0.228	0.304			U
Other	PFMOAA	674-13-5	ND	ND	ND	1.44	1.44			U
	PFMOPrA	377-73-1	ND	ND	ND	0.240	0.304			U
	PFO2HxA	39492-88-1	ND	ND	ND	1.44	1.44			U
	PFO3OA	39492-89-2	ND	ND	ND	1.44	1.44			U
	PFO4DA	39492-90-5	ND	ND	ND	1.52	1.52			U
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.304	0.304			U
	ADONA	919005-14-4	ND	ND	ND	0.120	0.288			U
	9Cl-PF3ONS	756426-58-1	ND	ND	ND	0.120	0.283			U
	11Cl-PF3OUdS	763051-92-9	ND	ND	ND	0.120	0.286			U
	10:2 FTS	120226-60-0	ND	ND	ND	0.240	0.304			U
	EVE Acid	69087-46-3	50.66	0.0811	0.0811	1.44	1.44			L
	FBSA	30334-69-1	ND	ND	ND	0.240	0.304			U
	Hydro-EVE Acid	773804-62-9	55.73	0.0892	0.0892	1.44	1.44			L
	Hydrolyzed PSDA	2416366-19-1	ND	ND	ND	1.44	1.44			U
	Nafion Byproduct 2	749836-20-2	ND	ND	ND	0.304	0.304			U
	N-EiFOSA	4151-50-2	ND	ND	ND	0.240	0.304			U
	N-EiFOSE	1691-99-2	ND	ND	ND	7.20	7.20			U
	NFDHA	151772-58-6	ND	ND	ND	0.240	0.304			U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.240	0.304			U
	N-MeFOSE	24448-09-7	ND	ND	ND	7.20	7.20			U
	NVHOS	1132933-86-8	ND	ND	ND	1.44	1.44			U
	PEPA	267239-61-2	ND	ND	ND	1.44	1.44			U
	PFEGA-G	801212-59-9	ND	ND	ND	0.304	1.44			U
	PFEESA	113507-82-7	ND	ND	ND	0.240	0.304			U
	PFHxDA	67905-19-5	ND	ND	ND	1.44	1.44			U
	PFMOBA	863090-89-5	ND	ND	ND	1.44	1.44			U
	PFO5DA	39492-91-6	ND	ND	ND	1.52	1.52			U
	PMPA	13140-29-9	ND	ND	ND	1.44	1.44			U
	R-EVE	2416366-22-6	ND	ND	ND	1.44	1.44			U
	R-PSDA	2416366-18-0	ND	ND	ND	1.44	1.44			U
	R-PSDCA	2416366-21-5	55.27	0.0884	0.0884	1.44	1.44			L
	ES	MPFBA		3529.93	5.65				20-150%	70.6%
M5PFPeA			4007.43	6.41				20-150%	80.1%	
M3PFBS			2994.14	4.79				20-150%	59.9%	
M2-4:2 FTS			3416.93	5.47				20-150%	68.3%	
M5PFHxA			5186.62	8.30				20-150%	103.7%	
M3HFPO-DA			3373.84	5.40				20-150%	67.5%	
M4PFHpA			3636.19	5.82				20-150%	72.7%	
M3PFHxS			3998.83	6.40				20-150%	80.0%	
M2-6:2 FTS			2551.51	4.08				20-150%	51.0%	
M8PFOA			3991.76	6.39				20-150%	79.8%	
M9PFNA			3452.46	5.52				20-150%	69.0%	
M8PFOS			3425.99	5.48				20-150%	68.5%	
M2-8:2 FTS			4516.96	7.23				20-150%	90.3%	
M8FOSA-I			3642.22	5.83				20-150%	72.8%	
M6PFDA			3974.96	6.36				20-150%	79.5%	
d3-N-MeFOSAA			3153.88	5.05				20-150%	63.1%	
d5-N-EiFOSAA			3155.49	5.05				20-150%	63.1%	
M7PFUDA			3956.64	6.33				20-150%	79.1%	
MPFDaA			3432.34	5.49				20-150%	68.6%	
M2PFTeDA			2302.38	3.68				20-150%	46.0%	

Enthalpy Analytical

Job No.: 1221-748-1 PFAS by Isotope Dilution (non-potable water)

County of Brunswick Client Proj.: N/A Site: Northwest Water Plant, Leland, NC

Enthalpy ID	OPR-12711-PFAS	Prep Batch	EU12711	Sample Vol (mL)	250
Sample Name	OPR-12711-PFAS	Prep Date	2021-12-13 15:33	Extract Vol (mL)	0.4
Matrix	Aqueous	Analysis Date	2021-12-14 06:18	Dilution Factor	1
Sampling Date		Analyst	brneff/itbrooker	Method Code	WM-026
Received Date		Instrument	Kili	Sample Type	Control

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	12458.18	19.9	19.9	0.153	0.304	73-129%	99.7%	
	PFPeA	2706-90-3	13712.72	21.9	21.9	0.170	0.304	72-129%	109.7%	
	PFHxA	307-24-4	13463.56	21.5	21.5	0.193	0.304	72-129%	107.7%	
	PFHpA	375-85-9	12423.13	19.9	19.9	0.122	0.304	72-130%	99.4%	
	PFOA	335-67-1	13211.26	21.1	21.1	0.177	0.304	71-133%	105.7%	
	PFNA	375-95-1	13264.52	21.2	21.2	0.0761	0.304	69-130%	106.1%	
	PFDA	335-76-2	12717.74	20.3	20.3	0.0845	0.304	71-129%	101.7%	
	PFUnDA	2058-94-8	13650.80	21.8	21.8	0.185	0.304	69-133%	109.2%	
	PFDoDA	307-55-1	13456.44	21.5	21.5	0.202	0.304	72-134%	107.7%	
	PFTeDA	72629-94-8	19073.64	30.5	30.5	0.151	0.304	65-144%	152.6%	Q
Sulfonates	PFBS	375-73-5	12984.26	20.8	20.8	0.355	0.355	72-134%	117.1%	
	PFPeS	2706-91-4	14454.48	23.1	23.1	0.206	0.286	71-127%	122.9%	
	PFHxS	355-46-4	14353.97	23.0	23.0	0.191	0.278	68-131%	125.6%	
	PFHpS	375-92-8	13569.20	21.7	21.7	0.135	0.290	69-134%	113.9%	
	PFOS	1763-23-1	11658.25	18.7	18.7	0.160	0.282	65-140%	100.5%	
	PFNS	68259-12-1	10971.53	17.6	17.6	0.0864	0.293	69-127%	91.2%	
	PFDS	335-77-3	10216.76	16.3	16.3	0.192	0.293	53-142%	84.7%	
	4:2 FTS	757124-72-4	10925.49	17.5	17.5	0.118	0.285	63-143%	93.3%	
6:2 FTS	27619-97-2	13844.48	22.2	22.2	0.116	0.290	64-140%	116.5%		
8:2 FTS	39108-34-4	11256.66	18.0	18.0	0.171	0.291	67-138%	93.8%		
Other	PFOSA	754-91-6	15094.20	24.2	24.2	0.130	0.304	67-137%	120.8%	
	N-MeFOSAA	2355-31-9	13852.84	22.2	22.2	0.144	0.304	65-136%	110.8%	
	N-EtFOSAA	2991-50-6	12615.03	20.2	20.2	0.109	0.304	61-135%	100.9%	
	HFPO-DA	13252-13-6	11553.12	18.5	18.5	0.228	0.304	70-130%	92.4%	
ES	MPFBA		3845.95	6.15				20-150%	76.9%	
	M5PFPeA		3782.06	6.05				20-150%	75.6%	
	M3PFBS		3136.67	5.02				20-150%	62.7%	
	M2-4:2 FTS		4101.49	6.56				20-150%	82.0%	
	M5PFHxA		4075.78	6.52				20-150%	81.5%	
	M3HFPO-DA		4846.05	7.75				20-150%	96.9%	
	M4PFHpA		3924.38	6.28				20-150%	78.5%	
	M3PFHxS		3381.54	5.41				20-150%	67.6%	
	M2-6:2 FTS		2759.12	4.41				20-150%	55.2%	
	M8PFOA		3908.17	6.25				20-150%	78.2%	
	M9PFNA		3561.04	5.70				20-150%	71.2%	
	M8PFOS		3971.15	6.35				20-150%	79.4%	
	M2-8:2 FTS		2575.71	4.12				20-150%	51.5%	
	M8FOSA-I		2905.33	4.65				20-150%	58.1%	
	M6PFDA		3779.84	6.05				20-150%	75.6%	
	d3-N-MeFOSAA		3497.42	5.60				20-150%	69.9%	
	d5-N-EtFOSAA		3626.81	5.80				20-150%	72.5%	
	M7PFUdA		3222.56	5.16				20-150%	64.5%	
MPFDoA		2987.59	4.78				20-150%	59.8%		
M2PFTeDA		1811.11	2.90				20-150%	36.2%		

Narrative Summary



Enthalpy Analytical Narrative Summary

Company	County of Brunswick
Job No.	1221-712-1 PFAS by Isotope Dilution (non-potable water)
Client ID.	N/A Site: Northwest Water Plant, Leland, NC

1. Custody

Dallas King received the samples on December 02, 2021 at 2.9 °C after being relinquished by County of Brunswick. The samples were received in good condition.

Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC

Table 1 - Sample Inventory

EU Lab Sample ID	Client Sample ID	Matrix
1221-712-001-1	120221-SO1	Aqueous
1221-712-002-1	120221-EO1	Aqueous

2. Methods and Analytes

A list of analytes of interest and corresponding methods of analysis is shown in Table 3. Abbreviations are defined in the listed Appendices.

Table 3 - Methods and Analytes

EU Method	Analytes	Cleanup Method
EU-047	Brunswick PFAS List	ENVI-Carb

3. Analysis

The samples were analyzed using Waters Acquity UPLC equipped with Xevo TQ MS (LC/MS/MS "Kili").

For aqueous samples, the sample volume was measured gravimetrically by the laboratory, and spiked with Extraction Standard (ES). The sample was then mixed well and centrifuged, if needed. The samples were then extracted via SPE, and the extracts were cleaned up using ENVI-Carb.

Each final sample extract was transferred to an autosampler vial, spiked with Injection Standard (IS), and brought to a final volume of 400µL prior to analysis.

Samples were run in more than one sequence in order to analyze all analytes of interest.

4. Calibration

In the initial calibration, the reported analytes exhibited R^2 of ≥ 0.99 . The reported analytes in the calibration standards, continuing calibration (concal) and Initial Calibration Verification (ICV) met the accuracy criterion for native analytes.

5. QC Notes

QC sample analyses passed all method criteria.

Enthalpy Analytical Narrative Summary

Company	County of Brunswick
Job No.	1221-712-1 PFAS by Isotope Dilution (non-potable water)
Client ID.	N/A Site: Northwest Water Plant, Leland, NC

*OPR-12681-PFAS PFHpA - Met method criteria. Recovery Limits are set to DoD QSM criteria but will follow SOP method criteria if the analytes fall outside QSM criteria but within SOP criteria.

The samples were extracted within the 28-day from collection holding time and analyzed within the 28-day from extraction to analysis holding time required by the method.

6. Reporting Notes

Some labeled extraction standards in the sample analyses fell outside the control limits for ES recovery, as denoted by the "Q" qualifier. The target analytes are quantified based on their ratio to their labeled standard analogs. As a result, low or high labeled standard recovery do not cause any change to ratios or contribute any additional error in the measurement of the target analytes. The data have been accepted and reported with no further actions.

The results presented in this report are representative of the samples as provided to the laboratory.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, LLC in Wilmington NC is accredited by the Louisiana Department of Environmental Quality to the 2009 TNI Standard under certificate number 05075.



General Reporting Notes – Data Qualifiers

The following are general reporting notes that are applicable to all Enthalpy Analytical, LLC - Wilmington, NC data reports, unless specifically noted otherwise.

General Data Qualifiers

- B – The analyte was found in the method blank, at a concentration that was at least 10% of the concentration in the sample.
- Cxx – Two or more congeners co-elute. In EDDs, C denotes the lowest IUPAC congener in a co-elution group and additional co-eluters for the group ('xx') are shown with the number of the lowest IUPAC co-eluter.
- E – The reported concentration exceeds the calibration range (upper point of the calibration curve). For HRMS data, this condition does not imply additional measurement uncertainty. For LC-MS/MS data, these values should be considered as having measurement uncertainty higher than values within the calibration range.
- EDL – Estimated Detection Level. Specific to Dioxin/Furan tests and equivalent to MDL
- EMPC – Estimated Maximum Possible Concentration Specific to Dioxin/Furan tests to indicate the signal/noise ratio was not sufficient for peak identification (the determined ion-abundance ratio was outside the allowed theoretical range), or where there was a co-eluting interference. Indicates that a peak was identified but did not meet the method specified ion-abundance ratio.
- IR – The ion ratio between the primary and secondary ions was observed to be outside the method criteria therefore the actual analyte concentration cannot be accurately determined as defined by DoD QSM Table B-15.
- J – The analyte has a concentration below the minimum calibration level (LOQ value) but greater than the LOD. These values should be considered as having measurement uncertainty higher than values within the calibration range
- L - Indicates that an analyte has a concentration below the Minimum Detection Limit (MDL). The reported concentration is not recommended for regulatory use as the analyte signal may have a signal-to-noise ratio less than the criteria deemed necessary to be considered a detected analyte.
- LOD – Limit of Detection: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOD. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the minimum detection limit (MDL). The LOD is adjusted for sample weight or volume.
- LOQ – Limit of Quantiation: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOQ. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the reporting limit (RL). The LOD is adjusted for sample weight or volume.
- <LOD() – Analyte was not found at a concentration high enough to be reported as detected. It is reported as less than the LOD, and the LOD is given in the parentheses.



General Reporting Notes – Data Qualifiers

- ND – Indicates a non-detect.
- NR – Indicates a value that is not reportable due to issues observed in sample preparation or analysis.
- PR – The associated congener(s) is(are) poorly resolved.
- QI – Indicates the presence of a quantitative interference.
- RL – Reporting Limit. Lowest reportable value. The level is higher than the MDL.
- SI – Denotes “Single Ion Mode” and is utilized for PCBs where the secondary ion trace has a significantly elevated noise level due to background PFK. Responses for such peaks are calculated using an EMPC approach based solely on the primary ion area(s) and may be considered estimates.
- U – The analyte was not detected.
- V / Q – The labeled standard recovery is not within method control limits.
- X – Results from re-injection/repeat/second-column analysis.

Lab Identifiers/ Data Attributes

- AR – Indicates use of the archived portion of the sample extract.
- CU – Indicates a sample that required additional clean-up prior to HRMS injection/processing.
- D – Dilution Data. Result was obtained from the analysis of a dilution. The number that follows the “D” indicates the dilution factor.
- DE – Indicates a dilution performed with the addition of ES (Extraction Standard) solution.
- DUP – Designation for a duplicate sample.
- MS – Designation for a matrix spike.
- MSD – Designation for a matrix spike duplicate.
- RJ – Indicates a reinjection of the sample extract.
- S – Indicates a sample split. The number that follows the “S” indicates the split factor.
- R – Indicates a re-extraction of the sample.

PFAS Compound Acronym List		
Acronym	CAS #	Compound Name
Target Analytes		
* Analyte is not accredited		
PFBA	375-22-4	Perfluorobutanoic Acid
PFPeA	2706-90-3	Perfluoropentanoic Acid
PFHxA	307-24-4	Perfluorohexanoic Acid
PFHpA	375-85-9	Perfluoroheptanoic Acid
PFOA	335-67-1	Perfluorooctanoic Acid
PFNA	375-95-1	Perfluorononanoic Acid
PFDA	335-76-2	Perfluorodecanoic acid
PFUnA (PFUnDA)	2058-94-8	Perfluoroundecanoic acid
PFDoA (PFDoDA)	307-55-1	Perfluorododecanoic acid
PFTriA (PFTriA)	72629-94-8	Perfluorotridecanoic acid
PFTeDA (PFTA)	376-06-7	Perfluorotetradecanoic acid
PFBS	375-73-5	Perfluorobutane sulfonic acid
PFPeS	2706-91-4	Perfluoropentane sulfonic acid
PFHxS	355-46-4	Perfluorohexane sulfonic acid
PFHpS	375-92-8	Perfluoroheptane sulfonic acid
PFOS	1763-23-1	Perfluorooctane sulfonic acid
PFNS	68259-12-1	Perfluorononane sulfonic acid
PFDS	335-77-3	Perfluorodecane sulfonic acid
4:2 FTS	757124-72-4	4:2 fluorotelomer sulfonic acid
6:2 FTS	27619-97-2	6:2 fluorotelomer sulfonic acid
8:2 FTS	39108-34-4	8:2 fluorotelomer sulfonic acid
PFOSA (FOSA)	754-91-6	Perfluorooctane sulfonamide
N-MeFOSAA	2355-31-9	N-methyl perfluorooctane sulfonamido acetic acid
N-EtFOSAA	2991-50-6	N-ethyl perfluorooctane sulfonamido acetic acid
HFPO-DA	13252-13-6	2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy)-propanoic acid (Gen-X)
11Cl-PF3OUdS	763051-92-9	11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid
9Cl-PF3ONS	756426-58-1	9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid
ADONA	919005-14-4	4,8-dioxa-3H-perfluorononanoic acid
* PFMOAA	674-13-5	Perfluoro-2-methoxyacetic acid
PFMOPrA (PFMPA)	377-73-1	Perfluoro-3-methoxypropanoic acid
* PFO2HxA	39492-88-1	Perfluoro (3,5-dioxahexanoic) acid
* PFO3OA	39492-89-2	Perfluoro (3,5,7-trioxaoctanoic) acid
* PFO4DA	39492-90-5	Perfluoro (3,5,7,9-tetraoxadecanoic) acid
* PFO5DA	39492-91-6	Perfluoro(3,5,7,9,11-pentaoxadodecanoic) acid
* Nafion Byproduct 1	29311-67-9	Nafion Byproduct 1
* Nafion Byproduct 2	749836-20-2	Nafion Byproduct 2
PFEESA	113507-82-7	Perfluoro(2-ethoxyethane)sulphonic acid
PFMOBA (PFMBA)	863090-89-5	Perfluoro-4-methoxybutanic acid
NFDHA	151772-58-6	Nonafluoro-3,6-dioxaheptanoic acid
* PEPA	267239-61-2	Perfluoro-2-ethoxypropanoic acid
* PMPA	13140-29-9	Perfluoro-2-methoxypropanoic acid
* 10:2 FTS	120226-60-0	Fluorotelomer sulfonate 10:2
* N-EtFOSA	4151-50-2	N-ethylperfluoro-1-octanesulfonamide
* N-EtFOSE	1691-99-2	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol
* N-MeFOSA	31506-32-8	N-methylperfluoro-1-octanesulfonamide
* N-MeFOSE	24448-09-7	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol
* PFECA-G	801212-59-9	4-(Heptafluoroisopropoxy)hexafluorobutanoic acid
* PFHxDA	67905-19-5	Perfluorohexadecanoic acid
* R-PSDA	2416366-18-0	Perfluoro-4-(2-sulfoethoxy)pentanoic acid

PFAS Compound Acronym List		
Acronym	CAS #	Compound Name
Target Analytes		
* Analyte is not accredited		
* Hydrolyzed PSDA	2416366-19-1	2-fluoro-2-[1,1,2,3,3,3-hexafluoro-2-(1,1,2,2-tetrafluoro-2-sulfoethoxy)propoxy]-acetic acid
* R-PSDCA	2416366-21-5	1,1,2,2-tetrafluoro-2-[1,2,2,3,3-pentafluoro-1-(trifluoromethyl)propoxy] ethanesulfonic acid
* EVE Acid	69087-46-3	2,2,3,3-tetrafluoro-3-({1,1,1,2,3,3-hexafluoro-3-[(1,2,2-trifluoroethenyl)oxy]propan-2-yl}oxy)propionic acid
* FBSA	30334-69-1	Perfluorobutylsulfonamide
* Hydro-EVE Acid	773804-62-9	2,2,3,3-Tetrafluoro-3-([1,1,1,2,3,3-hexafluoro-3-(1,2,2,2-tetrafluoroethoxy)propan-2-yl]oxy)propanoic acid
* R-EVE Acid	2416366-22-6	4-(2-carboxy-1,1,2,2-tetrafluoroethoxy)-2,2,3,3,4,5,5,5-octafluoro-pentanoic acid
* NVHOS	1132933-86-8	Perfluoroethoxysulfonic acid
* PFDoS	79780-39-5	Perfluorododecane sulfonic acid
* PFODA	16517-11-6	Perfluorooctadecanoic acid



Sample Custody



**This Is The Last Page
Of This Report.**

