

County of Brunswick


3954 Clearwell Dr NE
Leland, NC 28451

Northwest Water Plant

Leland, NC
Samples Received: 01/08/21

Analytical Report 0121-717

Isotope Dilution Method PFAS



Enthalpy Analytical, LLC – Ultratrace

Lindsay Boone
O: (910) 212-5855 / F: 910-212-5866
lboone@enthalpy.com / www.enthalpy.com
2714 Exchange Drive, Wilmington, NC 28405

I certify that to the best of my knowledge all analytical data presented in this report:

- Have been checked for completeness
- Are accurate, error-free, and legible
- Have been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s)

This analytical report was prepared in Portable Document Format (.PDF) and contains _____ pages.

....."Report Issued Date: _____"



Summary of Results



Enthalpy Analytical

Job No.: 0121-717-1 PFAS by Isotope Dilution (non-potable water)

County of Brunswick Site: Northwest Water Plant - Leland, NC

Summary

	Compound	CAS	010821-SO1 ng/L	010821-EO1 ng/L
Acids	PFBA	375-22-4	0.996	ND U
	PFPeA	2706-90-3	ND U	ND U
	PFHxA	307-24-4	2.05	2.34
	PFHpA	375-85-9	1.36	1.59
	PFOA	335-67-1	2.68	3.21
	PFNA	375-95-1	0.591	0.565
	PFDA	335-76-2	0.239 J	0.299
	PFUnDA	2058-94-8	0.177 J	0.101 J
	PFDoDA	307-55-1	ND U	ND U
	PFTTrDA	72629-94-8	ND U	ND U
	PFTeDA	376-06-7	ND U	ND U
Sulfonates	PFBS	375-73-5	1.50	2.04
	PFPeS	2706-91-4	0.330	0.426
	PFHxS	355-46-4	1.52	1.71
	PFHpS	375-92-8	0.109 J	0.158 J
	PFOS	1763-23-1	5.77	5.72
	PFNS	68259-12-1	ND U	ND U
	PFDS	335-77-3	ND U	ND U
	4:2 FTS	757124-72-4	ND U	ND U
	6:2 FTS	27619-97-2	0.111 J	0.169 J
8:2 FTS	39108-34-4	ND U	ND U	
other	PFOSA	754-91-6	ND U	ND U
	N-MeFOSAA	2355-31-9	ND U	ND U
	N-EtFOSAA	2991-50-6	ND U	ND U
	HFPO-DA	13252-13-6	2.74	2.31
	PFMOAA	674-13-5	8.25	7.69
	PFMOPrA	377-73-1	ND U	ND U
	PFO2HxA	39492-88-1	ND U	ND U
	PFO3OA	39492-89-2	ND U	ND U
	PFO4DA	39492-90-5	ND U	2.50 L
	Nafion Byproduct 1	29311-67-9	ND U	ND U
	ADONA	919005-14-4	ND U	ND U
	9CI-PF3ONS	756426-58-1	0.0409 L	ND U
	11CI-PF3OUdS	763051-92-9	0.0259 L	ND U
	10:2 FTS	120226-60-0	ND U	ND U
	FBSA	30334-69-1	0.141 L	0.117 L
	N-EtFOSA	4151-50-2	ND U	ND U
	N-EtFOSE	1691-99-2	ND U	ND U
	N-MeFOSA	31506-32-8	ND U	ND U
	N-MeFOSE	24448-09-7	ND U	ND U
	Nafion Byproduct 2	749836-20-2	ND U	0.221 L
	NFDHA	151772-58-6	ND U	ND U
	PEPA		ND U	ND U
	PFECA-G	801212-59-9	ND U	ND U
	PFEESA	113507-82-7	ND U	ND U
PFHxDA	67905-19-5	ND U	ND U	
PFMOBA	863090-89-5	ND U	ND U	
PFO5DA	39492-91-6	ND U	ND U	
PMPA	13140-29-9	4.17	ND U	

Detailed Results



Enthalpy Analytical

Job No.: 0121-717-1 PFAS by Isotope Dilution (non-potable water)

County of Brunswick Site: Northwest Water Plant - Leland, NC

Enthalpy ID 0121-717-001-1
 Sample Name 010821-SO1
 Matrix Aqueous
 Sampling Date 20210108 00:00

Prep Batch EU11429
 Prep Date 2021-01-12 13:59
 Analysis Date 2021-01-12 17:51

Sample Vol (mL) 288.44
 Extract Vol (mL) 0.4
 Dilution Factor 1

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	717.88	0.996	0.996	0.136	0.255			
	PFPeA	2706-90-3	ND	ND	ND	0.0778	0.255			U
	PFHxA	307-24-4	1478.26	2.05	2.05	0.137	0.255			
	PFHpA	375-85-9	980.14	1.36	1.36	0.0602	0.255			
	PFOA	335-67-1	1933.59	2.68	2.68	0.0689	0.255			
	PFNA	375-95-1	426.02	0.591	0.591	0.0441	0.255			
	PFDA	335-76-2	172.01	0.239	0.239	0.108	0.255			J
	PFUnDA	2058-94-8	127.81	0.177	0.177	0.0417	0.255			J
	PFDODA	307-55-1	ND	ND	ND	0.0412	0.255			U
	PFTDA	72629-94-8	ND	ND	ND	0.0646	0.255			U
Sulfonates	PFTeDA	376-06-7	ND	ND	ND	0.0719	0.255			U
	PFBS	375-73-5	1082.33	1.50	1.50	0.0719	0.226			
	PFPeS	2706-91-4	238.27	0.330	0.330	0.0858	0.240			
	PFHxS	355-46-4	1097.09	1.52	1.52	0.0717	0.233			
	PFHpS	375-92-8	78.79	0.109	0.109	0.0675	0.243			J
	PFOS	1763-23-1	4162.31	5.77	5.77	0.0408	0.237			
	PFNS	68259-12-1	ND	ND	ND	0.0567	0.245			U
	PFDS	335-77-3	ND	ND	ND	0.117	0.247			U
	4:2 FTS	757124-72-4	ND	ND	ND	0.0560	0.239			U
	6:2 FTS	27619-97-2	80.29	0.111	0.111	0.0627	0.243			J
other	8:2 FTS	39108-34-4	ND	ND	ND	0.0493	0.245			U
	PFOSA	754-91-6	ND	ND	ND	0.316	0.318			U
	N-MeFOSAA	2355-31-9	ND	ND	ND	0.0472	0.255			U
	N-EiFOSAA	2991-50-6	ND	ND	ND	0.0564	0.255			U
	HFPO-DA	13252-13-6	1976.58	2.74	2.74	0.0825	0.255			
	PFMOA	674-13-5	5950.88	8.25	8.25	1.28	1.28			
	PFMOPrA	377-73-1	ND	ND	ND	0.208	0.255			U
	PFO2HxA	39492-88-1	ND	ND	ND	1.28	1.28			U
	PFO3OA	39492-89-2	ND	ND	ND	1.28	1.28			U
	PFO4DA	39492-90-5	ND	ND	ND	2.61	2.61			U
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.208	0.255			U
	ADONA	919005-14-4	ND	ND	ND	0.104	0.241			U
	9Cl-PF3ONS	756426-58-1	29.46	0.0409	0.0409	0.104	0.239			L
	11Cl-PF3OUdS	763051-92-9	18.70	0.0259	0.0259	0.104	0.241			L
	10:2 FTS	120226-60-0	ND	ND	ND	0.208	0.255			U
	FBSA	30334-69-1	101.49	0.141	0.141	0.208	0.255			L
	N-EiFOSA	4151-50-2	ND	ND	ND	0.208	0.255			U
	N-EiFOSE	1691-99-2	ND	ND	ND	6.24	6.24			U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.208	0.255			U
	N-MeFOSE	24448-09-7	ND	ND	ND	6.24	6.24			U
Nafion Byproduct 2	749836-20-2	ND	ND	ND	1.28	1.28			U	
NFDHA	151772-58-6	ND	ND	ND	0.208	0.255			U	
PEPA	ND	ND	ND	ND	1.28	1.28			U	
PFECA-G	801212-59-9	ND	ND	ND	1.28	1.28			U	
PFEESA	113507-82-7	ND	ND	ND	0.208	0.255			U	
PFHxD	67905-19-5	ND	ND	ND	1.28	1.28			U	
PFMOBA	863090-89-5	ND	ND	ND	1.28	1.28			U	
PFO5DA	39492-91-6	ND	ND	ND	2.61	2.61			U	
PMPA	13140-29-9	3003.41	4.17	4.17	1.28	1.28				
ES	MPFBA		3857.28	5.35				20-150%	77.1%	
	M5PFPeA		12735.81	17.7				20-150%	254.7%	Q
	M3PFBS		20559.96	28.5				20-150%	411.2%	Q
	M2-4:2 FTS		55856.48	77.5				20-150%	1117.1%	Q
	M5PFHxA		3835.87	5.32				20-150%	76.7%	
	M3HFPO-DA		3232.76	4.48				20-150%	64.7%	
	M4PFHpA		4266.99	5.92				20-150%	85.3%	
	M3PFHxS		4206.78	5.83				20-150%	84.1%	
	M2-6:2 FTS		41740.39	57.9				20-150%	834.8%	Q
	M8PFOA		4518.02	6.27				20-150%	90.4%	
	M9PFNA		3790.43	5.26				20-150%	75.8%	
	M8PFOS		3766.24	5.22				20-150%	75.3%	
	M2-8:2 FTS		5303.27	7.35				20-150%	106.1%	
	M8FOSA-I		3045.57	4.22				20-150%	60.9%	
	M6PFDA		4026.77	5.58				20-150%	80.5%	
	d3-N-MeFOSAA		4276.24	5.93				20-150%	85.5%	
	d5-N-EiFOSAA		4040.43	5.60				20-150%	80.8%	
	M7PFUdA		3142.51	4.36				20-150%	62.9%	
MPFDOA		2183.43	3.03				20-150%	43.7%		
M2PFTeDA		776.82	1.08				20-150%	15.5%	Q	

Enthalpy Analytical

Job No.: 0121-717-1 PFAS by Isotope Dilution (non-potable water)
 County of Brunswick Site: Northwest Water Plant - Leland, NC

Enthalpy ID 0121-717-002-1 Prep Batch EU11429 Sample Vol (mL) 283.24
 Sample Name 010821-EO1 Prep Date 2021-01-12 13:59 Extract Vol (mL) 0.4
 Matrix Aqueous Analysis Date 2021-01-12 18:14 Dilution Factor 1
 Sampling Date 20210108 00:00

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	ND	ND	ND	0.139	0.260			U
	PFPeA	2706-90-3	ND	ND	ND	0.0792	0.260			U
	PFHxA	307-24-4	1653.84	2.34	2.34	0.139	0.260			
	PFHpA	375-85-9	1126.59	1.59	1.59	0.0613	0.260			
	PFOA	335-67-1	2274.95	3.21	3.21	0.0701	0.260			
	PFNA	375-95-1	400.36	0.565	0.565	0.0449	0.260			
	PFDA	335-76-2	212.04	0.299	0.299	0.110	0.260			
	PFUnDA	2058-94-8	71.85	0.101	0.101	0.0425	0.260			J
	PFDODA	307-55-1	ND	ND	ND	0.0419	0.260			U
	PFTrDA	72629-94-8	ND	ND	ND	0.0658	0.260			U
PFTeDA	376-06-7	ND	ND	ND	0.0733	0.260			U	
Sulfonates	PFBS	375-73-5	1442.86	2.04	2.04	0.0733	0.230			
	PFPeS	2706-91-4	301.98	0.426	0.426	0.0874	0.244			
	PFHxS	355-46-4	1211.18	1.71	1.71	0.0730	0.237			
	PFHpS	375-92-8	111.94	0.158	0.158	0.0688	0.247			J
	PFOS	1763-23-1	4052.43	5.72	5.72	0.0416	0.241			
	PFNS	68259-12-1	ND	ND	ND	0.0577	0.250			U
	PFDS	335-77-3	ND	ND	ND	0.119	0.251			U
	4:2 FTS	757124-72-4	ND	ND	ND	0.0570	0.243			U
other	6:2 FTS	27619-97-2	119.63	0.169	0.169	0.0638	0.247			J
	8:2 FTS	39108-34-4	ND	ND	ND	0.0502	0.250			U
	PFOSA	754-91-6	ND	ND	ND	0.322	0.323			U
	N-MeFOSAA	2355-31-9	ND	ND	ND	0.0480	0.260			U
	N-EiFOSAA	2991-50-6	ND	ND	ND	0.0575	0.260			U
	HFPO-DA	13252-13-6	1633.94	2.31	2.31	0.0840	0.260			
	PFMOAA	674-13-5	5446.39	7.69	7.69	1.31	1.31			
	PFMOPrA	377-73-1	ND	ND	ND	0.212	0.260			U
	PFO2HxA	39492-88-1	ND	ND	ND	1.31	1.31			U
	PFO3OA	39492-89-2	ND	ND	ND	1.31	1.31			U
	PFO4DA	39492-90-5	1770.26	2.50	2.50	2.65	2.65			L
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.212	0.260			U
	ADONA	919005-14-4	ND	ND	ND	0.106	0.246			U
	9Cl-PF3ONS	756426-58-1	ND	ND	ND	0.106	0.243			U
	11Cl-PF3OUdS	763051-92-9	ND	ND	ND	0.106	0.246			U
	10:2 FTS	120226-60-0	ND	ND	ND	0.212	0.260			U
	FBSA	30334-69-1	83.05	0.117	0.117	0.212	0.260			L
	N-EiFOSA	4151-50-2	ND	ND	ND	0.212	0.260			U
	N-EiFOSE	1691-99-2	ND	ND	ND	6.36	6.36			U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.212	0.260			U
	N-MeFOSE	24448-09-7	ND	ND	ND	6.36	6.36			U
	Nafion Byproduct 2	749836-20-2	156.14	0.221	0.221	1.31	1.31			L
	NFDHA	151772-58-6	ND	ND	ND	0.212	0.260			U
	PEPA	ND	ND	ND	ND	1.31	1.31			U
	PFECA-G	801212-59-9	ND	ND	ND	1.31	1.31			U
	PFEESA	113507-82-7	ND	ND	ND	0.212	0.260			U
	PFHxDA	67905-19-5	ND	ND	ND	1.31	1.31			U
	PFMOBA	863090-89-5	ND	ND	ND	1.31	1.31			U
	PFO5DA	39492-91-6	ND	ND	ND	2.65	2.65			U
	PMPA	13140-29-9	ND	ND	ND	1.31	1.31			U
ES	MPPFA		3440.34	4.86				20-150%	68.8%	
	M5PPPeA		9351.65	13.2				20-150%	187.0%	Q
	M3PFBS		19338.21	27.3				20-150%	386.8%	Q
	M2-4:2 FTS		44341.95	62.6				20-150%	886.8%	Q
	M5PFHxA		3405.33	4.81				20-150%	68.1%	
	M3HFPO-DA		2646.00	3.74				20-150%	52.9%	
	M4PFHpA		3754.39	5.30				20-150%	75.1%	
	M3PFHxS		4195.30	5.92				20-150%	83.9%	
	M2-6:2 FTS		39643.96	56.0				20-150%	792.9%	Q
	M8PFOA		3765.98	5.32				20-150%	75.3%	
	M9PFNA		3623.68	5.12				20-150%	72.5%	
	M8PFOS		3877.04	5.48				20-150%	77.5%	
	M2-8:2 FTS		5206.10	7.35				20-150%	104.1%	
	M8FOSA-I		1970.68	2.78				20-150%	39.4%	
	M6PFDA		3476.82	4.91				20-150%	69.5%	
	d3-N-MeFOSAA		3955.89	5.59				20-150%	79.1%	
	d5-N-EiFOSAA		3980.12	5.62				20-150%	79.6%	
	M7PFUdA		3371.03	4.76				20-150%	67.4%	
MPFDOA		2871.71	4.06				20-150%	57.4%		
M2PFTeDA		3256.21	4.60				20-150%	65.1%		

QC Data



Enthalpy Analytical

Job No.: 0121-717-1 PFAS by Isotope Dilution (non-potable water)
 County of Brunswick Site: Northwest Water Plant - Leland, NC

Enthalpy ID MB-11429-PFAS Prep Batch EU11429 Sample Vol (mL) 250
 Sample Name MB-11429-PFAS Prep Date 2021-01-12 13:59 Extract Vol (mL) 0.4
 Matrix Aqueous Analysis Date 2021-01-12 15:56 Dilution Factor 1
 Sampling Date

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	ND	ND	ND	0.157	0.294			U
	PFPeA	2706-90-3	ND	ND	ND	0.0898	0.294			U
	PFHxA	307-24-4	28.83	0.0461	0.0461	0.158	0.294			L
	PFHpA	375-85-9	ND	ND	ND	0.0695	0.294			U
	PFOA	335-67-1	49.88	0.0798	0.0798	0.0795	0.294			J
	PFNA	375-95-1	12.15	0.0194	0.0194	0.0509	0.294			L
	PFDA	335-76-2	ND	ND	ND	0.125	0.294			U
	PFUnDA	2058-94-8	ND	ND	ND	0.0481	0.294			U
	PFDODA	307-55-1	ND	ND	ND	0.0475	0.294			U
	PFTDA	72629-94-8	ND	ND	ND	0.0745	0.294			U
Sulfonates	PFTeDA	376-06-7	ND	ND	ND	0.0830	0.294			U
	PFBS	375-73-5	46.31	0.0741	0.0741	0.0830	0.261			L
	PFPeS	2706-91-4	ND	ND	ND	0.0990	0.277			U
	PFHxS	355-46-4	ND	ND	ND	0.0827	0.269			U
	PFHpS	375-92-8	ND	ND	ND	0.0779	0.280			U
	PFOS	1763-23-1	ND	ND	ND	0.0471	0.274			U
	PFNS	68259-12-1	ND	ND	ND	0.0654	0.283			U
	PFDS	335-77-3	ND	ND	ND	0.135	0.285			U
	4:2 FTS	757124-72-4	ND	ND	ND	0.0646	0.275			U
	6:2 FTS	27619-97-2	72.75	0.116	0.116	0.0723	0.280			J
other	8:2 FTS	39108-34-4	ND	ND	ND	0.0569	0.283			U
	PFOSA	754-91-6	ND	ND	ND	0.365	0.366			U
	N-MeFOSAA	2355-31-9	ND	ND	ND	0.0544	0.294			U
	N-EiFOSAA	2991-50-6	ND	ND	ND	0.0651	0.294			U
	HFPO-DA	13252-13-6	ND	ND	ND	0.0951	0.294			U
	PFMOA	674-13-5	ND	ND	ND	1.48	1.48			U
	PFMOPrA	377-73-1	ND	ND	ND	0.240	0.294			U
	PFO2HxA	39492-88-1	ND	ND	ND	1.48	1.48			U
	PFO3OA	39492-89-2	ND	ND	ND	1.48	1.48			U
	PFO4DA	39492-90-5	ND	ND	ND	3.01	3.01			U
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.240	0.294			U
	ADONA	919005-14-4	ND	ND	ND	0.120	0.278			U
	9Cl-PF3ONS	756426-58-1	ND	ND	ND	0.120	0.275			U
	11Cl-PF3OUdS	763051-92-9	ND	ND	ND	0.120	0.278			U
	10:2 FTS	120226-60-0	ND	ND	ND	0.240	0.294			U
	FBSA	30334-69-1	ND	ND	ND	0.240	0.294			U
	N-EiFOSA	4151-50-2	ND	ND	ND	0.240	0.294			U
	N-EiFOSE	1691-99-2	ND	ND	ND	7.20	7.20			U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.240	0.294			U
	N-MeFOSE	24448-09-7	ND	ND	ND	7.20	7.20			U
Nafion Byproduct 2	749836-20-2	ND	ND	ND	1.48	1.48			U	
NFDHA	151772-58-6	ND	ND	ND	0.240	0.294			U	
ES	PEPA		ND	ND	ND	1.48	1.48			U
	PFECA-G	801212-59-9	ND	ND	ND	1.48	1.48			U
	PFEESA	113507-82-7	ND	ND	ND	0.240	0.294			U
	PFHxD	67905-19-5	ND	ND	ND	1.48	1.48			U
	PFMOBA	863090-89-5	ND	ND	ND	1.48	1.48			U
	PFO5DA	39492-91-6	ND	ND	ND	3.01	3.01			U
	PMPA	13140-29-9	ND	ND	ND	1.48	1.48			U
	MPPFA		4310.38	6.90				20-150%	86.2%	
	M5PFPeA		5471.70	8.75				20-150%	109.4%	
	M3PFBS		4798.05	7.68				20-150%	96.0%	
	M2-4:2 FTS		4891.56	7.83				20-150%	97.8%	
	M5PFHxA		3888.24	6.22				20-150%	77.8%	
	M3HFPO-DA		3242.47	5.19				20-150%	64.8%	
	M4PFHpA		4045.09	6.47				20-150%	80.9%	
	M3PFHxS		4239.13	6.78				20-150%	84.8%	
	M2-6:2 FTS		13062.61	20.9				20-150%	261.3%	Q
	M8PFOA		4215.39	6.74				20-150%	84.3%	
	M9PFNA		3662.31	5.86				20-150%	73.2%	
	M8PFOS		3936.54	6.30				20-150%	78.7%	
	M2-8:2 FTS		5806.21	9.29				20-150%	116.1%	
M8FOSA-I		3159.05	5.05				20-150%	63.2%		
M6PFDA		3455.88	5.53				20-150%	69.1%		
d3-N-MeFOSAA		4003.04	6.40				20-150%	80.1%		
d5-N-EiFOSAA		4091.78	6.55				20-150%	81.8%		
M7PFUdA		3031.88	4.85				20-150%	60.6%		
MPPFOA		2479.06	3.97				20-150%	49.6%		
M2PFTeDA		2419.35	3.87				20-150%	48.4%		

Enthalpy Analytical

Job No.: 0121-717-1 PFAS by Isotope Dilution (non-potable water)

County of Brunswick Site: Northwest Water Plant - Leland, NC

Enthalpy ID	OPR-11429-PFAS	Prep Batch	EU11429	Sample Vol (mL)	250
Sample Name	OPR-11429-PFAS	Prep Date	2021-01-12 13:59	Extract Vol (mL)	0.4
Matrix	Aqueous	Analysis Date	2021-01-12 16:19	Dilution Factor	1
Sampling Date					

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	13005.97	20.8	20.8	0.157	0.294		104.0%	
	PFPeA	2706-90-3	13821.75	22.1	22.1	0.0898	0.294		110.6%	
	PFHxA	307-24-4	13530.59	21.6	21.6	0.158	0.294		108.2%	
	PFHpA	375-85-9	13407.78	21.5	21.5	0.0695	0.294		107.3%	
	PFOA	335-67-1	13452.77	21.5	21.5	0.0795	0.294		107.6%	
	PFNA	375-95-1	13713.01	21.9	21.9	0.0509	0.294		109.7%	
	PFDA	335-76-2	14877.30	23.8	23.8	0.125	0.294		119.0%	
	PFUnDA	2058-94-8	13896.67	22.2	22.2	0.0481	0.294		111.2%	
	PFDoDA	307-55-1	13602.40	21.8	21.8	0.0475	0.294		108.8%	
	PFTTrDA	72629-94-8	17701.06	28.3	28.3	0.0745	0.294		141.6%	
PFTeDA	376-06-7	13290.97	21.3	21.3	0.0830	0.294		106.3%		
Sulfonates	PFBS	375-73-5	10815.02	17.3	17.3	0.0830	0.261		97.5%	
	PFPeS	2706-91-4	12972.40	20.8	20.8	0.0990	0.277		110.3%	
	PFHxS	355-46-4	11646.24	18.6	18.6	0.0827	0.269		101.9%	
	PFHpS	375-92-8	13813.85	22.1	22.1	0.0779	0.280		116.0%	
	PFOS	1763-23-1	13540.79	21.7	21.7	0.0471	0.274		116.7%	
	PFNS	68259-12-1	12457.79	19.9	19.9	0.0654	0.283		103.6%	
	PFDS	335-77-3	12886.15	20.6	20.6	0.135	0.285		106.8%	
	4:2 FTS	757124-72-4	13209.21	21.1	21.1	0.0646	0.275		112.8%	
6:2 FTS	27619-97-2	12120.81	19.4	19.4	0.0723	0.280		102.0%		
8:2 FTS	39108-34-4	13603.95	21.8	21.8	0.0569	0.283		113.4%		
Other	PFOSA	754-91-6	14394.80	23.0	23.0	0.365	0.366		115.2%	
	N-MeFOSAA	2355-31-9	13251.58	21.2	21.2	0.0544	0.294		106.0%	
	N-EtFOSAA	2991-50-6	13351.83	21.4	21.4	0.0651	0.294		106.8%	
	HFPO-DA	13252-13-6	14181.02	22.7	22.7	0.0951	0.294		113.4%	
ES	MPFBA		4324.82	6.92				20-150%	86.5%	
	M5PFPeA		4944.57	7.91				20-150%	98.9%	
	M3PFBS		4516.19	7.23				20-150%	90.3%	
	M2-4:2 FTS		4211.18	6.74				20-150%	84.2%	
	M5PFHxA		4175.34	6.68				20-150%	83.5%	
	M3HFPO-DA		3819.44	6.11				20-150%	76.4%	
	M4PFHpA		3987.12	6.38				20-150%	79.7%	
	M3PFHxS		4587.01	7.34				20-150%	91.7%	
	M2-6:2 FTS		5541.57	8.87				20-150%	110.8%	
	M8PFOA		3855.11	6.17				20-150%	77.1%	
	M9PFNA		3545.37	5.67				20-150%	70.9%	
	M8PFOS		3612.91	5.78				20-150%	72.3%	
	M2-8:2 FTS		4841.98	7.75				20-150%	96.8%	
	M8FOSA-I		3420.91	5.47				20-150%	68.4%	
	M6PFDA		3650.01	5.84				20-150%	73.0%	
	d3-N-MeFOSAA		3866.26	6.19				20-150%	77.3%	
	d5-N-EtFOSAA		4185.30	6.70				20-150%	83.7%	
	M7PFUdA		3630.94	5.81				20-150%	72.6%	
MPFDoA		3003.94	4.81				20-150%	60.1%		
M2PFTeDA		2211.22	3.54				20-150%	44.2%		

Narrative Summary



Enthalpy Analytical Narrative Summary

Company Job No. Client ID.	County of Brunswick 0121-717-1 PFAS by Isotope Dilution (non-potable water)
----------------------------------	--

1. Custody

Caity Hayes received the samples on January 08, 2021 at 8.9 °C after being relinquished by County of Brunswick. The samples were received in good condition.

Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC

Table 1 - Sample Inventory

EU Lab Sample ID	Client Sample ID	Matrix
0121-717-001-1	010821-SO1	Aqueous
0121-717-002-1	010821-EO1	Aqueous

2. Methods and Analytes

A list of analytes of interest and corresponding methods of analysis is shown in Table 3. Abbreviations are defined in the listed Appendices.

Table 3 - Methods and Analytes

EU Method	Analytes	Cleanup Method
EU-047	PFAS	ENVI-Carb

3. Analysis

The samples were analyzed using Waters Acquity UPLC equipped with Xevo TQ MS (LC/MS/MS "Kili").

For aqueous samples, the sample volume was measured gravimetrically by the laboratory, and spiked with Extraction Standard (ES). The sample was then mixed well and centrifuged, if needed. The samples were then extracted via SPE, and the extracts were cleaned up using ENVI-Carb.

Each final sample extract was transferred to an autosampler vial, spiked with Injection Standard (IS), and brought to a final volume of 400µL prior to analysis.

4. Calibration

In the initial calibration, the reported analytes exhibited R^2 of ≥ 0.99 . The reported analytes in the calibration standards and Initial Calibration Verification (ICV) met the 30% accuracy criterion for native analytes. The reported analytes in the continuing calibration standards meet the 30% accuracy criterion with the following exception. The analyte recovery for N-EtFFOSE was outside the high recovery limit at 139% in the opening continuing calibration verification. The analyst was not detected in the samples.

Enthalpy Analytical Narrative Summary

Company
Job No.
Client ID.

County of Brunswick
0121-717-1 PFAS by Isotope Dilution (non-potable water)

5. QC Notes

The QC sample analyses met all method criteria with the following exceptions. The analyte recovery for PFTrDA in the laboratory control sample (OPR) was outside the upper control limit at 142%. The extraction standard recovery in the method blank for M2-6:2-FTS was outside the upper control limit at 261%. Neither analyte was detected in the samples.

The samples were extracted within the 28-day from collection holding time and analyzed within the 28-day from extraction to analysis holding time required by the method.

6. Reporting Notes

Some labeled standards in the samples fell outside the limits for ES recoveries, as noted by a Q qualifier. The target analytes are quantified based on their ratio to the labeled standards, therefore, undergo the same losses as the labeled standards. As a result, low or high recoveries do not cause any change to ratios or contribute any additional error in the measurement of the target analytes. Therefore, the data are considered acceptable.

The results presented in this report are representative of the samples as provided to the laboratory.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, LLC in Wilmington NC is accredited by the Louisiana Department of Environmental Quality to the 2009 TNI Standard under certificate number 05075.



General Reporting Notes – Data Qualifiers

The following are general reporting notes that are applicable to all Enthalpy Analytical, LLC - Wilmington, NC data reports, unless specifically noted otherwise.

General Data Qualifiers

- B – The analyte was found in the method blank, at a concentration that was at least 10% of the concentration in the sample.
- Cxx – Two or more congeners co-elute. In EDDs, C denotes the lowest IUPAC congener in a co-elution group and additional co-eluters for the group ('xx') are shown with the number of the lowest IUPAC co-eluter.
- E – The reported concentration exceeds the calibration range (upper point of the calibration curve). For HRMS data, this condition does not imply additional measurement uncertainty. For LC-MS/MS data, these values should be considered as having measurement uncertainty higher than values within the calibration range.
- EDL – Estimated Detection Level. Specific to Dioxin/Furan tests and equivalent to MDL
- EMPC – Estimated Maximum Possible Concentration Specific to Dioxin/Furan tests to indicate the signal/noise ratio was not sufficient for peak identification (the determined ion-abundance ratio was outside the allowed theoretical range), or where there was a co-eluting interference. Indicates that a peak was identified but did not meet the method specified ion-abundance ratio.
- IR – The ion ratio between the primary and secondary ions was observed to be outside the method criteria therefore the actual analyte concentration cannot be accurately determined as defined by DoD QSM Table B-15.
- J – The analyte has a concentration below the minimum calibration level (LOQ value) but greater than the LOD. These values should be considered as having measurement uncertainty higher than values within the calibration range
- L - Indicates that an analyte has a concentration below the Minimum Detection Limit (MDL). The reported concentration is not recommended for regulatory use as the analyte signal may have a signal-to-noise ratio less than the criteria deemed necessary to be considered a detected analyte.
- LOD – Limit of Detection: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOD. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the minimum detection limit (MDL). The LOD is adjusted for sample weight or volume.
- LOQ – Limit of Quantiation: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOQ. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the reporting limit (RL). The LOD is adjusted for sample weight or volume.
- <LOD() – Analyte was not found at a concentration high enough to be reported as detected. It is reported as less than the LOD, and the LOD is given in the parentheses.



General Reporting Notes – Data Qualifiers

- ND – Indicates a non-detect.
- NR – Indicates a value that is not reportable due to issues observed in sample preparation or analysis.
- PR – The associated congener(s) is(are) poorly resolved.
- QI – Indicates the presence of a quantitative interference.
- RL – Reporting Limit. Lowest reportable value. The level is higher than the MDL.
- SI – Denotes “Single Ion Mode” and is utilized for PCBs where the secondary ion trace has a significantly elevated noise level due to background PFK. Responses for such peaks are calculated using an EMPC approach based solely on the primary ion area(s) and may be considered estimates.
- U – The analyte was not detected.
- V – The labeled standard recovery is not within method control limits.
- X – Results from re-injection/repeat/second-column analysis.

Lab Identifiers/ Data Attributes

- AR – Indicates use of the archived portion of the sample extract.
- CU – Indicates a sample that required additional clean-up prior to HRMS injection/processing.
- D – Dilution Data. Result was obtained from the analysis of a dilution. The number that follows the “D” indicates the dilution factor.
- DE – Indicates a dilution performed with the addition of ES (Extraction Standard) solution.
- DUP – Designation for a duplicate sample.
- MS – Designation for a matrix spike.
- MSD – Designation for a matrix spike duplicate.
- RJ – Indicates a reinjection of the sample extract.
- S – Indicates a sample split. The number that follows the “S” indicates the split factor.
- R – Indicates a re-extraction of the sample.

PFAS Compound Acronym List

Acronym	Compound Name	CAS #
Target Analytes		
PFBA	Perfluorobutanoic Acid	375-22-4
PFPeA	Perfluoropentanoic Acid	2706-90-3
PFHxA	Perfluorohexanoic Acid	307-24-4
PFHpA	Perfluoroheptanoic Acid	375-85-9
PFOA	Perfluorooctanoic Acid	335-67-1
PFNA	Perfluorononanoic Acid	375-95-1
PFDA	Perfluorodecanoic acid	335-76-2
PFUnA (PFUnDA)	Perfluoroundecanoic acid	2058-94-8
PFDoA (PFDoDA)	Perfluorododecanoic acid	307-55-1
PFTrDA (PFTriA)	Perfluorotridecanoic acid	72629-94-8
PFTeDA (PFTA)	Perfluorotetradecanoic acid	376-06-7
PFBS	Perfluorobutane sulfonic acid	375-73-5
PFPeS	Perfluoropentane sulfonic acid	2706-91-4
PFHxS	Perfluorohexane sulfonic acid	355-46-4
PFHpS	Perfluoroheptane sulfonic acid	375-92-8
PFOS	Perfluorooctane sulfonic acid	1763-23-1
PFNS	Perfluorononane sulfonic acid	68259-12-1
PFDS	Perfluorodecane sulfonic acid	757124-72-4
4:2 FTS	4:2 fluorotelomer sulfonic acid	27619-97-2
6:2 FTS	6:2 fluorotelomer sulfonic acid	39108-34-4
8:2 FTS	8:2 fluorotelomer sulfonic acid	13252-13-6
PFOSA (FOSA)	Perfluorooctane sulfonamide	754-91-6
N-MeFOSAA	N-methyl perfluorooctane sulfonamido acetic acid	2355-31-9
N-EtFOSAA	N-ethyl perfluorooctane sulfonamido acetic acid	2991-50-6
* HFPO-DA	2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy)-propanoic acid (Gen-X)	13252-13-6
* 11CI-PF3OUdS	11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	674-13-5
* 9CI-PF3ONS	9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	919005-14-4
* ADONA	4,8-dioxa-3H-perfluorononanoic acid	756426-58-1
* PFMOAA	Perfluoro-2-methoxyacetic acid	763051-92-9
* PFMOPrA	Perfluoro-3-methoxypropanoic acid	377-73-1
* PFO2HxA	Perfluoro (3,5-dioxahexanoic) acid	39492-88-1
* PFO3OA	Perfluoro (3,5,7-trioxaoctanoic) acid	39492-89-2
* PFO4DA	Perfluoro (3,5,7,9-tetraoxadecanoic) acid	39492-90-5
* PFO5DA	Perfluoro(3,5,7,9,11-pentaoxadodecanoic) acid	39492-91-6
* Nafion Byproduct 1	Nafion Byproduct 1	29311-67-9
* Nafion Byproduct 2	Nafion Byproduct 2	749836-20-2
* PFEESA	Perfluoro(2-ethoxyethane)sulphonic acid	113507-82-7
* PFMOBA	Perfluoro-4-methoxybutanic acid	863090-89-5
* PEPA	Perfluoro-2-ethoxypropanoic acid	N/A
* PMPA	Perfluoro-2-methoxypropanoic acid	13140-29-9
* 10:2 FTS	Fluorotelomer sulfonate 10:2	120226-60-0
* N-EtFOSA	N-ethylperfluoro-1-octanesulfonamide	4151-50-2
* N-EtFOSE	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol	1691-99-2
* N-MeFOSA	N-methylperfluoro-1-octanesulfonamide	31506-32-8
* N-MeFOSE	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol	24448-09-7
* PFECA-G	4-(Heptafluoroisopropoxy)hexafluorobutanoic acid	801212-59-9
* PFHxDA	Perfluorohexadecanoic acid	67905-19-5



Extraction Standards		
MPFBA	Perfluoro-n-[13C4]butanoic acid	
M5PFPeA	Perfluoro-n-[13C5]pentanoic acid	
M3PFBS	Sodium perfluoro-1-[2,3,4-13C3]-butanesulfonic acid	
M2-4:2 FTS	Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]-hexane sulfonic acid	
M5PFHxA	Perfluoro-n-[1,2,3,4,6-13C5]hexanoic acid	
M3HFPO-DA	2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy)-13C3-propanoic acid	
M4PFHpA	Perfluoro-n-[1,2,3,4-13C4]heptanoic acid	
M3PFHxS	Sodium perfluoro-1-[1,2,3-13C3]-hexanesulfonic acid	
M2-6:2 FTS	Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]-octane sulfonic acid	
M8PFOA	Perfluoro-n-[13C8]octanoic acid	
M9PFNA	Perfluoro-n-[13C9]nonanoic acid	
M8PFOS	Sodium perfluoro-1-[13C8]-octanesulfonic acid	
M2-8:2 FTS	Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]-decane sulfonic acid	
M8FOSA	Perfluoro-1-[13C8]octanesulfonamide	
M6PFDA	Perfluoro-n-[1,2,3,4,5,6-13C6]decanoic acid	
d3-N-MeFOSAA	N-methyl-d3-perfluoro-1-octanesulfonamide	
d5-N-EtFOSAA	N-ethyl-d5-perfluoro-1-octanesulfonamide	
M7PFUnDA (M7PFUdA)	Perfluoro-n-[1,2,3,4,5,6,7-13C7]undecanoic acid	
MPFDoA	Perfluoro-n-[1,2-13C2]dodecanoic acid	
M2PFTeDA	Perfluoro-n-[1,2-13C2]tetradecanoic acid	
Injection Standards		
M3PFBA	Perfluoro-n-[2,3,4-13C3]butanoic acid	
M2PFOA	Perfluoro-n-[1,2-13C2]octanoic acid	
MPFDA	Perfluoro-n-[1,2-13C2]decanoic acid	
MPFOS	Sodium perfluoro-1-[1,2,3,4-13C4]-octanesulfonic acid	

* Analytes are currently not accredited under TNI Standard. Accreditation pending.

Sample Custody



**This Is The Last Page
Of This Report.**

