

Brunswick County Public Utilities - NC

PO Box 249
Bolivia, NC 28422-0249

Leland, NC

Client Project# NORTHWEST WATER PLANT
Samples Received: 1/23/2026

Analytical Report 0126-854

PFAS by Isotope Dilution (non-potable water)

Report Issue Date: 2/18/2026

I certify that to the best of my knowledge all analytical data presented in this report have been checked for completeness, accuracy, errors and legibility in addition to having been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s). This analytical report was prepared in Portable Document Format (.PDF) and contains 38 pages. This report shall not be reproduced except in full without approval of the laboratory. This will provide assurance that parts of the report are not taken out of context.

Amendment(s):

Signature:



Laura Boivin, QA Associate II



Enthalpy Analytical, LLC – Wilmington
Amanda Valois, Project Manager
amvalois@montrose-env.com / www.enthalpy.com
O: (910) 212-5858
2714 Exchange Drive, Wilmington, NC 28405

Table of Contents

PFAS by Isotope Dilution (non-potable water)	3
Case Narrative	3
General Reporting Notes	6
Results	12
Summary of Results	13
Detailed Results	14
QC Data	20
Blanks	21
Controls	24
PFAS by Isotope Dilution (non-potable water)	26
Case Narrative	26
Results	29
Summary of Results	30
Detailed Results	31
QC Data	33
Blanks	34
Controls	35
Sample Custody	36
Chain of Custody	37

Narrative Summary



Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0126-854-1
Client ID.	NORTHWEST WATER PLANT

1. Custody

Isabelle Martin received the samples at 2.1 °C after being relinquished by Brunswick County Public Utilities - NC.

The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC.

Table 1 - Sample Inventory

EU Lab Sample ID	Client Sample ID	Matrix	Received
0126-854-001-2	012326-S01	aqueous	2026-01-23
0126-854-001-3	012326-S01	aqueous	2026-01-23
0126-854-002-2	012326-E01	aqueous	2026-01-23
0126-854-002-3	012326-E01	aqueous	2026-01-23

2. Methods and Analytes

A list of analytes of interest and corresponding methods of analysis is shown in Table 3. Abbreviations are defined in the listed Appendices.

Table 3 - Methods and Analytes

EU Method	Analytes	Cleanup Method
EU047	Brunswick List	ENVI-Carb

3. Analysis

The samples were analyzed using LC/MS/MS instrument Frodo.

Samples were analyzed using more than one extraction batch to meet method acceptance criteria.

4. Calibration

In the initial calibration, the reported analytes exhibited R^2 of ≥ 0.99 . The reported analytes in the calibration standards, Initial Calibration Verification (ICV) and continuing calibration (concal) met the accuracy criterion for native analytes.

Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0126-854-1
Client ID.	NORTHWEST WATER PLANT

5. QC Notes

The QC sample analyses passed all method criteria.

Ongoing Precision Recovery (OPR) control limits have not been established for some analytes of interest.

PFAS by Isotope Dilution (non-potable water) samples were extracted within 28 days, and extracts analyzed within 28 days.

6. Reporting Notes

The results presented in this report are representative of the samples as provided to the laboratory.

This report provides all results including detections below LOD following client instruction.

Some labeled extraction standards (ES) in the analyses recovered outside method control limits for ES recovery, as denoted by the "Q" qualifier. The target analytes are quantified based on their ratio to their labeled standard analogs. As a result, low or high labeled standard recovery do not cause any change to ratios or contribute any additional error in the measurement of the target analytes. When detected at a signal-to-noise above 10:1 the ES peak area is used to quantify its respective target analyte using accepted isotope dilution principles. The data is reported without adverse impact.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, LLC in Wilmington NC is accredited by the Louisiana Department of Environmental Quality to the 2016 TNI Standard under certificate number 05075.

General Reporting Notes – Data Qualifiers

The following are general reporting notes that are applicable to all Enthalpy Analytical, LLC - Wilmington, NC data reports, unless specifically noted otherwise.

General Data Qualifiers

- Ac - Alternate calculation flag indicates the es recovery was calculated using the opening concal when either of the following situations is encountered in the data processing software: the ES recovery is over 400% or the JS is not detected.
- B – The analyte was found in the method blank, at a concentration that was at least 10% of the amount in the sample.
- Cxx – Two or more congeners co-elute. In EDDs, C denotes the lowest IUPAC congener in a co-elution group and additional co-eluters for the group ('xx') are shown with the number of the lowest IUPAC co-eluter.
- E – The reported concentration exceeds the calibration range (upper point of the calibration curve). For HRMS data, this condition does not imply additional measurement uncertainty. For LC-MS/MS data, these values should be considered as having measurement uncertainty higher than values within the calibration range.
- EDL – Estimated Detection Level: The EDL is unique to isotope dilution methods and reflects the conditions of analysis at the time of analysis, including the equipment used. Where the MDL is a static value, the EDL is a dynamic value.
- EMPC – Estimated Maximum Possible Concentration: EMPC is specific to Dioxin/Furan tests to indicate the determined ion-abundance ratio was outside the allowed theoretical range (usually due to being near the detection limit, although it can very rarely be caused by a co-eluting interference). The EMPC concentration is adjusted to reflect the value at the theoretical ion-abundance ratio.
- I/IR – The ion ratio between the primary and secondary ions was observed to be outside the method criteria. The analyte concentration may be inaccurate due to interference.
- J – The analyte has a concentration below the minimum calibration level (LOQ value) but greater than the LOD. These values should be considered as having measurement uncertainty higher than values within the calibration range
- L - For reports containing PFAS analytes only, this flag indicates that an analyte has a concentration below the Minimum Detection Limit (MDL) . The reported concentration is not recommended for regulatory use as the analyte signal may have a signal-to-noise ratio less than the criteria deemed necessary to be considered a detected analyte.
- LOD – Limit of Detection: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOD. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the minimum detection limit (MDL). The LOD is adjusted for sample weight or volume.

General Reporting Notes – Data Qualifiers

- LOQ – Limit of Quantitation: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOQ. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the reporting limit (RL). The LOQ is adjusted for sample weight or volume.
- <LOD() – Analyte was not found at a concentration high enough to be reported as detected. It is reported as less than the LOD, and the LOD is given in the parentheses.
- <LOQ() – Analyte was not found at a concentration high enough to be reported as above the QSM-defined LOQ or TNI defined Reporting Limit. It is reported as less than the LOQ, and the LOQ is given in the parentheses.
- ND – Indicates a non-detect.
- NR – Indicates a value that is not reportable due to issues observed in sample preparation or analysis.
- PR – The associated congener(s) is(are) poorly resolved.
- QI – Indicates the presence of a quantitative interference.
- RL – Reporting Limit. Lowest reportable value. The level is higher than the MDL.
- SI – Denotes “Single Ion Mode” and is utilized for PCBs where the secondary ion trace has a significantly elevated noise level due to background PFK. Responses for such peaks are calculated using an EMPC approach based solely on the primary ion area(s) and may be considered estimates.
- U – The analyte was not detected.
- V / Q – The labeled standard recovery is not within method control limits.
- X – Indicates the result is from re-injection/repeat/second-column analysis.

Lab Identifiers/ Data Attributes

- AR – Indicates use of the archived portion of the sample extract.
- CU – Indicates a sample that required additional clean-up prior to HRMS injection/processing.
- D – Dilution Data. Result was obtained from the analysis of a dilution. The number that follows the “D” indicates the dilution factor.
- DE – Indicates a dilution performed with the addition of ES (Extraction Standard) solution.
- DUP – Designation for a duplicate sample.
- MS – Designation for a matrix spike.
- MSD – Designation for a matrix spike duplicate.



General Reporting Notes – Data Qualifiers

- R – Indicates a re-extraction of the sample.
- RJ – Indicates a reinjection of the sample extract.
- S – Indicates a sample split. The number that follows the “S” indicates the split factor.
- SAT – Indicates an analyte saturated the detector.

PFAS Compound Acronym List			Methods					
Acronym	CAS #	Compound Name	SOP EU047	EPA 1633 (B-24)	EPA 1633X	EPA 537.1	EPA 533	EPA 8327*
Target Analytes								
PFBA	375-22-4	Perfluorobutanoic Acid	X	X	X		X	X
PFPeA	2706-90-3	Perfluoropentanoic Acid	X	X	X		X	X
PFHxA	307-24-4	Perfluorohexanoic Acid	X	X	X	X	X	X
PFHpA	375-85-9	Perfluoroheptanoic Acid	X	X	X	X	X	X
PFOA	335-67-1	Perfluorooctanoic Acid	X	X	X	X	X	X
PFNA	375-95-1	Perfluorononanoic Acid	X	X	X	X	X	X
PFDA	335-76-2	Perfluorodecanoic acid	X	X	X	X	X	X
PFUnA (PFUnDA)	2058-94-8	Perfluoroundecanoic acid	X	X	X	X	X	X
PFDoA (PFDoDA)	307-55-1	Perfluorododecanoic acid	X	X	X	X		X
PFTrDA (PFTriA, PFTrDA)	72629-94-8	Perfluorotridecanoic acid	X	X	X	X		X
PFTeDA (PFTA, PFTreA)	376-06-7	Perfluorotetradecanoic acid	X	X	X	X		X
PFBS	375-73-5	Perfluorobutane sulfonic acid	X	X	X	X	X	X
PFPeS	2706-91-4	Perfluoropentane sulfonic acid	X	X	X		X	X
PFHxS	355-46-4	Perfluorohexane sulfonic acid	X	X	X	X	X	X
PFHpS	375-92-8	Perfluoroheptane sulfonic acid	X	X	X		X	X
PFOS	1763-23-1	Perfluorooctane sulfonic acid	X	X	X	X	X	X
PFNS	68259-12-1	Perfluorononane sulfonic acid	X	X	X			X
PFDS	335-77-3	Perfluorodecane sulfonic acid	X	X	X			X
4:2 FTS	757124-72-4	4:2 fluorotelomer sulfonic acid	X	X	X		X	X
6:2 FTS	27619-97-2	6:2 fluorotelomer sulfonic acid	X	X	X		X	X
8:2 FTS	39108-34-4	8:2 fluorotelomer sulfonic acid	X	X	X		X	X
10:2 FTS	120226-60-0	Fluorotelomer sulfonate 10:2						X
FHxSA	41997-13-1	Perfluorohexanesulfonamide			X			X
PFOSA (FOSA)	754-91-6	Perfluorooctane sulfonamide	X	X	X			X
N-MeFOSAA	2355-31-9	N-methyl perfluorooctane sulfonamido acetic acid	X	X	X	X		X
N-MeFOSA	31506-32-8	N-methylperfluoro-1-octanesulfonamide	X	X	X			X
N-MeFOSE	24448-09-7	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol	X	X	X			X
N-EtFOSAA	2991-50-6	N-ethyl perfluorooctane sulfonamido acetic acid	X	X	X	X		X
N-EtFOSA	4151-50-2	N-ethylperfluoro-1-octanesulfonamide	X	X	X			X
N-EtFOSE	1691-99-2	2-(N-Ethylperfluoro-1-octanesulfonamido)-ethanol	X	X	X			X
HFPO-DA	13252-13-6	Hexafluoropropyleneoxide dimer acid (GenX)	X	X	X	X	X	X
11Cl-PF3OUds	763051-92-9	11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	X	X	X	X	X	X
9Cl-PF3ONS	756426-58-1	9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	X	X	X	X	X	X
ADONA	919005-14-4	4,8-dioxa-3H-perfluorononanoic acid	X	X	X	X	X	X
PFESA	113507-82-7	Perfluoro(2-ethoxyethane)sulfonic acid		X	X		X	X
PFMOBA (PFMBA)	863090-89-5	Perfluoro-4-methoxybutanoic acid		X	X		X	X
NFDHA	151772-58-6	Nonafluoro-3,6-dioxaheptanoic acid		X	X		X	X
PFMOPrA (PFMPA)	377-73-1	Perfluoro-3-methoxypropanoic acid		X	X		X	X
PFPrA	422-64-0	Perfluoropropionic acid, 2,2,3,3,3-Pentafluoropropionic acid			X			X
PFPrS (PFPS)	423-41-6	Perfluoropropanesulfonic acid			X			X



PFAS Compound Acronym List			Methods					
Acronym	CAS #	Compound Name	SOP EU047	EPA 1633 (B-24)	EPA 1633X	EPA 537.1	EPA 533	EPA 8327*
PFMOAA	674-13-5	Perfluoro-2-methoxyacetic acid;			X			X
PFO2HxA	39492-88-1	Perfluoro (3,5-dioxahexanoic) acid			X			X
PFO3OA	39492-89-2	Perfluoro (3,5,7-trioxaoctanoic) acid			X			X
PFO4DA	39492-90-5	Perfluoro (3,5,7,9-tetraoxadecanoic) acid			X			X
PFO5DA	39492-91-6	Perfluoro(3,5,7,9,11-pentaoxadodecanoic) acid			X			X
Nafion Byproduct 1 (PS Acid)	29311-67-9	1,1,2,2-tetrafluoro-2-[1,1,1,2,3,3-hexafluoro-3-(1,2,2-trifluoroethenoxy)propan-2-yl]oxyethanesulfonic acid			X			X
Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	Perfluoro-2-[[perfluoro-3-(perfluoroethoxy)-2-propanyl]oxy]ethanesulfonic acid (Hydro-PS Acid)			X			X
PEPA	267239-61-2	Perfluoro-2-ethoxypropanoic acid			X			X
PMPA	13140-29-9	Perfluoro-2-methoxypropanoic acid			X			X
PFECA-G, (PFPE-1)	801212-59-9	4-(Heptafluoroisopropoxy)hexafluorobutanoic acid, Perfluoro-4-isopropoxybutanoic acid			X			X
PFHxDA	67905-19-5	Perfluorohexadecanoic acid			X			
R-PSDA (Nafion Byproduct 4)	2416366-18-0	Perfluoro-4-(2-sulfoethoxy)pentanoic acid; 2,2,3,3,4,5,5-Octafluoro-4-(1,1,2,2-tetrafluoro-2-sulfoethoxy)pentanoic acid			X			X
Hydrolyzed PSDA (Nafion Byproduct 5)	2416366-19-1	2-fluoro-2-[1,1,2,3,3,3-hexafluoro-2-(1,1,2,2-tetrafluoro-2-sulfoethoxy)propoxy]-acetic acid			X			X
R-PSDCA (Nafion Byproduct 6)	2416366-21-5	1,1,2,2-tetrafluoro-2-[1,2,2,3,3-pentafluoro-1-(trifluoromethyl)propoxy] ethanesulfonic acid			X			X
EVE Acid	69087-46-3	2,2,3,3-tetrafluoro-3-((1,1,1,2,3,3-hexafluoro-3-[(1,2,2-trifluoroethenyl)oxy]propan-2-yl)oxy)propionic acid			X			X
FBSA	30334-69-1	Perfluorobutylsulfonamide			X			X
MeFBSA	68298-12-4	1-Butanesulfonamide; (N-(Methyl)nonafluorobutanesulfonamide); 1,1,2,2,3,3,4,4,4-nonafluoro-N-methyl-1-Butanesulfonamide			X			X
Hydro-EVE Acid	773804-62-9	2,2,3,3-Tetrafluoro-3-[[1,1,1,2,3,3-hexafluoro-3-(1,2,2,2-tetrafluoroethoxy)propan-2-yl]oxy}propanoic acid			X			X
R-EVE Acid	2416366-22-6	4-(2-carboxy-1,1,2,2-tetrafluoroethoxy)-2,2,3,3,4,5,5,5-octafluoro-pentanoic acid			X			X
NVHOS	1132933-86-8	Perfluoroethoxysulfonic acid; 1,1,2,2-Tetrafluoro-2-(1,2,2,2-tetrafluoroethoxy)ethane-1-sulfonic acid			X			X

PFAS Compound Acronym List			Methods					
Acronym	CAS #	Compound Name	SOP EU047	EPA 1633 (B-24)	EPA 1633X	EPA 537.1	EPA 533	EPA 8327*
PFDoS	79780-39-5	Perfluorododecane sulfonic acid		X	X			X
PFOA	16517-11-6	Perfluorooctadecanoic acid			X			
3:3 FTCA	356-02-5	2H,2H,3H,3H-Perfluorohexanoic acid		X	X			X
5:3 FTCA	914637-49-3	2H,2H,3H,3H-Perfluorooctanoic acid		X	X			X
7:3 FTCA	812-70-4	2H,2H,3H,3H-Perfluorodecanoic acid		X	X			X
N-AP-FHxSA	50598-28-2	N-(3-(Dimethylamino)propyl)tridecafluoro-1-hexanesulfonamide			X			X
N-CMAmP-6:2 FOSA	34455-29-3	N-(Carboxymethyl)-N,N-dimethyl-3-(((3,3,4,4,5,5,6,6,7,7,8,8,8-tridecafluorooctyl)sulfonyl)amino)1-propanaminium			X			X
BPAF	1478-61-1	Bisphenol AF			X			X
HQ-115	90076-65-6	Bis(trifluoromethane)sulfonimide lithium salt			X			X

* Accreditation pending

Results

Enthalpy Analytical

Job No.: 0126-854-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Summary

	Compound	CAS	012326-S01 ng/L	012326-E01 ng/L	
Acids	PFBA	375-22-4	4.53	5.69	
	PFPeA	2706-90-3	9.73	10.4	
	PFHxA	307-24-4	7.74	8.22	
	PFHpA	375-85-9	3.88	4.35	
	PFOA	335-67-1	5.18	5.27	
	PFNA	375-95-1	0.510 J	0.492 J	
	PFDA	335-76-2	0.184 J	0.175 J	
	PFUnDA	2058-94-8	ND U	ND U	
	PFDODA	307-55-1	ND U	ND U	
	PFTrDA	72629-94-8	ND U	ND U	
	PFTeDA	376-06-7	ND U	ND U	
	PFHxDA	67905-19-5	ND U	ND U	
	Sulfonates	PFBS	375-73-5	4.00	4.14
		PFPeS	2706-91-4	0.722	0.774
PFHxS		355-46-4	4.63	4.69	
PFHpS		375-92-8	0.120 L	0.0884 L	
PFOS		1763-23-1	8.39	8.63	
PFNS		68259-12-1	ND U	ND U	
PFDS		335-77-3	ND U	ND U	
4:2 FTS		757124-72-4	ND U	ND U	
6:2 FTS		27619-97-2	0.433 J	0.556	
8:2 FTS		39108-34-4	ND U	ND U	
10:2 FTS	120226-60-0	ND U	ND U		
Sulfonamidos	FBSA	30334-69-1	0.741	0.765	
	N-EtFOSA	4151-50-2	ND U	ND U	
	N-EtFOSAA	2991-50-6	ND U	ND U	
	N-EtFOSE	1691-99-2	ND U	ND U	
	N-MeFOSA	31506-32-8	ND U	ND U	
	N-MeFOSAA	2355-31-9	ND U	ND U	
	N-MeFOSE	24448-09-7	ND U	ND U	
PFOSA	754-91-6	0.0151 L	ND U		
PFECAs	ADONA	919005-14-4	ND U	ND U	
	EVE Acid	69087-46-3	0.127 L	0.114 L	
	HFPO-DA	13252-13-6	4.15	3.74	
	Hydro-EVE Acid	773804-62-9	0.214 J	0.254 J	
	NFDHA	151772-58-6	ND U	ND U	
	PEPA	267239-61-2	3.80	3.31	
	PFECA-G	801212-59-9	ND U	ND U	
	PFMOAA	674-13-5	3.67	2.25	
	PFMOBA	863090-89-5	ND U	ND U	
	PFMOPrA	377-73-1	0.166 L	0.198 L	
	PFO2HxA	39492-88-1	5.45	4.93	
	PFO3OA	39492-89-2	1.23	1.34	
	PFO4DA	39492-90-5	ND U	ND U	
	PFO5DA	39492-91-6	ND U	0.0330 L	
PMPA	13140-29-9	8.09	7.60		
R-EVE	2416366-22-6	2.95	3.78		
PFESAs	11Cl-PF3OUdS	763051-92-9	ND U	ND U	
	9Cl-PF3ONS	756426-58-1	ND U	ND U	
	Hydrolyzed PSDA	2416366-19-1	5.03	4.38	
	Nafion Byproduct 1 (PS Acid)	29311-67-9	0.291 L	0.199 L	
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	0.404 L	0.432 L	
	NVHOS	1132933-86-8	4.21	5.09	
	PFEESA	113507-82-7	ND U	ND U	
	R-PSDA	2416366-18-0	3.71	3.85	
R-PSDCA	2416366-21-5	ND U	ND U		

Enthalpy Analytical

Job No.: 0126-854-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	012326-S01	Prep Batch	EU119846
Sampling Site		Analyst	bmay
Enthalpy ID	0126-854-001-2	Instrument	Frodo
Matrix	aqueous	Sample Vol mL	290
Sampling Date	2026-01-23 08:30	Extract Vol mL	0.4
Received Date	2026-01-23	Split Factor	N/A
Prep Date	2026-01-26 14:49	Method Code	EU-047-NPW
AnalysisDate	2026-01-27 16:30		
SampleType	Sample		
Bottle ID	A		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFBA	375-22-4	FR270126023	4.53	0.219	0.552			
	PFFPeA	2706-90-3	FR270126023	9.73	0.158	0.552			
	PFFHxA	307-24-4	FR270126023	7.74	0.184	0.552			
	PFFHpA	375-85-9	FR270126023	3.88	0.193	0.552			
	PFOA	335-67-1	FR270126023	5.18	0.126	0.552			
	PFNA	375-95-1	FR270126023	0.510	0.125	0.552			J
	PFDA	335-76-2	FR270126023	0.184	0.158	0.552			J
	PFUnDA	2058-94-8	FR270126023	ND	0.125	0.552			U
	PFDODA	307-55-1	FR270126023	ND	0.224	0.552			U
	PFFTrDA	72629-94-8	FR270126023	ND	0.183	0.552			U
	PFFTeDA	376-06-7	FR270126023	ND	0.210	0.552			U
	PFFHxDA	67905-19-5	FR270126023	ND	0.293	0.552			U
	Sulfonates	PFFPeS	2706-91-4	FR270126023	0.722	0.113	0.520		
PFFHxS		355-46-4	FR270126023	4.63	0.426	0.505			
PFFHpS		375-92-8	FR270126023	0.120	0.267	0.526			L
PFOs		1763-23-1	FR270126023	8.39	0.291	0.511			
PFNS		68259-12-1	FR270126023	ND	0.171	0.531			U
PFDS		335-77-3	FR270126023	ND	0.290	0.531			U
4:2 FTS		757124-72-4	FR270126023	ND	0.0716	0.517			U
6:2 FTS		27619-97-2	FR270126023	0.433	0.260	0.526			J
8:2 FTS		39108-34-4	FR270126023	ND	0.124	0.528			U
10:2 FTS		120226-60-0	FR270126023	ND	0.422	0.552			U
Sulfonamidos	FBSA	30334-69-1	FR270126023	0.741	0.262	0.552			
	N-EtFOSA	4151-50-2	FR270126023	ND	0.341	0.552			U
	N-EtFOSAA	2991-50-6	FR270126023	ND	0.224	0.552			U
	N-EtFOSE	1691-99-2	FR270126023	ND	0.845	2.48			U
	N-MeFOSA	31506-32-8	FR270126023	ND	0.228	0.552			U
	N-MeFOSAA	2355-31-9	FR270126023	ND	0.155	0.552			U
	N-MeFOSE	24448-09-7	FR270126023	ND	0.524	2.48			U
	PFOSA	754-91-6	FR270126023	0.0151	0.0774	0.552			L
	ADONA	919005-14-4	FR270126023	ND	0.149	0.523			U
PFECAs	EVE Acid	69087-46-3	FR270126023	0.127	0.176	1.24			L
	HFPO-DA	13252-13-6	FR270126023	4.15	0.0584	0.552			
	Hydro-EVE Acid	773804-62-9	FR270126023	0.214	0.181	0.552			J
	NFDHA	151772-58-6	FR270126023	ND	0.116	0.552			U
	PEPA	267239-61-2	FR270126023	3.80	0.103	0.552			
	PFECA-G	801212-59-9	FR270126023	ND	0.0736	0.552			U
	PFMOAA	674-13-5	FR270126023	3.67	0.279	0.552			
	PFMOBA	863090-89-5	FR270126023	ND	0.926	1.24			U
	PFMOPrA	377-73-1	FR270126023	0.166	0.197	0.552			L
	PFO2HxA	39492-88-1	FR270126023	5.45	0.178	0.552			
	PFO3OA	39492-89-2	FR270126023	1.23	0.253	0.552			
	PFO4DA	39492-90-5	FR270126023	ND	0.436	2.76			U
	PFO5DA	39492-91-6	FR270126023	ND	0.441	2.76			U
	PMPA	13140-29-9	FR270126023	8.09	0.130	0.552			
	R-EVE	2416366-22-6	FR270126023	2.95	0.916	1.24			
	PFESAs	11Cl-PF3OUds	763051-92-9	FR270126023	ND	0.260	0.520		
9Cl-PF3ONS		756426-58-1	FR270126023	ND	0.353	0.514			U
Hydrolyzed PSDA		2416366-19-1	FR270126023	5.03	0.367	0.552			
Nafion Byproduct 1 (PS Acid)		29311-67-9	FR270126023	0.291	0.295	0.552			L
Nafion Byproduct 2 (Hydro-PS Acid)		749836-20-2	FR270126023	0.404	0.457	0.552			L
NVHOS		1132933-86-8	FR270126023	4.21	0.0850	0.552			
PFEESA		113507-82-7	FR270126023	ND	0.166	0.552			U
R-PSDA		2416366-18-0	FR270126023	3.71	2.43	2.43			
ES	R-PSDCA	2416366-21-5	FR270126023	ND	0.233	0.552			U
	MPFBA		FR270126023				20-150%	59.6%	
	M5PFFPeA		FR270126023				20-150%	141%	
	M3PFBS		FR270126023				20-150%	152%	Q
	M2-4:2 FTS		FR270126023				20-150%	100%	
	M5PFFHxA		FR270126023				20-150%	93.5%	
	M3HFPO-DA		FR270126023				20-150%	100%	
	M4PFFHpA		FR270126023				20-150%	94.9%	

Enthalpy Analytical

Job No.: 0126-854-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	012326-S01		
Sampling Site			
Enthalpy ID	0126-854-001-2	Prep Batch	EU119846
Matrix	aqueous	Analyst	bmay
Sampling Date	2026-01-23 08:30	Instrument	Frodo
Received Date	2026-01-23	Sample Vol mL	290
Prep Date	2026-01-26 14:49	Extract Vol mL	0.4
AnalysisDate	2026-01-27 16:30	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	A		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M3PFHxS		FR270126023				20-150%	92.7%	
M2-6:2 FTS		FR270126023				20-150%	90.0%	
M8PFOA		FR270126023				20-150%	94.1%	
M9PFNA		FR270126023				20-150%	97.1%	
M8PFOS		FR270126023				20-150%	88.6%	
M2-8:2 FTS		FR270126023				20-150%	77.7%	
M8FOSA-I		FR270126023				20-150%	88.8%	
M6PFDA		FR270126023				20-150%	89.1%	
d3-N-MeFOSAA		FR270126023				20-150%	78.0%	
d5-N-EtFOSAA		FR270126023				20-150%	77.7%	
M7PFUdA		FR270126023				20-150%	83.8%	
MPFDoA		FR270126023				20-150%	74.1%	
M2PFTeDA		FR270126023				20-150%	52.0%	
d3-N-MeFOSA		FR270126023				10-200%	52.3%	
d5-N-EtFOSA		FR270126023				10-200%	46.0%	
d7-N-MeFOSE		FR270126023				10-200%	64.8%	
d9-N-EtFOSE		FR270126023				10-200%	59.1%	

Enthalpy Analytical

Job No.: 0126-854-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	012326-S01		
Sampling Site			
Enthalpy ID	0126-854-001-3	Prep Batch	EU119979
Matrix	aqueous	Analyst	bmay
Sampling Date	2026-01-23 08:30	Instrument	Frodo
Received Date	2026-01-23	Sample Vol mL	295.19
Prep Date	2026-02-14 07:45	Extract Vol mL	0.4
AnalysisDate	2026-02-15 14:11	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	B		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Sulfonates	PFBS	375-73-5	FR150226005	4.00	0.288	0.542			
ES	M3PFBS		FR150226005				20-150%	230%	Q

Enthalpy Analytical

Job No.: 0126-854-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	012326-E01	Prep Batch	EU119846
Sampling Site		Analyst	bmay
Enthalpy ID	0126-854-002-2	Instrument	Frodo
Matrix	aqueous	Sample Vol mL	281.6
Sampling Date	2026-01-23 08:30	Extract Vol mL	0.4
Received Date	2026-01-23	Split Factor	N/A
Prep Date	2026-01-26 14:49	Method Code	EU-047-NPW
AnalysisDate	2026-01-27 16:52		
SampleType	Sample		
Bottle ID	A		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	FR270126024	5.69	0.225	0.568				
	PFFPeA	2706-90-3	FR270126024	10.4	0.162	0.568				
	PFFHxA	307-24-4	FR270126024	8.22	0.190	0.568				
	PFFHpA	375-85-9	FR270126024	4.35	0.199	0.568				
	PFOA	335-67-1	FR270126024	5.27	0.130	0.568				
	PFNA	375-95-1	FR270126024	0.492	0.128	0.568			J	
	PFDA	335-76-2	FR270126024	0.175	0.162	0.568			J	
	PFUnDA	2058-94-8	FR270126024	ND	0.128	0.568			U	
	PFDODA	307-55-1	FR270126024	ND	0.231	0.568			U	
	PFTTrDA	72629-94-8	FR270126024	ND	0.188	0.568			U	
	PFTTeDA	376-06-7	FR270126024	ND	0.217	0.568			U	
	PFFHxDA	67905-19-5	FR270126024	ND	0.302	0.568			U	
	Sulfonates	PFFPeS	2706-91-4	FR270126024	0.774	0.117	0.535			
PFFHxS		355-46-4	FR270126024	4.69	0.439	0.520				
PFFHpS		375-92-8	FR270126024	0.0884	0.275	0.541			L	
PFOS		1763-23-1	FR270126024	8.63	0.300	0.526				
PFNS		68259-12-1	FR270126024	ND	0.176	0.547			U	
PFDS		335-77-3	FR270126024	ND	0.298	0.547			U	
4:2 FTS		757124-72-4	FR270126024	ND	0.0737	0.532			U	
6:2 FTS		27619-97-2	FR270126024	0.556	0.268	0.541				
8:2 FTS		39108-34-4	FR270126024	ND	0.127	0.544			U	
10:2 FTS		120226-60-0	FR270126024	ND	0.435	0.568			U	
Sulfonamidos	FBSA	30334-69-1	FR270126024	0.765	0.270	0.568				
	N-EtFOSA	4151-50-2	FR270126024	ND	0.352	0.568			U	
	N-EtFOSAA	2991-50-6	FR270126024	ND	0.231	0.568			U	
	N-EtFOSE	1691-99-2	FR270126024	ND	0.870	2.56			U	
	N-MeFOSA	31506-32-8	FR270126024	ND	0.234	0.568			U	
	N-MeFOSAA	2355-31-9	FR270126024	ND	0.160	0.568			U	
	N-MeFOSE	24448-09-7	FR270126024	ND	0.540	2.56			U	
	PFOSA	754-91-6	FR270126024	ND	0.0797	0.568			U	
	ADONA	919005-14-4	FR270126024	ND	0.154	0.538			U	
	PFECAs	EVE Acid	69087-46-3	FR270126024	0.114	0.181	1.28			L
HFPO-DA		13252-13-6	FR270126024	3.74	0.0602	0.568				
Hydro-EVE Acid		773804-62-9	FR270126024	0.254	0.186	0.568			J	
NFDHA		151772-58-6	FR270126024	ND	0.119	0.568			U	
PEPA		267239-61-2	FR270126024	3.31	0.107	0.568				
PFECA-G		801212-59-9	FR270126024	ND	0.0758	0.568			U	
PFMOAA		674-13-5	FR270126024	2.25	0.288	0.568				
PFMOBA		863090-89-5	FR270126024	ND	0.953	1.28			U	
PFMOPrA		377-73-1	FR270126024	0.198	0.202	0.568			L	
PFO2HxA		39492-88-1	FR270126024	4.93	0.183	0.568				
PFO3OA		39492-89-2	FR270126024	1.34	0.261	0.568				
PFO4DA		39492-90-5	FR270126024	ND	0.449	2.84			U	
PFO5DA		39492-91-6	FR270126024	0.0330	0.455	2.84			L	
PMPA		13140-29-9	FR270126024	7.60	0.134	0.568				
R-EVE		2416366-22-6	FR270126024	3.78	0.943	1.28				
PFESAs		11Cl-PF3OUds	763051-92-9	FR270126024	ND	0.268	0.535			U
		9Cl-PF3ONS	756426-58-1	FR270126024	ND	0.364	0.529			U
	Hydrolyzed PSDA	2416366-19-1	FR270126024	4.38	0.378	0.568				
	Nafion Byproduct 1 (PS Acid)	29311-67-9	FR270126024	0.199	0.304	0.568			L	
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	FR270126024	0.432	0.471	0.568			L	
	NVHOS	1132933-86-8	FR270126024	5.09	0.0875	0.568				
	PFEESA	113507-82-7	FR270126024	ND	0.171	0.568			U	
	R-PSDA	2416366-18-0	FR270126024	3.85	2.50	2.50				
ES	R-PSDCA	2416366-21-5	FR270126024	ND	0.240	0.568			U	
	MPFBA		FR270126024				20-150%	55.2%		
	M5PFFPeA		FR270126024				20-150%	133%		
	M3PFBS		FR270126024				20-150%	140%		
	M2-4:2 FTS		FR270126024				20-150%	99.2%		
	M5PFFHxA		FR270126024				20-150%	88.0%		
	M3HFPO-DA		FR270126024				20-150%	95.9%		
	M4PFHpA		FR270126024				20-150%	87.9%		

Enthalpy Analytical

Job No.: 0126-854-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	012326-E01		
Sampling Site			
Enthalpy ID	0126-854-002-2	Prep Batch	EU119846
Matrix	aqueous	Analyst	bmay
Sampling Date	2026-01-23 08:30	Instrument	Frodo
Received Date	2026-01-23	Sample Vol mL	281.6
Prep Date	2026-01-26 14:49	Extract Vol mL	0.4
AnalysisDate	2026-01-27 16:52	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	A		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M3PFHxS		FR270126024				20-150%	91.3%	
M2-6:2 FTS		FR270126024				20-150%	84.8%	
M8PFOA		FR270126024				20-150%	88.1%	
M9PFNA		FR270126024				20-150%	93.1%	
M8PFOS		FR270126024				20-150%	85.0%	
M2-8:2 FTS		FR270126024				20-150%	75.9%	
M8FOSA-I		FR270126024				20-150%	87.1%	
M6PFDA		FR270126024				20-150%	87.9%	
d3-N-MeFOSAA		FR270126024				20-150%	73.4%	
d5-N-EtFOSAA		FR270126024				20-150%	76.0%	
M7PFUdA		FR270126024				20-150%	81.4%	
MPFDoA		FR270126024				20-150%	74.4%	
M2PFTeDA		FR270126024				20-150%	67.0%	
d3-N-MeFOSA		FR270126024				10-200%	50.1%	
d5-N-EtFOSA		FR270126024				10-200%	45.7%	
d7-N-MeFOSE		FR270126024				10-200%	72.4%	
d9-N-EtFOSE		FR270126024				10-200%	70.6%	

Enthalpy Analytical

Job No.: 0126-854-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	012326-E01		
Sampling Site			
Enthalpy ID	0126-854-002-3	Prep Batch	EU119979
Matrix	aqueous	Analyst	bmay
Sampling Date	2026-01-23 08:30	Instrument	Frodo
Received Date	2026-01-23	Sample Vol mL	288.28
Prep Date	2026-02-14 07:45	Extract Vol mL	0.4
AnalysisDate	2026-02-15 14:34	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	B		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Sulfonates	PFBS	375-73-5	FR150226006	4.14	0.295	0.555			
ES	M3PFBS		FR150226006				20-150%	256%	Q

QC Data

Enthalpy Analytical

Job No.: 0126-854-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	MB_119846_PFAS	Prep Batch	EU119846
Sampling Site		Analyst	bmay
Enthalpy ID	MB_119846_PFAS	Instrument	Frodo
Matrix	aqueous	Sample Vol mL	250
Sampling Date		Extract Vol mL	0.4
Received Date		Split Factor	N/A
Prep Date	2026-01-26 14:49	Method Code	EU-047-NPW
AnalysisDate	2026-01-27 15:44		
SampleType	Blank		
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	FR270126021	ND	0.254	0.640			U	
	PFPeA	2706-90-3	FR270126021	ND	0.183	0.640			U	
	PFHxA	307-24-4	FR270126021	ND	0.214	0.640			U	
	PFHpA	375-85-9	FR270126021	ND	0.224	0.640			U	
	PFOA	335-67-1	FR270126021	ND	0.146	0.640			U	
	PFNA	375-95-1	FR270126021	ND	0.145	0.640			U	
	PFDA	335-76-2	FR270126021	ND	0.183	0.640			U	
	PFUnDA	2058-94-8	FR270126021	ND	0.145	0.640			U	
	PFDoDA	307-55-1	FR270126021	ND	0.260	0.640			U	
	PFTrDA	72629-94-8	FR270126021	ND	0.212	0.640			U	
	PFTeDA	376-06-7	FR270126021	ND	0.244	0.640			U	
	PFHxDA	67905-19-5	FR270126021	ND	0.340	0.640			U	
	Sulfonates	PFPeS	2706-91-4	FR270126021	ND	0.131	0.603			U
		PFHxS	355-46-4	FR270126021	ND	0.494	0.586			U
PFHpS		375-92-8	FR270126021	ND	0.310	0.610			U	
PFOS		1763-23-1	FR270126021	ND	0.338	0.593			U	
PFNS		68259-12-1	FR270126021	ND	0.199	0.616			U	
PFDS		335-77-3	FR270126021	ND	0.336	0.616			U	
4:2 FTS		757124-72-4	FR270126021	ND	0.0830	0.600			U	
6:2 FTS		27619-97-2	FR270126021	ND	0.302	0.610			U	
8:2 FTS		39108-34-4	FR270126021	ND	0.143	0.613			U	
10:2 FTS		120226-60-0	FR270126021	ND	0.490	0.640			U	
Sulfonamidos	FBSA	30334-69-1	FR270126021	ND	0.304	0.640			U	
	N-EtFOSA	4151-50-2	FR270126021	ND	0.396	0.640			U	
	N-EtFOSAA	2991-50-6	FR270126021	ND	0.260	0.640			U	
	N-EtFOSE	1691-99-2	FR270126021	ND	0.980	2.88			U	
	N-MeFOSA	31506-32-8	FR270126021	ND	0.264	0.640			U	
	N-MeFOSAA	2355-31-9	FR270126021	ND	0.180	0.640			U	
	N-MeFOSE	24448-09-7	FR270126021	ND	0.608	2.88			U	
	PFOSA	754-91-6	FR270126021	ND	0.0898	0.640			U	
	PFECAs	ADONA	919005-14-4	FR270126021	ND	0.173	0.606			U
		EVE Acid	69087-46-3	FR270126021	ND	0.204	1.44			U
HFPO-DA		13252-13-6	FR270126021	ND	0.0678	0.640			U	
Hydro-EVE Acid		773804-62-9	FR270126021	ND	0.210	0.640			U	
NFDHA		151772-58-6	FR270126021	ND	0.135	0.640			U	
PEPA		267239-61-2	FR270126021	ND	0.120	0.640			U	
PFECA-G		801212-59-9	FR270126021	ND	0.0854	0.640			U	
PFMOAA		674-13-5	FR270126021	ND	0.324	0.640			U	
PFMOBA		863090-89-5	FR270126021	ND	1.07	1.44			U	
PFMOPrA		377-73-1	FR270126021	ND	0.228	0.640			U	
PFO2HxA		39492-88-1	FR270126021	ND	0.206	0.640			U	
PFO3OA		39492-89-2	FR270126021	ND	0.294	0.640			U	
PFO4DA		39492-90-5	FR270126021	ND	0.506	3.20			U	
PFO5DA		39492-91-6	FR270126021	ND	0.512	3.20			U	
PMPA		13140-29-9	FR270126021	ND	0.151	0.640			U	
R-EVE		2416366-22-6	FR270126021	ND	1.06	1.44			U	
PFESAs		11Cl-PF3OUdS	763051-92-9	FR270126021	ND	0.302	0.603			U
		9Cl-PF3ONS	756426-58-1	FR270126021	ND	0.410	0.596			U
	Hydrolyzed PSDA	2416366-19-1	FR270126021	ND	0.426	0.640			U	
	Nafion Byproduct 1 (PS Acid)	29311-67-9	FR270126021	ND	0.342	0.640			U	
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	FR270126021	ND	0.530	0.640			U	
	NVHOS	1132933-86-8	FR270126021	ND	0.0986	0.640			U	
	PFEEA	113507-82-7	FR270126021	ND	0.192	0.640			U	
	R-PSDA	2416366-18-0	FR270126021	ND	2.82	2.82			U	
ES	R-PSDCA	2416366-21-5	FR270126021	ND	0.270	0.640			U	
	MPFBA		FR270126021				20-150%	82.2%		
	M5PFPeA		FR270126021				20-150%	85.9%		
	M3PFBS		FR270126021				20-150%	81.0%		
	M2-4:2 FTS		FR270126021				20-150%	91.4%		
	M5PFHxA		FR270126021				20-150%	92.3%		
	M3HFPO-DA		FR270126021				20-150%	100%		
M4PFHpA		FR270126021				20-150%	89.2%			

Enthalpy Analytical

Job No.: 0126-854-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	MB_119846_PFAS	Prep Batch	EU119846
Sampling Site		Analyst	bmay
Enthalpy ID	MB_119846_PFAS	Instrument	Frodo
Matrix	aqueous	Sample Vol mL	250
Sampling Date		Extract Vol mL	0.4
Received Date		Split Factor	N/A
Prep Date	2026-01-26 14:49	Method Code	EU-047-NPW
AnalysisDate	2026-01-27 15:44		
SampleType	Blank		
Bottle ID	-		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M3PFHxS		FR270126021				20-150%	90.9%	
M2-6:2 FTS		FR270126021				20-150%	82.9%	
M8PFOA		FR270126021				20-150%	87.0%	
M9PFNA		FR270126021				20-150%	90.3%	
M8PFOS		FR270126021				20-150%	88.9%	
M2-8:2 FTS		FR270126021				20-150%	73.7%	
M8FOSA-I		FR270126021				20-150%	83.4%	
M6PFDA		FR270126021				20-150%	86.6%	
d3-N-MeFOSAA		FR270126021				20-150%	75.1%	
d5-N-EtFOSAA		FR270126021				20-150%	73.2%	
M7PFUdA		FR270126021				20-150%	81.4%	
MPFDaA		FR270126021				20-150%	75.2%	
M2PFTeDA		FR270126021				20-150%	72.7%	
d3-N-MeFOSA		FR270126021				10-200%	34.2%	
d5-N-EtFOSA		FR270126021				10-200%	32.2%	
d7-N-MeFOSE		FR270126021				10-200%	69.9%	
d9-N-EtFOSE		FR270126021				10-200%	68.3%	

Enthalpy Analytical

Job No.: 0126-854-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	MB_119979_PFAS		
Sampling Site			
Enthalpy ID	MB_119979_PFAS	Prep Batch	EU119979
Matrix	aqueous	Analyst	bmay
Sampling Date		Instrument	Frodo
Received Date		Sample Vol mL	250
Prep Date	2026-02-14 07:45	Extract Vol mL	0.4
AnalysisDate	2026-02-15 13:26	Split Factor	N/A
SampleType	Blank	Method Code	EU-047-NPW
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Sulfonates	PFBS	375-73-5	FR150226003	ND	0.340	0.640			U
ES	M3PFBS		FR150226003				20-150%	103%	

Enthalpy Analytical

Job No.: 0126-854-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Enthalpy ID	OPR_119846_PFAS	Prep Batch	EU119846	Sample Vol (mL)	250
Sample Name	OPR_119846_PFAS	Prep Date	2026-01-26 14:49	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	2026-01-27 16:07	Split Factor	N/A
Sampling Date		Analyst	bmay	Method Code	EU-047-NPW
Received Date		Instrument	Frodo	Sample Type	Control
		Bottle ID	-		

	Compound	CAS	InjFileName	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	FR270126022	18.5	0.254	0.640	47.9-144%	92.5%		
	PFPeA	2706-90-3	FR270126022	18.8	0.183	0.640	41.7-159%	94.0%		
	PFHxA	307-24-4	FR270126022	18.6	0.214	0.640	43.2-154%	92.8%		
	PFHpA	375-85-9	FR270126022	20.0	0.224	0.640	42.1-155%	100%		
	PFOA	335-67-1	FR270126022	19.0	0.146	0.640	51.1-148%	94.9%		
	PFNA	375-95-1	FR270126022	18.1	0.145	0.640	51.6-153%	90.3%		
	PFDA	335-76-2	FR270126022	18.4	0.183	0.640	44.5-156%	92.2%		
	PFUnDA	2058-94-8	FR270126022	19.0	0.145	0.640	40.3-156%	94.8%		
	PFDoDA	307-55-1	FR270126022	19.1	0.260	0.640	40.4-158%	95.6%		
	PFTriDA	72629-94-8	FR270126022	21.1	0.212	0.640	42.2-201%	105%		
	PFTeDA	376-06-7	FR270126022	19.7	0.244	0.640	43-162%	98.5%		
	Sulfonates	PFPeS	2706-91-4	FR270126022	17.4	0.131	0.603	40.3-152%	92.2%	
		PFHxS	355-46-4	FR270126022	15.7	0.494	0.586	45-148%	85.6%	
PFHpS		375-92-8	FR270126022	18.0	0.310	0.610	39.8-166%	94.3%		
PFOS		1763-23-1	FR270126022	17.3	0.338	0.593	59.2-132%	93.3%		
PFNS		68259-12-1	FR270126022	17.9	0.199	0.616	38.1-153%	93.0%		
PFDS		335-77-3	FR270126022	17.5	0.336	0.616	28.6-148%	90.7%		
4:2 FTS		757124-72-4	FR270126022	18.0	0.0830	0.600	41.5-157%	96.0%		
6:2 FTS		27619-97-2	FR270126022	18.0	0.302	0.610	44.5-160%	94.4%		
8:2 FTS		39108-34-4	FR270126022	18.7	0.143	0.613	39.4-166%	97.5%		
Sulfonamidos	N-EtFOSA	4151-50-2	FR270126022	18.2	0.396	0.640	26.7-172%	90.9%		
	N-EtFOSAA	2991-50-6	FR270126022	18.2	0.260	0.640	42.8-156%	91.2%		
	N-EtFOSE	1691-99-2	FR270126022	79.6	0.980	2.88	38.9-161%	88.4%		
	N-MeFOSA	31506-32-8	FR270126022	16.8	0.264	0.640	26.4-183%	84.0%		
	N-MeFOSAA	2355-31-9	FR270126022	17.9	0.180	0.640	42-155%	89.3%		
	N-MeFOSE	24448-09-7	FR270126022	77.2	0.608	2.88	37.6-155%	85.7%		
	PFOSA	754-91-6	FR270126022	18.8	0.0898	0.640	39.1-158%	94.2%		
PFECAs	ADONA	919005-14-4	FR270126022	17.9	0.173	0.606	32.2-151%	89.6%		
	HFPO-DA	13252-13-6	FR270126022	15.6	0.0678	0.640	61.8-131%	77.8%		
PFESAs	11Cl-PF3OUdS	763051-92-9	FR270126022	15.8	0.302	0.603	21.8-141%	79.1%		
	9Cl-PF3ONS	756426-58-1	FR270126022	17.1	0.410	0.596	37.6-146%	85.3%		
ES	MPFBA		FR270126022				20-150%	86.8%		
	M5PFPeA		FR270126022				20-150%	86.7%		
	M3PFBS		FR270126022				20-150%	85.7%		
	M2-4:2 FTS		FR270126022				20-150%	90.7%		
	M5PFHxA		FR270126022				20-150%	93.9%		
	M3HFPO-DA		FR270126022				20-150%	105%		
	M4PFHpA		FR270126022				20-150%	89.6%		
	M3PFHxS		FR270126022				20-150%	92.5%		
	M2-6:2 FTS		FR270126022				20-150%	84.2%		
	M8PFOA		FR270126022				20-150%	88.4%		
	M9PFNA		FR270126022				20-150%	91.9%		
	M8PFOS		FR270126022				20-150%	85.5%		
	M2-8:2 FTS		FR270126022				20-150%	76.3%		
	M8FOSA-I		FR270126022				20-150%	85.7%		
	M6PFDA		FR270126022				20-150%	88.5%		
	d3-N-MeFOSAA		FR270126022				20-150%	76.3%		
	d5-N-EtFOSAA		FR270126022				20-150%	76.8%		
	M7PFUdA		FR270126022				20-150%	80.4%		
	MPFDoA		FR270126022				20-150%	78.8%		
	M2PFTeDA		FR270126022				20-150%	74.0%		
	d3-N-MeFOSA		FR270126022				10-200%	38.3%		
d5-N-EtFOSA		FR270126022				10-200%	34.1%			
d7-N-MeFOSE		FR270126022				10-200%	74.3%			
d9-N-EtFOSE		FR270126022				10-200%	71.0%			

Enthalpy Analytical

Job No.: 0126-854-1 PFAS by Isotope Dilution (non-potable water)

Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Enthalpy ID	OPR_119979_PFAS	Prep Batch	EU119979	Sample Vol (mL)	250
Sample Name	OPR_119979_PFAS	Prep Date	2026-02-14 07:45	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	2026-02-15 13:48	Split Factor	N/A
Sampling Date		Analyst	bmay	Method Code	EU-047-NPW
Received Date		Instrument	Frodo	Sample Type	Control
		Bottle ID	-		

	Compound	CAS	InjFileName	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Sulfonates	PFBS	375-73-5	FR150226004	17.1	0.340	0.640	42.7-155%	96.2%	
ES	M3PFBS		FR150226004				20-150%	100%	

Narrative Summary

Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0126-854-2
Client ID.	NORTHWEST WATER PLANT

1. Custody

Isabelle Martin received the samples at 2.1 °C after being relinquished by Brunswick County Public Utilities - NC.

The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC.

Table 1 - Sample Inventory

EU Lab Sample ID	Client Sample ID	Matrix	Received
0126-854-001-1	012326-S01	aqueous	2026-01-23
0126-854-002-1	012326-E01	aqueous	2026-01-23

2. Methods and Analytes

A list of analytes of interest and corresponding methods of analysis is shown in Table 3. Abbreviations are defined in the listed Appendices.

Table 3 - Methods and Analytes

EU Method	Analytes	Cleanup Method
EU047	PFPrA	n/a

3. Analysis

The samples were analyzed by direct injection using LC/MS/MS instrument Bumblebee.

4. Calibration

In the initial calibration, the reported analytes exhibited R^2 of ≥ 0.99 . The reported analytes in the calibration standards, Initial Calibration Verification (ICV) and continuing calibration (concal) met the accuracy criterion for native analytes.

5. QC Notes

The QC sample analyses passed all method criteria.

PFAS by Isotope Dilution (non-potable water) samples were extracted within 28 days, and extracts analyzed within 28 days.

Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0126-854-2
Client ID.	NORTHWEST WATER PLANT

6. Reporting Notes

The results presented in this report are representative of the samples as provided to the laboratory.

Some labeled extraction standards (ES) in the analyses recovered outside method control limits for ES recovery, as denoted by the "Q" qualifier. The target analytes are quantified based on their ratio to their labeled standard analogs. As a result, low or high labeled standard recovery do not cause any change to ratios or contribute any additional error in the measurement of the target analytes. When detected at a signal-to-noise above 10:1 the ES peak area is used to quantify its respective target analyte using accepted isotope dilution principles. The data is reported without adverse impact.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, LLC in Wilmington NC is accredited by the Louisiana Department of Environmental Quality to the 2016 TNI Standard under certificate number 05075.

Results

Enthalpy Analytical

Job No.: 0126-854-2 PFAS by Isotope Dilution (non-potable water)
Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Summary

	Compound	CAS	012326-S01 ng/L	012326-E01 ng/L
Acids	PFPrA	422-64-0	344 L	325 L

Enthalpy Analytical

Job No.: 0126-854-2 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	012326-S01		
Sampling Site			
Enthalpy ID	0126-854-001-1	Prep Batch	EU119839
Matrix	aqueous	Analyst	zoeardt
Sampling Date	2026-01-23 08:30	Instrument	Bumblebee
Received Date	2026-01-23	Sample Vol mL	0.1
Prep Date	2026-01-24 10:30	Extract Vol mL	0.2
AnalysisDate	2026-01-25 15:50	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	A		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	B250126-01251550	344	700	1530			L
ES	13C3-PFPrA		B250126-01251550				20-150%	115%	

Enthalpy Analytical

Job No.: 0126-854-2 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	012326-E01		
Sampling Site			
Enthalpy ID	0126-854-002-1	Prep Batch	EU119839
Matrix	aqueous	Analyst	zoeardt
Sampling Date	2026-01-23 08:30	Instrument	Bumblebee
Received Date	2026-01-23	Sample Vol mL	0.1
Prep Date	2026-01-24 10:30	Extract Vol mL	0.2
AnalysisDate	2026-01-25 16:02	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	A		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	B250126-01251602	325	700	1530			L
ES	¹³ C3-PFPrA		B250126-01251602				20-150%	203%	Q

QC Data

Enthalpy Analytical

Job No.: 0126-854-2 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	MB_119839_PFAS	Prep Batch	EU119839
Sampling Site		Analyst	zoeardt
Enthalpy ID	MB_119839_PFAS	Instrument	Bumblebee
Matrix	aqueous	Sample Vol mL	0.1
Sampling Date		Extract Vol mL	0.2
Received Date		Split Factor	N/A
Prep Date	2026-01-24 10:30	Method Code	EU-047-NPW
AnalysisDate	2026-01-25 15:27		
SampleType	Blank		
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPtA	422-64-0	B250126-01251527	340	700	1530			L
ES	¹³ C3-PFPtA		B250126-01251527				20-150%	109%	

Enthalpy Analytical

Job No.: 0126-854-2 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	OPR_119839_PFAS		
Sampling Site			
Enthalpy ID	OPR_119839_PFAS	Prep Batch	EU119839
Matrix	aqueous	Analyst	zoeardt
Sampling Date		Instrument	Bumblebee
Received Date		Sample Vol mL	0.1
Prep Date	2026-01-24 10:30	Extract Vol mL	0.2
AnalysisDate	2026-01-25 15:39	Split Factor	N/A
SampleType	Control	Method Code	EU-047-NPW
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	B250126-01251539	17200	700	1530	40-150%	86.1%	
ES	13C3-PFPrA		B250126-01251539				20-150%	106%	

Sample Custody

0126-854



Chain of Custody Record

Enthalpy Ultratrace Job#: _____ COC Page 1 of 1

Special Handling:

- Standard Turn Around Time
- Rush Turn Around Time -- Date Needed _____
- All Fast TATs Subject to Approval by Enthalpy Analytical, Inc.
- All Samples Disposed of After 6 months Unless Otherwise Instructed.

Enthalpy Analytical-Wilmington, NC has added enhancements to standard methods to improve accuracy, precision and permit an assessment of laboratory performance in the context of your specific data needs. For more information email Cindy.James@enthalpy.com.

Client Name: <u>BRUNSWICK COUNTY UTILITIES</u>	Project Number: _____	PO#: _____
Project Manager: <u>CHRIS GIESTING</u>	Site Name: <u>NORTHWEST WATER PLANT</u>	Telephone#: _____
Report To: <u>SAME</u>	Location: <u>LELAND N.C.</u>	Email: _____

This Chain of Custody is applicable to Non-Air samples. Standard TAT differ per analysis and are provided by request.

Client Special Instructions:						Sample Containers				Analyses:							Notes:		
Sample ID	Date	Time	Sample Volume	Type	Matrix	# of Bottles	# of Jars	# of Bags	# Other	Method 1613	Method 8290	Method 1668A/B/C PCE	PFAS by LC/MS/MS	PAHs by HRGC/HRMS	Sample on Hold	Method 23		ALL PFAS	
012326-S01	1/23/2026	8:30 AM	250 ml	G	NW	2												X	Please Add PFPrA and
012326-E01	1/23/2026	8:30 PM	250 ml	G	DW	2												X	PFHpA To The Testing.
																			Mark Hager Knows About
																			This If you Have Questions

ORIGINAL
IF NOT RED,
DESTROY THIS COPY AFTER USE

Relinquished By:	Date:	Received By:	Date:	Time:	Sample Temperature Upon Receipt:
PHIL MCCULLOCH	1/23/2026	<i>Isabelle Morito</i>	1/23/26	14:39	<input checked="" type="checkbox"/> Iced <input type="checkbox"/> Ambient °C <u>2.1</u>
					<input type="checkbox"/> Iced <input type="checkbox"/> Ambient °C _____
					<input type="checkbox"/> Iced <input type="checkbox"/> Ambient °C _____

COOLER RECEIPT LOG

JOB ID: <input type="text"/>	Date / Time: <u>1/23/24 14:39</u>	Initials: <u>IVM</u>
OR		
Client: <u>Brunswick Co. Utilities</u>		

Temp °C: <u>2.1°</u>	Thermometer ID: <u>T10</u>	Cooler <u>1</u> of <u>1</u> Bottle Order #: <u>N/A</u>												
Received via FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input checked="" type="checkbox"/> Other <input type="checkbox"/>	Check one On ice: <input checked="" type="checkbox"/> Melted ice: <input type="checkbox"/> Ambient: <input type="checkbox"/>	Check one in a Box: <input type="checkbox"/> in a Cooler: <input checked="" type="checkbox"/> Cooler in Box: <input type="checkbox"/>												
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th style="text-align: center;">Yes</th> <th style="text-align: center;">No</th> </tr> </thead> <tbody> <tr> <td>Cooler seals:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input checked="" type="checkbox"/></td> </tr> <tr> <td>Sample seals:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input checked="" type="checkbox"/></td> </tr> <tr> <td>Good condition:</td> <td style="text-align: center;"><input checked="" type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> </tbody> </table>		Yes	No	Cooler seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Sample seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Good condition:	<input checked="" type="checkbox"/>	<input type="checkbox"/>
	Yes	No												
Cooler seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>												
Sample seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>												
Good condition:	<input checked="" type="checkbox"/>	<input type="checkbox"/>												
Comment: <input style="width: 100%;" type="text"/>														

Temp °C: <input type="text"/>	Thermometer ID: <input type="text"/>	Cooler <input type="text"/> of <input type="text"/> Bottle Order #: <input type="text"/>												
Received via FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input type="checkbox"/> Other <input type="checkbox"/>	Check one On ice: <input type="checkbox"/> Melted ice: <input type="checkbox"/> Ambient: <input type="checkbox"/>	Check one in a Box: <input type="checkbox"/> in a Cooler: <input type="checkbox"/> Cooler in Box: <input type="checkbox"/>												
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th style="text-align: center;">Yes</th> <th style="text-align: center;">No</th> </tr> </thead> <tbody> <tr> <td>Cooler seals:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td>Sample seals:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td>Good condition:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> </tbody> </table>		Yes	No	Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>	Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>	Good condition:	<input type="checkbox"/>	<input type="checkbox"/>
	Yes	No												
Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>												
Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>												
Good condition:	<input type="checkbox"/>	<input type="checkbox"/>												
Comment: <input style="width: 100%;" type="text"/>														

1/23/24 IVM

Temp °C: <input type="text"/>	Thermometer ID: <input type="text"/>	Cooler <input type="text"/> of <input type="text"/> Bottle Order #: <input type="text"/>												
Received via FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input type="checkbox"/> Other <input type="checkbox"/>	Check one On ice: <input type="checkbox"/> Melted ice: <input type="checkbox"/> Ambient: <input type="checkbox"/>	Check one in a Box: <input type="checkbox"/> in a Cooler: <input type="checkbox"/> Cooler in Box: <input type="checkbox"/>												
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th style="text-align: center;">Yes</th> <th style="text-align: center;">No</th> </tr> </thead> <tbody> <tr> <td>Cooler seals:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td>Sample seals:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td>Good condition:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> </tbody> </table>		Yes	No	Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>	Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>	Good condition:	<input type="checkbox"/>	<input type="checkbox"/>
	Yes	No												
Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>												
Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>												
Good condition:	<input type="checkbox"/>	<input type="checkbox"/>												
Comment: <input style="width: 100%;" type="text"/>														