

Brunswick County Public Utilities - NC

PO Box 249
Bolivia, NC 28422-0249

Client Project# NORTHWEST WATER PLANT
Samples Received: 2/6/2026

Analytical Report 0226-767

PFAS by Isotope Dilution (non-potable water)

Report Issue Date: 2/26/2026

I certify that to the best of my knowledge all analytical data presented in this report have been checked for completeness, accuracy, errors and legibility in addition to having been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s). This analytical report was prepared in Portable Document Format (.PDF) and contains 35 pages. This report shall not be reproduced except in full without approval of the laboratory. This will provide assurance that parts of the report are not taken out of context.

Amendment(s):

Signature:



Alexandra Mejia, Quality Assurance Associate I



Enthalpy Analytical, LLC – Wilmington
Amanda Valois, Project Manager
amvalois@montrose-env.com / www.enthalpy.com
O: (910) 212-5858
2714 Exchange Drive, Wilmington, NC 28405

Table of Contents

PFAS by Isotope Dilution (non-potable water)	3
Case Narrative	3
General Reporting Notes	6
Results	12
Summary of Results	13
Detailed Results	15
QC Data	19
Blanks	20
Controls	22
PFAS by Isotope Dilution (non-potable water)	23
Case Narrative	23
Results	26
Summary of Results	27
Detailed Results	28
QC Data	30
Blanks	31
Controls	32
Sample Custody	33
Chain of Custody	34

Narrative Summary

Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0226-767-1
Client ID.	NORTHWEST WATER PLANT

1. Custody

Isabelle Martin received the samples at 1.6 °C after being relinquished by Brunswick County Public Utilities - NC.

The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC.

Table 1 - Sample Inventory

EU Lab Sample ID	Client Sample ID	Matrix	Received
0226-767-001-1	020626-S01	aqueous	2026-02-06
0226-767-002-1	020626-E01	aqueous	2026-02-06

2. Methods and Analytes

A list of analytes of interest and corresponding methods of analysis is shown in Table 3. Abbreviations are defined in the listed Appendices.

Table 3 - Methods and Analytes

EU Method	Analytes	Cleanup Method
EU047	Brunswick List	ENVI-Carb

3. Analysis

The samples were analyzed using LC/MS/MS instrument Frodo.

4. Calibration

In the initial calibration, the reported analytes exhibited R^2 of ≥ 0.99 . The reported analytes in the calibration standards, Initial Calibration Verification (ICV) and continuing calibration (concal) met the accuracy criterion for native analytes.

5. QC Notes

Ongoing Precision Recovery (OPR) control limits have not been established for some analytes of interest.

The QC sample analyses passed all method criteria.

PFAS by Isotope Dilution (non-potable water) samples were extracted within 28 days, and extracts analyzed within 28 days.

Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0226-767-1
Client ID.	NORTHWEST WATER PLANT

6. Reporting Notes

The results presented in this report are representative of the samples as provided to the laboratory.

This report provides all results including detections below LOD following client instruction.

Some labeled extraction standards (ES) in the analyses recovered outside method control limits for ES recovery, as denoted by the "Q" qualifier. The target analytes are quantified based on their ratio to their labeled standard analogs. As a result, low or high labeled standard recovery do not cause any change to ratios or contribute any additional error in the measurement of the target analytes. When detected at a signal-to-noise above 10:1 the ES peak area is used to quantify its respective target analyte using accepted isotope dilution principles. The data is reported without adverse impact.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, LLC in Wilmington NC is accredited by the Louisiana Department of Environmental Quality to the 2016 TNI Standard under certificate number 05075.

General Reporting Notes – Data Qualifiers

The following are general reporting notes that are applicable to all Enthalpy Analytical, LLC - Wilmington, NC data reports, unless specifically noted otherwise.

General Data Qualifiers

- Ac - Alternate calculation flag indicates the es recovery was calculated using the opening concal when either of the following situations is encountered in the data processing software: the ES recovery is over 400% or the JS is not detected.
- B – The analyte was found in the method blank, at a concentration that was at least 10% of the amount in the sample.
- Cxx – Two or more congeners co-elute. In EDDs, C denotes the lowest IUPAC congener in a co-elution group and additional co-eluters for the group ('xx') are shown with the number of the lowest IUPAC co-eluter.
- E – The reported concentration exceeds the calibration range (upper point of the calibration curve). For HRMS data, this condition does not imply additional measurement uncertainty. For LC-MS/MS data, these values should be considered as having measurement uncertainty higher than values within the calibration range.
- EDL – Estimated Detection Level: The EDL is unique to isotope dilution methods and reflects the conditions of analysis at the time of analysis, including the equipment used. Where the MDL is a static value, the EDL is a dynamic value.
- EMPC – Estimated Maximum Possible Concentration: EMPC is specific to Dioxin/Furan tests to indicate the determined ion-abundance ratio was outside the allowed theoretical range (usually due to being near the detection limit, although it can very rarely be caused by a co-eluting interference). The EMPC concentration is adjusted to reflect the value at the theoretical ion-abundance ratio.
- I/IR – The ion ratio between the primary and secondary ions was observed to be outside the method criteria. The analyte concentration may be inaccurate due to interference.
- J – The analyte has a concentration below the minimum calibration level (LOQ value) but greater than the LOD. These values should be considered as having measurement uncertainty higher than values within the calibration range
- L - For reports containing PFAS analytes only, this flag indicates that an analyte has a concentration below the Minimum Detection Limit (MDL) . The reported concentration is not recommended for regulatory use as the analyte signal may have a signal-to-noise ratio less than the criteria deemed necessary to be considered a detected analyte.
- LOD – Limit of Detection: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOD. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the minimum detection limit (MDL). The LOD is adjusted for sample weight or volume.

General Reporting Notes – Data Qualifiers

- LOQ – Limit of Quantitation: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOQ. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the reporting limit (RL). The LOQ is adjusted for sample weight or volume.
- <LOD() – Analyte was not found at a concentration high enough to be reported as detected. It is reported as less than the LOD, and the LOD is given in the parentheses.
- <LOQ() – Analyte was not found at a concentration high enough to be reported as above the QSM-defined LOQ or TNI defined Reporting Limit. It is reported as less than the LOQ, and the LOQ is given in the parentheses.
- ND – Indicates a non-detect.
- NR – Indicates a value that is not reportable due to issues observed in sample preparation or analysis.
- PR – The associated congener(s) is(are) poorly resolved.
- QI – Indicates the presence of a quantitative interference.
- RL – Reporting Limit. Lowest reportable value. The level is higher than the MDL.
- SI – Denotes “Single Ion Mode” and is utilized for PCBs where the secondary ion trace has a significantly elevated noise level due to background PFK. Responses for such peaks are calculated using an EMPC approach based solely on the primary ion area(s) and may be considered estimates.
- U – The analyte was not detected.
- V / Q – The labeled standard recovery is not within method control limits.
- X – Indicates the result is from re-injection/repeat/second-column analysis.

Lab Identifiers/ Data Attributes

- AR – Indicates use of the archived portion of the sample extract.
- CU – Indicates a sample that required additional clean-up prior to HRMS injection/processing.
- D – Dilution Data. Result was obtained from the analysis of a dilution. The number that follows the “D” indicates the dilution factor.
- DE – Indicates a dilution performed with the addition of ES (Extraction Standard) solution.
- DUP – Designation for a duplicate sample.
- MS – Designation for a matrix spike.
- MSD – Designation for a matrix spike duplicate.



General Reporting Notes – Data Qualifiers

- R – Indicates a re-extraction of the sample.
- RJ – Indicates a reinjection of the sample extract.
- S – Indicates a sample split. The number that follows the “S” indicates the split factor.
- SAT – Indicates an analyte saturated the detector.

PFAS Compound Acronym List			Methods					
Acronym	CAS #	Compound Name	SOP EU047	EPA 1633 (B-24)	EPA 1633X	EPA 537.1	EPA 533	EPA 8327*
Target Analytes								
PFBA	375-22-4	Perfluorobutanoic Acid	X	X	X		X	X
PFPeA	2706-90-3	Perfluoropentanoic Acid	X	X	X		X	X
PFHxA	307-24-4	Perfluorohexanoic Acid	X	X	X	X	X	X
PFHpA	375-85-9	Perfluoroheptanoic Acid	X	X	X	X	X	X
PFOA	335-67-1	Perfluorooctanoic Acid	X	X	X	X	X	X
PFNA	375-95-1	Perfluorononanoic Acid	X	X	X	X	X	X
PFDA	335-76-2	Perfluorodecanoic acid	X	X	X	X	X	X
PFUnA (PFUnDA)	2058-94-8	Perfluoroundecanoic acid	X	X	X	X	X	X
PFDoA (PFDoDA)	307-55-1	Perfluorododecanoic acid	X	X	X	X		X
PFTrDA (PFTriA, PFTrDA)	72629-94-8	Perfluorotridecanoic acid	X	X	X	X		X
PFTeDA (PFTA, PFTreA)	376-06-7	Perfluorotetradecanoic acid	X	X	X	X		X
PFBS	375-73-5	Perfluorobutane sulfonic acid	X	X	X	X	X	X
PFPeS	2706-91-4	Perfluoropentane sulfonic acid	X	X	X		X	X
PFHxS	355-46-4	Perfluorohexane sulfonic acid	X	X	X	X	X	X
PFHpS	375-92-8	Perfluoroheptane sulfonic acid	X	X	X		X	X
PFOS	1763-23-1	Perfluorooctane sulfonic acid	X	X	X	X	X	X
PFNS	68259-12-1	Perfluorononane sulfonic acid	X	X	X			X
PFDS	335-77-3	Perfluorodecane sulfonic acid	X	X	X			X
4:2 FTS	757124-72-4	4:2 fluorotelomer sulfonic acid	X	X	X		X	X
6:2 FTS	27619-97-2	6:2 fluorotelomer sulfonic acid	X	X	X		X	X
8:2 FTS	39108-34-4	8:2 fluorotelomer sulfonic acid	X	X	X		X	X
10:2 FTS	120226-60-0	Fluorotelomer sulfonate 10:2						X
FHxSA	41997-13-1	Perfluorohexanesulfonamide			X			X
PFOSA (FOSA)	754-91-6	Perfluorooctane sulfonamide	X	X	X			X
N-MeFOSAA	2355-31-9	N-methyl perfluorooctane sulfonamido acetic acid	X	X	X	X		X
N-MeFOSA	31506-32-8	N-methylperfluoro-1-octanesulfonamide	X	X	X			X
N-MeFOSE	24448-09-7	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol	X	X	X			X
N-EtFOSAA	2991-50-6	N-ethyl perfluorooctane sulfonamido acetic acid	X	X	X	X		X
N-EtFOSA	4151-50-2	N-ethylperfluoro-1-octanesulfonamide	X	X	X			X
N-EtFOSE	1691-99-2	2-(N-Ethylperfluoro-1-octanesulfonamido)-ethanol	X	X	X			X
HFPO-DA	13252-13-6	Hexafluoropropyleneoxide dimer acid (GenX)	X	X	X	X	X	X
11Cl-PF3OUds	763051-92-9	11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	X	X	X	X	X	X
9Cl-PF3ONS	756426-58-1	9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	X	X	X	X	X	X
ADONA	919005-14-4	4,8-dioxa-3H-perfluorononanoic acid	X	X	X	X	X	X
PFESA	113507-82-7	Perfluoro(2-ethoxyethane)sulfonic acid		X	X		X	X
PFMOBA (PFMBA)	863090-89-5	Perfluoro-4-methoxybutanoic acid		X	X		X	X
NFDHA	151772-58-6	Nonafluoro-3,6-dioxaheptanoic acid		X	X		X	X
PFMOPrA (PFMPA)	377-73-1	Perfluoro-3-methoxypropanoic acid		X	X		X	X
PFPrA	422-64-0	Perfluoropropionic acid, 2,2,3,3,3-Pentafluoropropionic acid			X			X
PFPrS (PFPS)	423-41-6	Perfluoropropanesulfonic acid			X			X



PFAS Compound Acronym List			Methods					
Acronym	CAS #	Compound Name	SOP EU047	EPA 1633 (B-24)	EPA 1633X	EPA 537.1	EPA 533	EPA 8327*
PFMOAA	674-13-5	Perfluoro-2-methoxyacetic acid;			X			X
PFO2HxA	39492-88-1	Perfluoro (3,5-dioxahexanoic) acid			X			X
PFO3OA	39492-89-2	Perfluoro (3,5,7-trioxaoctanoic) acid			X			X
PFO4DA	39492-90-5	Perfluoro (3,5,7,9-tetraoxadecanoic) acid			X			X
PFO5DA	39492-91-6	Perfluoro(3,5,7,9,11-pentaoxadodecanoic) acid			X			X
Nafion Byproduct 1 (PS Acid)	29311-67-9	1,1,2,2-tetrafluoro-2-[1,1,1,2,3,3-hexafluoro-3-(1,2,2-trifluoroethenoxy)propan-2-yl]oxyethanesulfonic acid			X			X
Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	Perfluoro-2-[[perfluoro-3-(perfluoroethoxy)-2-propanyl]oxy]ethanesulfonic acid (Hydro-PS Acid)			X			X
PEPA	267239-61-2	Perfluoro-2-ethoxypropanoic acid			X			X
PMPA	13140-29-9	Perfluoro-2-methoxypropanoic acid			X			X
PFECA-G, (PFPE-1)	801212-59-9	4-(Heptafluoroisopropoxy)hexafluorobutanoic acid, Perfluoro-4-isopropoxybutanoic acid			X			X
PFHxDA	67905-19-5	Perfluorohexadecanoic acid			X			
R-PSDA (Nafion Byproduct 4)	2416366-18-0	Perfluoro-4-(2-sulfoethoxy)pentanoic acid; 2,2,3,3,4,5,5-Octafluoro-4-(1,1,2,2-tetrafluoro-2-sulfoethoxy)pentanoic acid			X			X
Hydrolyzed PSDA (Nafion Byproduct 5)	2416366-19-1	2-fluoro-2-[1,1,2,3,3,3-hexafluoro-2-(1,1,2,2-tetrafluoro-2-sulfoethoxy)propoxy]-acetic acid			X			X
R-PSDCA (Nafion Byproduct 6)	2416366-21-5	1,1,2,2-tetrafluoro-2-[1,2,2,3,3-pentafluoro-1-(trifluoromethyl)propoxy] ethanesulfonic acid			X			X
EVE Acid	69087-46-3	2,2,3,3-tetrafluoro-3-((1,1,1,2,3,3-hexafluoro-3-[(1,2,2-trifluoroethenyl)oxy]propan-2-yl)oxy)propionic acid			X			X
FBSA	30334-69-1	Perfluorobutylsulfonamide			X			X
MeFBSA	68298-12-4	1-Butanesulfonamide; (N-(Methyl)nonafluorobutanesulfonamide); 1,1,2,2,3,3,4,4,4-nonafluoro-N-methyl-1-Butanesulfonamide			X			X
Hydro-EVE Acid	773804-62-9	2,2,3,3-Tetrafluoro-3-[[1,1,1,2,3,3-hexafluoro-3-(1,2,2,2-tetrafluoroethoxy)propan-2-yl]oxy}propanoic acid			X			X
R-EVE Acid	2416366-22-6	4-(2-carboxy-1,1,2,2-tetrafluoroethoxy)-2,2,3,3,4,5,5,5-octafluoro-pentanoic acid			X			X
NVHOS	1132933-86-8	Perfluoroethoxysulfonic acid; 1,1,2,2-Tetrafluoro-2-(1,2,2,2-tetrafluoroethoxy)ethane-1-sulfonic acid			X			X

PFAS Compound Acronym List			Methods					
Acronym	CAS #	Compound Name	SOP EU047	EPA 1633 (B-24)	EPA 1633X	EPA 537.1	EPA 533	EPA 8327*
PFDoS	79780-39-5	Perfluorododecane sulfonic acid		X	X			X
PFOA	16517-11-6	Perfluorooctadecanoic acid			X			
3:3 FTCA	356-02-5	2H,2H,3H,3H-Perfluorohexanoic acid		X	X			X
5:3 FTCA	914637-49-3	2H,2H,3H,3H-Perfluorooctanoic acid		X	X			X
7:3 FTCA	812-70-4	2H,2H,3H,3H-Perfluorodecanoic acid		X	X			X
N-AP-FHxSA	50598-28-2	N-(3-(Dimethylamino)propyl)tridecafluoro-1-hexanesulfonamide			X			X
N-CMAmP-6:2 FOSA	34455-29-3	N-(Carboxymethyl)-N,N-dimethyl-3-(((3,3,4,4,5,5,6,6,7,7,8,8,8-tridecafluorooctyl)sulfonyl)amino)1-propanaminium			X			X
BPAF	1478-61-1	Bisphenol AF			X			X
HQ-115	90076-65-6	Bis(trifluoromethane)sulfonimide lithium salt			X			X

* Accreditation pending

Results

Enthalpy Analytical

Job No.: 0226-767-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Summary

	Compound	CAS	020626-S01 ng/L	020626-E01 ng/L
Acids	PFBA	375-22-4	4.61	3.57
	PFPeA	2706-90-3	8.90	8.04
	PFHxA	307-24-4	6.67	6.36
	PFHpA	375-85-9	3.21	3.14
	PFOA	335-67-1	4.41	4.24
	PFNA	375-95-1	0.467 J	0.397 J
	PFDA	335-76-2	0.125 L	0.117 L
	PFUnDA	2058-94-8	ND U	ND U
	PFDoDA	307-55-1	ND U	ND U
	PFTTrDA	72629-94-8	ND U	ND U
	PFTeDA	376-06-7	ND U	ND U
	PFHxDA	67905-19-5	ND U	ND U
	Sulfonates	PFBS	375-73-5	3.95
PFPeS		2706-91-4	0.366 J	0.521 J
PFHxS		355-46-4	4.03	3.68
PFHpS		375-92-8	0.167 L	0.180 L
PFOS		1763-23-1	7.48	7.11
PFNS		68259-12-1	ND U	ND U
PFDS		335-77-3	ND U	ND U
4:2 FTS		757124-72-4	ND U	ND U
6:2 FTS		27619-97-2	0.700	0.689
8:2 FTS		39108-34-4	ND U	ND U
10:2 FTS	120226-60-0	ND U	ND U	
Sulfonamidos	FBSA	30334-69-1	0.369 J	0.503 J
	N-EtFOSA	4151-50-2	ND U	ND U
	N-EtFOSAA	2991-50-6	ND U	ND U
	N-EtFOSE	1691-99-2	ND U	ND U
	N-MeFOSA	31506-32-8	ND U	ND U
	N-MeFOSAA	2355-31-9	ND U	ND U
	N-MeFOSE	24448-09-7	ND U	ND U
	PFOSA	754-91-6	ND U	ND U
PFECAs	ADONA	919005-14-4	ND U	ND U
	EVE Acid	69087-46-3	0.0132 L	0.0237 L
	HFPO-DA	13252-13-6	3.14	3.67
	Hydro-EVE Acid	773804-62-9	0.172 L	0.190 J
	NFDHA	151772-58-6	ND U	ND U
	PEPA	267239-61-2	1.39	2.37
	PFECA-G	801212-59-9	ND U	ND U
	PFMOAA	674-13-5	7.96	5.24
	PFMOBA	863090-89-5	ND U	ND U
	PFMOPrA	377-73-1	ND U	ND U
	PFO2HxA	39492-88-1	3.14	3.07

Enthalpy Analytical

Job No.: 0226-767-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Summary

	Compound	CAS	020626-S01 ng/L	020626-E01 ng/L
PFECAs	PFO3OA	39492-89-2	ND U	0.883
	PFO4DA	39492-90-5	ND U	ND U
	PFO5DA	39492-91-6	ND U	ND U
	PMPA	13140-29-9	5.78	7.63
	R-EVE	2416366-22-6	2.85	5.25
PFESAs	11Cl-PF3OUdS	763051-92-9	ND U	ND U
	9Cl-PF3ONS	756426-58-1	ND U	ND U
	Hydrolyzed PSDA	2416366-19-1	2.08	4.19
	Nafion Byproduct 1 (PS Acid)	29311-67-9	ND U	ND U
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	0.522 J	0.496 J
	NVHOS	1132933-86-8	7.71	7.19
	PFEESA	113507-82-7	ND U	ND U
	R-PSDA	2416366-18-0	4.24	8.03
	R-PSDCA	2416366-21-5	ND U	ND U

Enthalpy Analytical

Job No.: 0226-767-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	020626-S01	Prep Batch	EU119928
Sampling Site		Analyst	zoearndt
Enthalpy ID	0226-767-001-1	Instrument	Frodo
Matrix	aqueous	Sample Vol mL	285.4
Sampling Date	2026-02-06 10:30	Extract Vol mL	0.4
Received Date	2026-02-06	Split Factor	N/A
Prep Date	2026-02-09 08:16	Method Code	EU-047-NPW
AnalysisDate	2026-02-11 10:26		
SampleType	Sample		
Bottle ID	A		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	FR100226048	4.61	0.222	0.561				
	PFPeA	2706-90-3	FR100226048	8.90	0.160	0.561				
	PFFhxA	307-24-4	FR100226048	6.67	0.187	0.561				
	PFFhPA	375-85-9	FR100226048	3.21	0.196	0.561				
	PFOA	335-67-1	FR100226048	4.41	0.128	0.561				
	PFNA	375-95-1	FR100226048	0.467	0.127	0.561			J	
	PFDA	335-76-2	FR100226048	0.125	0.160	0.561			L	
	PFUnDA	2058-94-8	FR100226048	ND	0.127	0.561			U	
	PFDODA	307-55-1	FR100226048	ND	0.228	0.561			U	
	PFTrDA	72629-94-8	FR100226048	ND	0.186	0.561			U	
	PFTeDA	376-06-7	FR100226048	ND	0.214	0.561			U	
	PFFhxDA	67905-19-5	FR100226048	ND	0.298	0.561			U	
	Sulfonates	PFBS	375-73-5	FR100226048	3.95	0.298	0.561			
		PFPeS	2706-91-4	FR100226048	0.366	0.115	0.528			J
PFFhXS		355-46-4	FR100226048	4.03	0.433	0.513				
PFFhPS		375-92-8	FR100226048	0.167	0.272	0.534			L	
PFOS		1763-23-1	FR100226048	7.48	0.296	0.519				
PFNS		68259-12-1	FR100226048	ND	0.174	0.540			U	
PFDS		335-77-3	FR100226048	ND	0.294	0.540			U	
4:2 FTS		757124-72-4	FR100226048	ND	0.0727	0.525			U	
6:2 FTS		27619-97-2	FR100226048	0.700	0.265	0.534				
8:2 FTS		39108-34-4	FR100226048	ND	0.126	0.537			U	
10:2 FTS		120226-60-0	FR100226048	ND	0.429	0.561			U	
Sulfonamidos	FBSA	30334-69-1	FR100226048	0.369	0.266	0.561			J	
	N-EtFOSA	4151-50-2	FR100226048	ND	0.347	0.561			U	
	N-EtFOSAA	2991-50-6	FR100226048	ND	0.228	0.561			U	
	N-EtFOSE	1691-99-2	FR100226048	ND	0.858	2.52			U	
	N-MeFOSA	31506-32-8	FR100226048	ND	0.231	0.561			U	
	N-MeFOSAA	2355-31-9	FR100226048	ND	0.157	0.561			U	
	N-MeFOSE	24448-09-7	FR100226048	ND	0.533	2.52			U	
	PFOSA	754-91-6	FR100226048	ND	0.0787	0.561			U	
	PFECAs	ADONA	919005-14-4	FR100226048	ND	0.152	0.531			U
EVE Acid		69087-46-3	FR100226048	0.0132	0.179	1.26			L	
HFFO-DA		13252-13-6	FR100226048	3.14	0.0594	0.561				
Hydro-EVE Acid		773804-62-9	FR100226048	0.172	0.184	0.561			L	
NFDHA		151772-58-6	FR100226048	ND	0.118	0.561			U	
PEPA		267239-61-2	FR100226048	1.39	0.105	0.561				
PFECA-G		801212-59-9	FR100226048	ND	0.0748	0.561			U	
PFMOAA		674-13-5	FR100226048	7.96	0.284	0.561				
PFMOBA		863090-89-5	FR100226048	ND	0.941	1.26			U	
PFMOPrA		377-73-1	FR100226048	ND	0.200	0.561			U	
PFO2HxA		39492-88-1	FR100226048	3.14	0.180	0.561				
PFO3OA		39492-89-2	FR100226048	ND	0.258	0.561			U	
PFO4DA		39492-90-5	FR100226048	ND	0.443	2.80			U	
PFO5DA		39492-91-6	FR100226048	ND	0.448	2.80			U	
PMPA		13140-29-9	FR100226048	5.78	0.132	0.561				
R-EVE		2416366-22-6	FR100226048	2.85	0.930	1.26				
PFESAs		11Cl-PF3OUdS	763051-92-9	FR100226048	ND	0.265	0.528			U
	9Cl-PF3ONS	756426-58-1	FR100226048	ND	0.359	0.522			U	
	Hydrolyzed PSDA	2416366-19-1	FR100226048	2.08	0.373	0.561				
	Nafion Byproduct 1 (PS Acid)	29311-67-9	FR100226048	ND	0.300	0.561			U	
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	FR100226048	0.522	0.464	0.561			J	
	NVHOS	1132933-86-8	FR100226048	7.71	0.0864	0.561				
	PFEESA	113507-82-7	FR100226048	ND	0.169	0.561			U	
R-PSDA	R-PSDA	2416366-18-0	FR100226048	4.24	2.47	2.47				
	R-PSDCA	2416366-21-5	FR100226048	ND	0.237	0.561			U	
ES	MPFBA		FR100226048				20-150%	51.2%		
	M5PFPeA		FR100226048				20-150%	76.6%		
	M3PFBS		FR100226048				20-150%	83.0%		
	M2-4:2 FTS		FR100226048				20-150%	55.9%		
	M5PFFhxA		FR100226048				20-150%	52.5%		
	M3HFPO-DA		FR100226048				20-150%	46.8%		

Enthalpy Analytical

Job No.: 0226-767-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	020626-S01		
Sampling Site			
Enthalpy ID	0226-767-001-1	Prep Batch	EU119928
Matrix	aqueous	Analyst	zoearndt
Sampling Date	2026-02-06 10:30	Instrument	Frodo
Received Date	2026-02-06	Sample Vol mL	285.4
Prep Date	2026-02-09 08:16	Extract Vol mL	0.4
AnalysisDate	2026-02-11 10:26	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	A		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M4PFHpA		FR100226048				20-150%	58.7%	
M3PFHxS		FR100226048				20-150%	63.5%	
M2-6:2 FTS		FR100226048				20-150%	69.7%	
M8PFOA		FR100226048				20-150%	57.9%	
M9PFNA		FR100226048				20-150%	53.8%	
M8PFOS		FR100226048				20-150%	52.6%	
M2-8:2 FTS		FR100226048				20-150%	47.9%	
M8FOSA-I		FR100226048				20-150%	23.6%	
M6PFDA		FR100226048				20-150%	49.3%	
d3-N-MeFOSAA		FR100226048				20-150%	43.5%	
d5-N-EtFOSAA		FR100226048				20-150%	40.0%	
M7PFUdA		FR100226048				20-150%	42.0%	
MPFDoA		FR100226048				20-150%	34.2%	
M2PFTeDA		FR100226048				20-150%	15.7%	Q
d3-N-MeFOSA		FR100226048				10-200%	2.00%	Q
d5-N-EtFOSA		FR100226048				10-200%	2.02%	Q
d7-N-MeFOSE		FR100226048				10-200%	11.1%	
d9-N-EtFOSE		FR100226048				10-200%	8.93%	Q

Enthalpy Analytical

Job No.: 0226-767-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	020626-E01	Prep Batch	EU119928
Sampling Site		Analyst	zoearndt
Enthalpy ID	0226-767-002-1	Instrument	Frodo
Matrix	aqueous	Sample Vol mL	283.82
Sampling Date	2026-02-06 10:30	Extract Vol mL	0.4
Received Date	2026-02-06	Split Factor	N/A
Prep Date	2026-02-09 08:16	Method Code	EU-047-NPW
AnalysisDate	2026-02-11 10:48		
SampleType	Sample		
Bottle ID	A		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	FR100226049	3.57	0.224	0.564				
	PFPeA	2706-90-3	FR100226049	8.04	0.161	0.564				
	PFHxA	307-24-4	FR100226049	6.36	0.188	0.564				
	PFHpA	375-85-9	FR100226049	3.14	0.197	0.564				
	PFOA	335-67-1	FR100226049	4.24	0.129	0.564				
	PFNA	375-95-1	FR100226049	0.397	0.127	0.564			J	
	PFDA	335-76-2	FR100226049	0.117	0.161	0.564			L	
	PFUnDA	2058-94-8	FR100226049	ND	0.127	0.564			U	
	PFDODA	307-55-1	FR100226049	ND	0.229	0.564			U	
	PFTrDA	72629-94-8	FR100226049	ND	0.187	0.564			U	
	PFTeDA	376-06-7	FR100226049	ND	0.215	0.564			U	
	PFHxDA	67905-19-5	FR100226049	ND	0.299	0.564			U	
	Sulfonates	PFBS	375-73-5	FR100226049	3.91	0.299	0.564			
		PFPeS	2706-91-4	FR100226049	0.521	0.116	0.531			J
PFHxS		355-46-4	FR100226049	3.68	0.435	0.516				
PFHpS		375-92-8	FR100226049	0.180	0.273	0.537			L	
PFOS		1763-23-1	FR100226049	7.11	0.298	0.522				
PFNS		68259-12-1	FR100226049	ND	0.175	0.543			U	
PFDS		335-77-3	FR100226049	ND	0.296	0.543			U	
4:2 FTS		757124-72-4	FR100226049	ND	0.0731	0.528			U	
6:2 FTS		27619-97-2	FR100226049	0.689	0.266	0.537				
8:2 FTS		39108-34-4	FR100226049	ND	0.126	0.540			U	
10:2 FTS	120226-60-0	FR100226049	ND	0.432	0.564			U		
Sulfonamidos	FBSA	30334-69-1	FR100226049	0.503	0.268	0.564			J	
	N-EtFOSA	4151-50-2	FR100226049	ND	0.349	0.564			U	
	N-EtFOSAA	2991-50-6	FR100226049	ND	0.229	0.564			U	
	N-EtFOSE	1691-99-2	FR100226049	ND	0.863	2.54			U	
	N-MeFOSA	31506-32-8	FR100226049	ND	0.233	0.564			U	
	N-MeFOSAA	2355-31-9	FR100226049	ND	0.158	0.564			U	
	N-MeFOSE	24448-09-7	FR100226049	ND	0.536	2.54			U	
	PFOSA	754-91-6	FR100226049	ND	0.0791	0.564			U	
PFECAs	ADONA	919005-14-4	FR100226049	ND	0.153	0.534			U	
	EVE Acid	69087-46-3	FR100226049	0.0237	0.180	1.27			L	
	HFPO-DA	13252-13-6	FR100226049	3.67	0.0597	0.564				
	Hydro-EVE Acid	773804-62-9	FR100226049	0.190	0.185	0.564			J	
	NFDHA	151772-58-6	FR100226049	ND	0.119	0.564			U	
	PEPA	267239-61-2	FR100226049	2.37	0.106	0.564				
	PFECA-G	801212-59-9	FR100226049	ND	0.0752	0.564			U	
	PFMOAA	674-13-5	FR100226049	5.24	0.285	0.564				
	PFMOBA	863090-89-5	FR100226049	ND	0.946	1.27			U	
	PFMOPrA	377-73-1	FR100226049	ND	0.201	0.564			U	
	PFO2HxA	39492-88-1	FR100226049	3.07	0.181	0.564				
	PFO3OA	39492-89-2	FR100226049	0.883	0.259	0.564				
	PFO4DA	39492-90-5	FR100226049	ND	0.446	2.82			U	
	PFO5DA	39492-91-6	FR100226049	ND	0.451	2.82			U	
	PMPA	13140-29-9	FR100226049	7.63	0.133	0.564				
	R-EVE	2416366-22-6	FR100226049	5.25	0.935	1.27				
PFESAs	11Cl-PF3OUds	763051-92-9	FR100226049	ND	0.266	0.531			U	
	9Cl-PF3ONS	756426-58-1	FR100226049	ND	0.361	0.525			U	
	Hydrolyzed PSDA	2416366-19-1	FR100226049	4.19	0.375	0.564				
	Nafion Byproduct 1 (PS Acid)	29311-67-9	FR100226049	ND	0.301	0.564			U	
	Nafion Byproduct 2 (Hydro-PS Acid)	749836-20-2	FR100226049	0.496	0.467	0.564			J	
	NVHOS	1132933-86-8	FR100226049	7.19	0.0869	0.564				
	PFEESA	113507-82-7	FR100226049	ND	0.169	0.564			U	
R-PSDA	2416366-18-0	FR100226049	8.03	2.48	2.48					
R-PSDCA	2416366-21-5	FR100226049	ND	0.238	0.564			U		
ES	MPFBA		FR100226049				20-150%	62.3%		
	M5PFPeA		FR100226049				20-150%	125%		
	M3PFBS		FR100226049				20-150%	152%	Q	
	M2-4:2 FTS		FR100226049				20-150%	92.0%		
	M5PFHxA		FR100226049				20-150%	69.4%		
	M3HFPO-DA		FR100226049				20-150%	63.3%		

Enthalpy Analytical


Job No.: 0226-767-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	020626-E01	Prep Batch	EU119928
Sampling Site		Analyst	zoearndt
Enthalpy ID	0226-767-002-1	Instrument	Frodo
Matrix	aqueous	Sample Vol mL	283.82
Sampling Date	2026-02-06 10:30	Extract Vol mL	0.4
Received Date	2026-02-06	Split Factor	N/A
Prep Date	2026-02-09 08:16	Method Code	EU-047-NPW
AnalysisDate	2026-02-11 10:48		
SampleType	Sample		
Bottle ID	A		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M4PFHpA		FR100226049				20-150%	77.0%	
M3PFHxS		FR100226049				20-150%	85.0%	
M2-6:2 FTS		FR100226049				20-150%	89.7%	
M8PFOA		FR100226049				20-150%	76.8%	
M9PFNA		FR100226049				20-150%	73.6%	
M8PFOS		FR100226049				20-150%	70.8%	
M2-8:2 FTS		FR100226049				20-150%	69.0%	
M8FOSA-I		FR100226049				20-150%	64.9%	
M6PFDA		FR100226049				20-150%	71.3%	
d3-N-MeFOSAA		FR100226049				20-150%	58.9%	
d5-N-EiFOSAA		FR100226049				20-150%	56.2%	
M7PFUdA		FR100226049				20-150%	65.9%	
MPFDoA		FR100226049				20-150%	57.1%	
M2PFTeDA		FR100226049				20-150%	39.5%	
d3-N-MeFOSA		FR100226049				10-200%	14.1%	
d5-N-EiFOSA		FR100226049				10-200%	14.4%	
d7-N-MeFOSE		FR100226049				10-200%	37.9%	
d9-N-EiFOSE		FR100226049				10-200%	25.5%	

QC Data



Enthalpy Analytical

Job No.: 0226-767-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	MB_119928_PFAS	Prep Batch	EU119928
Sampling Site		Analyst	zoeardt
Enthalpy ID	MB_119928_PFAS	Instrument	Frodo
Matrix	aqueous	Sample Vol mL	250
Sampling Date		Extract Vol mL	0.4
Received Date		Split Factor	N/A
Prep Date	2026-02-09 08:16	Method Code	EU-047-NPW
AnalysisDate	2026-02-11 03:58		
SampleType	Blank		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	FR100226031	ND	0.254	0.640			U	
	PFPeA	2706-90-3	FR100226031	ND	0.183	0.640			U	
	PFHxA	307-24-4	FR100226031	ND	0.214	0.640			U	
	PFHpA	375-85-9	FR100226031	ND	0.224	0.640			U	
	PFOA	335-67-1	FR100226031	ND	0.146	0.640			U	
	PFNA	375-95-1	FR100226031	ND	0.145	0.640			U	
	PFDA	335-76-2	FR100226031	ND	0.183	0.640			U	
	PFUnDA	2058-94-8	FR100226031	ND	0.145	0.640			U	
	PFDoDA	307-55-1	FR100226031	ND	0.260	0.640			U	
	PFTrDA	72629-94-8	FR100226031	ND	0.212	0.640			U	
	PFTeDA	376-06-7	FR100226031	ND	0.244	0.640			U	
	PFHxDA	67905-19-5	FR100226031	ND	0.340	0.640			U	
	Sulfonates	PFBS	375-73-5	FR100226031	ND	0.340	0.640			U
		PFPeS	2706-91-4	FR100226031	ND	0.131	0.603			U
		PFHxS	355-46-4	FR100226031	ND	0.494	0.586			U
PFHpS		375-92-8	FR100226031	ND	0.310	0.610			U	
PFOS		1763-23-1	FR100226031	ND	0.338	0.593			U	
PFNS		68259-12-1	FR100226031	ND	0.199	0.616			U	
PFDS		335-77-3	FR100226031	ND	0.336	0.616			U	
4:2 FTS		757124-72-4	FR100226031	ND	0.0830	0.600			U	
6:2 FTS		27619-97-2	FR100226031	ND	0.302	0.610			U	
8:2 FTS		39108-34-4	FR100226031	ND	0.143	0.613			U	
10:2 FTS	120226-60-0	FR100226031	ND	0.490	0.640			U		
Sulfonamidos	FBSA	30334-69-1	FR100226031	ND	0.304	0.640			U	
	N-EtFOSA	4151-50-2	FR100226031	ND	0.396	0.640			U	
	N-EtFOSAA	2991-50-6	FR100226031	ND	0.260	0.640			U	
	N-EtFOSE	1691-99-2	FR100226031	ND	0.980	2.88			U	
	N-MeFOSA	31506-32-8	FR100226031	ND	0.264	0.640			U	
	N-MeFOSAA	2355-31-9	FR100226031	ND	0.180	0.640			U	
	N-MeFOSE	24448-09-7	FR100226031	ND	0.608	2.88			U	
	PFOSA	754-91-6	FR100226031	ND	0.0898	0.640			U	
PFECAs	ADONA	919005-14-4	FR100226031	ND	0.173	0.606			U	
	EVE Acid	69087-46-3	FR100226031	0.00481	0.204	1.44			L	
	HFFO-DA	13252-13-6	FR100226031	ND	0.0678	0.640			U	
	Hydro-EVE Acid	773804-62-9	FR100226031	ND	0.210	0.640			U	
	NFDHA	151772-58-6	FR100226031	ND	0.135	0.640			U	
	PEPA	267239-61-2	FR100226031	ND	0.120	0.640			U	
	PFECA-G	801212-59-9	FR100226031	0.0248	0.0854	0.640			L	
	PFMOAA	674-13-5	FR100226031	ND	0.324	0.640			U	
	PFMOBA	863090-89-5	FR100226031	ND	1.07	1.44			U	
	PFMOPrA	377-73-1	FR100226031	ND	0.228	0.640			U	
	PFO2HxA	39492-88-1	FR100226031	ND	0.206	0.640			U	
	PFO3OA	39492-89-2	FR100226031	ND	0.294	0.640			U	
	PFO4DA	39492-90-5	FR100226031	ND	0.506	3.20			U	
	PFO5DA	39492-91-6	FR100226031	ND	0.512	3.20			U	
	PMPA	13140-29-9	FR100226031	ND	0.151	0.640			U	
	R-EVE	2416366-22-6	FR100226031	ND	1.06	1.44			U	
	PFESAs	11Cl-PF3OUdS	763051-92-9	FR100226031	ND	0.302	0.603			U
		9Cl-PF3ONS	756426-58-1	FR100226031	ND	0.410	0.596			U
Hydrolyzed PSDA		2416366-19-1	FR100226031	ND	0.426	0.640			U	
Nafion Byproduct 1 (PS Acid)		29311-67-9	FR100226031	ND	0.342	0.640			U	
Nafion Byproduct 2 (Hydro-PS Acid)		749836-20-2	FR100226031	ND	0.530	0.640			U	
NVHOS		1132933-86-8	FR100226031	ND	0.0986	0.640			U	
PFEESA		113507-82-7	FR100226031	ND	0.192	0.640			U	
R-PSDA		2416366-18-0	FR100226031	ND	2.82	2.82			U	
R-PSDCA	2416366-21-5	FR100226031	ND	0.270	0.640			U		
ES	MFPFBA		FR100226031				20-150%	80.3%		
	M5PFPeA		FR100226031				20-150%	71.0%		
	M3PFBS		FR100226031				20-150%	84.6%		
	M2-4:2 FTS		FR100226031				20-150%	86.2%		
	M5PFPeA		FR100226031				20-150%	71.6%		
	M3HFPO-DA		FR100226031				20-150%	58.4%		

Enthalpy Analytical

Job No.: 0226-767-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	MB_119928_PFAS	Prep Batch	EU119928
Sampling Site		Analyst	zoeardt
Enthalpy ID	MB_119928_PFAS	Instrument	Frodo
Matrix	aqueous	Sample Vol mL	250
Sampling Date		Extract Vol mL	0.4
Received Date		Split Factor	N/A
Prep Date	2026-02-09 08:16	Method Code	EU-047-NPW
AnalysisDate	2026-02-11 03:58		
SampleType	Blank		

Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
M4PFHpA		FR100226031				20-150%	72.3%	
M3PFHxS		FR100226031				20-150%	77.5%	
M2-6:2 FTS		FR100226031				20-150%	79.5%	
M8PFOA		FR100226031				20-150%	68.0%	
M9PFNA		FR100226031				20-150%	57.1%	
M8PFOS		FR100226031				20-150%	57.0%	
M2-8:2 FTS		FR100226031				20-150%	50.3%	
M8FOSA-I		FR100226031				20-150%	53.5%	
M6PFDA		FR100226031				20-150%	53.1%	
d3-N-MeFOSAA		FR100226031				20-150%	43.4%	
d5-N-EiFOSAA		FR100226031				20-150%	40.1%	
M7PFUdA		FR100226031				20-150%	47.4%	
MPFDoA		FR100226031				20-150%	40.2%	
M2PFTeDA		FR100226031				20-150%	30.7%	
d3-N-MeFOSA		FR100226031				10-200%	23.6%	
d5-N-EiFOSA		FR100226031				10-200%	26.5%	
d7-N-MeFOSE		FR100226031				10-200%	38.9%	
d9-N-EiFOSE		FR100226031				10-200%	29.4%	

Enthalpy Analytical

Job No.: 0226-767-1 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Enthalpy ID	OPR_119928_PFAS	Prep Batch	EU119928	Sample Vol (mL)	250
Sample Name	OPR_119928_PFAS	Prep Date	2026-02-09 08:16	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	2026-02-11 04:21	Split Factor	N/A
Sampling Date		Analyst	zoeamdt	Method Code	EU-047-NPW
Received Date		Instrument	Frodo	Sample Type	Control

	Compound	CAS	InjFileName	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags	
Acids	PFBA	375-22-4	FR100226032	17.8	0.254	0.640	47.9-144%	89.0%		
	PFPeA	2706-90-3	FR100226032	18.2	0.183	0.640	41.7-159%	90.8%		
	PFHxA	307-24-4	FR100226032	17.9	0.214	0.640	43.2-154%	89.6%		
	PFHpA	375-85-9	FR100226032	18.0	0.224	0.640	42.1-155%	90.1%		
	PFOA	335-67-1	FR100226032	18.3	0.146	0.640	51.1-148%	91.7%		
	PFNA	375-95-1	FR100226032	17.8	0.145	0.640	51.6-153%	88.9%		
	PFDA	335-76-2	FR100226032	18.1	0.183	0.640	44.5-156%	90.3%		
	PFUnDA	2058-94-8	FR100226032	18.7	0.145	0.640	40.3-156%	93.3%		
	PFDoDA	307-55-1	FR100226032	18.6	0.260	0.640	40.4-158%	92.8%		
	PFTriDA	72629-94-8	FR100226032	21.4	0.212	0.640	42.2-201%	107%		
	PFTeDA	376-06-7	FR100226032	18.0	0.244	0.640	43-162%	89.8%		
	Sulfonates	PFBS	375-73-5	FR100226032	15.6	0.340	0.640	42.7-155%	87.9%	
		PFPeS	2706-91-4	FR100226032	15.8	0.131	0.603	40.3-152%	84.2%	
		PFHxS	355-46-4	FR100226032	16.3	0.494	0.586	45-148%	89.4%	
PFHpS		375-92-8	FR100226032	19.0	0.310	0.610	39.8-166%	99.7%		
PFOS		1763-23-1	FR100226032	15.9	0.338	0.593	59.2-132%	85.8%		
PFNS		68259-12-1	FR100226032	15.6	0.199	0.616	38.1-153%	81.1%		
PFDS		335-77-3	FR100226032	14.7	0.336	0.616	28.6-148%	76.1%		
4:2 FTS		757124-72-4	FR100226032	16.9	0.0830	0.600	41.5-157%	90.4%		
6:2 FTS		27619-97-2	FR100226032	18.3	0.302	0.610	44.5-160%	96.3%		
8:2 FTS		39108-34-4	FR100226032	18.4	0.143	0.613	39.4-166%	95.7%		
Sulfonamidos	N-EiFOSA	4151-50-2	FR100226032	20.2	0.396	0.640	26.7-172%	101%		
	N-EiFOSAA	2991-50-6	FR100226032	19.0	0.260	0.640	42.8-156%	95.2%		
	N-EiFOSE	1691-99-2	FR100226032	86.0	0.980	2.88	38.9-161%	95.5%		
	N-MeFOSA	31506-32-8	FR100226032	21.3	0.264	0.640	26.4-183%	107%		
	N-MeFOSAA	2355-31-9	FR100226032	18.7	0.180	0.640	42-155%	93.4%		
	N-MeFOSE	24448-09-7	FR100226032	78.9	0.608	2.88	37.6-155%	87.7%		
	PFOSA	754-91-6	FR100226032	18.6	0.0898	0.640	39.1-158%	93.1%		
PFECAs	ADONA	919005-14-4	FR100226032	17.9	0.173	0.606	32.2-151%	89.5%		
	HFPO-DA	13252-13-6	FR100226032	18.2	0.0678	0.640	61.8-131%	91.0%		
PFESAs	11Cl-PF3OUdS	763051-92-9	FR100226032	13.5	0.302	0.603	21.8-141%	67.6%		
	9Cl-PF3ONS	756426-58-1	FR100226032	17.6	0.410	0.596	37.6-146%	88.1%		
ES	MPFBA		FR100226032				20-150%	83.7%		
	M5PFPeA		FR100226032				20-150%	75.3%		
	M3PFBS		FR100226032				20-150%	86.1%		
	M2-4:2 FTS		FR100226032				20-150%	85.4%		
	M5PFHxA		FR100226032				20-150%	75.6%		
	M3HFPO-DA		FR100226032				20-150%	63.7%		
	M4PFHpA		FR100226032				20-150%	77.3%		
	M3PFHxS		FR100226032				20-150%	80.4%		
	M2-6:2 FTS		FR100226032				20-150%	85.3%		
	M8PFOA		FR100226032				20-150%	71.7%		
	M9PFNA		FR100226032				20-150%	64.1%		
	M8PFOS		FR100226032				20-150%	62.7%		
	M2-8:2 FTS		FR100226032				20-150%	58.2%		
	M8FOSA-I		FR100226032				20-150%	54.8%		
	M6PFDA		FR100226032				20-150%	58.4%		
	d3-N-MeFOSAA		FR100226032				20-150%	48.6%		
	d5-N-EiFOSAA		FR100226032				20-150%	45.7%		
	M7PFUdA		FR100226032				20-150%	53.7%		
	MPFDoA		FR100226032				20-150%	45.3%		
	M2PFTeDA		FR100226032				20-150%	35.6%		
	d3-N-MeFOSA		FR100226032				10-200%	26.3%		
d5-N-EiFOSA		FR100226032				10-200%	26.3%			
d7-N-MeFOSE		FR100226032				10-200%	43.0%			
d9-N-EiFOSE		FR100226032				10-200%	33.9%			

Narrative Summary

Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0226-767-2
Client ID.	NORTHWEST WATER PLANT

1. Custody

Isabelle Martin received the samples at 1.6 °C after being relinquished by Brunswick County Public Utilities - NC.

The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC.

Table 1 - Sample Inventory

EU Lab Sample ID	Client Sample ID	Matrix	Received
0226-767-001-2	020626-S01	aqueous	2026-02-06
0226-767-002-2	020626-E01	aqueous	2026-02-06

2. Methods and Analytes

A list of analytes of interest and corresponding methods of analysis is shown in Table 3. Abbreviations are defined in the listed Appendices.

Table 3 - Methods and Analytes

EU Method	Analytes	Cleanup Method
EU047	PFPrA	n/a

3. Analysis

The samples were analyzed using LC/MS/MS instrument Starscream.

Polar compound PFPrA in the samples, including the method blank (MB) and Ongoing Precision Recovery (OPR) samples, was analyzed by direct inject calibration.

4. Calibration

In the initial calibration, the reported analytes exhibited R^2 of ≥ 0.99 . The reported analytes in the calibration standards, Initial Calibration Verification (ICV) and continuing calibration (concal) met the accuracy criterion for native analytes.

5. QC Notes

The QC sample analyses passed all method criteria.

PFAS by Isotope Dilution (non-potable water) samples were extracted within 28 days, and extracts analyzed within 28 days.

Enthalpy Analytical Narrative Summary

Company	Brunswick County Public Utilities - NC
Job No.	0226-767-2
Client ID.	NORTHWEST WATER PLANT

6. Reporting Notes

The results presented in this report are representative of the samples as provided to the laboratory.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, LLC in Wilmington NC is accredited by the Louisiana Department of Environmental Quality to the 2016 TNI Standard under certificate number 05075.

Results

Enthalpy Analytical

Job No.: 0226-767-2 PFAS by Isotope Dilution (non-potable water)
Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Summary

	Compound	CAS	020626-S01 ng/L	020626-E01 ng/L
Acids	PFPrA	422-64-0	ND U	ND U

Enthalpy Analytical

Job No.: 0226-767-2 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	020626-S01		
Sampling Site			
Enthalpy ID	0226-767-001-2	Prep Batch	EU119948
Matrix	aqueous	Analyst	zoeardt
Sampling Date	2026-02-06 10:30	Instrument	Starscream
Received Date	2026-02-06	Sample Vol mL	0.1
Prep Date	2026-02-10 12:05	Extract Vol mL	0.2
AnalysisDate	2026-02-10 17:03	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	B		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	ST100226-02101703	ND	700	1530			U
ES	13C3-PFPrA		ST100226-02101703				20-150%	75.2%	

Enthalpy Analytical

Job No.: 0226-767-2 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	020626-E01		
Sampling Site			
Enthalpy ID	0226-767-002-2	Prep Batch	EU119948
Matrix	aqueous	Analyst	zoeardt
Sampling Date	2026-02-06 10:30	Instrument	Starscream
Received Date	2026-02-06	Sample Vol mL	0.1
Prep Date	2026-02-10 12:05	Extract Vol mL	0.2
AnalysisDate	2026-02-10 17:15	Split Factor	N/A
SampleType	Sample	Method Code	EU-047-NPW
Bottle ID	B		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	ST100226-02101715	ND	700	1530			U
ES	13C3-PFPrA		ST100226-02101715				20-150%	87.2%	

QC Data

Enthalpy Analytical

Job No.: 0226-767-2 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	MB_119948_PFAS	Prep Batch	EU119948
Sampling Site		Analyst	zoeardt
Enthalpy ID	MB_119948_PFAS	Instrument	Starscream
Matrix	aqueous	Sample Vol mL	0.1
Sampling Date		Extract Vol mL	0.2
Received Date		Split Factor	N/A
Prep Date	2026-02-10 12:05	Method Code	EU-047-NPW
AnalysisDate	2026-02-10 16:17		
SampleType	Blank		
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	ST100226-02101617	ND	700	1530			U
ES	13C3-PFPrA		ST100226-02101617				20-150%	84.3%	

Enthalpy Analytical

Job No.: 0226-767-2 PFAS by Isotope Dilution (non-potable water)
 Brunswick County Public Utilities - NC NORTHWEST WATER PLANT

Details

Sample Name	OPR_119948_PFAS		
Sampling Site			
Enthalpy ID	OPR_119948_PFAS	Prep Batch	EU119948
Matrix	aqueous	Analyst	zoeardt
Sampling Date		Instrument	Starscream
Received Date		Sample Vol mL	0.1
Prep Date	2026-02-10 12:05	Extract Vol mL	0.2
AnalysisDate	2026-02-10 16:28	Split Factor	N/A
SampleType	Control	Method Code	EU-047-NPW
Bottle ID	-		

	Compound	CAS	Injection File Name	Sample Concentration ng/L	LOD ng/L	LOQ ng/L	Recovery Limits	Recovery	Flags
Acids	PFPrA	422-64-0	ST100226-02101628	21000	700	1530	40-150%	105%	
ES	13C3-PFPrA		ST100226-02101628				20-150%	93.8%	

Sample Custody

COOLER RECEIPT LOG

JOB ID: <input type="text"/>	Date / Time: <u>2/6/26 14:30</u>	Initials: <u>IWM</u>
OR		
Client: <u>Brunswick Co. Utilities</u>		

Temp °C: <u>1.6</u>	Thermometer ID: <u>T10</u>	Cooler <u>1</u> of <u>1</u> Bottle Order #: <u>1821</u>												
Received via	<div style="border: 1px solid black; padding: 5px; margin-bottom: 5px;"> <i>Check one</i> On ice: <input checked="" type="checkbox"/> Melted ice: <input type="checkbox"/> Ambient: <input type="checkbox"/> </div> <div style="border: 1px solid black; padding: 5px; margin-bottom: 5px;"> <i>Check one</i> in a Box: <input type="checkbox"/> in a Cooler: <input checked="" type="checkbox"/> Cooler in Box: <input type="checkbox"/> </div>	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th style="text-align: center;">Yes</th> <th style="text-align: center;">No</th> </tr> </thead> <tbody> <tr> <td>Cooler seals:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input checked="" type="checkbox"/></td> </tr> <tr> <td>Sample seals:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input checked="" type="checkbox"/></td> </tr> <tr> <td>Good condition:</td> <td style="text-align: center;"><input checked="" type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> </tbody> </table>		Yes	No	Cooler seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Sample seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Good condition:	<input checked="" type="checkbox"/>	<input type="checkbox"/>
	Yes	No												
Cooler seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>												
Sample seals:	<input type="checkbox"/>	<input checked="" type="checkbox"/>												
Good condition:	<input checked="" type="checkbox"/>	<input type="checkbox"/>												
FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input checked="" type="checkbox"/> Other <input type="checkbox"/>	Comment: <input style="width: 100%;" type="text"/>													

Temp °C: <input type="text"/>	Thermometer ID: <input type="text"/>	Cooler <input type="text"/> of <input type="text"/> Bottle Order #: <input type="text"/>												
Received via	<div style="border: 1px solid black; padding: 5px; margin-bottom: 5px;"> <i>Check one</i> On ice: <input type="checkbox"/> Melted ice: <input type="checkbox"/> Ambient: <input type="checkbox"/> </div> <div style="border: 1px solid black; padding: 5px; margin-bottom: 5px;"> <i>Check one</i> in a Box: <input type="checkbox"/> in a Cooler: <input type="checkbox"/> Cooler in Box: <input type="checkbox"/> </div>	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th style="text-align: center;">Yes</th> <th style="text-align: center;">No</th> </tr> </thead> <tbody> <tr> <td>Cooler seals:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td>Sample seals:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td>Good condition:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> </tbody> </table>		Yes	No	Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>	Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>	Good condition:	<input type="checkbox"/>	<input type="checkbox"/>
	Yes	No												
Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>												
Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>												
Good condition:	<input type="checkbox"/>	<input type="checkbox"/>												
FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input type="checkbox"/> Other <input type="checkbox"/>	Comment: <input style="width: 100%;" type="text"/>	<u>2/6/26 IWM</u>												

Temp °C: <input type="text"/>	Thermometer ID: <input type="text"/>	Cooler <input type="text"/> of <input type="text"/> Bottle Order #: <input type="text"/>												
Received via	<div style="border: 1px solid black; padding: 5px; margin-bottom: 5px;"> <i>Check one</i> On ice: <input type="checkbox"/> Melted ice: <input type="checkbox"/> Ambient: <input type="checkbox"/> </div> <div style="border: 1px solid black; padding: 5px; margin-bottom: 5px;"> <i>Check one</i> in a Box: <input type="checkbox"/> in a Cooler: <input type="checkbox"/> Cooler in Box: <input type="checkbox"/> </div>	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th style="text-align: center;">Yes</th> <th style="text-align: center;">No</th> </tr> </thead> <tbody> <tr> <td>Cooler seals:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td>Sample seals:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td>Good condition:</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> </tbody> </table>		Yes	No	Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>	Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>	Good condition:	<input type="checkbox"/>	<input type="checkbox"/>
	Yes	No												
Cooler seals:	<input type="checkbox"/>	<input type="checkbox"/>												
Sample seals:	<input type="checkbox"/>	<input type="checkbox"/>												
Good condition:	<input type="checkbox"/>	<input type="checkbox"/>												
FedEx <input type="checkbox"/> UPS <input type="checkbox"/> DHL <input type="checkbox"/> USPS <input type="checkbox"/> Courier <input type="checkbox"/> Other <input type="checkbox"/>	Comment: <input style="width: 100%;" type="text"/>													