

County of Brunswick

3954 Clearwell Dr NE
Leland, NC 28451

Northwest Water Plant

Leland, NC
Samples Received: 09/10/20

Analytical Report 0920-719

Isotope Dilution Method PFAS



Enthalpy Analytical, LLC – Ultratrace

Lindsay Boone
O: (910) 212-5855 / F: 910-212-5866
lboone@enthalpy.com / www.enthalpy.com
2714 Exchange Drive, Wilmington, NC 28405

I certify that to the best of my knowledge all analytical data presented in this report:

- Have been checked for completeness
- Are accurate, error-free, and legible
- Have been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s)

This analytical report was prepared in Portable Document Format (.PDF) and contains _____ pages.

Report Issued Date: _____



Summary of Results

Enthalpy Analytical

Job No.: 0920-719 - PFAS by Isotope Dilution (non-potable water)

County of Brunswick Site: NorthWest Water Plant, Leland, NC

Summary

	Compound	CAS	091020-SO1 ng/L	091020-EO1 ng/L
Acids	PFBA	375-22-4	ND U	ND U
	PFPeA	2706-90-3	ND U	ND U
	PFHxA	307-24-4	3.99	3.88
	PFHpA	375-85-9	2.12	2.07
	PFOA	335-67-1	3.69	3.64
	PFNA	375-95-1	0.743	0.657
	PFDA	335-76-2	0.277	0.227 J
	PFUnDA	2058-94-8	0.0993 J	0.0712 J
	PFDoDA	307-55-1	ND U	ND U
	PFTTrDA	72629-94-8	ND U	ND U
	PFTeDA	376-06-7	ND U	ND U
Sulfonates	PFBS	375-73-5	ND U	ND U
	PFPeS	2706-91-4	0.528	0.449
	PFHxS	355-46-4	2.89	2.71
	PFHpS	375-92-8	0.238 J	ND U
	PFOS	1763-23-1	6.47	6.01
	PFNS	68259-12-1	ND U	ND U
	PFDS	335-77-3	ND U	ND U
	4:2 FTS	757124-72-4	ND U	ND U
	6:2 FTS	27619-97-2	0.105 J	0.148 J
	8:2 FTS	39108-34-4	ND U	ND U
other	PFOSA	754-91-6	ND U	ND U
	N-MeFOSAA	2355-31-9	ND U	ND U
	N-EtFOSAA	2991-50-6	ND U	ND U
	HFPO-DA	13252-13-6	8.19	6.53
	PFMOAA	674-13-5	57.3	42.4
	PFMOPrA	377-73-1	ND U	ND U
	PFO2HxA	39492-88-1	2.32	2.61
	PFO3OA	39492-89-2	ND U	3.69
	PFO4DA	39492-90-5	ND U	0.863 L
	Nafion Byproduct 1	29311-67-9	ND U	ND U
	ADONA	919005-14-4	ND U	ND U
	9Cl-PF3ONS	756426-58-1	ND U	ND U
	11Cl-PF3OUdS	763051-92-9	ND U	ND U
	10:2 FTS	120226-60-0	ND U	ND U
	FBSA	30334-69-1	0.337	0.311
	N-EtFOSA	4151-50-2	ND U	ND U
	N-EtFOSE	1691-99-2	ND U	ND U
	N-MeFOSA	31506-32-8	ND U	ND U
	N-MeFOSE	24448-09-7	ND U	ND U
	Nafion Byproduct 2	749836-20-2	0.574 L	0.618 L
	NFDHA	151772-58-6	ND U	ND U
	PEPA		ND U	ND U
	PFECA-G	801212-59-9	ND U	ND U
	PFEESA	113507-82-7	ND U	ND U
	PFHxDA	67905-19-5	ND U	ND U
	PFMOBA	863090-89-5	ND U	ND U
	PFO5DA	39492-91-6	ND U	ND U
	PMPA	13140-29-9	ND U	ND U

Detailed Results

Enthalpy Analytical

Job No.: 0920-719 - PFAS by Isotope Dilution (non-potable water)
County of Brunswick Site: NorthWest Water Plant, Leland, NC

Enthalpy ID	0920-719-001-1	Prep Batch	EU11147	Sample Vol (mL)	294.76
Sample Name	091020-SO1	Prep Date	2020-09-11	Extract Vol (mL)	0.4
Matrix	Aqueous	Analysis Date	2020-09-11	Dilution Factor	1
Sampling Date	20200910 00:00				

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery	Flags
Acids	PFBA	375-22-4	ND	ND	ND	0.134	0.250		U
	PFPeA	2706-90-3	ND	ND	ND	0.0761	0.250		U
	PFHxA	307-24-4	2942.83	3.99	3.99	0.134	0.250		
	PFHpA	375-85-9	1558.55	2.12	2.12	0.0589	0.250		
	PFOA	335-67-1	2716.34	3.69	3.69	0.0674	0.250		
	PFNA	375-95-1	547.69	0.743	0.743	0.0432	0.250		
	PFDA	335-76-2	204.16	0.277	0.277	0.106	0.250		
	PFUnDA	2058-94-8	73.17	0.0993	0.0993	0.0408	0.250		J
	PFDoDA	307-55-1	ND	ND	ND	0.0403	0.250		U
	PFTriDA	72629-94-8	ND	ND	ND	0.0632	0.250		U
Sulfonates	PFTeDA	376-06-7	ND	ND	ND	0.0704	0.250		U
	PFBS	375-73-5	ND	ND	ND	0.0704	0.250		U
	PFPeS	2706-91-4	389.09	0.528	0.528	0.0840	0.250		
	PFHxS	355-46-4	2126.41	2.89	2.89	0.0701	0.250		
	PFHpS	375-92-8	175.12	0.238	0.238	0.0661	0.250		J
	PFOS	1763-23-1	4768.45	6.47	6.47	0.0399	0.250		
	PFNS	68259-12-1	ND	ND	ND	0.0555	0.250		U
	PFDS	335-77-3	ND	ND	ND	0.114	0.250		U
	4:2 FTS	757124-72-4	ND	ND	ND	0.0548	0.250		U
	6:2 FTS	27619-97-2	77.67	0.105	0.105	0.0613	0.250		J
other	8:2 FTS	39108-34-4	ND	ND	ND	0.0483	0.250		U
	PFOSA	754-91-6	ND	ND	ND	0.310	0.311		U
	N-MeFOSAA	2355-31-9	ND	ND	ND	0.0461	0.250		U
	N-EiFOSAA	2991-50-6	ND	ND	ND	0.0552	0.250		U
	HFPO-DA	13252-13-6	6032.60	8.19	8.19	0.0807	0.250		
	PFMOAA	674-13-5	42196.33	57.3	57.3	1.26	1.26		
	PFMOPrA	377-73-1	ND	ND	ND	0.204	0.250		U
	PFO2HxA	39492-88-1	1709.83	2.32	2.32	1.26	1.26		
	PFO3OA	39492-89-2	ND	ND	ND	1.26	1.26		U
	PFO4DA	39492-90-5	ND	ND	ND	1.26	1.26		U
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.204	0.250		U
	ADONA	919005-14-4	ND	ND	ND	0.102	0.250		U
	9Cl-PF3ONS	756426-58-1	ND	ND	ND	0.102	0.250		U
	11Cl-PF3OUdS	763051-92-9	ND	ND	ND	0.102	0.250		U
	10:2 FTS	120226-60-0	ND	ND	ND	0.204	0.250		U
	FBSA	30334-69-1	248.25	0.337	0.337	0.204	0.250		
	N-EiFOSA	4151-50-2	ND	ND	ND	0.204	0.250		U
	N-EiFOSE	1691-99-2	ND	ND	ND	6.11	6.11		U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.204	0.250		U
	N-MeFOSE	24448-09-7	ND	ND	ND	6.11	6.11		U
	Nafion Byproduct 2	749836-20-2	422.64	0.574	0.574	1.26	1.26		L
	NFDHA	151772-58-6	ND	ND	ND	0.204	0.250		U
	PEPA		ND	ND	ND	1.26	1.26		U
	PFECA-G	801212-59-9	ND	ND	ND	1.26	1.26		U
	PFEEA	113507-82-7	ND	ND	ND	0.204	0.250		U
	PFHxDA	67905-19-5	ND	ND	ND	1.26	1.26		U
	PFMOBA	863090-89-5	ND	ND	ND	1.26	1.26		U
	PFO5DA	39492-91-6	ND	ND	ND	2.55	2.55		U
	PMPA	13140-29-9	ND	ND	ND	1.26	1.26		U
ES	MPFBA		5143.37	6.98				9.4%	Q
	M5PFPeA		14168.00	19.2				25.6%	
	M3PFBS		31084.46	42.2				54.7%	
	M2-4:2 FTS		63292.63	85.9				143.2%	
	M5PFHxA		3890.87	5.28				45.4%	
	M3HFPO-DA		3140.43	4.26				38.4%	
	M4PFHpA		4420.62	6.00				53.5%	
	M3PFHxS		4330.99	5.88				53.8%	
	M2-6:2 FTS		41713.72	56.6				98.1%	
	M8PFOA		4236.36	5.75				48.2%	
	M9PFNA		3661.62	4.97				39.8%	
	M8PFOS		3714.63	5.04				48.3%	
	M2-8:2 FTS		3107.91	4.22				38.3%	
	M8FOSA-I		2024.04	2.75				26.5%	
	M6PFDA		3972.21	5.39				36.5%	
	d3-N-MeFOSAA		2541.27	3.45				34.8%	
	d5-N-EiFOSAA		2793.23	3.79				38.1%	
	M7PFUDa		3009.13	4.08				26.4%	
	MPFDoA		2207.81	3.00				22.3%	
	M2PFTeDA		783.84	1.06				9.2%	Q

Enthalpy Analytical

Job No.: 0920-719 - PFAS by Isotope Dilution (non-potable water)

County of Brunswick Site: NorthWest Water Plant, Leland, NC

Enthalpy ID	0920-719-002-1	Prep Batch	EU11147	Sample Vol (mL)	284.93
Sample Name	091020-EO1	Prep Date	2020-09-11	Extract Vol (mL)	0.4
Matrix	Aqueous	Analysis Date	2020-09-11	Dilution Factor	1
Sampling Date	20200910 00:00				

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery	Flags
Acids	PFBA	375-22-4	ND	ND	ND	0.138	0.258		U
	PFPeA	2706-90-3	ND	ND	ND	0.0788	0.258		U
	PFHxA	307-24-4	2763.53	3.88	3.88	0.138	0.258		
	PFHpA	375-85-9	1473.42	2.07	2.07	0.0610	0.258		
	PFOA	335-67-1	2595.57	3.64	3.64	0.0697	0.258		
	PFNA	375-95-1	467.86	0.657	0.657	0.0447	0.258		
	PFDA	335-76-2	162.02	0.227	0.227	0.110	0.258		J
	PFUnDA	2058-94-8	50.75	0.0712	0.0712	0.0422	0.258		J
	PFDoDA	307-55-1	ND	ND	ND	0.0417	0.258		U
	PFTriDA	72629-94-8	ND	ND	ND	0.0654	0.258		U
	PFTeDA	376-06-7	ND	ND	ND	0.0728	0.258		U
	PFBS	375-73-5	ND	ND	ND	0.0728	0.258		U
Sulfonates	PFPeS	2706-91-4	319.88	0.449	0.449	0.0869	0.258		
	PFHxS	355-46-4	1927.35	2.71	2.71	0.0726	0.258		
	PFHpS	375-92-8	ND	ND	ND	0.0684	0.258		U
	PFOS	1763-23-1	4280.58	6.01	6.01	0.0413	0.258		
	PFNS	68259-12-1	ND	ND	ND	0.0574	0.258		U
	PFDS	335-77-3	ND	ND	ND	0.118	0.258		U
	4:2 FTS	757124-72-4	ND	ND	ND	0.0567	0.258		U
	6:2 FTS	27619-97-2	105.17	0.148	0.148	0.0634	0.258		J
	8:2 FTS	39108-34-4	ND	ND	ND	0.0499	0.258		U
	PFOSA	754-91-6	ND	ND	ND	0.320	0.321		U
other	N-MeFOSAA	2355-31-9	ND	ND	ND	0.0477	0.258		U
	N-EiFOSAA	2991-50-6	ND	ND	ND	0.0571	0.258		U
	HFPO-DA	13252-13-6	4650.76	6.53	6.53	0.0835	0.258		
	PFMOAA	674-13-5	30220.27	42.4	42.4	1.30	1.30		
	PFMOPrA	377-73-1	ND	ND	ND	0.211	0.258		U
	PFO2HxA	39492-88-1	1864.81	2.61	2.61	1.30	1.30		
	PFO3OA	39492-89-2	2627.47	3.69	3.69	1.30	1.30		
	PFO4DA	39492-90-5	614.72	0.863	0.863	1.30	1.30		L
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.211	0.258		U
	ADONA	919005-14-4	ND	ND	ND	0.105	0.258		U
	9Cl-PF3ONS	756426-58-1	ND	ND	ND	0.105	0.258		U
	11Cl-PF3OUdS	763051-92-9	ND	ND	ND	0.105	0.258		U
	10:2 FTS	120226-60-0	ND	ND	ND	0.211	0.258		U
	FBSA	30334-69-1	221.27	0.311	0.311	0.211	0.258		
	N-EiFOSA	4151-50-2	ND	ND	ND	0.211	0.258		U
	N-EiFOSE	1691-99-2	ND	ND	ND	6.32	6.32		U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.211	0.258		U
	N-MeFOSE	24448-09-7	ND	ND	ND	6.32	6.32		U
	Nafion Byproduct 2	749836-20-2	440.41	0.618	0.618	1.30	1.30		L
	NFDHA	151772-58-6	ND	ND	ND	0.211	0.258		U
	PEPA		ND	ND	ND	1.30	1.30		U
	PFECA-G	801212-59-9	ND	ND	ND	1.30	1.30		U
	PFEEA	113507-82-7	ND	ND	ND	0.211	0.258		U
	PFHxDA	67905-19-5	ND	ND	ND	1.30	1.30		U
	PFMOBA	863090-89-5	ND	ND	ND	1.30	1.30		U
	PFO5DA	39492-91-6	ND	ND	ND	2.64	2.64		U
	PMPA	13140-29-9	ND	ND	ND	1.30	1.30		U
ES	MPFBA		4897.86	6.88				14.3%	Q
	M5PFPeA		11694.82	16.4				33.6%	
	M3PFBS		28175.83	39.6				78.8%	
	M2-4:2 FTS		59438.29	83.4				213.5%	Q
	M5PFHxA		4199.12	5.89				64.5%	
	M3HFPO-DA		4419.15	6.20				71.1%	
	M4PFHpA		4682.39	6.57				74.6%	
	M3PFHxS		4542.99	6.38				82.3%	
	M2-6:2 FTS		42164.59	59.2				157.4%	Q
	M8PFOA		4628.56	6.50				69.3%	
	M9PFNA		4891.01	6.87				70.0%	
	M8PFOS		4328.68	6.08				82.0%	
	M2-8:2 FTS		4085.22	5.74				73.3%	
	M8FOSA-I		3264.46	4.58				62.2%	
	M6PFDA		4267.89	5.99				60.7%	
	d3-N-MeFOSAA		3163.70	4.44				63.1%	
	d5-N-EiFOSAA		2948.99	4.14				58.6%	
	M7PFUDa		2870.13	4.03				38.9%	
	MPFDoA		2717.53	3.82				42.6%	
	M2PFTeDA		1274.45	1.79				20.2%	

QC Data

Enthalpy Analytical

Job No.: 0920-719 - PFAS by Isotope Dilution (non-potable water)

County of Brunswick Site: NorthWest Water Plant, Leland, NC

Enthalpy ID	MB-11147-PFAS	Prep Batch	EU11147	Sample Vol (mL)	250
Sample Name	MB-11147-PFAS	Prep Date	2020-09-11	Extract Vol (mL)	0.4
Matrix	Aqueous	Analysis Date	2020-09-11	Dilution Factor	1
Sampling Date					

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery	Flags
Acids	PFBA	375-22-4	ND	ND	ND	0.157	0.294		U
	PFPeA	2706-90-3	ND	ND	ND	0.0898	0.294		U
	PFHxA	307-24-4	ND	ND	ND	0.158	0.294		U
	PFHpA	375-85-9	ND	ND	ND	0.0695	0.294		U
	PFOA	335-67-1	ND	ND	ND	0.0795	0.294		U
	PFNA	375-95-1	9.53	0.0152	0.0152	0.0509	0.294		L
	PFDA	335-76-2	ND	ND	ND	0.125	0.294		U
	PFUnDA	2058-94-8	ND	ND	ND	0.0481	0.294		U
	PFDoDA	307-55-1	ND	ND	ND	0.0475	0.294		U
	PFTriDA	72629-94-8	ND	ND	ND	0.0745	0.294		U
	PFTeDA	376-06-7	ND	ND	ND	0.0830	0.294		U
	PFBs	375-73-5	ND	ND	ND	0.0830	0.294		U
Sulfonates	PFPeS	2706-91-4	ND	ND	ND	0.0990	0.294		U
	PFHxS	355-46-4	ND	ND	ND	0.0827	0.294		U
	PFHpS	375-92-8	ND	ND	ND	0.0779	0.294		U
	PFOS	1763-23-1	ND	ND	ND	0.0471	0.294		U
	PFNS	68259-12-1	ND	ND	ND	0.0654	0.294		U
	PFDS	335-77-3	ND	ND	ND	0.135	0.294		U
	4:2 FTS	757124-72-4	ND	ND	ND	0.0646	0.294		U
	6:2 FTS	27619-97-2	ND	ND	ND	0.0723	0.294		U
	8:2 FTS	39108-34-4	ND	ND	ND	0.0569	0.294		U
	PFOSA	754-91-6	ND	ND	ND	0.365	0.366		U
	N-MeFOSAA	2355-31-9	ND	ND	ND	0.0544	0.294		U
	N-EiFOSAA	2991-50-6	ND	ND	ND	0.0651	0.294		U
other	HFPO-DA	13252-13-6	ND	ND	ND	0.0951	0.294		U
	PFMOAA	674-13-5	ND	ND	ND	1.48	1.48		U
	PFMOPrA	377-73-1	ND	ND	ND	0.240	0.294		U
	PFO2HxA	39492-88-1	ND	ND	ND	1.48	1.48		U
	PFO3OA	39492-89-2	ND	ND	ND	1.48	1.48		U
	PFO4DA	39492-90-5	ND	ND	ND	1.48	1.48		U
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.240	0.294		U
	ADONA	919005-14-4	ND	ND	ND	0.120	0.294		U
	9Cl-PF3ONS	756426-58-1	ND	ND	ND	0.120	0.294		U
	11Cl-PF3OUdS	763051-92-9	ND	ND	ND	0.120	0.294		U
	10:2 FTS	120226-60-0	ND	ND	ND	0.240	0.294		U
	FBSA	30334-69-1	ND	ND	ND	0.240	0.294		U
	N-EiFOSA	4151-50-2	ND	ND	ND	0.240	0.294		U
	N-EiFOSE	1691-99-2	ND	ND	ND	7.20	7.20		U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.240	0.294		U
	N-MeFOSE	24448-09-7	ND	ND	ND	7.20	7.20		U
	Nafion Byproduct 2	749836-20-2	ND	ND	ND	1.48	1.48		U
	NFDHA	151772-58-6	ND	ND	ND	0.240	0.294		U
	PEPA		ND	ND	ND	1.48	1.48		U
	PFECA-G	801212-59-9	ND	ND	ND	1.48	1.48		U
	PFEEA	113507-82-7	ND	ND	ND	0.240	0.294		U
	PFHxDA	67905-19-5	ND	ND	ND	1.48	1.48		U
	PFMOBA	863090-89-5	ND	ND	ND	1.48	1.48		U
	PFO5DA	39492-91-6	ND	ND	ND	3.01	3.01		U
	PMPA	13140-29-9	ND	ND	ND	1.48	1.48		U
ES	MPFBA		5063.84	8.10				64.3%	
	M5PFPeA		5460.25	8.74				68.3%	
	M3PFBS		5518.92	8.83				67.2%	
	M2-4:2 FTS		4981.09	7.97				78.0%	
	M5PFHxA		4513.11	7.22				62.0%	
	M3HFPO-DA		4456.08	7.13				64.1%	
	M4PFHpA		4712.90	7.54				67.2%	
	M3PFHxS		5128.90	8.21				72.7%	
	M2-6:2 FTS		5729.06	9.17				93.2%	
	M8PFOA		4181.20	6.69				56.0%	
	M9PFNA		4199.83	6.72				53.8%	
	M8PFOS		4203.41	6.73				62.3%	
	M2-8:2 FTS		3678.33	5.89				51.6%	
	M8FOSA-I		2833.12	4.53				42.2%	
	M6PFDA		4584.29	7.33				52.7%	
	d3-N-MeFOSAA		3171.28	5.07				49.5%	
	d5-N-EiFOSAA		2964.69	4.74				46.1%	
	M7PFUDa		3321.03	5.31				36.4%	
	MPFDoA		2873.98	4.60				36.4%	
	M2PFTeDA		1406.92	2.25				19.7%	Q

Enthalpy Analytical

Job No.: 0920-719 - PFAS by Isotope Dilution (non-potable water)

County of Brunswick Site: NorthWest Water Plant, Leland, NC

Enthalpy ID	OPR-11147-PFAS	Prep Batch	EU11147	Sample Vol (mL)	250
Sample Name	OPR-11147-PFAS	Prep Date	2020-09-11	Extract Vol (mL)	0.4
Matrix	Aqueous	Analysis Date	2020-09-11	Dilution Factor	1
Sampling Date					

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery	Flags
Acids	PFBA	375-22-4	9912.88	15.9	15.9	0.157	0.294	79.3%	
	PFPeA	2706-90-3	9192.33	14.7	14.7	0.0898	0.294	73.5%	
	PFHxA	307-24-4	9482.01	15.2	15.2	0.158	0.294	75.9%	
	PFHpA	375-85-9	9122.26	14.6	14.6	0.0695	0.294	73.0%	
	PFOA	335-67-1	9955.14	15.9	15.9	0.0795	0.294	79.6%	
	PFNA	375-95-1	9479.45	15.2	15.2	0.0509	0.294	75.8%	
	PFDA	335-76-2	9809.08	15.7	15.7	0.125	0.294	78.5%	
	PFUnDA	2058-94-8	9646.37	15.4	15.4	0.0481	0.294	77.2%	
	PFDoDA	307-55-1	9024.31	14.4	14.4	0.0475	0.294	72.2%	
	PFTTrDA	72629-94-8	11789.66	18.9	18.9	0.0745	0.294	94.3%	
	PFTeDA	376-06-7	10916.68	17.5	17.5	0.0830	0.294	87.3%	
Sulfonates	PFBS	375-73-5	7310.64	11.7	11.7	0.0830	0.294	65.9%	
	PFPeS	2706-91-4	8892.30	14.2	14.2	0.0990	0.294	75.6%	
	PFHxS	355-46-4	8599.66	13.8	13.8	0.0827	0.294	75.3%	
	PFHpS	375-92-8	8088.36	12.9	12.9	0.0779	0.294	67.9%	
	PFOS	1763-23-1	8282.93	13.3	13.3	0.0471	0.294	71.4%	
	PFNS	68259-12-1	8461.82	13.5	13.5	0.0654	0.294	70.4%	
	PFDS	335-77-3	9340.87	14.9	14.9	0.135	0.294	77.4%	
	4:2 FTS	757124-72-4	9842.22	15.7	15.7	0.0646	0.294	84.0%	
	6:2 FTS	27619-97-2	7972.38	12.8	12.8	0.0723	0.294	67.1%	
	8:2 FTS	39108-34-4	9062.68	14.5	14.5	0.0569	0.294	75.5%	
Other	PFOSA	754-91-6	9035.07	14.5	14.5	0.365	0.366	72.3%	
	N-MeFOSAA	2355-31-9	8918.16	14.3	14.3	0.0544	0.294	71.3%	
	N-EtFOSAA	2991-50-6	9948.96	15.9	15.9	0.0651	0.294	79.6%	
	HFPO-DA	13252-13-6	9719.81	15.6	15.6	0.0951	0.294	77.8%	
ES	MPFBA		4990.19	7.98				74.9%	
	M5PFPeA		5277.84	8.44				78.0%	
	M3PFBS		5469.42	8.75				78.7%	
	M2-4:2 FTS		4337.11	6.94				80.3%	
	M5PFHxA		4886.63	7.82				76.5%	
	M3HFPO-DA		4160.63	6.66				68.3%	
	M4PFHpA		4679.81	7.49				76.0%	
	M3PFHxS		5262.82	8.42				82.9%	
	M2-6:2 FTS		5947.40	9.52				114.4%	
	M8PFOA		4474.93	7.16				68.3%	
	M9PFNA		4737.41	7.58				69.1%	
	M8PFOS		5047.90	8.08				83.2%	
	M2-8:2 FTS		4044.58	6.47				63.1%	
	M8FOSA-I		4126.47	6.60				68.4%	
	M6PFDA		4357.97	6.97				60.3%	
	d3-N-MeFOSAA		3865.46	6.18				67.1%	
	d5-N-EtFOSAA		3772.02	6.04				65.2%	
	M7PFUDa		4109.99	6.58				54.3%	
	MPFDoA		3815.54	6.10				58.2%	
	M2PFTeDA		1809.44	2.90				27.5%	

Narrative Summary

Enthalpy Analytical Narrative Summary

Company	County of Brunswick
Job No.	0920-719 PFAS by Isotope Dilution (non-potable water)
Client ID.	N/A Site: NorthWest Water Plant, Leland, NC

1. Custody

Robin Appelle received the samples on 09/10/10 at 2.9°C after being relinquished by County of Brunswick. The samples were received in good condition.

Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC.

Table 1 - Sample Inventory

EU Lab Sample ID	Client Sample ID	Matrix
0920-719-002-1	091020-EO1	Aqueous
0920-719-001-1	091020-SO1	Aqueous

2. Methods and analytes

A list of analytes of interest and corresponding methods of analysis is shown in Table 3. Abbreviations are defined in the listed Appendices.

Table 3 - Methods and Analytes

EU Method	Analytes	Cleanup Method
EU-047	PFAS Full List	ENVI-Carb

3. Analysis

The samples were analyzed using Waters Acquity UPLC equipped with Xevo TQ MS (LC/MS/MS "Fili").

For aqueous samples, the sample volume was measured gravimetrically by the laboratory, and spiked with Extraction Standard (ES). The sample was then mixed well and centrifuged.

Cleanup procedures were performed on the supernatant and then extracted via SPE. Each final sample extract was transferred to an autosampler vial and spiked with 80µL of Injection Standard (IS), prior to analysis.



Enthalpy Analytical Narrative Summary

Company	County of Brunswick
Job No.	0920-719 PFAS by Isotope Dilution (non-potable water)
Client ID.	N/A Site: NorthWest Water Plant, Leland, NC

4. Calibration

In the initial calibration, the analytes exhibited R^2 of ≥ 0.99 . The calibration standard analytes, continuing calibration (concal) and Initial Calibration Verification (ICV) met the 30% criterion for native analytes.

5. QC Notes

Except where noted below, the QC sample analyses passed all method criteria.

QC samples that did not meet method acceptance criteria were: MB-11147-PFAS M2PFTeDA

ES analyte M2PFTeDA in the MB did not meet method criteria. There is no impact on data due to this analyte meeting method criteria in the OPR.

For OPR % Recovery, several analytes (PFBS, PFHpS, 6:2 FTS) did not meet method criteria but met marginal exceedance criteria. Therefore, data is accepted

The samples were extracted within the 28-day from collection holding time and analyzed within the 28-day from extraction to analysis holding time required by the method.

6. Reporting Notes

Some labeled standards in the samples fell outside the limits for ES recoveries notated by the Q qualifier. The target analytes are quantified based on their ratio to the labeled standards, therefore, undergo the same losses as the labeled standards. As a result, low or high recoveries do not cause any change to ratios or contribute any additional error in the measurement of the target analytes. Therefore, the data are considered acceptable.

The results presented in this report are representative of the samples as provided to the laboratory.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, LLC in Wilmington, NC is accredited by the Louisiana Department of Environmental Quality to the 2009 TNI Standard under certificate number 05075.



General Reporting Notes – Data Qualifiers

The following are general reporting notes that are applicable to all Enthalpy Analytical, LLC - Wilmington, NC data reports, unless specifically noted otherwise.

General Data Qualifiers

- B – The analyte was found in the method blank, at a concentration that was at least 10% of the concentration in the sample.
- Cxx – Two or more congeners co-elute. In EDDs, C denotes the lowest IUPAC congener in a co-elution group and additional co-eluters for the group ('xx') are shown with the number of the lowest IUPAC co-eluter.
- E – The reported concentration exceeds the calibration range (upper point of the calibration curve). For HRMS data, this condition does not imply additional measurement uncertainty. For LC-MS/MS data, these values should be considered as having measurement uncertainty higher than values within the calibration range.
- EDL – Estimated Detection Level. Specific to Dioxin/Furan tests and equivalent to MDL
- EMPC – Estimated Maximum Possible Concentration Specific to Dioxin/Furan tests to indicate the signal/noise ratio was not sufficient for peak identification (the determined ion-abundance ratio was outside the allowed theoretical range), or where there was a co-eluting interference. Indicates that a peak was identified but did not meet the method specified ion-abundance ratio.
- IR – The ion ratio between the primary and secondary ions was observed to be outside the method criteria therefore the actual analyte concentration cannot be accurately determined as defined by DoD QSM Table B-15.
- J – The analyte has a concentration below the minimum calibration level (LOQ value) but greater than the LOD. These values should be considered as having measurement uncertainty higher than values within the calibration range
- L - Indicates that an analyte has a concentration below the Minimum Detection Limit (MDL). The reported concentration is not recommended for regulatory use as the analyte signal may have a signal-to-noise ratio less than the criteria deemed necessary to be considered a detected analyte.
- LOD – Limit of Detection: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOD. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the minimum detection limit (MDL). The LOD is adjusted for sample weight or volume.
- LOQ – Limit of Quantiation: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOQ. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the reporting limit (RL). The LOD is adjusted for sample weight or volume.
- <LOD() – Analyte was not found at a concentration high enough to be reported as detected. It is reported as less than the LOD, and the LOD is given in the parentheses.

General Reporting Notes – Data Qualifiers

- ND – Indicates a non-detect.
- NR – Indicates a value that is not reportable due to issues observed in sample preparation or analysis.
- PR – The associated congener(s) is(are) poorly resolved.
- QI – Indicates the presence of a quantitative interference.
- RL – Reporting Limit. Lowest reportable value. The level is higher than the MDL.
- SI – Denotes “Single Ion Mode” and is utilized for PCBs where the secondary ion trace has a significantly elevated noise level due to background PFK. Responses for such peaks are calculated using an EMPC approach based solely on the primary ion area(s) and may be considered estimates.
- U – The analyte was not detected.
- V – The labeled standard recovery is not within method control limits.
- X – Results from re-injection/repeat/second-column analysis.

Lab Identifiers/ Data Attributes

- AR – Indicates use of the archived portion of the sample extract.
- CU – Indicates a sample that required additional clean-up prior to HRMS injection/processing.
- D – Dilution Data. Result was obtained from the analysis of a dilution. The number that follows the “D” indicates the dilution factor.
- DE – Indicates a dilution performed with the addition of ES (Extraction Standard) solution.
- DUP – Designation for a duplicate sample.
- MS – Designation for a matrix spike.
- MSD – Designation for a matrix spike duplicate.
- RJ – Indicates a reinjection of the sample extract.
- S – Indicates a sample split. The number that follows the “S” indicates the split factor.
- R – Indicates a re-extraction of the sample.

PFAS Compound Acronym List		
Acronym	Compound Name	CAS #
Target Analytes		
PFBA	Perfluorobutanoic Acid	375-22-4
PFPeA	Perfluoropentanoic Acid	2706-90-3
PFHxA	Perfluorohexanoic Acid	307-24-4
PFHpA	Perfluoroheptanoic Acid	375-85-9
PFOA	Perfluorooctanoic Acid	335-67-1
PFNA	Perfluorononanoic Acid	375-95-1
PFDA	Perfluorodecanoic acid	335-76-2
PFUnA (PFUnDA)	Perfluoroundecanoic acid	2058-94-8
PFDoA (PFDoDA)	Perfluorododecanoic acid	307-55-1
PFTriA (PFTriA)	Perfluorotridecanoic acid	72629-94-8
PFTeDA (PFTA)	Perfluorotetradecanoic acid	376-06-7
PFBS	Perfluorobutane sulfonic acid	375-73-5
PFPeS	Perfluoropentane sulfonic acid	2706-91-4
PFHxS	Perfluorohexane sulfonic acid	355-46-4
PFHpS	Perfluoroheptane sulfonic acid	375-92-8
PFOS	Perfluorooctane sulfonic acid	1763-23-1
PFNS	Perfluorononane sulfonic acid	68259-12-1
PFDS	Perfluorodecane sulfonic acid	757124-72-4
4:2 FTS	4:2 fluorotelomer sulfonic acid	27619-97-2
6:2 FTS	6:2 fluorotelomer sulfonic acid	39108-34-4
8:2 FTS	8:2 fluorotelomer sulfonic acid	13252-13-6
PFOSA (FOSA)	Perfluorooctane sulfonamide	754-91-6
N-MeFOSAA	N-methyl perfluorooctane sulfonamido acetic acid	2355-31-9
N-EtFOSAA	N-ethyl perfluorooctane sulfonamido acetic acid	2991-50-6
* HFPO-DA	2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy)-propanoic acid (Gen-X)	13252-13-6
* 11CI-PF3OUdS	11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	674-13-5
* 9CI-PF3ONS	9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	919005-14-4
* ADONA	4,8-dioxia-3H-perfluorononanoic acid	756426-58-1
* PFMOAA	Perfluoro-2-methoxyacetic acid	763051-92-9
* PFMOPrA	Perfluoro-3-methoxypropanoic acid	377-73-1
* PFO2HxA	Perfluoro (3,5-dioxahexanoic) acid	39492-88-1
* PFO3OA	Perfluoro (3,5,7-trioxaoctanoic) acid	39492-89-2
* PFO4DA	Perfluoro (3,5,7,9-tetraoxadecanoic) acid	39492-90-5
* PFO5DA	Perfluoro(3,5,7,9,11-pentaoxadodecanoic) acid	39492-91-6
* Nafion Byproduct 1	Nafion Byproduct 1	29311-67-9
* Nafion Byproduct 2	Nafion Byproduct 2	749836-20-2
* PFEESA	Perfluoro(2-ethoxyethane)sulphonic acid	113507-82-7
* PFMOBA	Perfluoro-4-methoxybutanic acid	863090-89-5
* PEPA	Perfluoro-2-ethoxypropanoic acid	N/A
* PMPA	Perfluoro-2-methoxypropanoic acid	13140-29-9
* 10:2 FTS	Fluorotelomer sulfonate 10:2	120226-60-0
* N-EtFOSA	N-ethylperfluoro-1-octanesulfonamide	4151-50-2
* N-EtFOSE	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol	1691-99-2
* N-MeFOSA	N-methylperfluoro-1-octanesulfonamide	31506-32-8
* N-MeFOSE	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol	24448-09-7
* PFECA-G	4-(Heptafluoroisopropoxy)hexafluorobutanoic acid	801212-59-9
* PFHxDA	Perfluorohexadecanoic acid	67905-19-5

Extraction Standards		
MPFBA	Perfluoro-n-[13C4]butanoic acid	
M5PFPeA	Perfluoro-n-[13C5]pentanoic acid	
M3PFBS	Sodium perfluoro-1-[2,3,4-13C3]-butanesulfonic acid	
M2-4:2 FTS	Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]-hexane sulfonic acid	
M5PFHxA	Perfluoro-n-[1,2,3,4,6-13C5]hexanoic acid	
M3HFPO-DA	2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy)-13C3-propanoic acid	
M4PFHpA	Perfluoro-n-[1,2,3,4-13C4]heptanoic acid	
M3PFHxS	Sodium perfluoro-1-[1,2,3-13C3]-hexanesulfonic acid	
M2-6:2 FTS	Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]-octane sulfonic acid	
M8PFOA	Perfluoro-n-[13C8]octanoic acid	
M9PFNA	Perfluoro-n-[13C9]nonanoic acid	
M8PFOS	Sodium perfluoro-1-[13C8]-octanesulfonic acid	
M2-8:2 FTS	Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]-decane sulfonic acid	
M8FOSA	Perfluoro-1-[13C8]octanesulfonamide	
M6PFDA	Perfluoro-n-[1,2,3,4,5,6-13C6]decanoic acid	
d3-N-MeFOSAA	N-methyl-d3-perfluoro-1-octanesulfonamide	
d5-N-EtFOSAA	N-ethyl-d5-perfluoro-1-octanesulfonamide	
M7PFUnDA (M7PFUDa)	Perfluoro-n-[1,2,3,4,5,6,7-13C7]undecanoic acid	
MPFDoA	Perfluoro-n-[1,2-13C2]dodecanoic acid	
M2PFTeDA	Perfluoro-n-[1,2-13C2]tetradecanoic acid	
Injection Standards		
M3PFBA	Perfluoro-n-[2,3,4-13C3]butanoic acid	
M2PFOA	Perfluoro-n-[1,2-13C2]octanoic acid	
MPFDA	Perfluoro-n-[1,2-13C2]decanoic acid	
MPFOS	Sodium perfluoro-1-[1,2,3,4-13C4]-octanesulfonic acid	

* Analytes are currently not accredited under TNI Standard. Accreditation pending.

Sample Custody

Client Courier, on 7/19/2019, No results Good condition 9/1/20

**This Is The Last Page
Of This Report.**