

County of Brunswick

3954 Clearwell Dr NE
Leland, NC 28451

Northwest Water Plant

Leland, NC
Samples Received: 09/17/20

Analytical Report 0920-740

Isotope Dilution Method PFAS



Enthalpy Analytical, LLC – Ultratrace

Lindsay Boone
O: (910) 212-5855 / F: 910-212-5866
lboone@enthalpy.com / www.enthalpy.com
2714 Exchange Drive, Wilmington, NC 28405

I certify that to the best of my knowledge all analytical data presented in this report:

- Have been checked for completeness
- Are accurate, error-free, and legible
- Have been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s)

This analytical report was prepared in Portable Document Format (.PDF) and contains _____ pages.

Report Issued Date: _____



Summary of Results

Enthalpy Analytical

Job No.: 0920-740 PFAS by Isotope Dilution (non-potable water)

County of Brunswick Site: Northwest Water Plant, Leland N.C.

Summary

	Compound	CAS	091720-S01 ng/L	091720-E01 ng/L
Acids	PFBA	375-22-4	ND U	ND U
	PFPeA	2706-90-3	9.86	9.00
	PFHxA	307-24-4	7.99	8.39
	PFHpA	375-85-9	4.69	3.76
	PFOA	335-67-1	7.22	6.90
	PFNA	375-95-1	0.947	0.968
	PFDA	335-76-2	0.618	0.471
	PFUnDA	2058-94-8	0.116 J	0.209 J
	PFDoDA	307-55-1	ND U	ND U
	PFTTrDA	72629-94-8	ND U	ND U
	PFTeDA	376-06-7	ND U	ND U
Sulfonates	PFBS	375-73-5	ND U	ND U
	PFPeS	2706-91-4	ND U	ND U
	PFHxS	355-46-4	4.99	5.05
	PFHpS	375-92-8	ND U	ND U
	PFOS	1763-23-1	11.8	12.3
	PFNS	68259-12-1	ND U	ND U
	PFDS	335-77-3	ND U	ND U
	4:2 FTS	757124-72-4	ND U	ND U
	6:2 FTS	27619-97-2	0.171 J	0.306
other	8:2 FTS	39108-34-4	ND U	ND U
	PFOSA	754-91-6	ND U	ND U
	N-MeFOSAA	2355-31-9	ND U	ND U
	N-EtFOSAA	2991-50-6	ND U	ND U
	HFPO-DA	13252-13-6	18.9	15.3
	PFMOAA	674-13-5	130	110
	PFMOPrA	377-73-1	ND U	ND U
	PFO2HxA	39492-88-1	4.79	7.47
	PFO3OA	39492-89-2	ND U	7.38
	PFO4DA	39492-90-5	3.63	3.18
	Nafion Byproduct 1	29311-67-9	ND U	0.489
	ADONA	919005-14-4	ND U	ND U
	9Cl-PF3ONS	756426-58-1	ND U	ND U
	11Cl-PF3OUdS	763051-92-9	ND U	ND U
	10:2 FTS	120226-60-0	ND U	ND U
	FBSA	30334-69-1	0.737	1.12
	N-EtFOSA	4151-50-2	ND U	ND U
	N-EtFOSE	1691-99-2	ND U	ND U
	N-MeFOSA	31506-32-8	ND U	ND U
	N-MeFOSE	24448-09-7	ND U	ND U
	Nafion Byproduct 2	749836-20-2	1.30	1.08 L
	NFDHA	151772-58-6	ND U	ND U
	PEPA		ND U	ND U
	PFECA-G	801212-59-9	ND U	ND U
	PFEESA	113507-82-7	ND U	ND U
	PFHxDA	67905-19-5	ND U	ND U
	PFMOBA	863090-89-5	ND U	ND U
	PFO5DA	39492-91-6	ND U	ND U
	PMPA	13140-29-9	ND U	ND U

Detailed Results

Enthalpy Analytical

Job No.: 0920-740 PFAS by Isotope Dilution (non-potable water)

County of Brunswick Site: Northwest Water Plant, Leland N.C.

Enthalpy ID	0920-740-001-1	Prep Batch	EU11165	Sample Vol (mL)	294.61
Sample Name	091720-S01	Prep Date	2020-09-19	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	2020-09-20	Dilution Factor	1
Sampling Date	20200917 00:00				

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery	Flags
Acids	PFBA	375-22-4	ND	ND	ND	0.134	0.250		U
	PFPeA	2706-90-3	7259.76	9.86	9.86	0.0762	0.250		
	PFHxA	307-24-4	5882.79	7.99	7.99	0.134	0.250		
	PFHpA	375-85-9	3455.70	4.69	4.69	0.0590	0.250		
	PFOA	335-67-1	5316.65	7.22	7.22	0.0674	0.250		
	PFNA	375-95-1	697.27	0.947	0.947	0.0432	0.250		
	PFDA	335-76-2	455.47	0.618	0.618	0.106	0.250		
	PFUnDA	2058-94-8	85.46	0.116	0.116	0.0408	0.250		J
	PFDaDA	307-55-1	ND	ND	ND	0.0403	0.250		U
	PFTriDA	72629-94-8	ND	ND	ND	0.0632	0.250		U
	PFTeDA	376-06-7	ND	ND	ND	0.0704	0.250		U
Sulfonates	PFBS	375-73-5	ND	ND	ND	0.0704	0.250		U
	PFPeS	2706-91-4	ND	ND	ND	0.0840	0.250		U
	PFHxS	355-46-4	3678.23	4.99	4.99	0.0702	0.250		
	PFHpS	375-92-8	ND	ND	ND	0.0661	0.250		U
	PFOS	1763-23-1	8714.71	11.8	11.8	0.0400	0.250		
	PFNS	68259-12-1	ND	ND	ND	0.0555	0.250		U
	PFDS	335-77-3	ND	ND	ND	0.115	0.250		U
	4:2 FTS	757124-72-4	ND	ND	ND	0.0548	0.250		U
	6:2 FTS	27619-97-2	126.30	0.171	0.171	0.0614	0.250		J
	8:2 FTS	39108-34-4	ND	ND	ND	0.0483	0.250		U
other	PFOSA	754-91-6	ND	ND	ND	0.310	0.311		U
	N-MeFOSAA	2355-31-9	ND	ND	ND	0.0462	0.250		U
	N-EiFOSAA	2991-50-6	ND	ND	ND	0.0552	0.250		U
	HFPO-DA	13252-13-6	13949.15	18.9	18.9	0.0807	0.250		
	PFMOAA	674-13-5	95674.29	130	130	1.26	1.26		
	PFMOPrA	377-73-1	ND	ND	ND	0.204	0.250		U
	PFO2HxA	39492-88-1	3525.26	4.79	4.79	1.26	1.26		
	PFO3OA	39492-89-2	ND	ND	ND	1.26	1.26		U
	PFO4DA	39492-90-5	2673.44	3.63	3.63	1.26	1.26		
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.204	0.250		U
	ADONA	919005-14-4	ND	ND	ND	0.102	0.250		U
	9Cl-PF3ONS	756426-58-1	ND	ND	ND	0.102	0.250		U
	11Cl-PF3OUdS	763051-92-9	ND	ND	ND	0.102	0.250		U
	10:2 FTS	120226-60-0	ND	ND	ND	0.204	0.250		U
	FBSA	30334-69-1	543.09	0.737	0.737	0.204	0.250		
	N-EiFOSA	4151-50-2	ND	ND	ND	0.204	0.250		U
	N-EiFOSE	1691-99-2	ND	ND	ND	6.11	6.11		U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.204	0.250		U
	N-MeFOSE	24448-09-7	ND	ND	ND	6.11	6.11		U
	Nafion Byproduct 2	749836-20-2	956.87	1.30	1.30	1.26	1.26		
	NFDHA	151772-58-6	ND	ND	ND	0.204	0.250		U
	PEPA		ND	ND	ND	1.26	1.26		U
	PFECA-G	801212-59-9	ND	ND	ND	1.26	1.26		U
	PFEEA	113507-82-7	ND	ND	ND	0.204	0.250		U
	PFHxDA	67905-19-5	ND	ND	ND	1.26	1.26		U
	PFMOBA	863090-89-5	ND	ND	ND	1.26	1.26		U
	PFO5DA	39492-91-6	ND	ND	ND	2.55	2.55		U
	PMPA	13140-29-9	ND	ND	ND	1.26	1.26		U
ES	MPFBA		5363.71	7.28				12.3%	Q
	M5PFPeA		12369.53	16.8				28.0%	
	M3PFBS		30572.31	41.5				67.4%	
	M2-4:2 FTS		61835.13	84.0				175.3%	Q
	M5PFHxA		4554.86	6.18				58.7%	
	M3HFPO-DA		3777.22	5.13				51.0%	
	M4PFHpA		4593.29	6.24				61.4%	
	M3PFHxS		4632.37	6.29				67.6%	
	M2-6:2 FTS		41829.24	56.8				123.3%	
	M8PFOA		4720.34	6.41				59.3%	
	M9PFNA		5123.52	6.96				61.5%	
	M8PFOS		4710.32	6.40				71.9%	
	M2-8:2 FTS		4109.87	5.58				59.4%	
	M8FOSA-I		2127.09	2.89				32.7%	
	M6PFDA		4117.69	5.59				52.8%	
	d3-N-MeFOSAA		3603.89	4.89				57.9%	
	d5-N-EiFOSAA		3918.26	5.32				62.8%	
	M7PFUDa		4230.97	5.74				51.7%	
	MPFDaA		3496.77	4.75				49.4%	
	M2PFTeDA		2245.53	3.05				30.5%	

Enthalpy Analytical

Job No.: 0920-740 PFAS by Isotope Dilution (non-potable water)

County of Brunswick Site: Northwest Water Plant, Leland N.C.

Enthalpy ID	0920-740-002-1	Prep Batch	EU11165	Sample Vol (mL)	298.06
Sample Name	091720-E01	Prep Date	2020-09-19	Extract Vol (mL)	0.4
Matrix	aqueous	Analysis Date	2020-09-20	Dilution Factor	1
Sampling Date	20200917 00:00				

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery	Flags
Acids	PFBA	375-22-4	ND	ND	ND	0.132	0.247		U
	PFPeA	2706-90-3	6703.10	9.00	9.00	0.0753	0.247		
	PFHxA	307-24-4	6250.29	8.39	8.39	0.132	0.247		
	PFHpA	375-85-9	2798.74	3.76	3.76	0.0583	0.247		
	PFOA	335-67-1	5142.42	6.90	6.90	0.0666	0.247		
	PFNA	375-95-1	721.28	0.968	0.968	0.0427	0.247		
	PFDA	335-76-2	350.82	0.471	0.471	0.105	0.247		
	PFUnDA	2058-94-8	155.73	0.209	0.209	0.0403	0.247		J
	PFDoDA	307-55-1	ND	ND	ND	0.0398	0.247		U
	PFTriDA	72629-94-8	ND	ND	ND	0.0625	0.247		U
Sulfonates	PFTeDA	376-06-7	ND	ND	ND	0.0696	0.247		U
	PFBS	375-73-5	ND	ND	ND	0.0696	0.247		U
	PFPeS	2706-91-4	ND	ND	ND	0.0830	0.247		U
	PFHxS	355-46-4	3762.91	5.05	5.05	0.0694	0.247		
	PFHpS	375-92-8	ND	ND	ND	0.0653	0.247		U
	PFOS	1763-23-1	9131.21	12.3	12.3	0.0395	0.247		
	PFNS	68259-12-1	ND	ND	ND	0.0549	0.247		U
	PFDS	335-77-3	ND	ND	ND	0.113	0.247		U
	4:2 FTS	757124-72-4	ND	ND	ND	0.0542	0.247		U
	6:2 FTS	27619-97-2	228.14	0.306	0.306	0.0606	0.247		
other	8:2 FTS	39108-34-4	ND	ND	ND	0.0477	0.247		U
	PFOSA	754-91-6	ND	ND	ND	0.306	0.307		U
	N-MeFOSAA	2355-31-9	ND	ND	ND	0.0456	0.247		U
	N-EiFOSAA	2991-50-6	ND	ND	ND	0.0546	0.247		U
	HFPO-DA	13252-13-6	11422.23	15.3	15.3	0.0798	0.247		
	PFMOAA	674-13-5	81629.63	110	110	1.24	1.24		
	PFMOPrA	377-73-1	ND	ND	ND	0.201	0.247		U
	PFO2HxA	39492-88-1	5563.23	7.47	7.47	1.24	1.24		
	PFO3OA	39492-89-2	5495.51	7.38	7.38	1.24	1.24		
	PFO4DA	39492-90-5	2368.16	3.18	3.18	1.24	1.24		
	Nafion Byproduct 1	29311-67-9	364.36	0.489	0.489	0.201	0.247		
	ADONA	919005-14-4	ND	ND	ND	0.101	0.247		U
	9Cl-PF3ONS	756426-58-1	ND	ND	ND	0.101	0.247		U
	11Cl-PF3OUdS	763051-92-9	ND	ND	ND	0.101	0.247		U
	10:2 FTS	120226-60-0	ND	ND	ND	0.201	0.247		U
	FBSA	30334-69-1	838.22	1.12	1.12	0.201	0.247		
	N-EiFOSA	4151-50-2	ND	ND	ND	0.201	0.247		U
	N-EiFOSE	1691-99-2	ND	ND	ND	6.04	6.04		U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.201	0.247		U
	N-MeFOSE	24448-09-7	ND	ND	ND	6.04	6.04		U
	Nafion Byproduct 2	749836-20-2	806.85	1.08	1.08	1.24	1.24		L
	NFDHA	151772-58-6	ND	ND	ND	0.201	0.247		U
	PEPA		ND	ND	ND	1.24	1.24		U
	PFECA-G	801212-59-9	ND	ND	ND	1.24	1.24		U
	PFEEA	113507-82-7	ND	ND	ND	0.201	0.247		U
	PFHxDA	67905-19-5	ND	ND	ND	1.24	1.24		U
	PFMOBA	863090-89-5	ND	ND	ND	1.24	1.24		U
	PFO5DA	39492-91-6	ND	ND	ND	2.52	2.52		U
	PMPA	13140-29-9	ND	ND	ND	1.24	1.24		U
ES	MPFBA		5393.86	7.24				13.0%	Q
	M5PFPeA		14282.72	19.2				33.9%	
	M3PFBS		29747.79	39.9				68.8%	
	M2-4:2 FTS		48238.12	64.7				143.5%	
	M5PFHxA		4578.78	6.14				59.8%	
	M3HFPO-DA		4793.33	6.43				65.6%	
	M4PFHpA		5271.76	7.07				71.5%	
	M3PFHxS		4551.40	6.11				70.3%	
	M2-6:2 FTS		37580.47	50.4				116.2%	
	M8PFOA		5165.56	6.93				65.8%	
	M9PFNA		5109.52	6.86				62.2%	
	M8PFOS		4392.85	5.90				71.0%	
	M2-8:2 FTS		3996.25	5.36				61.1%	
	M8FOSA-I		3040.66	4.08				49.4%	
	M6PFDA		4387.05	5.89				59.6%	
	d3-N-MeFOSAA		3695.49	4.96				62.8%	
	d5-N-EiFOSAA		3663.50	4.92				62.1%	
	M7PFUDa		3956.20	5.31				51.2%	
	MPFDoA		2744.76	3.68				41.0%	
	M2PFTeDA		2260.41	3.03				32.5%	

QC Data

Enthalpy Analytical

Job No.: 0920-740 PFAS by Isotope Dilution (non-potable water)

County of Brunswick Site: Northwest Water Plant, Leland N.C.

Enthalpy ID	MB-11165-PFAS	Prep Batch	EU11165	Sample Vol (mL)	250
Sample Name	MB-11165-PFAS	Prep Date	2020-09-19	Extract Vol (mL)	0.4
Matrix		Analysis Date	2020-09-19	Dilution Factor	1
Sampling Date					

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery	Flags
Acids	PFBA	375-22-4	ND	ND	ND	0.157	0.294		U
	PFPeA	2706-90-3	ND	ND	ND	0.0898	0.294		U
	PFHxA	307-24-4	ND	ND	ND	0.158	0.294		U
	PFHpA	375-85-9	ND	ND	ND	0.0695	0.294		U
	PFOA	335-67-1	ND	ND	ND	0.0795	0.294		U
	PFNA	375-95-1	ND	ND	ND	0.0509	0.294		U
	PFDA	335-76-2	ND	ND	ND	0.125	0.294		U
	PFUnDA	2058-94-8	ND	ND	ND	0.0481	0.294		U
	PFDoDA	307-55-1	ND	ND	ND	0.0475	0.294		U
	PFTriDA	72629-94-8	ND	ND	ND	0.0745	0.294		U
Sulfonates	PFTeDA	376-06-7	ND	ND	ND	0.0830	0.294		U
	PFBS	375-73-5	ND	ND	ND	0.0830	0.294		U
	PFPeS	2706-91-4	ND	ND	ND	0.0990	0.294		U
	PFHxS	355-46-4	ND	ND	ND	0.0827	0.294		U
	PFHpS	375-92-8	ND	ND	ND	0.0779	0.294		U
	PFOS	1763-23-1	ND	ND	ND	0.0471	0.294		U
	PFNS	68259-12-1	ND	ND	ND	0.0654	0.294		U
	PFDS	335-77-3	ND	ND	ND	0.135	0.294		U
	4:2 FTS	757124-72-4	ND	ND	ND	0.0646	0.294		U
	6:2 FTS	27619-97-2	ND	ND	ND	0.0723	0.294		U
other	8:2 FTS	39108-34-4	ND	ND	ND	0.0569	0.294		U
	PFOSA	754-91-6	ND	ND	ND	0.365	0.366		U
	N-MeFOSAA	2355-31-9	ND	ND	ND	0.0544	0.294		U
	N-EiFOSAA	2991-50-6	ND	ND	ND	0.0651	0.294		U
	HFPO-DA	13252-13-6	ND	ND	ND	0.0951	0.294		U
	PFMOAA	674-13-5	ND	ND	ND	1.48	1.48		U
	PFMOPrA	377-73-1	ND	ND	ND	0.240	0.294		U
	PFO2HxA	39492-88-1	ND	ND	ND	1.48	1.48		U
	PFO3OA	39492-89-2	ND	ND	ND	1.48	1.48		U
	PFO4DA	39492-90-5	ND	ND	ND	1.48	1.48		U
	Nafion Byproduct 1	29311-67-9	ND	ND	ND	0.240	0.294		U
	ADONA	919005-14-4	ND	ND	ND	0.120	0.294		U
	9Cl-PF3ONS	756426-58-1	ND	ND	ND	0.120	0.294		U
	11Cl-PF3OUdS	763051-92-9	ND	ND	ND	0.120	0.294		U
	10:2 FTS	120226-60-0	ND	ND	ND	0.240	0.294		U
	FBSA	30334-69-1	ND	ND	ND	0.240	0.294		U
	N-EiFOSA	4151-50-2	ND	ND	ND	0.240	0.294		U
	N-EiFOSE	1691-99-2	ND	ND	ND	7.20	7.20		U
	N-MeFOSA	31506-32-8	ND	ND	ND	0.240	0.294		U
	N-MeFOSE	24448-09-7	ND	ND	ND	7.20	7.20		U
	Nafion Byproduct 2	749836-20-2	ND	ND	ND	1.48	1.48		U
	NFDHA	151772-58-6	ND	ND	ND	0.240	0.294		U
	PEPA		ND	ND	ND	1.48	1.48		U
	PFECA-G	801212-59-9	ND	ND	ND	1.48	1.48		U
	PFEEA	113507-82-7	ND	ND	ND	0.240	0.294		U
	PFHxDA	67905-19-5	ND	ND	ND	1.48	1.48		U
	PFMOBA	863090-89-5	ND	ND	ND	1.48	1.48		U
	PFO5DA	39492-91-6	ND	ND	ND	3.01	3.01		U
	PMPA	13140-29-9	ND	ND	ND	1.48	1.48		U
ES	MPFBA		5329.83	8.53				51.2%	
	M5PFPeA		6070.01	9.71				57.4%	
	M3PFBS		6531.40	10.5				60.1%	
	M2-4:2 FTS		5692.70	9.11				67.3%	
	M5PFHxA		4900.61	7.84				59.9%	
	M3HFPO-DA		4777.07	7.64				61.2%	
	M4PFHpA		4785.31	7.66				60.7%	
	M3PFHxS		4609.58	7.38				62.1%	
	M2-6:2 FTS		6100.71	9.76				75.0%	
	M8PFOA		5078.40	8.13				60.6%	
	M9PFNA		5508.09	8.81				62.8%	
	M8PFOS		4780.09	7.65				67.4%	
	M2-8:2 FTS		4190.75	6.71				55.9%	
	M8FOSA-I		2866.45	4.59				40.7%	
	M6PFDA		4992.56	7.99				54.5%	
	d3-N-MeFOSAA		3786.76	6.06				56.2%	
	d5-N-EiFOSAA		3627.29	5.80				53.7%	
	M7PFUDa		5051.10	8.08				52.6%	
	MPFDoA		3970.51	6.35				47.8%	
	M2PFTeDA		3078.48	4.93				35.7%	

Enthalpy Analytical

Job No.: 0920-740 PFAS by Isotope Dilution (non-potable water)

County of Brunswick Site: Northwest Water Plant, Leland N.C.

Enthalpy ID	OPR-11165-PFAS	Prep Batch	EU11165	Sample Vol (mL)	250
Sample Name	OPR-11165-PFAS	Prep Date	2020-09-19	Extract Vol (mL)	0.4
Matrix		Analysis Date	2020-09-19	Dilution Factor	1
Sampling Date					

	Compound	CAS	Extract Concentration ng/L	Sample Concentration ng/L	Formatted Result ng/L	LOD ng/L	LOQ ng/L	Recovery	Flags
Acids	PFBA	375-22-4	10809.05	17.3	17.3	0.157	0.294	86.5%	
	PFPeA	2706-90-3	9636.32	15.4	15.4	0.0898	0.294	77.1%	
	PFHxA	307-24-4	10519.45	16.8	16.8	0.158	0.294	84.2%	
	PFHpA	375-85-9	10798.28	17.3	17.3	0.0695	0.294	86.4%	
	PFOA	335-67-1	10217.84	16.3	16.3	0.0795	0.294	81.7%	
	PFNA	375-95-1	9604.21	15.4	15.4	0.0509	0.294	76.8%	
	PFDA	335-76-2	11073.99	17.7	17.7	0.125	0.294	88.6%	
	PFUnDA	2058-94-8	11537.83	18.5	18.5	0.0481	0.294	92.3%	
	PFDoDA	307-55-1	9015.45	14.4	14.4	0.0475	0.294	72.1%	
	PFTTrDA	72629-94-8	10711.43	17.1	17.1	0.0745	0.294	85.7%	
	PFTeDA	376-06-7	9577.30	15.3	15.3	0.0830	0.294	76.6%	
Sulfonates	PFBS	375-73-5	8855.42	14.2	14.2	0.0830	0.294	79.9%	
	PFPeS	2706-91-4	9806.78	15.7	15.7	0.0990	0.294	83.4%	
	PFHxS	355-46-4	9485.85	15.2	15.2	0.0827	0.294	83.0%	
	PFHpS	375-92-8	8984.34	14.4	14.4	0.0779	0.294	75.4%	
	PFOS	1763-23-1	9278.14	14.8	14.8	0.0471	0.294	80.0%	
	PFNS	68259-12-1	9385.93	15.0	15.0	0.0654	0.294	78.1%	
	PFDS	335-77-3	13044.48	20.9	20.9	0.135	0.294	108.1%	
	4:2 FTS	757124-72-4	9400.74	15.0	15.0	0.0646	0.294	80.3%	
	6:2 FTS	27619-97-2	8894.06	14.2	14.2	0.0723	0.294	74.8%	
Other	8:2 FTS	39108-34-4	8503.30	13.6	13.6	0.0569	0.294	70.9%	
	PFOSA	754-91-6	10409.54	16.7	16.7	0.365	0.366	83.3%	
	N-MeFOSAA	2355-31-9	8867.10	14.2	14.2	0.0544	0.294	70.9%	
	N-EtFOSAA	2991-50-6	9979.75	16.0	16.0	0.0651	0.294	79.8%	
ES	HFPO-DA	13252-13-6	8821.74	14.1	14.1	0.0951	0.294	70.6%	
	MPFBA		4729.10	7.57				45.1%	
	M5PFPeA		5440.46	8.70				51.1%	
	M3PFBS		5693.04	9.11				52.1%	
	M2-4:2 FTS		5542.31	8.87				65.2%	
	M5PFHxA		4800.95	7.68				55.8%	
	M3HFPO-DA		4544.26	7.27				55.4%	
	M4PFHpA		4351.88	6.96				52.5%	
	M3PFHxS		4348.62	6.96				57.9%	
	M2-6:2 FTS		6023.88	9.64				73.6%	
	M8PFOA		4688.88	7.50				53.2%	
	M9PFNA		4986.29	7.98				54.0%	
	M8PFOS		4396.76	7.03				61.3%	
	M2-8:2 FTS		3895.12	6.23				51.4%	
	M8FOSA-I		3130.12	5.01				43.9%	
	M6PFDA		3981.40	6.37				43.7%	
	d3-N-MeFOSAA		3410.12	5.46				50.0%	
	d5-N-EtFOSAA		3074.95	4.92				45.0%	
	M7PFUDa		3610.97	5.78				37.8%	
	MPFDoA		3870.08	6.19				46.8%	
	M2PFTeDA		3175.73	5.08				37.0%	

Narrative Summary

Enthalpy Analytical Narrative Summary

Company	County of Brunswick
Job No.	0920-740 PFAS by Isotope Dilution (non-potable water)
Client ID.	N/A Site: Northwest Water Plant, Leland N.C.

1. Custody

Braidy May received the samples on 09/17/20 at 4.3°C after being relinquished by County of Brunswick. The samples were received in good condition.

Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC.

Table 1 - Sample Inventory

EU Lab Sample ID	Client Sample ID	Matrix
0920-740-002-1	091720-E01	aqueous
0920-740-001-1	091720-S01	aqueous

2. Methods and analytes

A list of analytes of interest and corresponding methods of analysis is shown in Table 3. Abbreviations are defined in the listed Appendices.

Table 3 - Methods and Analytes

EU Method	Analytes	Cleanup Method
EU-047	PFAS - Full List	ENVI-Carb

3. Analysis

The samples were analyzed using Waters Acquity UPLC equipped with Xevo TQ MS (LC/MS/MS "Fili").

For aqueous samples, the sample volume was measured gravimetrically by the laboratory, and spiked with Extraction Standard (ES). The sample was then mixed well and centrifuged.

Cleanup procedures were performed on the supernatant and then extracted via SPE. Each final sample extract was transferred to an autosampler vial and spiked with 400µL of Injection Standard (IS), prior to analysis.



Enthalpy Analytical Narrative Summary

Company	County of Brunswick
Job No.	0920-740 PFAS by Isotope Dilution (non-potable water)
Client ID.	N/A Site: Northwest Water Plant, Leland N.C.

4. Calibration

In the initial calibration, the analytes exhibited R^2 of ≥ 0.99 . The calibration standard analytes, continuing calibration (concal) and Initial Calibration Verification (ICV) met the 30% criterion for native analytes.

5. QC Notes

QC sample analyses passed all method criteria.

The samples were extracted within the 28-day from collection holding time and analyzed within the 28-day from extraction to analysis holding time required by the method.

6. Reporting Notes

Some labeled standards in the samples fell outside the limits for ES recoveries, notated with a Q qualifier. The target analytes are quantified based on their ratio to the labeled standards, therefore, undergo the same losses as the labeled standards. As a result, low or high recoveries do not cause any change to ratios or contribute any additional error in the measurement of the target analytes. Therefore, the data are considered acceptable.

The results presented in this report are representative of the samples as provided to the laboratory.

These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, LLC in Wilmington NC is accredited by the Louisiana Department of Environmental Quality to the 2009 TNI Standard under certificate number 05075.





General Reporting Notes – Data Qualifiers

The following are general reporting notes that are applicable to all Enthalpy Analytical, LLC - Wilmington, NC data reports, unless specifically noted otherwise.

General Data Qualifiers

- B – The analyte was found in the method blank, at a concentration that was at least 10% of the concentration in the sample.
- Cxx – Two or more congeners co-elute. In EDDs, C denotes the lowest IUPAC congener in a co-elution group and additional co-eluters for the group ('xx') are shown with the number of the lowest IUPAC co-eluter.
- E – The reported concentration exceeds the calibration range (upper point of the calibration curve). For HRMS data, this condition does not imply additional measurement uncertainty. For LC-MS/MS data, these values should be considered as having measurement uncertainty higher than values within the calibration range.
- EDL – Estimated Detection Level. Specific to Dioxin/Furan tests and equivalent to MDL
- EMPC – Estimated Maximum Possible Concentration Specific to Dioxin/Furan tests to indicate the signal/noise ratio was not sufficient for peak identification (the determined ion-abundance ratio was outside the allowed theoretical range), or where there was a co-eluting interference. Indicates that a peak was identified but did not meet the method specified ion-abundance ratio.
- IR – The ion ratio between the primary and secondary ions was observed to be outside the method criteria therefore the actual analyte concentration cannot be accurately determined as defined by DoD QSM Table B-15.
- J – The analyte has a concentration below the minimum calibration level (LOQ value) but greater than the LOD. These values should be considered as having measurement uncertainty higher than values within the calibration range
- L - Indicates that an analyte has a concentration below the Minimum Detection Limit (MDL). The reported concentration is not recommended for regulatory use as the analyte signal may have a signal-to-noise ratio less than the criteria deemed necessary to be considered a detected analyte.
- LOD – Limit of Detection: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOD. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the minimum detection limit (MDL). The LOD is adjusted for sample weight or volume.
- LOQ – Limit of Quantitation: For reports conforming to the DOD ELAP QSM, this is the QSM-defined LOQ. For reports conforming to TNI requirements (but not DOD ELAP QSM requirements), this value is the reporting limit (RL). The LOD is adjusted for sample weight or volume.
- <LOD() – Analyte was not found at a concentration high enough to be reported as detected. It is reported as less than the LOD, and the LOD is given in the parentheses.

General Reporting Notes – Data Qualifiers

- ND – Indicates a non-detect.
- NR – Indicates a value that is not reportable due to issues observed in sample preparation or analysis.
- PR – The associated congener(s) is(are) poorly resolved.
- QI – Indicates the presence of a quantitative interference.
- RL – Reporting Limit. Lowest reportable value. The level is higher than the MDL.
- SI – Denotes “Single Ion Mode” and is utilized for PCBs where the secondary ion trace has a significantly elevated noise level due to background PFK. Responses for such peaks are calculated using an EMPC approach based solely on the primary ion area(s) and may be considered estimates.
- U – The analyte was not detected.
- V – The labeled standard recovery is not within method control limits.
- X – Results from re-injection/repeat/second-column analysis.

Lab Identifiers/ Data Attributes

- AR – Indicates use of the archived portion of the sample extract.
- CU – Indicates a sample that required additional clean-up prior to HRMS injection/processing.
- D – Dilution Data. Result was obtained from the analysis of a dilution. The number that follows the “D” indicates the dilution factor.
- DE – Indicates a dilution performed with the addition of ES (Extraction Standard) solution.
- DUP – Designation for a duplicate sample.
- MS – Designation for a matrix spike.
- MSD – Designation for a matrix spike duplicate.
- RJ – Indicates a reinjection of the sample extract.
- S – Indicates a sample split. The number that follows the “S” indicates the split factor.
- R – Indicates a re-extraction of the sample.

PFAS Compound Acronym List		
Acronym	Compound Name	CAS #
Target Analytes		
PFBA	Perfluorobutanoic Acid	375-22-4
PFPeA	Perfluoropentanoic Acid	2706-90-3
PFHxA	Perfluorohexanoic Acid	307-24-4
PFHpA	Perfluoroheptanoic Acid	375-85-9
PFOA	Perfluorooctanoic Acid	335-67-1
PFNA	Perfluorononanoic Acid	375-95-1
PFDA	Perfluorodecanoic acid	335-76-2
PFUnA (PFUnDA)	Perfluoroundecanoic acid	2058-94-8
PFDoA (PFDoDA)	Perfluorododecanoic acid	307-55-1
PFTriA (PFTriA)	Perfluorotridecanoic acid	72629-94-8
PFTeDA (PFTA)	Perfluorotetradecanoic acid	376-06-7
PFBS	Perfluorobutane sulfonic acid	375-73-5
PFPeS	Perfluoropentane sulfonic acid	2706-91-4
PFHxS	Perfluorohexane sulfonic acid	355-46-4
PFHpS	Perfluoroheptane sulfonic acid	375-92-8
PFOS	Perfluorooctane sulfonic acid	1763-23-1
PFNS	Perfluorononane sulfonic acid	68259-12-1
PFDS	Perfluorodecane sulfonic acid	757124-72-4
4:2 FTS	4:2 fluorotelomer sulfonic acid	27619-97-2
6:2 FTS	6:2 fluorotelomer sulfonic acid	39108-34-4
8:2 FTS	8:2 fluorotelomer sulfonic acid	13252-13-6
PFOSA (FOSA)	Perfluorooctane sulfonamide	754-91-6
N-MeFOSAA	N-methyl perfluorooctane sulfonamido acetic acid	2355-31-9
N-EtFOSAA	N-ethyl perfluorooctane sulfonamido acetic acid	2991-50-6
* HFPO-DA	2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy)-propanoic acid (Gen-X)	13252-13-6
* 11CI-PF3OUdS	11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	674-13-5
* 9CI-PF3ONS	9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	919005-14-4
* ADONA	4,8-dioxia-3H-perfluorononanoic acid	756426-58-1
* PFMOAA	Perfluoro-2-methoxyacetic acid	763051-92-9
* PFMOPrA	Perfluoro-3-methoxypropanoic acid	377-73-1
* PFO2HxA	Perfluoro (3,5-dioxahexanoic) acid	39492-88-1
* PFO3OA	Perfluoro (3,5,7-trioxaoctanoic) acid	39492-89-2
* PFO4DA	Perfluoro (3,5,7,9-tetraoxadecanoic) acid	39492-90-5
* PFO5DA	Perfluoro(3,5,7,9,11-pentaoxadodecanoic) acid	39492-91-6
* Nafion Byproduct 1	Nafion Byproduct 1	29311-67-9
* Nafion Byproduct 2	Nafion Byproduct 2	749836-20-2
* PFEESA	Perfluoro(2-ethoxyethane)sulphonic acid	113507-82-7
* PFMOBA	Perfluoro-4-methoxybutanic acid	863090-89-5
* PEPA	Perfluoro-2-ethoxypropanoic acid	N/A
* PMPA	Perfluoro-2-methoxypropanoic acid	13140-29-9
* 10:2 FTS	Fluorotelomer sulfonate 10:2	120226-60-0
* N-EtFOSA	N-ethylperfluoro-1-octanesulfonamide	4151-50-2
* N-EtFOSE	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol	1691-99-2
* N-MeFOSA	N-methylperfluoro-1-octanesulfonamide	31506-32-8
* N-MeFOSE	2-(N-methylperfluoro-1-octanesulfonamido)-ethanol	24448-09-7
* PFECA-G	4-(Heptafluoroisopropoxy)hexafluorobutanoic acid	801212-59-9
* PFHxDA	Perfluorohexadecanoic acid	67905-19-5

Extraction Standards		
MPFBA	Perfluoro-n-[13C4]butanoic acid	
M5PFPeA	Perfluoro-n-[13C5]pentanoic acid	
M3PFBS	Sodium perfluoro-1-[2,3,4-13C3]-butanesulfonic acid	
M2-4:2 FTS	Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]-hexane sulfonic acid	
M5PFHxA	Perfluoro-n-[1,2,3,4,6-13C5]hexanoic acid	
M3HFPO-DA	2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy)-13C3-propanoic acid	
M4PFHpA	Perfluoro-n-[1,2,3,4-13C4]heptanoic acid	
M3PFHxS	Sodium perfluoro-1-[1,2,3-13C3]-hexanesulfonic acid	
M2-6:2 FTS	Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]-octane sulfonic acid	
M8PFOA	Perfluoro-n-[13C8]octanoic acid	
M9PFNA	Perfluoro-n-[13C9]nonanoic acid	
M8PFOS	Sodium perfluoro-1-[13C8]-octanesulfonic acid	
M2-8:2 FTS	Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]-decane sulfonic acid	
M8FOSA	Perfluoro-1-[13C8]octanesulfonamide	
M6PFDA	Perfluoro-n-[1,2,3,4,5,6-13C6]decanoic acid	
d3-N-MeFOSAA	N-methyl-d3-perfluoro-1-octanesulfonamide	
d5-N-EtFOSAA	N-ethyl-d5-perfluoro-1-octanesulfonamide	
M7PFUnDA (M7PFUdA)	Perfluoro-n-[1,2,3,4,5,6,7-13C7]undecanoic acid	
MPFDoA	Perfluoro-n-[1,2-13C2]dodecanoic acid	
M2PFTeDA	Perfluoro-n-[1,2-13C2]tetradecanoic acid	
Injection Standards		
M3PFBA	Perfluoro-n-[2,3,4-13C3]butanoic acid	
M2PFOA	Perfluoro-n-[1,2-13C2]octanoic acid	
MPFDA	Perfluoro-n-[1,2-13C2]decanoic acid	
MPFOS	Sodium perfluoro-1-[1,2,3,4-13C4]-octanesulfonic acid	

* Analytes are currently not accredited under TNI Standard. Accreditation pending.

Sample Custody



Enthalpy
analytical

Enthalpy Ultratrace Job#: _____ COC Page 1 of 1

X Standard Turn Around Time

☐ Rush Turn Around Time -- Date Needed _____

• All Fast TATs Subject to Approval by Enthalpy Analytical, Inc.

- All Samples Disposed of After 6 months Unless Otherwise Instructed.

Enthalpy Analytical-Wilmington, NC has added enhancements to standard methods to improve accuracy, precision and permit an assessment of laboratory performance in the context of your specific data needs. For more information email Cindy.James@enthalpy.com.

Client Name: **BRUNSWICK COUNTY UTILITIES**

Project Number: _____

PO#: _____

Project Manager: GLENN WALKER

Site Name: NORTHWEST WATER PLANT

Telephone#:

Report To: **SAME**

Location: **LELAND N.C.**

Email:

This Chain of Custody is applicable to Non-Air samples. Standard TAT differ per analysis and are provided by request.

Client Special Instructions:

Matrix: GW-Groundwater, WW-Wastewater, NW-Non-Potable Water, DW-Drinking Water, S-Soil, SL-Sludge, BT-Biological Tissue, O-Other

Type: G=Grab C=Composite Q=Quality Control

[illegible]

Relinquished By:

Date:

Received By:

Date: _____

Time:

Sample Temperature Upon Receipt:

9/17/2020

Brachylogy noly

9/17/20 11:30

☒ Iced ☐ Ambient °C 4.3 T9☐ Iced ☐ Ambient °C☐ Iced ☐ Ambient °C

2714 Exchange Drive • Wilmington, NC 28405 • (910) 212-5858 • www.enthalpy.com

EE
B
9/17/20

~~Brunswick~~ Courier, On Ice, cooler, no seal, good condition, Bu 9/17/20
Brunswick

**This Is The Last Page
Of This Report.**